

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102721\
 Data File : BG050730.D
 Acq On : 26 Oct 2021 14:10
 Operator : CG/JU
 Sample : SST00547
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST005447

Quant Time: Oct 26 16:22:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 26 15:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.252	152	35636	20.000	ng/u1	0.00
20) Naphthalene-d8	11.078	136	151440	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.880	164	102176	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.624	188	225561	20.000	ng/u1	0.00
79) Chrysene-d12	21.930	240	184227	20.000	ng/u1	# 0.00
88) Perylene-d12	25.368	264	179918	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.611	96	1637	1.688	ng/uL	0.00
4) Pyridine-d5	4.040	84	12709	4.397	ng/u1	0.00
7) Phenol-d5	7.395	99	15445	4.961	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.565	67	9938	5.072	ng/u1	0.00
11) 2-Chlorophenol-d4	7.776	132	11174	5.122	ng/u1	0.00
15) 4-Methylphenol-d8	8.952	113	11776	4.997	ng/u1	0.00
21) Nitrobenzene-d5	9.422	128	5501	5.192	ng/u1	0.00
24) 2-Nitrophenol-d4	10.150	143	6160	5.418	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.697	165	10942	4.873	ng/u1	0.00
31) 4-Chloroaniline-d4	11.214	131	15405	4.803	ng/u1	0.00
46) Dimethylphthalate-d6	14.269	166	35172	5.030	ng/u1	0.00
49) Acenaphthylene-d8	14.574	160	43947	5.004	ng/u1	0.00
54) 4-Nitrophenol-d4	15.068	143	5153	3.826	ng/u1	0.00
60) Fluorene-d10	15.867	176	30009	4.763	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.985	200	4725	4.200	ng/u1	0.00
73) Anthracene-d10	17.724	188	49237	5.167	ng/u1	0.00
81) Pyrene-d10	20.003	212	52772	5.387	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.121	264	46447	5.134	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.646	88	2414	2.057	ng/uL#	84
5) Pyridine	4.057	79	14139	4.829	ng/u1	98
6) Benzaldehyde	7.383	77	12353	6.016	ng/u1	98
8) Phenol	7.418	94	15777	4.944	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.653	93	12404	5.182	ng/u1	97
12) 2-Chlorophenol	7.812	128	11567	5.182	ng/u1#	89
13) 2-Methylphenol	8.687	108	11509	4.894	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.764	45	13080	3.642	ng/u1	96
16) Acetophenone	9.081	105	19756	5.276	ng/u1	99
17) N-Nitroso-di-n-propyla...	9.051	70	11728	5.212	ng/u1	98
18) 4-Methylphenol	9.010	108	12519	5.026	ng/u1	100
19) Hexachloroethane	9.339	117	5057	5.272	ng/u1	94
22) Nitrobenzene	9.469	77	16499	5.196	ng/u1	99
23) Isophorone	9.986	82	30753	5.040	ng/u1	100
25) 2-Nitrophenol	10.185	139	6405	5.481	ng/u1	94
26) 2,4-Dimethylphenol	10.227	107	14456	4.983	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.467	93	16911	5.133	ng/u1#	96
29) 2,4-Dichlorophenol	10.726	162	10445	4.699	ng/u1	95
30) Naphthalene	11.131	128	39294	5.054	ng/u1	97
32) 4-Chloroaniline	11.237	127	15413	4.773	ng/u1#	85
33) Hexachlorobutadiene	11.402	225	7110	4.401	ng/u1	96
34) Caprolactam	11.989	113	3908	4.551	ng/u1	94
35) 4-Chloro-3-methylphenol	12.336	107	13093	4.996	ng/u1	98
36) 2-Methylnaphthalene	12.718	142	25359	4.797	ng/u1	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102721\
 Data File : BG050730.D
 Acq On : 26 Oct 2021 14:10
 Operator : CG/JU
 Sample : SST00547
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SST005447


Quant Time: Oct 26 16:22:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 26 15:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.935	142	26010	4.888	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	13.082	216	13865	4.702	ng/ul	99
40) Hexachlorocyclopentadiene	13.053	237	3861	1.974	ng/ul#	90
41) 2,4,6-Trichlorophenol	13.317	196	8605	4.630	ng/ul	97
42) 2,4,5-Trichlorophenol	13.393	196	9084	4.468	ng/ul	99
43) 1,1'-Biphenyl	13.711	154	35282	5.131	ng/ul	97
44) 2-Chloronaphthalene	13.764	162	27534	5.111	ng/ul	95
45) 2-Nitroaniline	13.957	65	10230	5.801	ng/ul	92
47) Dimethylphthalate	14.316	163	36261	5.154	ng/ul	98
48) 2,6-Dinitrotoluene	14.445	165	6661	5.033	ng/ul	99
50) Acenaphthylene	14.604	152	45847	5.147	ng/ul	97
51) 3-Nitroaniline	14.780	138	7636	5.304	ng/ul	93
52) Acenaphthene	14.945	153	29127	4.947	ng/ul	95
53) 2,4-Dinitrophenol	14.997	184	1666	2.020	ng/ul#	81
55) 4-Nitrophenol	15.080	109	4807	3.609	ng/ul	95
56) Dibenzofuran	15.274	168	41871	4.890	ng/ul	97
57) 2,4-Dinitrotoluene	15.232	165	10075	5.243	ng/ul#	87
58) 2,3,4,6-Tetrachlorophenol	15.503	232	6939	3.969	ng/ul	91
59) Diethylphthalate	15.673	149	38360	5.225	ng/ul	97
61) Fluorene	15.926	166	33837	4.967	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.908	204	17200	4.662	ng/ul	93
63) 4-Nitroaniline	15.938	138	7568	5.206	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.996	198	4605	4.141	ng/ul#	96
67) N-Nitrosodiphenylamine	16.120	169	30362	5.324	ng/ul	96
68) 4-Bromophenyl-phenylether	16.801	248	10325	4.917	ng/ul	92
69) Hexachlorobenzene	16.925	284	10538	4.765	ng/ul	97
70) Atrazine	17.060	200	12133	4.997	ng/ul	95
71) Pentachlorophenol	17.283	266	2266	1.565	ng/ul	94
72) Phenanthrene	17.671	178	56899	5.054	ng/ul	98
74) Anthracene	17.759	178	57826	5.186	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.681	216	14813	4.981	ng/ul	94
76) Pentachlorobenzene	15.191	250	14192	5.224	ng/ul	95
77) Carbazole	18.029	167	52858	5.296	ng/ul	99
78) Di-n-butylphthalate	18.564	149	65409	5.530	ng/ul	99
80) Fluoranthene	19.668	202	67800	5.511	ng/ul	97
82) Pyrene	20.033	202	69049	5.692	ng/ul	99
83) Butylbenzylphthalate	20.896	149	28153	6.296	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.819	252	21825	5.748	ng/ul	92
85) Benzo(a)anthracene	21.913	228	60100	5.411	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.784	149	40276	6.268	ng/ul	100
87) Chrysene	21.983	228	56984	5.354	ng/ul	98
89) Di-n-octyl phthalate	23.064	149	68091	6.586	ng/ul	100
90) Benzo(b)fluoranthene	24.263	252	57616	5.258	ng/ul	98
91) Benzo(k)fluoranthene	24.263	252	57616	5.650	ng/ul	99
93) Benzo(a)pyrene	25.203	252	57180	5.469	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.304	276	55891	4.761	ng/ul#	95
95) Dibenzo(a,h)anthracene	29.363	278	53069	5.312	ng/ul	98
96) Benzo(g,h,i)perylene	30.538	276	50898	5.143	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 26 16:22:59 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 26 15:37:22 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 2.057 ng/uL
RT: 3.646 min Scan# 27
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion: 88 Resp: 2414
Ion Ratio Lower Upper
88 100
43 49.6 23.8 35.8#
58 72.6 53.7 80.5



#5
Pyridine
Concen: 4.829 ng/uL
RT: 4.057 min Scan# 97
Delta R.T. 0.008 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion: 79 Resp: 14139
Ion Ratio Lower Upper
79 100
52 77.2 61.8 92.8
51 41.2 30.6 45.8









#6
Benzaldehyde
Concen: 6.016 ng/ul
RT: 7.383 min Scan# 663
Delta R.T. 0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:	77	Resp:	12353
Ion	Ratio	Lower	Upper
77	100		
105	77.7	64.2	96.4
106	79.0	62.1	93.1



#8
Phenol
Concen: 4.944 ng/ul
RT: 7.418 min Scan# 669
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:	94	Resp:	15777
Ion	Ratio	Lower	Upper
94	100		
65	31.3	25.3	37.9
66	46.5	34.9	52.3









#10

Bis(2-Chloroethyl)ether
Concen: 5.182 ng/ul
RT: 7.653 min Scan# 709
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Instrument :
BNA_G
ClientSampleId :
SSTD005447




Tgt Ion:	93	Resp:	12404
Ion	Ratio	Lower	Upper
93	100		
63	86.5	66.9	100.3
95	31.1	26.1	39.1




#12

2-Chlorophenol
Concen: 5.182 ng/ul
RT: 7.812 min Scan# 736
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:	128	Resp:	11567
Ion	Ratio	Lower	Upper
128	100		
64	59.8	43.2	64.8
130	36.6	22.4	33.6









#13
2-Methylphenol
Concen: 4.894 ng/ul
RT: 8.687 min Scan# 885
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:108 Resp: 11509
Ion Ratio Lower Upper
108 100
107 90.6 75.5 113.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 3.642 ng/ul
RT: 8.764 min Scan# 898
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion: 45 Resp: 13080
Ion Ratio Lower Upper
45 100
77 11.8 11.4 17.2
79 10.9 8.6 13.0









#16
Acetophenone
Concen: 5.276 ng/ul
RT: 9.081 min Scan# 952
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:	105	Resp:	19756
Ion Ratio	Lower	Upper	
105	100		
77	88.9	72.1	108.1
51	30.9	24.6	36.8



#17
N-Nitroso-di-n-propylamine
Concen: 5.212 ng/ul
RT: 9.051 min Scan# 947
Delta R.T. 0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:	70	Resp:	11728
Ion Ratio	Lower	Upper	
70	100		
42	60.7	50.4	75.6
101	9.0	7.8	11.6
130	19.6	15.9	23.9









#18
4-Methylphenol
Concen: 5.026 ng/ul
RT: 9.010 min Scan# 940
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:108 Resp: 12519
Ion Ratio Lower Upper
108 100
107 117.2 93.8 140.8



#19
Hexachloroethane
Concen: 5.272 ng/ul
RT: 9.339 min Scan# 996
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:117 Resp: 5057
Ion Ratio Lower Upper
117 100
201 105.0 83.0 124.4
199 56.1 53.8 80.6









#22
Nitrobenzene
Concen: 5.196 ng/ul
RT: 9.469 min Scan# 1018
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion: 77 Resp: 16499
Ion Ratio Lower Upper
77 100
123 37.9 29.9 44.9
65 13.6 11.0 16.4



#23
Isophorone
Concen: 5.040 ng/ul
RT: 9.986 min Scan# 1106
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion: 82 Resp: 30753
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 14.6 11.6 17.4









#25
2-Nitrophenol
Concen: 5.481 ng/ul
RT: 10.185 min Scan# 1140
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:139 Resp: 6405
Ion Ratio Lower Upper
139 100
65 67.4 49.1 73.7
109 34.3 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 4.983 ng/ul
RT: 10.227 min Scan# 1147
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:107 Resp: 14456
Ion Ratio Lower Upper
107 100
121 47.1 39.0 58.6
122 83.9 64.6 97.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 5.133 ng/ul
 RT: 10.467 min Scan# 1188
 Delta R.T. 0.003 min
 Lab File: BG050730.D
 Acq: 26 Oct 2021 14:10




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005447


Tgt Ion:	93	Resp:	16911
Ion	Ratio	Lower	Upper
93	100		
95	32.3	24.6	37.0
123	13.3	8.3	12.5#



#29
 2,4-Dichlorophenol
 Concen: 4.699 ng/ul
 RT: 10.726 min Scan# 1232
 Delta R.T. 0.008 min
 Lab File: BG050730.D
 Acq: 26 Oct 2021 14:10

Tgt Ion:	162	Resp:	10445
Ion	Ratio	Lower	Upper
162	100		
164	58.8	49.4	74.0
98	40.2	29.0	43.4









#30
Naphthalene
Concen: 5.054 ng/ul
RT: 11.131 min Scan# 1301
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:128 Resp: 39294
Ion Ratio Lower Upper
128 100
129 11.5 8.2 12.2
127 12.3 11.0 16.4



#32
4-Chloroaniline
Concen: 4.773 ng/ul
RT: 11.237 min Scan# 1319
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:127 Resp: 15413
Ion Ratio Lower Upper
127 100
129 38.7 24.3 36.5#









#33
Hexachlorobutadiene
Concen: 4.401 ng/ul
RT: 11.402 min Scan# 1347
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:225 Resp: 7110
Ion Ratio Lower Upper
225 100
223 68.5 51.1 76.7
227 68.3 52.7 79.1



#34
Caprolactam
Concen: 4.551 ng/ul
RT: 11.989 min Scan# 1447
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:113 Resp: 3908
Ion Ratio Lower Upper
113 100
55 173.4 144.1 216.1
56 123.0 105.8 158.8









#35
4-Chloro-3-methylphenol
Concen: 4.996 ng/ul
RT: 12.336 min Scan# 1506
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:	107	Resp:	13093
Ion	Ratio	Lower	Upper
107	100		
144	20.1	16.8	25.2
142	67.8	55.4	83.2



#36
2-Methylnaphthalene
Concen: 4.797 ng/ul
RT: 12.718 min Scan# 1571
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:	142	Resp:	25359
Ion	Ratio	Lower	Upper
142	100		
141	95.0	72.0	108.0









#37
1-Methylnaphthalene
Concen: 4.888 ng/ul
RT: 12.935 min Scan# 1608
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:142 Resp: 26010
Ion Ratio Lower Upper
142 100
141 93.3 78.9 118.3
116 3.8 3.8 5.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.702 ng/ul
RT: 13.082 min Scan# 1633
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:216 Resp: 13865
Ion Ratio Lower Upper
216 100
214 80.5 65.0 97.4
179 20.3 16.5 24.7
108 17.5 12.2 18.4









#40

Hexachlorocyclopentadiene
Concen: 1.974 ng/ul
RT: 13.053 min Scan# 1628
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Instrument :
BNA_G
ClientSampleId :
SSTD005447




Tgt Ion:237 Resp: 3861
Ion Ratio Lower Upper
237 100
235 64.0 45.6 68.4
272 7.2 9.5 14.3#




#41

2,4,6-Trichlorophenol
Concen: 4.630 ng/ul
RT: 13.317 min Scan# 1673
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:196 Resp: 8605
Ion Ratio Lower Upper
196 100
198 96.0 75.1 112.7
200 27.6 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 4.468 ng/ul
RT: 13.393 min Scan# 1686
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:196 Resp: 9084
Ion Ratio Lower Upper
196 100
198 97.7 78.3 117.5
200 33.7 24.6 37.0



#43
1,1'-Biphenyl
Concen: 5.131 ng/ul
RT: 13.711 min Scan# 1740
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:154 Resp: 35282
Ion Ratio Lower Upper
154 100
153 39.7 33.4 50.2
76 15.3 12.4 18.6









#44
2-Chloronaphthalene
Concen: 5.111 ng/ul
RT: 13.764 min Scan# 1749
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:162 Resp: 27534
Ion Ratio Lower Upper
162 100
164 32.9 24.5 36.7
127 44.1 32.6 48.8



#45
2-Nitroaniline
Concen: 5.801 ng/ul
RT: 13.957 min Scan# 1782
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion: 65 Resp: 10230
Ion Ratio Lower Upper
65 100
92 58.8 54.8 82.2
138 90.0 76.1 114.1





#47

Dimethylphthalate

Concen: 5.154 ng/ul

RT: 14.316 min Scan# 1843

Delta R.T. 0.002 min

Lab File: BG050730.D

Acq: 26 Oct 2021 14:10





Tgt Ion:163 Resp: 36261

Ion Ratio Lower Upper

163 100

194 4.2 3.0 4.6

164 9.3 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 5.033 ng/ul

RT: 14.445 min Scan# 1865

Delta R.T. 0.002 min

Lab File: BG050730.D

Acq: 26 Oct 2021 14:10




Tgt Ion:165 Resp: 6661


Ion Ratio Lower Upper

165 100

89 77.7 62.3 93.5


121 28.4 20.8 31.2








#50
Acenaphthylene
Concen: 5.147 ng/ul
RT: 14.604 min Scan# 1892
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Instrument :
BNA_G
ClientSampleId :
SSTD005447







Tgt Ion:152 Resp: 45847
Ion Ratio Lower Upper
152 100
151 20.4 17.4 26.2
153 14.2 10.3 15.5



#51
3-Nitroaniline
Concen: 5.304 ng/ul
RT: 14.780 min Scan# 1922
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:138 Resp: 7636
Ion Ratio Lower Upper
138 100
108 12.7 8.6 12.8
92 128.8 97.0 145.4









#52
Acenaphthene
Concen: 4.947 ng/ul
RT: 14.945 min Scan# 1950
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:153 Resp: 29127
Ion Ratio Lower Upper
153 100
152 47.0 40.2 60.2
154 89.8 68.2 102.2



#53
2,4-Dinitrophenol
Concen: 2.020 ng/ul
RT: 14.997 min Scan# 1959
Delta R.T. 0.008 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:184 Resp: 1666
Ion Ratio Lower Upper
184 100
63 80.8 62.1 93.1
154 55.1 68.3 102.5#









#55
4-Nitrophenol
Concen: 3.609 ng/ul
RT: 15.080 min Scan# 1973
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:109 Resp: 4807
Ion Ratio Lower Upper
109 100
139 91.7 81.4 122.0
65 120.9 97.8 146.6



#56
Dibenzofuran
Concen: 4.890 ng/ul
RT: 15.274 min Scan# 2006
Delta R.T. 0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:168 Resp: 41871
Ion Ratio Lower Upper
168 100
139 39.6 30.2 45.4









#57
2,4-Dinitrotoluene
Concen: 5.243 ng/ul
RT: 15.232 min Scan# 1999
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:165 Resp: 10075
Ion Ratio Lower Upper
165 100
63 44.4 43.4 65.2
182 3.7 1.7 2.5#



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.969 ng/ul
RT: 15.503 min Scan# 2045
Delta R.T. 0.008 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:232 Resp: 6939
Ion Ratio Lower Upper
232 100
131 50.1 34.3 51.5
130 4.0 2.8 4.2
166 24.0 22.4 33.6









#59
Diethylphthalate
Concen: 5.225 ng/ul
RT: 15.673 min Scan# 2074
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:149 Resp: 38360
Ion Ratio Lower Upper
149 100
177 20.5 18.2 27.4
150 13.2 10.2 15.2



#61
Fluorene
Concen: 4.967 ng/ul
RT: 15.926 min Scan# 2117
Delta R.T. 0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:166 Resp: 33837
Ion Ratio Lower Upper
166 100
165 97.7 78.6 118.0
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 4.662 ng/ul
RT: 15.908 min Scan# 2114
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampled :
SSTD005447


Tgt Ion:204 Resp: 17200
Ion Ratio Lower Upper
204 100
206 36.1 24.8 37.2
141 58.7 44.1 66.1



#63
4-Nitroaniline
Concen: 5.206 ng/ul
RT: 15.938 min Scan# 2119
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:138 Resp: 7568
Ion Ratio Lower Upper
138 100
92 62.1 47.0 70.4
108 109.0 77.2 115.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 4.141 ng/ul
 RT: 15.996 min Scan# 2129
 Delta R.T. 0.002 min
 Lab File: BG050730.D
 Acq: 26 Oct 2021 14:10




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005447


Tgt Ion:198 Resp: 4605
 Ion Ratio Lower Upper
 198 100
 182 7.5 3.0 4.6#
 77 30.7 25.8 38.8



#67
 N-Nitrosodiphenylamine
 Concen: 5.324 ng/ul
 RT: 16.120 min Scan# 2150
 Delta R.T. 0.002 min
 Lab File: BG050730.D
 Acq: 26 Oct 2021 14:10

Tgt Ion:169 Resp: 30362
 Ion Ratio Lower Upper
 169 100
 168 70.7 58.3 87.5
 167 35.7 31.4 47.2









#68
4-Bromophenyl-phenylether
Concen: 4.917 ng/ul
RT: 16.801 min Scan# 2266
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:	248	Resp:	10325
Ion Ratio	Lower	Upper	
248	100		
250	96.8	73.3	109.9
141	85.2	60.2	90.2



#69
Hexachlorobenzene
Concen: 4.765 ng/ul
RT: 16.925 min Scan# 2287
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:	284	Resp:	10538
Ion Ratio	Lower	Upper	
284	100		
142	38.7	33.3	49.9
249	36.1	28.3	42.5





#70

Atrazine

Concen: 4.997 ng/ul

RT: 17.060 min Scan# 2310

Delta R.T. -0.003 min

Lab File: BG050730.D

Acq: 26 Oct 2021 14:10

Tgt Ion:200 Resp: 12133

Ion Ratio Lower Upper

200 100

173 28.3 19.9 29.9





215 53.7 40.4 60.6

Instrument :

BNA_G

ClientSampleId :

SSTD005447



#71

Pentachlorophenol

Concen: 1.565 ng/ul

RT: 17.283 min Scan# 2348

Delta R.T. 0.014 min

Lab File: BG050730.D

Acq: 26 Oct 2021 14:10




Tgt Ion:266 Resp: 2266


Ion Ratio Lower Upper

266 100

264 66.7 51.0 76.4

268 57.8 51.1 76.7









#72
Phenanthrene
Concen: 5.054 ng/ul
RT: 17.671 min Scan# 2414
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:178 Resp: 56899
Ion Ratio Lower Upper
178 100
179 15.5 12.7 19.1
176 18.9 16.2 24.2



#74
Anthracene
Concen: 5.186 ng/ul
RT: 17.759 min Scan# 2429
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:178 Resp: 57826
Ion Ratio Lower Upper
178 100
179 17.4 12.7 19.1
176 18.3 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.981 ng/uL
RT: 13.681 min Scan# 1735
Delta R.T. 0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampled :
SSTD005447


Tgt Ion:216 Resp: 14813
Ion Ratio Lower Upper
216 100
214 71.8 61.0 91.6
218 42.0 37.1 55.7



#76
Pentachlorobenzene
Concen: 5.224 ng/uL
RT: 15.191 min Scan# 1992
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:250 Resp: 14192
Ion Ratio Lower Upper
250 100
248 64.0 50.2 75.2
252 57.7 51.9 77.9









#77
 Carbazole
 Concen: 5.296 ng/ul
 RT: 18.029 min Scan# 2475
 Delta R.T. 0.002 min
 Lab File: BG050730.D
 Acq: 26 Oct 2021 14:10




Instrument :
 BNA_G
 ClientSampleId :
 SST005447


Tgt Ion:167	Resp:	52858
Ion Ratio	Lower	Upper
167	100	
166	22.0	17.6 26.4
139	14.1	10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 5.530 ng/ul
 RT: 18.564 min Scan# 2566
 Delta R.T. 0.002 min
 Lab File: BG050730.D
 Acq: 26 Oct 2021 14:10

Tgt Ion:149	Resp:	65409
Ion Ratio	Lower	Upper
149	100	
150	9.5	7.5 11.3
104	5.0	4.9 7.3









#80
Fluoranthene
Concen: 5.511 ng/u1
RT: 19.668 min Scan# 2754
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:202 Resp: 67800
Ion Ratio Lower Upper
202 100
101 11.1 7.7 11.5
100 9.0 6.6 10.0



#82
Pyrene
Concen: 5.692 ng/u1
RT: 20.033 min Scan# 2816
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:202 Resp: 69049
Ion Ratio Lower Upper
202 100
101 12.1 9.7 14.5
100 10.3 7.6 11.4









#83
Butylbenzylphthalate
Concen: 6.296 ng/ul
RT: 20.896 min Scan# 2963
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:149 Resp: 28153
Ion Ratio Lower Upper
149 100
91 78.0 65.8 98.6
206 19.2 17.2 25.8



#84
3,3'-Dichlorobenzidine
Concen: 5.748 ng/ul
RT: 21.819 min Scan# 3120
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:252 Resp: 21825
Ion Ratio Lower Upper
252 100
254 58.8 52.2 78.2
126 8.9 8.3 12.5









#85
Benzo(a)anthracene
Concen: 5.411 ng/ul
RT: 21.913 min Scan# 3136
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:228 Resp: 60100
Ion Ratio Lower Upper
228 100
229 19.2 15.9 23.9
226 25.2 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 6.268 ng/ul
RT: 21.784 min Scan# 3114
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:149 Resp: 40276
Ion Ratio Lower Upper
149 100
167 26.6 21.3 31.9
279 3.7 2.8 4.2





#87

Chrysene

Concen: 5.354 ng/ul

RT: 21.983 min Scan# 3148

Delta R.T. 0.002 min

Lab File: BG050730.D

Acq: 26 Oct 2021 14:10





Tgt Ion:228	Resp:	56984
Ion Ratio	Lower	Upper
228	100	
226	30.8	25.4 38.0
229	19.4	16.2 24.2

Instrument :

BNA_G

ClientSampleId :

SSTD005447



#89

Di-n-octyl phthalate

Concen: 6.586 ng/ul




RT: 23.064 min Scan# 3332


Delta R.T. 0.003 min

Lab File: BG050730.D

Acq: 26 Oct 2021 14:10

Tgt Ion:149 Resp: 68091









#90
Benzo(b)fluoranthene
Concen: 5.258 ng/ul
RT: 24.263 min Scan# 3536
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:252 Resp: 57616
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.8
125 9.6 6.7 10.1



#91
Benzo(k)fluoranthene
Concen: 5.650 ng/ul
RT: 24.263 min Scan# 3536
Delta R.T. -0.074 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:252 Resp: 57616
Ion Ratio Lower Upper
252 100
253 21.7 17.1 25.7
125 9.6 7.4 11.0









#93
Benzo(a)pyrene
Concen: 5.469 ng/ul
RT: 25.203 min Scan# 3696
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10




Instrument :
BNA_G
ClientSampleId :
SSTD005447


Tgt Ion:252 Resp: 57180
Ion Ratio Lower Upper
252 100
253 22.9 17.8 26.6
125 10.5 8.3 12.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.761 ng/ul
RT: 29.304 min Scan# 4394
Delta R.T. 0.002 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:276 Resp: 55891
Ion Ratio Lower Upper
276 100
138 19.4 12.4 18.6#
277 24.1 18.5 27.7









#95
Dibenzo(a,h)anthracene
Concen: 5.312 ng/ul
RT: 29.363 min Scan# 4404
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Instrument :
BNA_G
ClientSampleId :
SSTD005447

Tgt Ion:278 Resp: 53069
Ion Ratio Lower Upper
278 100
139 13.3 11.3 16.9
279 23.9 18.3 27.5



#96
Benzo(g,h,i)perylene
Concen: 5.143 ng/ul
RT: 30.538 min Scan# 4604
Delta R.T. -0.003 min
Lab File: BG050730.D
Acq: 26 Oct 2021 14:10

Tgt Ion:276 Resp: 50898
Ion Ratio Lower Upper
276 100
138 18.2 14.1 21.1
277 19.9 17.9 26.9

