

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102721\
 Data File : BG050732.D
 Acq On : 26 Oct 2021 15:32
 Operator : CG/JU
 Sample : SST02049
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020449

Quant Time: Oct 26 16:24:09 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 26 15:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.250	152	40869	20.000	ng/u1	0.00
20) Naphthalene-d8	11.076	136	186431	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.877	164	126729	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.627	188	277084	20.000	ng/u1	0.00
79) Chrysene-d12	21.934	240	223193	20.000	ng/u1	0.00
88) Perylene-d12	25.365	264	224632	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.608	96	6955	6.252	ng/uL	0.00
4) Pyridine-d5	4.031	84	51044	15.399	ng/u1	0.00
7) Phenol-d5	7.392	99	59639	16.702	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.562	67	39518	17.585	ng/u1	0.00
11) 2-Chlorophenol-d4	7.774	132	43308	17.310	ng/u1	0.00
15) 4-Methylphenol-d8	8.949	113	47687	17.644	ng/u1	0.00
21) Nitrobenzene-d5	9.425	128	23055	17.677	ng/u1	0.00
24) 2-Nitrophenol-d4	10.148	143	25660	18.332	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.688	165	44496	16.099	ng/u1	0.00
31) 4-Chloroaniline-d4	11.205	131	63673	16.126	ng/u1	0.00
46) Dimethylphthalate-d6	14.266	166	142108	16.385	ng/u1	0.00
49) Acenaphthylene-d8	14.572	160	178145	16.356	ng/u1	0.00
54) 4-Nitrophenol-d4	15.060	143	23997	14.366	ng/u1	0.00
60) Fluorene-d10	15.865	176	121747	15.581	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.982	200	23829	17.242	ng/u1	0.00
73) Anthracene-d10	17.727	188	188625	16.114	ng/u1	0.00
81) Pyrene-d10	20.001	212	206632	17.411	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.124	264	181540	16.072	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.644	88	7613	5.657	ng/uL#	96
5) Pyridine	4.049	79	54287	16.168	ng/u1	97
6) Benzaldehyde	7.380	77	46730	19.844	ng/u1	93
8) Phenol	7.416	94	63808	17.434	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.656	93	49229	17.933	ng/u1	98
12) 2-Chlorophenol	7.809	128	44584	17.418	ng/u1	89
13) 2-Methylphenol	8.685	108	47621	17.656	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.767	45	78184	18.982	ng/u1	99
16) Acetophenone	9.078	105	76054	17.711	ng/u1	98
17) N-Nitroso-di-n-propyla...	9.049	70	47316	18.335	ng/u1	93
18) 4-Methylphenol	9.014	108	50913	17.824	ng/u1	90
19) Hexachloroethane	9.343	117	18890	17.171	ng/u1	95
22) Nitrobenzene	9.466	77	65275	16.700	ng/u1	98
23) Isophorone	9.983	82	127630	16.990	ng/u1	99
25) 2-Nitrophenol	10.183	139	27262	18.950	ng/u1	97
26) 2,4-Dimethylphenol	10.224	107	58606	16.411	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.465	93	68138	16.801	ng/u1	100
29) 2,4-Dichlorophenol	10.718	162	43877	16.034	ng/u1	97
30) Naphthalene	11.129	128	150590	15.733	ng/u1	99
32) 4-Chloroaniline	11.235	127	64920	16.331	ng/u1	98
33) Hexachlorobutadiene	11.399	225	29516	14.842	ng/u1	96
34) Caprolactam	11.993	113	17437	16.493	ng/u1	97
35) 4-Chloro-3-methylphenol	12.333	107	55294	17.139	ng/u1	97
36) 2-Methylnaphthalene	12.721	142	103265	15.869	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102721\
 Data File : BG050732.D
 Acq On : 26 Oct 2021 15:32
 Operator : CG/JU
 Sample : SST02049
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SST020449


Quant Time: Oct 26 16:24:09 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 26 15:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.939	142	102592	15.660	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.080	216	55288	15.117	ng/ul	99
40) Hexachlorocyclopentadiene	13.050	237	24321	10.028	ng/ul#	84
41) 2,4,6-Trichlorophenol	13.315	196	36784	15.958	ng/ul	93
42) 2,4,5-Trichlorophenol	13.391	196	38626	15.318	ng/ul	98
43) 1,1'-Biphenyl	13.708	154	139716	16.382	ng/ul	98
44) 2-Chloronaphthalene	13.761	162	105620	15.808	ng/ul	97
45) 2-Nitroaniline	13.961	65	42405	19.386	ng/ul	94
47) Dimethylphthalate	14.313	163	142753	16.359	ng/ul	99
48) 2,6-Dinitrotoluene	14.443	165	29199	17.788	ng/ul	93
50) Acenaphthylene	14.601	152	181942	16.468	ng/ul	99
51) 3-Nitroaniline	14.778	138	31777	17.797	ng/ul	95
52) Acenaphthene	14.942	153	117313	16.066	ng/ul	97
53) 2,4-Dinitrophenol	14.989	184	13832	13.519	ng/ul#	84
55) 4-Nitrophenol	15.077	109	22911	13.870	ng/ul	92
56) Dibenzofuran	15.271	168	167120	15.736	ng/ul	95
57) 2,4-Dinitrotoluene	15.230	165	41607	17.456	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.494	232	30345	13.994	ng/ul#	87
59) Diethylphthalate	15.671	149	151048	16.589	ng/ul	97
61) Fluorene	15.923	166	133074	15.750	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.906	204	69327	15.151	ng/ul	96
63) 4-Nitroaniline	15.935	138	31981	17.736	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.994	198	23013	16.846	ng/ul#	96
67) N-Nitrosodiphenylamine	16.117	169	117125	16.718	ng/ul	98
68) 4-Bromophenyl-phenylether	16.805	248	41496	16.085	ng/ul	95
69) Hexachlorobenzene	16.922	284	42215	15.540	ng/ul	97
70) Atrazine	17.063	200	49803	16.696	ng/ul	92
71) Pentachlorophenol	17.269	266	17222	9.681	ng/ul	98
72) Phenanthrene	17.668	178	220995	15.978	ng/ul	99
74) Anthracene	17.756	178	220867	16.124	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.679	216	59142	16.190	ng/ul	97
76) Pentachlorobenzene	15.189	250	53705	16.092	ng/ul	99
77) Carbazole	18.027	167	207267	16.905	ng/ul	100
78) Di-n-butylphthalate	18.561	149	257903	17.751	ng/ul	100
80) Fluoranthene	19.666	202	264121	17.722	ng/ul	98
82) Pyrene	20.030	202	258970	17.621	ng/ul	96
83) Butylbenzylphthalate	20.894	149	107873	19.913	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.816	252	84844	18.443	ng/ul	98
85) Benzo(a)anthracene	21.910	228	226047	16.800	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.781	149	153961	19.778	ng/ul	99
87) Chrysene	21.981	228	217415	16.861	ng/ul	98
89) Di-n-octyl phthalate	23.062	149	262999	20.376	ng/ul	100
90) Benzo(b)fluoranthene	24.266	252	229031	16.739	ng/ul	99
91) Benzo(k)fluoranthene	24.337	252	207319	16.284	ng/ul	98
93) Benzo(a)pyrene	25.201	252	216215	16.562	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.302	276	237207	16.184	ng/ul	97
95) Dibenzo(a,h)anthracene	29.366	278	205190	16.451	ng/ul	99
96) Benzo(g,h,i)perylene	30.541	276	199253	16.126	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Oct 26 16:24:09 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 26 15:37:22 2021
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 5.657 ng/uL
RT: 3.644 min Scan# 26
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Instrument :
BNA_G
ClientSampleId :
SSTD020449



Tgt Ion: 88 Resp: 7613




Ion	Ratio	Lower	Upper
88	100		
43	36.6	23.8	35.8#
58	67.8	53.7	80.5




#5
Pyridine
Concen: 16.168 ng/uL
RT: 4.049 min Scan# 95
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion: 79 Resp: 54287

Ion	Ratio	Lower	Upper
79	100		
52	80.0	61.8	92.8
51	40.4	30.6	45.8









#6
Benzaldehyde
Concen: 19.844 ng/ul
RT: 7.380 min Scan# 662
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:	77	Resp:	46730
Ion	Ratio	Lower	Upper
77	100		
105	88.4	64.2	96.4
106	82.6	62.1	93.1



#8
Phenol
Concen: 17.434 ng/ul
RT: 7.416 min Scan# 668
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:	94	Resp:	63808
Ion	Ratio	Lower	Upper
94	100		
65	29.3	25.3	37.9
66	40.4	34.9	52.3









#10

Bis(2-Chloroethyl)ether
Concen: 17.933 ng/ul
RT: 7.656 min Scan# 709
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Instrument :
BNA_G
ClientSampleId :
SSTD020449




Tgt Ion: 93	Resp: 49229
Ion Ratio	Lower Upper
93 100	
63 83.4	66.9 100.3
95 29.6	26.1 39.1




#12

2-Chlorophenol
Concen: 17.418 ng/ul
RT: 7.809 min Scan# 735
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:128	Resp: 44584
Ion Ratio	Lower Upper
128 100	
64 62.1	43.2 64.8
130 33.4	22.4 33.6









#13
2-Methylphenol
Concen: 17.656 ng/ul
RT: 8.685 min Scan# 884
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:108 Resp: 47621
Ion Ratio Lower Upper
108 100
107 91.1 75.5 113.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.982 ng/ul
RT: 8.767 min Scan# 898
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion: 45 Resp: 78184
Ion Ratio Lower Upper
45 100
77 13.8 11.4 17.2
79 10.1 8.6 13.0









#16
Acetophenone
Concen: 17.711 ng/ul
RT: 9.078 min Scan# 951
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:	105	Resp:	76054
Ion Ratio	Lower	Upper	
105	100		
77	91.3	72.1	108.1
51	28.7	24.6	36.8



#17
N-Nitroso-di-n-propylamine
Concen: 18.335 ng/ul
RT: 9.049 min Scan# 946
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:	70	Resp:	47316
Ion Ratio	Lower	Upper	
70	100		
42	57.5	50.4	75.6
101	8.1	7.8	11.6
130	17.2	15.9	23.9









#18
4-Methylphenol
Concen: 17.824 ng/ul
RT: 9.014 min Scan# 940
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:108 Resp: 50913
Ion Ratio Lower Upper
108 100
107 106.3 93.8 140.8



#19
Hexachloroethane
Concen: 17.171 ng/ul
RT: 9.343 min Scan# 996
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:117 Resp: 18890
Ion Ratio Lower Upper
117 100
201 109.1 83.0 124.4
199 69.9 53.8 80.6









#22
Nitrobenzene
Concen: 16.700 ng/ul
RT: 9.466 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion: 77 Resp: 65275
Ion Ratio Lower Upper
77 100
123 38.6 29.9 44.9
65 13.8 11.0 16.4



#23
Isophorone
Concen: 16.990 ng/ul
RT: 9.983 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion: 82 Resp: 127630
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 15.0 11.6 17.4









#25
2-Nitrophenol
Concen: 18.950 ng/ul
RT: 10.183 min Scan# 1139
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:	139	Resp:	27262
Ion Ratio	Lower	Upper	
139	100		
65	59.9	49.1	73.7
109	32.3	28.6	42.8



#26
2,4-Dimethylphenol
Concen: 16.411 ng/ul
RT: 10.224 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:	107	Resp:	58606
Ion Ratio	Lower	Upper	
107	100		
121	48.6	39.0	58.6
122	80.7	64.6	97.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 16.801 ng/ul
 RT: 10.465 min Scan# 1187
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020449


Tgt Ion:	93	Resp:	68138
Ion	Ratio	Lower	Upper
93	100		
95	30.7	24.6	37.0
123	10.2	8.3	12.5



#29
 2,4-Dichlorophenol
 Concen: 16.034 ng/ul
 RT: 10.718 min Scan# 1230
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32

Tgt Ion:	162	Resp:	43877
Ion	Ratio	Lower	Upper
162	100		
164	61.8	49.4	74.0
98	41.3	29.0	43.4









#30
Naphthalene
Concen: 15.733 ng/ul
RT: 11.129 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:128 Resp: 150590
Ion Ratio Lower Upper
128 100
129 10.4 8.2 12.2
127 13.0 11.0 16.4



#32
4-Chloroaniline
Concen: 16.331 ng/ul
RT: 11.235 min Scan# 1318
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:127 Resp: 64920
Ion Ratio Lower Upper
127 100
129 31.5 24.3 36.5









#33
Hexachlorobutadiene
Concen: 14.842 ng/ul
RT: 11.399 min Scan# 1346
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:225 Resp: 29516
Ion Ratio Lower Upper
225 100
223 62.6 51.1 76.7
227 60.6 52.7 79.1



#34
Caprolactam
Concen: 16.493 ng/ul
RT: 11.993 min Scan# 1447
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:113 Resp: 17437
Ion Ratio Lower Upper
113 100
55 178.4 144.1 216.1
56 139.2 105.8 158.8









#35
 4-Chloro-3-methylphenol
 Concen: 17.139 ng/ul
 RT: 12.333 min Scan# 1505
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020449


Tgt Ion:	107	Resp:	55294
Ion	Ratio	Lower	Upper
107	100		
144	22.6	16.8	25.2
142	71.6	55.4	83.2



#36
 2-Methylnaphthalene
 Concen: 15.869 ng/ul
 RT: 12.721 min Scan# 1571
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32

Tgt Ion:	142	Resp:	103265
Ion	Ratio	Lower	Upper
142	100		
141	90.3	72.0	108.0









#37
1-Methylnaphthalene
Concen: 15.660 ng/ul
RT: 12.939 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:142 Resp: 102592
Ion Ratio Lower Upper
142 100
141 95.8 78.9 118.3
116 4.0 3.8 5.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 15.117 ng/ul
RT: 13.080 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:216 Resp: 55288
Ion Ratio Lower Upper
216 100
214 81.0 65.0 97.4
179 22.4 16.5 24.7
108 16.7 12.2 18.4









#40
Hexachlorocyclopentadiene
Concen: 10.028 ng/ul
RT: 13.050 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:237 Resp: 24321
Ion Ratio Lower Upper
237 100
235 70.7 45.6 68.4#
272 11.2 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 15.958 ng/ul
RT: 13.315 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:196 Resp: 36784
Ion Ratio Lower Upper
196 100
198 102.8 75.1 112.7
200 29.5 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 15.318 ng/ul
RT: 13.391 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:196 Resp: 38626
Ion Ratio Lower Upper
196 100
198 95.7 78.3 117.5
200 30.6 24.6 37.0



#43
1,1'-Biphenyl
Concen: 16.382 ng/ul
RT: 13.708 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:154 Resp: 139716
Ion Ratio Lower Upper
154 100
153 41.0 33.4 50.2
76 16.6 12.4 18.6









#44
2-Chloronaphthalene
Concen: 15.808 ng/ul
RT: 13.761 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:162 Resp: 105620
Ion Ratio Lower Upper
162 100
164 33.2 24.5 36.7
127 42.2 32.6 48.8



#45
2-Nitroaniline
Concen: 19.386 ng/ul
RT: 13.961 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion: 65 Resp: 42405
Ion Ratio Lower Upper
65 100
92 65.7 54.8 82.2
138 87.7 76.1 114.1









#47
Dimethylphthalate
Concen: 16.359 ng/ul
RT: 14.313 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:163 Resp: 142753
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 9.5 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 17.788 ng/ul
RT: 14.443 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:165 Resp: 29199
Ion Ratio Lower Upper
165 100
89 70.6 62.3 93.5
121 23.8 20.8 31.2









#50
Acenaphthylene
Concen: 16.468 ng/ul
RT: 14.601 min Scan# 1891
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:152 Resp: 181942
Ion Ratio Lower Upper
152 100
151 21.3 17.4 26.2
153 13.2 10.3 15.5



#51
3-Nitroaniline
Concen: 17.797 ng/ul
RT: 14.778 min Scan# 1921
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:138 Resp: 31777
Ion Ratio Lower Upper
138 100
108 11.1 8.6 12.8
92 126.6 97.0 145.4









#52
Acenaphthene
Concen: 16.066 ng/ul
RT: 14.942 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:153 Resp: 117313
Ion Ratio Lower Upper
153 100
152 47.7 40.2 60.2
154 88.2 68.2 102.2



#53
2,4-Dinitrophenol
Concen: 13.519 ng/ul
RT: 14.989 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:184 Resp: 13832
Ion Ratio Lower Upper
184 100
63 74.1 62.1 93.1
154 61.0 68.3 102.5#









#55
4-Nitrophenol
Concen: 13.870 ng/ul
RT: 15.077 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:	109	Resp:	22911
Ion	Ratio	Lower	Upper
109	100		
139	104.1	81.4	122.0
65	137.1	97.8	146.6



#56
Dibenzofuran
Concen: 15.736 ng/ul
RT: 15.271 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:	168	Resp:	167120
Ion	Ratio	Lower	Upper
168	100		
139	40.8	30.2	45.4









#57
2,4-Dinitrotoluene
Concen: 17.456 ng/ul
RT: 15.230 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:165 Resp: 41607
Ion Ratio Lower Upper
165 100
63 55.2 43.4 65.2
182 2.1 1.7 2.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 13.994 ng/ul
RT: 15.494 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:232 Resp: 30345
Ion Ratio Lower Upper
232 100
131 55.4 34.3 51.5#
130 2.9 2.8 4.2
166 30.9 22.4 33.6









#59
Diethylphthalate
Concen: 16.589 ng/ul
RT: 15.671 min Scan# 2073
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:149 Resp: 151048
Ion Ratio Lower Upper
149 100
177 20.8 18.2 27.4
150 12.5 10.2 15.2



#61
Fluorene
Concen: 15.750 ng/ul
RT: 15.923 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:166 Resp: 133074
Ion Ratio Lower Upper
166 100
165 102.5 78.6 118.0
167 14.0 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 15.151 ng/ul
RT: 15.906 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:204 Resp: 69327
Ion Ratio Lower Upper
204 100
206 31.9 24.8 37.2
141 58.5 44.1 66.1



#63
4-Nitroaniline
Concen: 17.736 ng/ul
RT: 15.935 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:138 Resp: 31981
Ion Ratio Lower Upper
138 100
92 60.1 47.0 70.4
108 97.4 77.2 115.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 16.846 ng/ul
 RT: 15.994 min Scan# 2128
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020449


Tgt Ion:198 Resp: 23013
 Ion Ratio Lower Upper
 198 100
 182 3.0 3.0 4.6#
 77 29.8 25.8 38.8



#67
 N-Nitrosodiphenylamine
 Concen: 16.718 ng/ul
 RT: 16.117 min Scan# 2149
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32

Tgt Ion:169 Resp: 117125
 Ion Ratio Lower Upper
 169 100
 168 70.5 58.3 87.5
 167 38.6 31.4 47.2









#68
4-Bromophenyl-phenylether
Concen: 16.085 ng/ul
RT: 16.805 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:248 Resp: 41496
Ion Ratio Lower Upper
248 100
250 97.9 73.3 109.9
141 77.6 60.2 90.2



#69
Hexachlorobenzene
Concen: 15.540 ng/ul
RT: 16.922 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:284 Resp: 42215
Ion Ratio Lower Upper
284 100
142 44.0 33.3 49.9
249 36.2 28.3 42.5









#70
Atrazine
Concen: 16.696 ng/ul
RT: 17.063 min Scan# 2310
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:200 Resp: 49803
Ion Ratio Lower Upper
200 100
173 22.8 19.9 29.9
215 44.2 40.4 60.6



#71
Pentachlorophenol
Concen: 9.681 ng/ul
RT: 17.269 min Scan# 2345
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:266 Resp: 17222
Ion Ratio Lower Upper
266 100
264 62.8 51.0 76.4
268 66.5 51.1 76.7









#72
Phenanthrene
Concen: 15.978 ng/ul
RT: 17.668 min Scan# 2413
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:178 Resp: 220995
Ion Ratio Lower Upper
178 100
179 15.9 12.7 19.1
176 19.6 16.2 24.2



#74
Anthracene
Concen: 16.124 ng/ul
RT: 17.756 min Scan# 2428
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:178 Resp: 220867
Ion Ratio Lower Upper
178 100
179 15.5 12.7 19.1
176 18.5 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 16.190 ng/uL
RT: 13.679 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:216 Resp: 59142
Ion Ratio Lower Upper
216 100
214 79.8 61.0 91.6
218 47.3 37.1 55.7



#76
Pentachlorobenzene
Concen: 16.092 ng/uL
RT: 15.189 min Scan# 1991
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:250 Resp: 53705
Ion Ratio Lower Upper
250 100
248 64.1 50.2 75.2
252 65.1 51.9 77.9









#77
 Carbazole
 Concen: 16.905 ng/ul
 RT: 18.027 min Scan# 2474
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32




Instrument :
 BNA_G
 ClientSampleId :
 SST020449


Tgt Ion:167	Resp:	207267
Ion	Ratio	Lower Upper
167	100	
166	22.1	17.6 26.4
139	13.1	10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 17.751 ng/ul
 RT: 18.561 min Scan# 2565
 Delta R.T. 0.000 min
 Lab File: BG050732.D
 Acq: 26 Oct 2021 15:32

Tgt Ion:149	Resp:	257903
Ion	Ratio	Lower Upper
149	100	
150	9.2	7.5 11.3
104	6.0	4.9 7.3









#80
Fluoranthene
Concen: 17.722 ng/u1
RT: 19.666 min Scan# 2753
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:202 Resp: 264121
Ion Ratio Lower Upper
202 100
101 10.3 7.7 11.5
100 8.8 6.6 10.0



#82
Pyrene
Concen: 17.621 ng/u1
RT: 20.030 min Scan# 2815
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:202 Resp: 258970
Ion Ratio Lower Upper
202 100
101 13.3 9.7 14.5
100 11.3 7.6 11.4









#83
Butylbenzylphthalate
Concen: 19.913 ng/ul
RT: 20.894 min Scan# 2962
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:149 Resp: 107873
Ion Ratio Lower Upper
149 100
91 83.0 65.8 98.6
206 19.4 17.2 25.8



#84
3,3'-Dichlorobenzidine
Concen: 18.443 ng/ul
RT: 21.816 min Scan# 3119
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:252 Resp: 84844
Ion Ratio Lower Upper
252 100
254 63.1 52.2 78.2
126 10.5 8.3 12.5





#85

Benzo(a)anthracene

Concen: 16.800 ng/ul

RT: 21.910 min Scan# 3135

Delta R.T. 0.000 min

Lab File: BG050732.D

Acq: 26 Oct 2021 15:32





Tgt Ion:228 Resp: 226047

Ion Ratio Lower Upper

228 100

229 19.1 15.9 23.9

226 26.9 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.778 ng/ul

RT: 21.781 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BG050732.D

Acq: 26 Oct 2021 15:32




Tgt Ion:149 Resp: 153961


Ion Ratio Lower Upper

149 100

167 26.9 21.3 31.9

279 3.7 2.8 4.2









#87
Chrysene
Concen: 16.861 ng/ul
RT: 21.981 min Scan# 3147
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:228 Resp: 217415
Ion Ratio Lower Upper
228 100
226 30.2 25.4 38.0
229 19.8 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 20.376 ng/ul
RT: 23.062 min Scan# 3331
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:149 Resp: 262999









#90
Benzo(b)fluoranthene
Concen: 16.739 ng/ul
RT: 24.266 min Scan# 3536
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:252 Resp: 229031
Ion Ratio Lower Upper
252 100
253 22.8 17.8 26.8
125 8.9 6.7 10.1



#91
Benzo(k)fluoranthene
Concen: 16.284 ng/ul
RT: 24.337 min Scan# 3548
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:252 Resp: 207319
Ion Ratio Lower Upper
252 100
253 20.1 17.1 25.7
125 9.7 7.4 11.0









#93
Benzo(a)pyrene
Concen: 16.562 ng/u1
RT: 25.201 min Scan# 3695
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32




Instrument :
BNA_G
ClientSampleId :
SSTD020449


Tgt Ion:252	Resp:	216215
Ion Ratio	Lower	Upper
252	100	
253	21.9	17.8 26.6
125	9.6	8.3 12.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 16.184 ng/u1
RT: 29.302 min Scan# 4393
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:276	Resp:	237207
Ion Ratio	Lower	Upper
276	100	
138	18.2	12.4 18.6
277	23.3	18.5 27.7









#95
Dibenzo(a,h)anthracene
Concen: 16.451 ng/ul
RT: 29.366 min Scan# 4404
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Instrument :
BNA_G
ClientSampleId :
SSTD020449

Tgt Ion:278 Resp: 205190
Ion Ratio Lower Upper
278 100
139 14.5 11.3 16.9
279 23.2 18.3 27.5



#96
Benzo(g,h,i)perylene
Concen: 16.126 ng/ul
RT: 30.541 min Scan# 4604
Delta R.T. 0.000 min
Lab File: BG050732.D
Acq: 26 Oct 2021 15:32

Tgt Ion:276 Resp: 199253
Ion Ratio Lower Upper
276 100
138 18.4 14.1 21.1
277 22.5 17.9 26.9

