

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102721\
 Data File : BG050734.D
 Acq On : 26 Oct 2021 16:55
 Operator : CG/JU
 Sample : SST08051
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST080402

Quant Time: Oct 26 17:34:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 26 15:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.252	152	35653	20.000	ng/u1	0.00
20) Naphthalene-d8	11.078	136	150676	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.880	164	95784	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.623	188	210046	20.000	ng/u1	0.00
79) Chrysene-d12	21.936	240	170136	20.000	ng/u1	0.00
88) Perylene-d12	25.361	264	168585	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.604	96	31331	32.286	ng/uL	0.00
4) Pyridine-d5	4.028	84	249697	86.350	ng/u1	0.00
7) Phenol-d5	7.394	99	291314	93.520	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.565	67	184086	93.901	ng/u1	0.00
11) 2-Chlorophenol-d4	7.782	132	207213	94.938	ng/u1	0.00
15) 4-Methylphenol-d8	8.957	113	222208	94.246	ng/u1	0.00
21) Nitrobenzene-d5	9.427	128	104812	99.435	ng/u1	0.00
24) 2-Nitrophenol-d4	10.156	143	123234	108.934	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.696	165	205709	92.086	ng/u1	0.00
31) 4-Chloroaniline-d4	11.213	131	287926	90.223	ng/u1	0.00
46) Dimethylphthalate-d6	14.274	166	604610	92.233	ng/u1	0.00
49) Acenaphthylene-d8	14.574	160	744852	90.479	ng/u1	0.00
54) 4-Nitrophenol-d4	15.073	143	111373	88.212	ng/u1	0.01
60) Fluorene-d10	15.867	176	519641	87.986	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.990	200	113299	108.145	ng/u1	0.00
73) Anthracene-d10	17.729	188	767396	86.481	ng/u1	0.00
81) Pyrene-d10	20.003	212	835272	92.329	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.138	264	775293	91.458	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.646	88	37541	31.975	ng/uL	92
5) Pyridine	4.051	79	256016	87.404	ng/u1	97
6) Benzaldehyde	7.382	77	112143	54.588	ng/u1	96
8) Phenol	7.424	94	298145	93.380	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.659	93	224638	93.800	ng/u1	96
12) 2-Chlorophenol	7.811	128	205518	92.035	ng/u1	90
13) 2-Methylphenol	8.687	108	220951	93.904	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.775	45	351251	97.755	ng/u1	97
16) Acetophenone	9.086	105	336666	89.871	ng/u1	98
17) N-Nitroso-di-n-propyla...	9.069	70	208176	92.469	ng/u1	95
18) 4-Methylphenol	9.028	108	229975	92.290	ng/u1	94
19) Hexachloroethane	9.345	117	88427	92.139	ng/u1	99
22) Nitrobenzene	9.468	77	299701	94.869	ng/u1	99
23) Isophorone	9.997	82	565155	93.088	ng/u1	99
25) 2-Nitrophenol	10.185	139	123682	106.374	ng/u1	98
26) 2,4-Dimethylphenol	10.232	107	260949	90.414	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.467	93	304863	93.009	ng/u1	97
29) 2,4-Dichlorophenol	10.720	162	197458	89.281	ng/u1	96
30) Naphthalene	11.131	128	668030	86.356	ng/u1	97
32) 4-Chloroaniline	11.237	127	288826	89.897	ng/u1	96
33) Hexachlorobutadiene	11.401	225	134367	83.601	ng/u1	99
34) Caprolactam	12.018	113	42946	50.261	ng/u1	81
35) 4-Chloro-3-methylphenol	12.341	107	244735	93.858	ng/u1	99
36) 2-Methylnaphthalene	12.717	142	451492	85.845	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102721\
 Data File : BG050734.D
 Acq On : 26 Oct 2021 16:55
 Operator : CG/JU
 Sample : SSTD08051
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD080402


Quant Time: Oct 26 17:34:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 26 15:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.935	142	447929	84.600	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.082	216	241361	87.317	ng/ul#	98
40) Hexachlorocyclopentadiene	13.052	237	145475	79.359	ng/ul	92
41) 2,4,6-Trichlorophenol	13.317	196	167481	96.130	ng/ul	95
42) 2,4,5-Trichlorophenol	13.393	196	176561	92.643	ng/ul	97
43) 1,1'-Biphenyl	13.716	154	580485	90.054	ng/ul	99
44) 2-Chloronaphthalene	13.763	162	457491	90.593	ng/ul	97
45) 2-Nitroaniline	13.963	65	193413	116.985	ng/ul	92
47) Dimethylphthalate	14.321	163	594533	90.142	ng/ul	99
48) 2,6-Dinitrotoluene	14.451	165	134211	108.178	ng/ul	95
50) Acenaphthylene	14.609	152	746531	89.400	ng/ul	99
51) 3-Nitroaniline	14.786	138	112779	83.568	ng/ul	88
52) Acenaphthene	14.944	153	502476	91.044	ng/ul	98
53) 2,4-Dinitrophenol	14.991	184	77849	100.669	ng/ul#	87
55) 4-Nitrophenol	15.091	109	109144	87.422	ng/ul	99
56) Dibenzofuran	15.279	168	688411	85.764	ng/ul	95
57) 2,4-Dinitrotoluene	15.238	165	185039	102.712	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.496	232	139945	85.390	ng/ul#	89
59) Diethylphthalate	15.679	149	630421	91.604	ng/ul	100
61) Fluorene	15.925	166	541005	84.718	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.908	204	291197	84.199	ng/ul	93
63) 4-Nitroaniline	15.949	138	99374	72.916	ng/ul	96
66) 4,6-Dinitro-2-methylph...	16.002	198	111212	107.389	ng/ul	98
67) N-Nitrosodiphenylamine	16.119	169	489392	92.150	ng/ul	98
68) 4-Bromophenyl-phenylether	16.801	248	182476	93.311	ng/ul	92
69) Hexachlorobenzene	16.924	284	186034	90.340	ng/ul	99
70) Atrazine	17.065	200	207887	91.936	ng/ul	97
71) Pentachlorophenol	17.271	266	103133	76.474	ng/ul	99
72) Phenanthrene	17.670	178	906738	86.481	ng/ul	98
74) Anthracene	17.764	178	883557	85.090	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.681	216	263398	95.117	ng/ul	97
76) Pentachlorobenzene	15.191	250	235948	93.265	ng/ul	98
77) Carbazole	18.029	167	829559	89.255	ng/ul	97
78) Di-n-butylphthalate	18.563	149	1032262	93.725	ng/ul	98
80) Fluoranthene	19.668	202	1051430	92.548	ng/ul	96
82) Pyrene	20.032	202	997905	89.074	ng/ul#	94
83) Butylbenzylphthalate	20.896	149	448927	108.714	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.818	252	318506	90.828	ng/ul	99
85) Benzo(a)anthracene	21.912	228	934392	91.100	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.783	149	635176	107.042	ng/ul	100
87) Chrysene	21.983	228	894583	91.011	ng/ul	99
89) Di-n-octyl phthalate	23.064	149	1068809	110.334	ng/ul	100
90) Benzo(b)fluoranthene	24.269	252	959675	93.458	ng/ul	98
91) Benzo(k)fluoranthene	24.345	252	862549	90.271	ng/ul	97
93) Benzo(a)pyrene	25.214	252	900074	91.867	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.321	276	543874	49.443	ng/ul#	94
95) Dibenzo(a,h)anthracene	29.398	278	873221	93.283	ng/ul	96
96) Benzo(g,h,i)perylene	30.579	276	860234	92.769	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 26 17:34:50 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 26 15:37:22 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 31.975 ng/uL
RT: 3.646 min Scan# 27
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:	88	Resp:	37541
Ion	Ratio	Lower	Upper
88	100		
43	33.3	23.8	35.8
58	73.9	53.7	80.5



#5
Pyridine
Concen: 87.404 ng/uL
RT: 4.051 min Scan# 96
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:	79	Resp:	256016
Ion	Ratio	Lower	Upper
79	100		
52	80.0	61.8	92.8
51	36.1	30.6	45.8









#6
Benzaldehyde
Concen: 54.588 ng/ul
RT: 7.382 min Scan# 663
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion: 77 Resp: 112143
Ion Ratio Lower Upper
77 100
105 84.7 64.2 96.4
106 79.6 62.1 93.1



#8
Phenol
Concen: 93.380 ng/ul
RT: 7.424 min Scan# 670
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion: 94 Resp: 298145
Ion Ratio Lower Upper
94 100
65 33.7 25.3 37.9
66 43.0 34.9 52.3









#10

Bis(2-Chloroethyl)ether
Concen: 93.800 ng/ul
RT: 7.659 min Scan# 710
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Instrument :
BNA_G
ClientSampleId :
SSTD080402




Tgt Ion:	93	Resp:	224638
Ion	Ratio	Lower	Upper
93	100		
63	88.5	66.9	100.3
95	32.9	26.1	39.1




#12

2-Chlorophenol
Concen: 92.035 ng/ul
RT: 7.811 min Scan# 736
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:	128	Resp:	205518
Ion	Ratio	Lower	Upper
128	100		
64	61.3	43.2	64.8
130	33.5	22.4	33.6









#13
2-Methylphenol
Concen: 93.904 ng/ul
RT: 8.687 min Scan# 885
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:108 Resp: 220951
Ion Ratio Lower Upper
108 100
107 91.4 75.5 113.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 97.755 ng/ul
RT: 8.775 min Scan# 900
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion: 45 Resp: 351251
Ion Ratio Lower Upper
45 100
77 13.0 11.4 17.2
79 10.1 8.6 13.0









#16
Acetophenone
Concen: 89.871 ng/ul
RT: 9.086 min Scan# 953
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:105 Resp: 336666
Ion Ratio Lower Upper
105 100
77 87.1 72.1 108.1
51 30.6 24.6 36.8



#17
N-Nitroso-di-n-propylamine
Concen: 92.469 ng/ul
RT: 9.069 min Scan# 950
Delta R.T. 0.020 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion: 70 Resp: 208176
Ion Ratio Lower Upper
70 100
42 59.2 50.4 75.6
101 8.5 7.8 11.6
130 16.4 15.9 23.9









#18
4-Methylphenol
Concen: 92.290 ng/ul
RT: 9.028 min Scan# 943
Delta R.T. 0.014 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:108 Resp: 229975
Ion Ratio Lower Upper
108 100
107 110.6 93.8 140.8



#19
Hexachloroethane
Concen: 92.139 ng/ul
RT: 9.345 min Scan# 997
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:117 Resp: 88427
Ion Ratio Lower Upper
117 100
201 103.2 83.0 124.4
199 67.8 53.8 80.6









#22
Nitrobenzene
Concen: 94.869 ng/ul
RT: 9.468 min Scan# 1018
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion: 77 Resp: 299701
Ion Ratio Lower Upper
77 100
123 36.6 29.9 44.9
65 13.9 11.0 16.4



#23
Isophorone
Concen: 93.088 ng/ul
RT: 9.997 min Scan# 1108
Delta R.T. 0.014 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion: 82 Resp: 565155
Ion Ratio Lower Upper
82 100
95 7.4 5.5 8.3
138 14.7 11.6 17.4









#25
2-Nitrophenol
Concen: 106.374 ng/ul
RT: 10.185 min Scan# 1140
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampled :
SSTD080402


Tgt Ion:139 Resp: 123682
Ion Ratio Lower Upper
139 100
65 63.6 49.1 73.7
109 35.1 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 90.414 ng/ul
RT: 10.232 min Scan# 1148
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:107 Resp: 260949
Ion Ratio Lower Upper
107 100
121 47.6 39.0 58.6
122 78.6 64.6 97.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 93.009 ng/ul
 RT: 10.467 min Scan# 1188
 Delta R.T. 0.002 min
 Lab File: BG050734.D
 Acq: 26 Oct 2021 16:55




Instrument :
 BNA_G
 ClientSampleId :
 SSTD080402


Tgt Ion: 93 Resp: 304863
 Ion Ratio Lower Upper
 93 100
 95 33.0 24.6 37.0
 123 9.9 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 89.281 ng/ul
 RT: 10.720 min Scan# 1231
 Delta R.T. 0.002 min
 Lab File: BG050734.D
 Acq: 26 Oct 2021 16:55

Tgt Ion: 162 Resp: 197458
 Ion Ratio Lower Upper
 162 100
 164 65.1 49.4 74.0
 98 38.6 29.0 43.4









#30
Naphthalene
Concen: 86.356 ng/ul
RT: 11.131 min Scan# 1301
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:128 Resp: 668030
Ion Ratio Lower Upper
128 100
129 11.8 8.2 12.2
127 14.2 11.0 16.4



#32
4-Chloroaniline
Concen: 89.897 ng/ul
RT: 11.237 min Scan# 1319
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:127 Resp: 288826
Ion Ratio Lower Upper
127 100
129 32.5 24.3 36.5









#33
Hexachlorobutadiene
Concen: 83.601 ng/ul
RT: 11.401 min Scan# 1347
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampled :
SSTD080402


Tgt Ion:225 Resp: 134367
Ion Ratio Lower Upper
225 100
223 63.8 51.1 76.7
227 64.3 52.7 79.1



#34
Caprolactam
Concen: 50.261 ng/ul
RT: 12.018 min Scan# 1452
Delta R.T. 0.026 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:113 Resp: 42946
Ion Ratio Lower Upper
113 100
55 210.1 144.1 216.1
56 150.8 105.8 158.8









#35
4-Chloro-3-methylphenol
Concen: 93.858 ng/ul
RT: 12.341 min Scan# 1507
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:107 Resp: 244735
Ion Ratio Lower Upper
107 100
144 22.7 16.8 25.2
142 69.6 55.4 83.2



#36
2-Methylnaphthalene
Concen: 85.845 ng/ul
RT: 12.717 min Scan# 1571
Delta R.T. -0.004 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:142 Resp: 451492
Ion Ratio Lower Upper
142 100
141 92.5 72.0 108.0









#37
1-Methylnaphthalene
Concen: 84.600 ng/ul
RT: 12.935 min Scan# 1608
Delta R.T. -0.004 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:142 Resp: 447929
Ion Ratio Lower Upper
142 100
141 95.4 78.9 118.3
116 5.0 3.8 5.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 87.317 ng/ul
RT: 13.082 min Scan# 1633
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:216 Resp: 241361
Ion Ratio Lower Upper
216 100
214 81.6 65.0 97.4
179 20.6 16.5 24.7
108 19.0 12.2 18.4#









#40

Hexachlorocyclopentadiene
Concen: 79.359 ng/ul
RT: 13.052 min Scan# 1628
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Instrument :
BNA_G
ClientSampleId :
SSTD080402




Tgt Ion:237 Resp: 145475
Ion Ratio Lower Upper
237 100
235 64.1 45.6 68.4
272 11.7 9.5 14.3




#41

2,4,6-Trichlorophenol
Concen: 96.130 ng/ul
RT: 13.317 min Scan# 1673
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:196 Resp: 167481
Ion Ratio Lower Upper
196 100
198 98.5 75.1 112.7
200 31.9 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 92.643 ng/ul
RT: 13.393 min Scan# 1686
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:196 Resp: 176561
Ion Ratio Lower Upper
196 100
198 94.3 78.3 117.5
200 30.8 24.6 37.0



#43
1,1'-Biphenyl
Concen: 90.054 ng/ul
RT: 13.716 min Scan# 1741
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:154 Resp: 580485
Ion Ratio Lower Upper
154 100
153 42.4 33.4 50.2
76 15.7 12.4 18.6









#44
2-Chloronaphthalene
Concen: 90.593 ng/ul
RT: 13.763 min Scan# 1749
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:162 Resp: 457491
Ion Ratio Lower Upper
162 100
164 33.5 24.5 36.7
127 41.6 32.6 48.8



#45
2-Nitroaniline
Concen: 116.985 ng/ul
RT: 13.963 min Scan# 1783
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion: 65 Resp: 193413
Ion Ratio Lower Upper
65 100
92 65.0 54.8 82.2
138 84.4 76.1 114.1









#47
Dimethylphthalate
Concen: 90.142 ng/ul
RT: 14.321 min Scan# 1844
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:163 Resp: 594533
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 10.5 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 108.178 ng/ul
RT: 14.451 min Scan# 1866
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55


Tgt Ion:165 Resp: 134211
Ion Ratio Lower Upper
165 100
89 72.8 62.3 93.5
121 24.3 20.8 31.2








#50
Acenaphthylene
Concen: 89.400 ng/ul
RT: 14.609 min Scan# 1893
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Instrument :
BNA_G
ClientSampleId :
SSTD080402







Tgt Ion:152 Resp: 746531
Ion Ratio Lower Upper
152 100
151 21.9 17.4 26.2
153 13.8 10.3 15.5



#51
3-Nitroaniline
Concen: 83.568 ng/ul
RT: 14.786 min Scan# 1923
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:138 Resp: 112779
Ion Ratio Lower Upper
138 100
108 10.3 8.6 12.8
92 135.2 97.0 145.4









#52
Acenaphthene
Concen: 91.044 ng/ul
RT: 14.944 min Scan# 1950
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:153 Resp: 502476
Ion Ratio Lower Upper
153 100
152 47.4 40.2 60.2
154 83.9 68.2 102.2



#53
2,4-Dinitrophenol
Concen: 100.669 ng/ul
RT: 14.991 min Scan# 1958
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:184 Resp: 77849
Ion Ratio Lower Upper
184 100
63 81.1 62.1 93.1
154 66.0 68.3 102.5#









#55
4-Nitrophenol
Concen: 87.422 ng/ul
RT: 15.091 min Scan# 1975
Delta R.T. 0.014 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:109 Resp: 109144
Ion Ratio Lower Upper
109 100
139 101.1 81.4 122.0
65 124.3 97.8 146.6



#56
Dibenzofuran
Concen: 85.764 ng/ul
RT: 15.279 min Scan# 2007
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:168 Resp: 688411
Ion Ratio Lower Upper
168 100
139 40.8 30.2 45.4









#57
2,4-Dinitrotoluene
Concen: 102.712 ng/ul
RT: 15.238 min Scan# 2000
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:165 Resp: 185039
Ion Ratio Lower Upper
165 100
63 55.4 43.4 65.2
182 2.6 1.7 2.5#



#58
2,3,4,6-Tetrachlorophenol
Concen: 85.390 ng/ul
RT: 15.496 min Scan# 2044
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:232 Resp: 139945
Ion Ratio Lower Upper
232 100
131 51.8 34.3 51.5#
130 2.9 2.8 4.2
166 32.0 22.4 33.6









#59
Diethylphthalate
Concen: 91.604 ng/ul
RT: 15.679 min Scan# 2075
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:149 Resp: 630421
Ion Ratio Lower Upper
149 100
177 22.9 18.2 27.4
150 12.7 10.2 15.2



#61
Fluorene
Concen: 84.718 ng/ul
RT: 15.925 min Scan# 2117
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:166 Resp: 541005
Ion Ratio Lower Upper
166 100
165 102.2 78.6 118.0
167 14.3 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 84.199 ng/ul
RT: 15.908 min Scan# 2114
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:204 Resp: 291197
Ion Ratio Lower Upper
204 100
206 33.6 24.8 37.2
141 61.1 44.1 66.1



#63
4-Nitroaniline
Concen: 72.916 ng/ul
RT: 15.949 min Scan# 2121
Delta R.T. 0.014 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:138 Resp: 99374
Ion Ratio Lower Upper
138 100
92 58.1 47.0 70.4
108 102.8 77.2 115.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 107.389 ng/ul
 RT: 16.002 min Scan# 2130
 Delta R.T. 0.008 min
 Lab File: BG050734.D
 Acq: 26 Oct 2021 16:55




Instrument :
 BNA_G
 ClientSampleID :
 SSTD080402


Tgt Ion:198 Resp: 111212
 Ion Ratio Lower Upper
 198 100
 182 3.1 3.0 4.6
 77 31.0 25.8 38.8



#67
 N-Nitrosodiphenylamine
 Concen: 92.150 ng/ul
 RT: 16.119 min Scan# 2150
 Delta R.T. 0.002 min
 Lab File: BG050734.D
 Acq: 26 Oct 2021 16:55

Tgt Ion:169 Resp: 489392
 Ion Ratio Lower Upper
 169 100
 168 71.1 58.3 87.5
 167 39.5 31.4 47.2









#68
4-Bromophenyl-phenylether
Concen: 93.311 ng/ul
RT: 16.801 min Scan# 2266
Delta R.T. -0.004 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:248 Resp: 182476
Ion Ratio Lower Upper
248 100
250 97.3 73.3 109.9
141 83.7 60.2 90.2



#69
Hexachlorobenzene
Concen: 90.340 ng/ul
RT: 16.924 min Scan# 2287
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:284 Resp: 186034
Ion Ratio Lower Upper
284 100
142 42.3 33.3 49.9
249 34.3 28.3 42.5





#70

Atrazine

Concen: 91.936 ng/ul

RT: 17.065 min Scan# 2311

Delta R.T. 0.002 min

Lab File: BG050734.D

Acq: 26 Oct 2021 16:55

Tgt Ion:200 Resp: 207887

Ion Ratio Lower Upper

200 100

173 25.6 19.9 29.9





215 47.6 40.4 60.6

Instrument :

BNA_G

ClientSampleId :

SSTD080402



#71

Pentachlorophenol

Concen: 76.474 ng/ul

RT: 17.271 min Scan# 2346

Delta R.T. 0.002 min

Lab File: BG050734.D

Acq: 26 Oct 2021 16:55




Tgt Ion:266 Resp: 103133


Ion Ratio Lower Upper

266 100

264 64.3 51.0 76.4


268 63.2 51.1 76.7








#72
Phenanthrene
Concen: 86.481 ng/u1
RT: 17.670 min Scan# 2414
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Instrument :
BNA_G
ClientSampleId :
SSTD080402







Tgt Ion:178 Resp: 906738
Ion Ratio Lower Upper
178 100
179 16.6 12.7 19.1
176 20.9 16.2 24.2



#74
Anthracene
Concen: 85.090 ng/u1
RT: 17.764 min Scan# 2430
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:178 Resp: 883557
Ion Ratio Lower Upper
178 100
179 17.1 12.7 19.1
176 20.5 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 95.117 ng/uL
RT: 13.681 min Scan# 1735
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:216 Resp: 263398
Ion Ratio Lower Upper
216 100
214 79.2 61.0 91.6
218 48.7 37.1 55.7



#76
Pentachlorobenzene
Concen: 93.265 ng/uL
RT: 15.191 min Scan# 1992
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:250 Resp: 235948
Ion Ratio Lower Upper
250 100
248 64.8 50.2 75.2
252 64.5 51.9 77.9









#77
 Carbazole
 Concen: 89.255 ng/ul
 RT: 18.029 min Scan# 2475
 Delta R.T. 0.002 min
 Lab File: BG050734.D
 Acq: 26 Oct 2021 16:55




Instrument :
 BNA_G
 ClientSampleId :
 SSTD080402


Tgt Ion:167 Resp: 829559
 Ion Ratio Lower Upper
 167 100
 166 23.5 17.6 26.4
 139 14.2 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 93.725 ng/ul
 RT: 18.563 min Scan# 2566
 Delta R.T. 0.002 min
 Lab File: BG050734.D
 Acq: 26 Oct 2021 16:55

Tgt Ion:149 Resp: 1032262
 Ion Ratio Lower Upper
 149 100
 150 10.2 7.5 11.3
 104 6.8 4.9 7.3









#80
Fluoranthene
Concen: 92.548 ng/u1
RT: 19.668 min Scan# 2754
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:202 Resp: 1051430
Ion Ratio Lower Upper
202 100
101 11.4 7.7 11.5
100 9.4 6.6 10.0



#82
Pyrene
Concen: 89.074 ng/u1
RT: 20.032 min Scan# 2816
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:202 Resp: 997905
Ion Ratio Lower Upper
202 100
101 13.8 9.7 14.5
100 12.2 7.6 11.4#









#83
Butylbenzylphthalate
Concen: 108.714 ng/ul
RT: 20.896 min Scan# 2963
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:149 Resp: 448927
Ion Ratio Lower Upper
149 100
91 81.7 65.8 98.6
206 20.7 17.2 25.8



#84
3,3'-Dichlorobenzidine
Concen: 90.828 ng/ul
RT: 21.818 min Scan# 3120
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:252 Resp: 318506
Ion Ratio Lower Upper
252 100
254 64.8 52.2 78.2
126 10.9 8.3 12.5









#85
Benzo(a)anthracene
Concen: 91.100 ng/ul
RT: 21.912 min Scan# 3136
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:228 Resp: 934392
Ion Ratio Lower Upper
228 100
229 21.1 15.9 23.9
226 29.3 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 107.042 ng/ul
RT: 21.783 min Scan# 3114
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:149 Resp: 635176
Ion Ratio Lower Upper
149 100
167 26.5 21.3 31.9
279 4.0 2.8 4.2





#87

Chrysene

Concen: 91.011 ng/ul

RT: 21.983 min Scan# 3148

Delta R.T. 0.002 min

Lab File: BG050734.D

Acq: 26 Oct 2021 16:55





Tgt Ion:228 Resp: 894583

Ion Ratio Lower Upper

228 100

226 31.7 25.4 38.0

229 21.5 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 110.334 ng/ul




RT: 23.064 min Scan# 3332

Delta R.T. 0.002 min

Lab File: BG050734.D

Acq: 26 Oct 2021 16:55

Tgt Ion:149 Resp: 1068809




Instrument :

BNA_G

ClientSampleId :





SSTD080402



#90
Benzo(b)fluoranthene
Concen: 93.458 ng/ul
RT: 24.269 min Scan# 3537
Delta R.T. 0.002 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:252 Resp: 959675
Ion Ratio Lower Upper
252 100
253 22.9 17.8 26.8
125 9.8 6.7 10.1



#91
Benzo(k)fluoranthene
Concen: 90.271 ng/ul
RT: 24.345 min Scan# 3550
Delta R.T. 0.008 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:252 Resp: 862549
Ion Ratio Lower Upper
252 100
253 23.3 17.1 25.7
125 9.7 7.4 11.0









#93
Benzo(a)pyrene
Concen: 91.867 ng/ul
RT: 25.214 min Scan# 3698
Delta R.T. 0.014 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55




Instrument :
BNA_G
ClientSampleId :
SSTD080402


Tgt Ion:252 Resp: 900074
Ion Ratio Lower Upper
252 100
253 23.0 17.8 26.6
125 10.9 8.3 12.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 49.443 ng/ul
RT: 29.321 min Scan# 4397
Delta R.T. 0.020 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:276 Resp: 543874
Ion Ratio Lower Upper
276 100
138 19.0 12.4 18.6#
277 25.0 18.5 27.7









#95
Dibenzo(a,h)anthracene
Concen: 93.283 ng/ul
RT: 29.398 min Scan# 4410
Delta R.T. 0.031 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Instrument :
BNA_G
ClientSampleId :
SSTD080402

Tgt Ion:278 Resp: 873221
Ion Ratio Lower Upper
278 100
139 14.9 11.3 16.9
279 25.2 18.3 27.5



#96
Benzo(g,h,i)perylene
Concen: 92.769 ng/ul
RT: 30.579 min Scan# 4611
Delta R.T. 0.037 min
Lab File: BG050734.D
Acq: 26 Oct 2021 16:55

Tgt Ion:276 Resp: 860234
Ion Ratio Lower Upper
276 100
138 18.6 14.1 21.1
277 24.7 17.9 26.9

