



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102721\
 Data File : BG050728.D
 Acq On : 26 Oct 2021 12:44
 Operator : CG/JU
 Sample : DFTPP425
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP425

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102721.M
 Title : SVOA CALIBRATION
 Last Update : Tue Oct 26 18:31:25 2021



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.3	31549	PASS
68	69	0.00	2	0.5	150	PASS
69	198	0.00	100	45.3	32277	PASS
70	69	0.00	2	0.8	248	PASS
127	198	10	80	49.8	35475	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	71179	PASS
199	198	5	9	7.1	5026	PASS
275	198	10	60	27.1	19263	PASS
365	198	1	100	3.7	2622	PASS
441	198	0.01	100	9.1	6487	PASS
442	198	50	100	57.1	40632	PASS
443	442	15	24	19.2	7803	PASS

m/z	Abundance
40.00	414.0
44.00	534.0
55.00	163.0
73.00	264.0
207.00	373.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

m/z	Abundance
37.00	172.0
38.00	502.0
39.00	3392.0
39.90	424.0
41.00	242.0
43.10	193.0
44.00	648.0
49.10	438.0
50.00	10104.0
51.00	38928.0
52.00	2041.0
55.00	288.0
56.00	1130.0
56.90	2631.0
60.10	152.0
61.00	412.0
62.00	630.0
63.00	1911.0
64.00	238.0
65.00	850.0
69.00	39776.0
70.00	269.0
73.00	496.0
74.00	4238.0
75.00	6740.0
76.00	2687.0
77.00	46992.0
78.00	3011.0
79.00	3383.0
80.00	2448.0
81.00	3265.0
81.90	1053.0
83.00	818.0
85.00	640.0
86.00	993.0
87.00	565.0
88.00	209.0
91.00	647.0
92.00	889.0
93.00	5117.0
93.90	331.0
95.00	179.0
98.00	4507.0
99.00	2964.0
99.90	299.0
100.90	1684.0
103.00	553.0
103.90	1168.0
105.00	1013.0
107.00	13394.0
107.90	2017.0
109.00	554.0
110.00	20456.0
111.00	3384.0
112.00	433.0
113.00	192.0
116.00	918.0
117.00	11920.0
118.00	995.0
120.00	217.0
122.00	748.0
123.00	1164.0
123.90	668.0
124.90	428.0
127.00	41440.0
128.00	3307.0
129.00	16920.0
130.00	1577.0
131.00	319.0
132.00	210.0
134.00	391.0
134.90	1220.0
135.90	595.0
137.00	723.0
138.00	203.0
139.00	357.0
139.90	440.0
141.00	2574.0
142.00	812.0
142.90	439.0
144.00	174.0
146.00	386.0
147.00	1210.0
148.00	2588.0
149.00	559.0
150.90	315.0
153.00	640.0
153.90	673.0
155.00	1516.0
156.00	2018.0
157.00	350.0
158.00	369.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

159.10	411.0
160.00	768.0
161.00	1039.0
162.00	282.0
163.00	178.0
163.80	160.0
165.00	1159.0
166.00	1145.0
167.00	6168.0
168.00	2175.0
168.90	350.0
169.90	166.0
170.90	270.0
171.90	404.0
172.90	574.0
174.00	1161.0
175.00	1798.0
176.00	545.0
177.00	864.0
178.00	392.0
178.90	3600.0
180.00	2419.0
181.00	1031.0
183.90	321.0
185.00	1542.0
186.00	12237.0
187.00	3907.0
188.00	353.0
189.00	938.0
190.10	212.0
191.00	374.0
192.00	1319.0
193.00	949.0
194.00	268.0
195.00	254.0
196.00	2436.0
197.90	77840.0
198.90	5823.0
199.90	573.0
201.30	342.0
202.90	653.0
204.00	3666.0
205.00	5529.0
206.00	23048.0
207.00	3213.0
208.00	878.0
208.90	192.0
210.00	370.0
211.00	1053.0
214.90	316.0
215.90	535.0
216.90	6822.0
218.00	969.0
220.20	268.0
221.00	2802.0
222.00	889.0
222.90	1338.0
224.00	12605.0
225.00	3118.0
226.00	546.0
227.00	5698.0
228.00	811.0
228.90	1027.0
229.90	236.0
231.00	383.0
233.90	422.0
235.00	501.0
236.00	318.0
237.00	394.0
238.90	190.0
240.00	197.0
240.90	385.0
242.00	661.0
243.00	807.0
244.00	10030.0
245.00	1462.0
246.00	2065.0
246.90	368.0
249.00	324.0
250.90	163.0
253.10	234.0
255.00	48456.0
256.00	7466.0
257.00	590.0
258.00	3534.0
258.90	644.0
264.90	1361.0
266.00	274.0
269.80	165.0
271.00	161.0
272.00	153.0
273.00	1514.0
274.00	3811.0
275.00	20576.0
276.00	2815.0
277.00	2294.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

278.00	241.0
283.00	236.0
285.00	382.0
293.00	570.0
296.00	6127.0
297.00	854.0
303.10	461.0
304.00	268.0
314.00	249.0
315.00	664.0
316.00	385.0
321.10	204.0
323.00	1679.0
324.00	280.0
326.90	396.0
328.00	297.0
332.00	168.0
332.90	161.0
334.00	1224.0
334.90	241.0
341.00	235.0
346.00	480.0
352.10	467.0
353.00	297.0
354.00	438.0
365.00	2616.0
366.00	482.0
372.10	482.0
373.10	158.0
383.00	162.0
402.10	361.0
403.00	230.0
421.00	411.0
422.10	249.0
423.00	1929.0
424.00	444.0
441.10	5964.0
442.10	36400.0
443.10	7517.0
444.10	761.0

Instrument :

BNA_G

ClientSampleId :

DFTPP425

m/z	Abundance
38.00	541.0
39.00	3030.0
40.00	487.0
41.10	347.0
43.00	202.0
44.00	562.0
48.90	310.0
50.00	9564.0
51.00	35656.0
52.00	1655.0
55.00	247.0
56.00	888.0
57.00	2366.0
61.00	467.0
62.00	632.0
63.00	1333.0
64.00	350.0
65.00	772.0
67.00	181.0
69.00	36568.0
69.80	229.0
71.10	194.0
73.00	404.0
74.00	4475.0
75.00	6205.0
76.00	1915.0
77.00	42200.0
78.00	2634.0
79.00	2586.0
80.00	2433.0
81.00	3370.0
82.00	788.0
83.00	609.0
83.80	227.0
84.90	658.0
86.00	773.0
87.10	513.0
88.00	169.0
91.00	753.0
92.00	860.0
93.00	5098.0
94.00	323.0
96.00	221.0
96.90	231.0
98.00	4114.0
99.00	2979.0
100.00	241.0
100.90	1625.0
102.90	515.0
104.00	1013.0
105.00	1038.0
107.00	12684.0
108.00	1858.0
110.00	20320.0
111.00	3822.0
112.00	417.0
112.80	271.0
116.00	710.0
117.00	11413.0
117.90	687.0
122.00	883.0
123.00	1385.0
123.90	704.0
124.90	553.0
127.00	41504.0
128.00	3370.0
129.00	17056.0
130.00	1507.0
131.00	336.0
133.90	500.0
135.00	1686.0
136.00	665.0
137.10	738.0
138.00	173.0
139.00	438.0
139.90	495.0
140.90	2460.0
142.00	809.0
143.00	537.0
145.90	506.0
147.00	1347.0
148.00	2729.0
149.00	554.0
151.00	383.0
153.00	934.0
154.00	487.0
154.90	1322.0
156.00	1991.0
157.00	424.0
157.90	396.0
159.00	267.0
160.00	873.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

160.90	957.0
161.90	321.0
164.00	152.0
165.00	775.0
166.00	917.0
167.00	7148.0
168.00	1865.0
169.00	377.0
170.00	265.0
170.90	184.0
171.90	399.0
173.00	586.0
174.00	989.0
175.00	1855.0
175.90	474.0
176.90	947.0
178.10	365.0
178.90	3743.0
180.00	2388.0
180.90	1000.0
184.00	379.0
185.00	1693.0
186.00	12331.0
187.00	3662.0
187.90	353.0
188.90	994.0
191.00	395.0
192.00	1108.0
193.00	1354.0
193.90	280.0
196.00	2778.0
197.90	84800.0
198.90	5381.0
200.00	400.0
201.50	538.0
203.00	807.0
204.00	3738.0
205.00	6188.0
206.00	24752.0
207.00	3608.0
208.00	974.0
209.00	349.0
210.10	452.0
211.00	995.0
211.90	170.0
215.00	351.0
216.10	762.0
217.00	6733.0
217.90	915.0
220.00	176.0
221.00	3409.0
221.90	803.0
223.00	1640.0
224.00	13629.0
225.00	3203.0
226.10	376.0
227.00	6666.0
228.00	845.0
228.90	1193.0
229.90	222.0
231.00	482.0
233.90	434.0
235.00	589.0
235.90	347.0
236.90	500.0
240.00	319.0
241.00	375.0
242.10	863.0
243.10	704.0
244.00	9469.0
245.00	1227.0
246.00	2050.0
247.00	413.0
248.00	150.0
248.90	457.0
249.80	175.0
253.00	310.0
255.00	53912.0
256.00	7907.0
257.00	506.0
258.00	3639.0
259.00	520.0
265.00	1521.0
265.90	306.0
273.00	1768.0
274.00	3831.0
275.00	22144.0
276.00	3349.0
277.00	2464.0
277.90	266.0
282.90	179.0
285.00	433.0
286.00	160.0
292.90	325.0
296.00	6826.0
297.00	1096.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

303.00	809.0
304.00	216.0
314.00	356.0
315.00	969.0
316.10	499.0
321.00	296.0
323.00	1942.0
324.00	362.0
327.00	408.0
327.90	184.0
332.00	159.0
333.20	255.0
334.00	1425.0
335.00	240.0
341.00	293.0
346.00	386.0
352.00	532.0
353.10	602.0
354.00	573.0
365.00	2931.0
366.00	475.0
372.00	952.0
372.90	194.0
383.00	387.0
402.10	305.0
403.10	532.0
404.00	162.0
421.10	271.0
422.00	324.0
423.00	2570.0
424.00	540.0
441.10	7922.0
442.10	47184.0
443.10	9065.0
444.00	869.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

m/z	Abundance
38.00	381.0
39.00	1544.0
40.00	428.0
41.00	199.0
44.00	446.0
50.00	5478.0
51.00	20064.0
52.00	878.0
55.10	174.0
56.00	579.0
57.00	1594.0
60.90	169.0
62.00	335.0
63.00	1040.0
64.00	167.0
65.00	424.0
68.00	450.0
69.00	20488.0
70.00	245.0
73.00	297.0
74.00	2032.0
75.00	3597.0
76.00	1287.0
77.00	24848.0
78.00	1456.0
79.00	1560.0
80.00	1197.0
80.90	1875.0
82.00	464.0
83.00	442.0
84.80	313.0
86.00	731.0
87.00	355.0
90.90	440.0
91.90	514.0
93.00	2598.0
94.00	153.0
98.00	2424.0
99.00	1799.0
101.00	1167.0
102.90	362.0
104.00	639.0
105.00	591.0
105.90	228.0
107.00	7631.0
108.00	1367.0
109.90	12563.0
111.00	2146.0
111.90	367.0
116.00	381.0
117.00	6374.0
117.90	414.0
122.00	464.0
123.00	821.0
124.00	350.0
124.90	345.0
127.00	23480.0
128.00	1953.0
128.90	10250.0
129.90	849.0
130.90	179.0
134.00	289.0
134.90	965.0
136.00	387.0
137.00	611.0
137.90	161.0
138.90	407.0
140.00	332.0
140.90	1319.0
141.90	490.0
143.00	241.0
146.00	205.0
147.00	671.0
147.90	1538.0
148.90	302.0
151.00	187.0
152.90	512.0
153.90	313.0
155.00	791.0
156.00	1122.0
158.00	327.0
159.90	422.0
161.00	780.0
162.00	277.0
165.00	478.0
166.10	883.0
167.00	4981.0
168.00	1594.0
168.90	213.0
171.00	153.0
172.00	378.0
172.90	334.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

174.00	565.0
175.00	1180.0
176.00	184.0
177.00	574.0
178.00	186.0
178.20	183.0
178.90	2254.0
180.00	1456.0
181.00	676.0
182.00	158.0
184.00	345.0
184.90	1094.0
186.00	7960.0
187.00	2206.0
188.00	203.0
189.00	616.0
191.10	174.0
192.00	551.0
192.90	645.0
193.90	168.0
196.00	1434.0
197.90	50896.0
198.90	3873.0
200.00	241.0
201.30	258.0
202.90	669.0
204.00	2244.0
205.00	3694.0
206.00	16096.0
207.00	2442.0
208.00	667.0
209.10	166.0
210.00	251.0
211.00	656.0
211.90	187.0
215.00	277.0
216.00	488.0
216.90	3914.0
217.90	417.0
221.00	1887.0
221.90	644.0
223.00	985.0
224.00	8371.0
225.00	2223.0
227.00	3949.0
227.90	750.0
229.00	783.0
231.00	352.0
234.00	243.0
234.90	413.0
236.00	185.0
236.90	268.0
241.00	161.0
241.90	475.0
243.00	571.0
244.00	6175.0
244.90	880.0
246.00	1658.0
247.00	298.0
249.00	277.0
255.00	35328.0
256.00	5196.0
257.00	375.0
258.00	2302.0
259.00	285.0
265.00	999.0
266.00	155.0
273.00	1203.0
274.00	2647.0
275.00	15069.0
276.00	1827.0
277.00	1495.0
278.00	187.0
283.90	266.0
284.90	217.0
293.00	397.0
296.00	4858.0
297.00	664.0
302.90	630.0
304.00	163.0
314.00	311.0
315.00	568.0
316.00	301.0
321.00	180.0
323.00	1326.0
324.00	310.0
326.90	223.0
328.00	165.0
333.00	184.0
334.00	762.0
335.10	248.0
345.90	417.0
352.00	506.0
353.00	283.0
354.00	320.0
365.00	2319.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425

366.00	232.0
372.00	834.0
401.90	280.0
403.00	273.0
421.00	274.0
422.10	305.0
423.10	1994.0
424.10	493.0
441.10	5575.0
442.10	38312.0
443.10	6827.0
444.10	718.0

Instrument :
BNA_G
ClientSampleId :
DFTPP425