

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102723\
 Data File : BG059513.D
 Acq On : 27 Oct 2023 16:08
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV440

Quant Time: Oct 28 04:04:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.364	152	74859	20.000	ng/ul	# 0.00
20) Naphthalene-d8	11.213	136	340126	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.991	164	215958	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.741	188	481473	20.000	ng/ul	0.00
79) Chrysene-d12	22.077	240	397881	20.000	ng/ul	0.00
88) Perylene-d12	25.684	264	437787	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.657	96	13299	6.647	ng/uL	0.00
4) Pyridine-d5	4.092	84	110206	19.711	ng/ul	0.00
7) Phenol-d5	7.465	99	156408	20.332	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.682	67	76055	21.525	ng/ul	0.00
11) 2-Chlorophenol-d4	7.876	132	120081	19.241	ng/ul	0.00
15) 4-Methylphenol-d8	9.039	113	127813	20.943	ng/ul	0.00
21) Nitrobenzene-d5	9.562	128	56472	23.351	ng/ul	0.00
24) 2-Nitrophenol-d4	10.279	143	55537	22.369	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.802	165	119369	21.823	ng/ul	0.00
31) 4-Chloroaniline-d4	11.348	131	186784	20.743	ng/ul	0.00
46) Dimethylphthalate-d6	14.386	166	382658	19.558	ng/ul	0.00
49) Acenaphthylene-d8	14.691	160	450054	19.601	ng/ul	0.00
54) 4-Nitrophenol-d4	15.150	143	75289	22.440	ng/ul	0.00
60) Fluorene-d10	15.978	176	320974	19.196	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.078	200	49162	21.561	ng/ul	0.00
73) Anthracene-d10	17.841	188	532017	20.396	ng/ul	0.00
81) Pyrene-d10	20.109	212	582088	21.047	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.432	264	532313	20.649	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.693	88	13740	6.136	ng/uL#	80
5) Pyridine	4.110	79	88512	16.049	ng/ul	87
6) Benzaldehyde	7.506	77	85322	24.625	ng/ul	88
8) Phenol	7.500	94	138314	19.072	ng/ul	94
10) Bis(2-Chloroethyl)ether	7.776	93	97241	18.285	ng/ul	97
12) 2-Chlorophenol	7.905	128	117842	19.047	ng/ul	93
13) 2-Methylphenol	8.775	108	115982	19.184	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.881	45	82624	16.509	ng/ul#	92
16) Acetophenone	9.204	105	198200	20.602	ng/ul	92
17) N-Nitroso-di-n-propyla...	9.168	70	83082	18.328	ng/ul#	92
18) 4-Methylphenol	9.104	108	128133	19.920	ng/ul	96
19) Hexachloroethane	9.456	117	46336	18.856	ng/ul	95
22) Nitrobenzene	9.603	77	116443	22.307	ng/ul	95
23) Isophorone	10.114	82	245257	18.657	ng/ul#	90
25) 2-Nitrophenol	10.314	139	58516	21.915	ng/ul	91
26) 2,4-Dimethylphenol	10.332	107	123517	17.587	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.596	93	137439	18.777	ng/ul	96
29) 2,4-Dichlorophenol	10.831	162	108461	20.495	ng/ul	91
30) Naphthalene	11.266	128	389665	18.381	ng/ul#	96
32) 4-Chloroaniline	11.378	127	175013	20.461	ng/ul#	84
33) Hexachlorobutadiene	11.495	225	58104	18.302	ng/ul	92
34) Caprolactam	12.147	113	39225	20.261	ng/ul	89
35) 4-Chloro-3-methylphenol	12.429	107	134265	21.264	ng/ul	87
36) 2-Methylnaphthalene	12.841	142	271592	18.622	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102723\
 Data File : BG059513.D
 Acq On : 27 Oct 2023 16:08
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SICV440


Quant Time: Oct 28 04:04:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.058	142	261215	17.982	ng/ul	92
39) 1,2,4,5-Tetrachloroben...	13.176	216	120459	19.339	ng/ul	96
40) Hexachlorocyclopentadiene	13.134	237	59791	16.467	ng/ul	91
41) 2,4,6-Trichlorophenol	13.416	196	81207	18.828	ng/ul	89
42) 2,4,5-Trichlorophenol	13.487	196	90845	19.455	ng/ul	96
43) 1,1'-Biphenyl	13.828	154	376858	19.290	ng/ul	99
44) 2-Chloronaphthalene	13.881	162	265353	17.476	ng/ul	96
45) 2-Nitroaniline	14.086	65	63149	22.351	ng/ul	92
47) Dimethylphthalate	14.433	163	350724	18.144	ng/ul#	95
48) 2,6-Dinitrotoluene	14.574	165	61852	21.212	ng/ul#	77
50) Acenaphthylene	14.721	152	476501	18.012	ng/ul	97
51) 3-Nitroaniline	14.909	138	85121	23.420	ng/ul#	81
52) Acenaphthene	15.056	153	310343	17.996	ng/ul	95
53) 2,4-Dinitrophenol	15.103	184	31697	19.131	ng/ul#	84
55) 4-Nitrophenol	15.161	109	80235	21.699	ng/ul	91
56) Dibenzofuran	15.385	168	425597	19.036	ng/ul	100
57) 2,4-Dinitrotoluene	15.344	165	95911	23.383	ng/ul#	94
58) 2,3,4,6-Tetrachlorophenol	15.590	232	76862	20.972	ng/ul	91
59) Diethylphthalate	15.778	149	392308	18.605	ng/ul	99
61) Fluorene	16.031	166	359181	18.610	ng/ul#	97
62) 4-Chlorophenyl-phenyle...	16.019	204	151953	18.422	ng/ul	95
63) 4-Nitroaniline	16.066	138	100080	25.946	ng/ul#	91
66) 4,6-Dinitro-2-methylph...	16.096	198	46934	19.531	ng/ul#	91
67) N-Nitrosodiphenylamine	16.237	169	311470	18.933	ng/ul	99
68) 4-Bromophenyl-phenylether	16.918	248	87610	18.184	ng/ul	94
69) Hexachlorobenzene	17.006	284	101761	18.275	ng/ul	96
70) Atrazine	17.171	200	108318	21.248	ng/ul	96
71) Pentachlorophenol	17.435	266	157	0.049	ng/ul	92
72) Phenanthrene	17.782	178	577597	18.453	ng/ul	96
74) Anthracene	17.876	178	589167	18.642	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.787	216	112556	17.555	ng/ul	94
76) Pentachlorobenzene	15.279	250	107854	17.727	ng/ul	95
77) Carbazole	18.146	167	565135	20.816	ng/ul	98
78) Di-n-butylphthalate	18.663	149	717763	19.241	ng/ul	98
80) Fluoranthene	19.774	202	651123	19.314	ng/ul	98
82) Pyrene	20.138	202	685400	19.573	ng/ul	99
83) Butylbenzylphthalate	21.013	149	320361	20.257	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.965	252	229363	22.168	ng/ul#	91
85) Benzo(a)anthracene	22.059	228	625525	19.104	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.912	149	456538	19.347	ng/ul	100
87) Chrysene	22.130	228	567509	18.632	ng/ul	96
89) Di-n-octyl phthalate	23.270	149	784719	20.473	ng/ul	100
90) Benzo(b)fluoranthene	24.515	252	618775	19.499	ng/ul	95
91) Benzo(k)fluoranthene	24.515	252	618775	19.521	ng/ul	96
93) Benzo(a)pyrene	25.520	252	534398	17.817	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	29.868	276	607571	15.982	ng/ul	99
95) Dibenzo(a,h)anthracene	29.956	278	227228	7.237	ng/ul	94
96) Benzo(g,h,i)perylene	31.196	276	350557	11.267	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 28 04:04:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 28 03:56:42 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.136 ng/uL
RT: 3.693 min Scan# 51
Delta R.T. -0.009 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion: 88 Resp: 13740
Ion Ratio Lower Upper
88 100
43 28.1 15.7 23.5#
58 58.7 36.2 54.2#



#5
Pyridine
Concen: 16.049 ng/uL
RT: 4.110 min Scan# 123
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion: 79 Resp: 88512
Ion Ratio Lower Upper
79 100
52 54.4 53.3 79.9
51 31.8 22.4 33.6









#6
Benzaldehyde
Concen: 24.625 ng/ul
RT: 7.506 min Scan# 701
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion: 77 Resp: 85322
Ion Ratio Lower Upper
77 100
105 90.6 83.6 125.4
106 87.3 78.6 118.0



#8
Phenol
Concen: 19.072 ng/ul
RT: 7.500 min Scan# 700
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion: 94 Resp: 138314
Ion Ratio Lower Upper
94 100
65 29.5 22.5 33.7
66 44.4 31.1 46.7









#10
Bis(2-Chloroethyl)ether
Concen: 18.285 ng/ul
RT: 7.776 min Scan# 747
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion: 93 Resp: 97241
Ion Ratio Lower Upper
93 100
63 63.2 52.5 78.7
95 35.0 26.5 39.7



#12
2-Chlorophenol
Concen: 19.047 ng/ul
RT: 7.905 min Scan# 769
Delta R.T. -0.009 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:128 Resp: 117842
Ion Ratio Lower Upper
128 100
64 32.7 30.2 45.2
130 31.3 27.9 41.9









#13
2-Methylphenol
Concen: 19.184 ng/ul
RT: 8.775 min Scan# 917
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:108 Resp: 115982
Ion Ratio Lower Upper
108 100
107 83.9 68.6 102.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 16.509 ng/ul
RT: 8.881 min Scan# 935
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion: 45 Resp: 82624
Ion Ratio Lower Upper
45 100
77 21.3 15.0 22.6
79 17.4 10.0 15.0#









#16
Acetophenone
Concen: 20.602 ng/ul
RT: 9.204 min Scan# 990
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:105 Resp: 198200
Ion Ratio Lower Upper
105 100
77 65.8 59.6 89.4
51 19.7 16.2 24.4



#17
N-Nitroso-di-n-propylamine
Concen: 18.328 ng/ul
RT: 9.168 min Scan# 984
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion: 70 Resp: 83082
Ion Ratio Lower Upper
70 100
42 48.0 33.8 50.8
101 10.5 5.4 8.0#
130 19.9 18.2 27.4









#18
4-Methylphenol
Concen: 19.920 ng/ul
RT: 9.104 min Scan# 91
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:108 Resp: 128133
Ion Ratio Lower Upper
108 100
107 106.6 88.9 133.3



#19
Hexachloroethane
Concen: 18.856 ng/ul
RT: 9.456 min Scan# 1033
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:117 Resp: 46336
Ion Ratio Lower Upper
117 100
201 88.8 75.8 113.6
199 51.0 42.3 63.5









#22
Nitrobenzene
Concen: 22.307 ng/ul
RT: 9.603 min Scan# 1058
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion: 77 Resp: 116443
Ion Ratio Lower Upper
77 100
123 48.5 42.6 63.8
65 13.8 11.0 16.4



#23
Isophorone
Concen: 18.657 ng/ul
RT: 10.114 min Scan# 1145
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion: 82 Resp: 245257
Ion Ratio Lower Upper
82 100
95 8.2 5.6 8.4
138 20.6 12.3 18.5









#25
2-Nitrophenol
Concen: 21.915 ng/ul
RT: 10.314 min Scan# 1179
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:139 Resp: 58516
Ion Ratio Lower Upper
139 100
65 42.7 39.7 59.5
109 30.1 21.0 31.6



#26
2,4-Dimethylphenol
Concen: 17.587 ng/ul
RT: 10.332 min Scan# 1182
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:107 Resp: 123517
Ion Ratio Lower Upper
107 100
121 59.9 42.7 64.1
122 95.3 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.777 ng/ul
 RT: 10.596 min Scan# 11
 Delta R.T. -0.004 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08




Instrument :
 BNA_G
 ClientSampleId :
 SICV440


Tgt Ion: 93 Resp: 137439
 Ion Ratio Lower Upper
 93 100
 95 31.1 27.1 40.7
 123 13.0 10.3 15.5



#29
 2,4-Dichlorophenol
 Concen: 20.495 ng/ul
 RT: 10.831 min Scan# 1267
 Delta R.T. 0.002 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08

Tgt Ion:162 Resp: 108461
 Ion Ratio Lower Upper
 162 100
 164 65.3 45.3 67.9
 98 36.6 32.2 48.4









#30
Naphthalene
Concen: 18.381 ng/ul
RT: 11.266 min Scan# 11266
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:128 Resp: 389665
Ion Ratio Lower Upper
128 100
129 12.3 7.4 11.0#
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 20.461 ng/ul
RT: 11.378 min Scan# 1360
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:127 Resp: 175013
Ion Ratio Lower Upper
127 100
129 28.3 30.4 45.6#









#33
Hexachlorobutadiene
Concen: 18.302 ng/ul
RT: 11.495 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:225 Resp: 58104
Ion Ratio Lower Upper
225 100
223 74.2 50.8 76.2
227 60.6 49.8 74.8



#34
Caprolactam
Concen: 20.261 ng/ul
RT: 12.147 min Scan# 1491
Delta R.T. -0.009 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:113 Resp: 39225
Ion Ratio Lower Upper
113 100
55 118.5 86.9 130.3
56 91.0 64.4 96.6









#35
4-Chloro-3-methylphenol
Concen: 21.264 ng/ul
RT: 12.429 min Scan# 1539
Delta R.T. -0.009 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:107 Resp: 134265
Ion Ratio Lower Upper
107 100
144 22.2 21.0 31.6
142 71.4 68.2 102.4



#36
2-Methylnaphthalene
Concen: 18.622 ng/ul
RT: 12.841 min Scan# 1609
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:142 Resp: 271592
Ion Ratio Lower Upper
142 100
141 89.0 71.7 107.5









#37
1-Methylnaphthalene
Concen: 17.982 ng/ul
RT: 13.058 min Scan# 1646
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:142 Resp: 261215
Ion Ratio Lower Upper
142 100
141 99.3 73.1 109.7
116 4.5 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.339 ng/ul
RT: 13.176 min Scan# 1666
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:216 Resp: 120459
Ion Ratio Lower Upper
216 100
214 75.2 63.7 95.5
179 21.1 19.2 28.8
108 20.6 16.5 24.7









#40
Hexachlorocyclopentadiene
Concen: 16.467 ng/ul
RT: 13.134 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:	237	Resp:	59791
Ion Ratio	Lower	Upper	
237	100		
235	65.2	46.6	69.8
272	13.2	13.0	19.4



#41
2,4,6-Trichlorophenol
Concen: 18.828 ng/ul
RT: 13.416 min Scan# 1707
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:	196	Resp:	81207
Ion Ratio	Lower	Upper	
196	100		
198	100.5	69.6	104.4
200	30.4	23.8	35.8









#42
2,4,5-Trichlorophenol
Concen: 19.455 ng/ul
RT: 13.487 min Scan# 1718
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:196 Resp: 90845
Ion Ratio Lower Upper
196 100
198 97.9 79.8 119.6
200 29.8 28.0 42.0



#43
1,1'-Biphenyl
Concen: 19.290 ng/ul
RT: 13.828 min Scan# 1777
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:154 Resp: 376858
Ion Ratio Lower Upper
154 100
153 37.7 30.7 46.1
76 10.1 8.2 12.4









#44
2-Chloronaphthalene
Concen: 17.476 ng/ul
RT: 13.881 min Scan# 1786
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:162 Resp: 265353
Ion Ratio Lower Upper
162 100
164 35.6 23.9 35.9
127 44.2 35.3 52.9



#45
2-Nitroaniline
Concen: 22.351 ng/ul
RT: 14.086 min Scan# 1821
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion: 65 Resp: 63149
Ion Ratio Lower Upper
65 100
92 74.2 59.6 89.4
138 121.1 109.0 163.6









#47
Dimethylphthalate
Concen: 18.144 ng/ul
RT: 14.433 min Scan# 1880
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:163 Resp: 350724
Ion Ratio Lower Upper
163 100
194 3.4 2.6 3.8
164 10.8 6.6 10.0#



#48
2,6-Dinitrotoluene
Concen: 21.212 ng/ul
RT: 14.574 min Scan# 1904
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:165 Resp: 61852
Ion Ratio Lower Upper
165 100
89 65.9 41.0 61.4#
121 27.6 12.3 18.5#









#50
Acenaphthylene
Concen: 18.012 ng/ul
RT: 14.721 min Scan# 1929
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:152 Resp: 476501
Ion Ratio Lower Upper
152 100
151 19.6 16.6 25.0
153 13.4 9.4 14.2



#51
3-Nitroaniline
Concen: 23.420 ng/ul
RT: 14.909 min Scan# 1961
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:138 Resp: 85121
Ion Ratio Lower Upper
138 100
108 13.5 17.1 25.7#
92 105.6 101.8 152.8









#52
Acenaphthene
Concen: 17.996 ng/ul
RT: 15.056 min Scan# 1986
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:153 Resp: 310343
Ion Ratio Lower Upper
153 100
152 44.0 37.6 56.4
154 91.4 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 19.131 ng/ul
RT: 15.103 min Scan# 1994
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:184 Resp: 31697
Ion Ratio Lower Upper
184 100
63 40.5 35.0 52.6
154 80.6 49.8 74.8#









#55
4-Nitrophenol
Concen: 21.699 ng/ul
RT: 15.161 min Scan# 2004
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:109 Resp: 80235
Ion Ratio Lower Upper
109 100
139 93.6 72.2 108.2
65 84.9 57.3 85.9



#56
Dibenzofuran
Concen: 19.036 ng/ul
RT: 15.385 min Scan# 2042
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:168 Resp: 425597
Ion Ratio Lower Upper
168 100
139 42.4 33.8 50.8









#57
2,4-Dinitrotoluene
Concen: 23.383 ng/ul
RT: 15.344 min Scan# 2035
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:165 Resp: 95911
Ion Ratio Lower Upper
165 100
63 42.5 30.9 46.3
182 1.7 1.8 2.6#



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.972 ng/ul
RT: 15.590 min Scan# 2077
Delta R.T. 0.002 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:232 Resp: 76862
Ion Ratio Lower Upper
232 100
131 61.5 44.2 66.2
130 3.0 2.3 3.5
166 34.6 32.6 48.8









#59
Diethylphthalate
Concen: 18.605 ng/ul
RT: 15.778 min Scan# 2109
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:149 Resp: 392308
Ion Ratio Lower Upper
149 100
177 19.9 16.3 24.5
150 11.9 9.1 13.7



#61
Fluorene
Concen: 18.610 ng/ul
RT: 16.031 min Scan# 2152
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:166 Resp: 359181
Ion Ratio Lower Upper
166 100
165 95.2 74.4 111.6
167 13.7 9.1 13.7#









#62
4-Chlorophenyl-phenylether
Concen: 18.422 ng/ul
RT: 16.019 min Scan# 2150
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:204 Resp: 151953
Ion Ratio Lower Upper
204 100
206 34.3 28.8 43.2
141 80.8 68.6 103.0



#63
4-Nitroaniline
Concen: 25.946 ng/ul
RT: 16.066 min Scan# 2158
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:138 Resp: 100080
Ion Ratio Lower Upper
138 100
92 47.4 49.8 74.8#
108 122.3 100.9 151.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.531 ng/ul
 RT: 16.096 min Scan# 2163
 Delta R.T. 0.002 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08




Instrument :
 BNA_G
 ClientSampleId :
 SICV440


Tgt Ion:198 Resp: 46934
 Ion Ratio Lower Upper
 198 100
 182 3.4 1.4 2.0#
 77 24.2 15.8 23.8#



#67
 N-Nitrosodiphenylamine
 Concen: 18.933 ng/ul
 RT: 16.237 min Scan# 2187
 Delta R.T. 0.002 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08

Tgt Ion:169 Resp: 311470
 Ion Ratio Lower Upper
 169 100
 168 68.3 54.2 81.2
 167 39.7 31.3 46.9









#68
 4-Bromophenyl-phenylether
 Concen: 18.184 ng/ul
 RT: 16.918 min Scan# 21
 Delta R.T. 0.002 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08




Instrument :
 BNA_G
 ClientSampleId :
 SICV440


Tgt Ion:248 Resp: 87610
 Ion Ratio Lower Upper
 248 100
 250 95.8 74.2 111.2
 141 118.5 88.5 132.7



#69
 Hexachlorobenzene
 Concen: 18.275 ng/ul
 RT: 17.006 min Scan# 2318
 Delta R.T. 0.002 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08

Tgt Ion:284 Resp: 101761
 Ion Ratio Lower Upper
 284 100
 142 46.9 39.6 59.4
 249 32.3 28.2 42.2









#70
Atrazine
Concen: 21.248 ng/ul
RT: 17.171 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:200 Resp: 108318
Ion Ratio Lower Upper
200 100
173 27.0 25.7 38.5
215 40.2 32.1 48.1



#71
Pentachlorophenol
Concen: 0.049 ng/ul
RT: 17.435 min Scan# 2391
Delta R.T. 0.079 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:266 Resp: 157
Ion Ratio Lower Upper
266 100
264 55.5 46.2 69.2
268 49.0 46.4 69.6





#72

Phenanthrene

Concen: 18.453 ng/ul

RT: 17.782 min Scan# 2450

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08

Instrument :

BNA_G

ClientSampleId :

SICV440





Tgt Ion:178 Resp: 577597

Ion Ratio Lower Upper

178 100

179 15.1 11.3 16.9

176 16.9 15.4 23.0



#74

Anthracene

Concen: 18.642 ng/ul

RT: 17.876 min Scan# 2466

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08




Tgt Ion:178 Resp: 589167


Ion Ratio Lower Upper

178 100

179 14.1 13.1 19.7

176 18.9 13.9 20.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.555 ng/uL
RT: 13.787 min Scan# 1770
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:216 Resp: 112556
Ion Ratio Lower Upper
216 100
214 79.9 70.2 105.2
218 50.8 39.9 59.9



#76
Pentachlorobenzene
Concen: 17.727 ng/uL
RT: 15.279 min Scan# 2024
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:250 Resp: 107854
Ion Ratio Lower Upper
250 100
248 68.5 48.7 73.1
252 64.6 51.2 76.8





#77

Carbazole

Concen: 20.816 ng/ul

RT: 18.146 min Scan# 2

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08

Instrument :

BNA_G

ClientSampleId :

SICV440





Tgt Ion:167 Resp: 565135

Ion Ratio Lower Upper

167 100

166 22.6 19.4 29.2

139 14.7 11.7 17.5



#78

Di-n-butylphthalate

Concen: 19.241 ng/ul

RT: 18.663 min Scan# 2600

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08




Tgt Ion:149 Resp: 717763


Ion Ratio Lower Upper

149 100

150 10.2 7.5 11.3

104 6.0 5.3 7.9









#80
Fluoranthene
Concen: 19.314 ng/ul
RT: 19.774 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:202 Resp: 651123
Ion Ratio Lower Upper
202 100
101 8.8 7.0 10.6
100 7.7 5.2 7.8



#82
Pyrene
Concen: 19.573 ng/ul
RT: 20.138 min Scan# 2851
Delta R.T. -0.004 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:202 Resp: 685400
Ion Ratio Lower Upper
202 100
101 10.3 7.9 11.9
100 9.8 7.9 11.9





#83

Butylbenzylphthalate

Concen: 20.257 ng/ul

RT: 21.013 min Scan# 3000

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08

Instrument :

BNA_G

ClientSampleId :

SICV440





Tgt Ion:149 Resp: 320361

Ion Ratio Lower Upper

149 100

91 65.2 52.2 78.2

206 17.5 15.1 22.7



#84

3,3'-Dichlorobenzidine

Concen: 22.168 ng/ul

RT: 21.965 min Scan# 3162

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08




Tgt Ion:252 Resp: 229363


Ion Ratio Lower Upper

252 100

254 68.4 48.9 73.3

126 11.4 11.5 17.3#





#85

Benzo(a)anthracene

Concen: 19.104 ng/ul

RT: 22.059 min Scan# 3177

Delta R.T. 0.002 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08

Instrument :

BNA_G

ClientSampleId :

SICV440





Tgt Ion:228 Resp: 625525

Ion Ratio Lower Upper

228 100

229 19.4 16.7 25.1

226 25.5 20.3 30.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.347 ng/ul

RT: 21.912 min Scan# 3153

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08




Tgt Ion:149 Resp: 456538


Ion Ratio Lower Upper

149 100

167 24.4 19.4 29.2

279 2.9 2.2 3.4





#87

Chrysene

Concen: 18.632 ng/ul

RT: 22.130 min Scan# 3190

Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08

Instrument :

BNA_G

ClientSampleId :

SICV440





Tgt Ion:228 Resp: 567509

Ion Ratio Lower Upper

228 100

226 29.3 21.0 31.4

229 18.3 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 20.473 ng/ul




RT: 23.270 min Scan# 3384


Delta R.T. -0.004 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08

Tgt Ion:149 Resp: 784719









#90
Benzo(b)fluoranthene
Concen: 19.499 ng/ul
RT: 24.515 min Scan# 3596
Delta R.T. -0.009 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08




Instrument :
BNA_G
ClientSampleId :
SICV440


Tgt Ion:252 Resp: 618775
Ion Ratio Lower Upper
252 100
253 19.5 17.8 26.8
125 9.6 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 19.521 ng/ul
RT: 24.515 min Scan# 3596
Delta R.T. -0.086 min
Lab File: BG059513.D
Acq: 27 Oct 2023 16:08

Tgt Ion:252 Resp: 618775
Ion Ratio Lower Upper
252 100
253 19.5 17.9 26.9
125 9.6 7.9 11.9





#93

Benzo(a)pyrene

Concen: 17.817 ng/ul

RT: 25.520 min Scan# 31

Delta R.T. 0.002 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08

Instrument :

BNA_G

ClientSampleId :

SICV440





Tgt Ion:252 Resp: 534398

Ion Ratio Lower Upper

252 100

253 21.0 18.5 27.7

125 11.7 7.6 11.4#



#94

Indeno(1,2,3-cd)pyrene

Concen: 15.982 ng/ul

RT: 29.868 min Scan# 4507

Delta R.T. -0.021 min

Lab File: BG059513.D

Acq: 27 Oct 2023 16:08




Tgt Ion:276 Resp: 607571


Ion Ratio Lower Upper

276 100

138 19.3 15.3 22.9

277 21.1 16.4 24.6









#95
 Dibenzo(a,h)anthracene
 Concen: 7.237 ng/ul
 RT: 29.956 min Scan# 4
 Delta R.T. -0.027 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08

Instrument :
 BNA_G
 ClientSampleId :
 SICV440

Tgt Ion:	278	Resp:	227228
Ion	Ratio	Lower	Upper
278	100		
139	11.3	11.0	16.6
279	25.7	18.3	27.5



#96
 Benzo(g,h,i)perylene
 Concen: 11.267 ng/ul
 RT: 31.196 min Scan# 4733
 Delta R.T. -0.021 min
 Lab File: BG059513.D
 Acq: 27 Oct 2023 16:08

Tgt Ion:	276	Resp:	350557
Ion	Ratio	Lower	Upper
276	100		
138	17.1	16.2	24.2
277	20.3	18.2	27.4

