

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102823\
 Data File : BG059516.D
 Acq On : 28 Oct 2023 11:06
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020459

Quant Time: Oct 30 02:12:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.365	152	64923	20.000	ng/ul	0.00
20) Naphthalene-d8	11.214	136	320493	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.992	164	199798	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.748	188	446518	20.000	ng/ul	0.00
79) Chrysene-d12	22.084	240	419712	20.000	ng/ul	0.00
88) Perylene-d12	25.691	264	440334	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.653	96	13178	7.594	ng/uL	0.00
4) Pyridine-d5	4.087	84	105827	21.824	ng/ul	0.00
7) Phenol-d5	7.466	99	139889	20.968	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.677	67	67339	21.975	ng/ul	0.00
11) 2-Chlorophenol-d4	7.877	132	109382	20.209	ng/ul	0.00
15) 4-Methylphenol-d8	9.040	113	106884	20.194	ng/ul	0.00
21) Nitrobenzene-d5	9.563	128	43512	19.094	ng/ul	0.00
24) 2-Nitrophenol-d4	10.280	143	45739	19.551	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.803	165	101728	19.737	ng/ul	0.00
31) 4-Chloroaniline-d4	11.349	131	180189	21.237	ng/ul	0.00
46) Dimethylphthalate-d6	14.387	166	355708	19.651	ng/ul	0.00
49) Acenaphthylene-d8	14.693	160	435986	20.524	ng/ul	0.00
54) 4-Nitrophenol-d4	15.151	143	62512	20.139	ng/ul	0.00
60) Fluorene-d10	15.979	176	302275	19.540	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.079	200	39785	18.815	ng/ul	0.00
73) Anthracene-d10	17.842	188	487702	20.161	ng/ul	0.00
81) Pyrene-d10	20.110	212	581209	19.923	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.439	264	541122	20.869	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.694	88	18078	9.308	ng/uL#	86
5) Pyridine	4.111	79	107937	22.567	ng/ul#	85
6) Benzaldehyde	7.507	77	73660	24.513	ng/ul	99
8) Phenol	7.495	94	132713	21.101	ng/ul	94
10) Bis(2-Chloroethyl)ether	7.777	93	105534	22.881	ng/ul	96
12) 2-Chlorophenol	7.906	128	112381	20.944	ng/ul	95
13) 2-Methylphenol	8.776	108	112752	21.504	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.870	45	63563	14.644	ng/ul	95
16) Acetophenone	9.205	105	182719	21.900	ng/ul	98
17) N-Nitroso-di-n-propyla...	9.170	70	77815	19.794	ng/ul#	97
18) 4-Methylphenol	9.105	108	118074	21.165	ng/ul	88
19) Hexachloroethane	9.452	117	41454	19.452	ng/ul	94
22) Nitrobenzene	9.604	77	101282	20.591	ng/ul	94
23) Isophorone	10.110	82	245297	19.803	ng/ul#	91
25) 2-Nitrophenol	10.315	139	49997	19.872	ng/ul#	84
26) 2,4-Dimethylphenol	10.333	107	132624	20.041	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.603	93	141279	20.484	ng/ul#	97
29) 2,4-Dichlorophenol	10.832	162	100273	20.109	ng/ul#	86
30) Naphthalene	11.267	128	409239	20.487	ng/ul#	97
32) 4-Chloroaniline	11.379	127	169889	21.079	ng/ul#	86
33) Hexachlorobutadiene	11.496	225	56370	18.844	ng/ul	97
34) Caprolactam	12.143	113	21601	11.841	ng/ul	93
35) 4-Chloro-3-methylphenol	12.436	107	127243	21.386	ng/ul	94
36) 2-Methylnaphthalene	12.842	142	280334	20.399	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102823\
 Data File : BG059516.D
 Acq On : 28 Oct 2023 11:06
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020459

Quant Time: Oct 30 02:12:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.059	142	278455	20.343	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	13.183	216	114774	19.917	ng/ul	98
40) Hexachlorocyclopentadiene	13.141	237	59622	17.748	ng/ul#	92
41) 2,4,6-Trichlorophenol	13.418	196	72204	18.095	ng/ul	88
42) 2,4,5-Trichlorophenol	13.488	196	80441	18.620	ng/ul	93
43) 1,1'-Biphenyl	13.829	154	361153	19.981	ng/ul	99
44) 2-Chloronaphthalene	13.882	162	271469	19.324	ng/ul#	95
45) 2-Nitroaniline	14.093	65	53159	20.337	ng/ul	96
47) Dimethylphthalate	14.434	163	357034	19.964	ng/ul#	92
48) 2,6-Dinitrotoluene	14.569	165	55243	20.477	ng/ul#	78
50) Acenaphthylene	14.722	152	484934	19.813	ng/ul#	97
51) 3-Nitroaniline	14.910	138	69880	20.781	ng/ul#	81
52) Acenaphthene	15.057	153	306215	19.193	ng/ul	99
53) 2,4-Dinitrophenol	15.098	184	30540	19.924	ng/ul	95
55) 4-Nitrophenol	15.169	109	66279	19.375	ng/ul	92
56) Dibenzofuran	15.386	168	417804	20.199	ng/ul	96
57) 2,4-Dinitrotoluene	15.345	165	77545	20.435	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.592	232	68965	20.340	ng/ul	92
59) Diethylphthalate	15.785	149	389487	19.965	ng/ul#	96
61) Fluorene	16.032	166	355491	19.909	ng/ul	98
62) 4-Chlorophenyl-phenyle...	16.020	204	148572	19.469	ng/ul	98
63) 4-Nitroaniline	16.067	138	72659	20.361	ng/ul#	92
66) 4,6-Dinitro-2-methylph...	16.091	198	44589	20.008	ng/ul#	87
67) N-Nitrosodiphenylamine	16.238	169	298950	19.594	ng/ul	98
68) 4-Bromophenyl-phenylether	16.919	248	55208	12.356	ng/ul	98
69) Hexachlorobenzene	17.008	284	105633	20.456	ng/ul	89
70) Atrazine	17.172	200	103322	21.854	ng/ul	95
71) Pentachlorophenol	17.360	266	53574	18.160	ng/ul	97
72) Phenanthrene	17.789	178	563959	19.428	ng/ul	99
74) Anthracene	17.877	178	592524	20.216	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.788	216	116157	19.535	ng/ul	95
76) Pentachlorobenzene	15.280	250	110033	19.501	ng/ul	96
77) Carbazole	18.147	167	557265	22.133	ng/ul	96
78) Di-n-butylphthalate	18.670	149	710291	20.532	ng/ul	98
80) Fluoranthene	19.781	202	690776	19.424	ng/ul	97
82) Pyrene	20.139	202	723508	19.587	ng/ul	99
83) Butylbenzylphthalate	21.015	149	328795	19.709	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.966	252	214394	19.643	ng/ul#	92
85) Benzo(a)anthracene	22.060	228	676752	19.594	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.914	149	474499	19.062	ng/ul	98
87) Chrysene	22.131	228	626059	19.485	ng/ul	98
89) Di-n-octyl phthalate	23.277	149	829102	21.506	ng/ul	100
90) Benzo(b)fluoranthene	24.528	252	657608	20.603	ng/ul	96
91) Benzo(k)fluoranthene	24.605	252	660529	20.718	ng/ul	97
93) Benzo(a)pyrene	25.527	252	602750	19.980	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	29.887	276	569264	14.888	ng/ul	94
95) Dibenzo(a,h)anthracene	29.981	278	619192	19.606	ng/ul	98
96) Benzo(g,h,i)perylene	31.209	276	604804	19.325	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.308 ng/uL
RT: 3.694 min Scan# 51
Delta R.T. -0.008 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion: 88 Resp: 18078
Ion Ratio Lower Upper
88 100
43 13.7 15.7 23.5#
58 36.0 36.2 54.2#



#5
Pyridine
Concen: 22.567 ng/uL
RT: 4.111 min Scan# 123
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion: 79 Resp: 107937
Ion Ratio Lower Upper
79 100
52 49.4 53.3 79.9#
51 27.9 22.4 33.6









#6
Benzaldehyde
Concen: 24.513 ng/ul
RT: 7.507 min Scan# 700
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion: 77 Resp: 73660
Ion Ratio Lower Upper
77 100
105 105.4 83.6 125.4
106 99.2 78.6 118.0



#8
Phenol
Concen: 21.101 ng/ul
RT: 7.495 min Scan# 699
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion: 94 Resp: 132713
Ion Ratio Lower Upper
94 100
65 26.8 22.5 33.7
66 44.1 31.1 46.7









#10
Bis(2-Chloroethyl)ether
Concen: 22.881 ng/ul
RT: 7.777 min Scan# 747
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion: 93 Resp: 105534
Ion Ratio Lower Upper
93 100
63 62.5 52.5 78.7
95 30.8 26.5 39.7



#12
2-Chlorophenol
Concen: 20.944 ng/ul
RT: 7.906 min Scan# 769
Delta R.T. -0.008 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion: 128 Resp: 112381
Ion Ratio Lower Upper
128 100
64 35.6 30.2 45.2
130 31.0 27.9 41.9









#13
2-Methylphenol
Concen: 21.504 ng/ul
RT: 8.776 min Scan# 917
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:108 Resp: 112752
Ion Ratio Lower Upper
108 100
107 88.3 68.6 102.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 14.644 ng/ul
RT: 8.870 min Scan# 933
Delta R.T. -0.008 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion: 45 Resp: 63563
Ion Ratio Lower Upper
45 100
77 20.8 15.0 22.6
79 10.4 10.0 15.0









#16
Acetophenone
Concen: 21.900 ng/ul
RT: 9.205 min Scan# 990
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:105 Resp: 182719
Ion Ratio Lower Upper
105 100
77 75.4 59.6 89.4
51 22.2 16.2 24.4



#17
N-Nitroso-di-n-propylamine
Concen: 19.794 ng/ul
RT: 9.170 min Scan# 984
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion: 70 Resp: 77815
Ion Ratio Lower Upper
70 100
42 43.6 33.8 50.8
101 11.6 5.4 8.0#
130 22.1 18.2 27.4









#18
4-Methylphenol
Concen: 21.165 ng/ul
RT: 9.105 min Scan# 91
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:108 Resp: 118074
Ion Ratio Lower Upper
108 100
107 98.2 88.9 133.3



#19
Hexachloroethane
Concen: 19.452 ng/ul
RT: 9.452 min Scan# 1032
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:117 Resp: 41454
Ion Ratio Lower Upper
117 100
201 89.0 75.8 113.6
199 57.9 42.3 63.5









#22
Nitrobenzene
Concen: 20.591 ng/ul
RT: 9.604 min Scan# 1058
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion: 77 Resp: 101282
Ion Ratio Lower Upper
77 100
123 48.6 42.6 63.8
65 12.2 11.0 16.4



#23
Isophorone
Concen: 19.803 ng/ul
RT: 10.110 min Scan# 1144
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion: 82 Resp: 245297
Ion Ratio Lower Upper
82 100
95 9.0 5.6 8.4#
138 20.0 12.3 18.5#









#25
2-Nitrophenol
Concen: 19.872 ng/ul
RT: 10.315 min Scan# 1179
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:139 Resp: 49997
Ion Ratio Lower Upper
139 100
65 43.2 39.7 59.5
109 40.4 21.0 31.6#



#26
2,4-Dimethylphenol
Concen: 20.041 ng/ul
RT: 10.333 min Scan# 1182
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:107 Resp: 132624
Ion Ratio Lower Upper
107 100
121 47.4 42.7 64.1
122 90.0 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.484 ng/ul
 RT: 10.603 min Scan# 1227
 Delta R.T. 0.003 min
 Lab File: BG059516.D
 Acq: 28 Oct 2023 11:06




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020459


Tgt Ion: 93 Resp: 141279
 Ion Ratio Lower Upper
 93 100
 95 32.7 27.1 40.7
 123 10.3 10.3 15.5#



#29
 2,4-Dichlorophenol
 Concen: 20.109 ng/ul
 RT: 10.832 min Scan# 1267
 Delta R.T. 0.003 min
 Lab File: BG059516.D
 Acq: 28 Oct 2023 11:06

Tgt Ion:162 Resp: 100273
 Ion Ratio Lower Upper
 162 100
 164 71.1 45.3 67.9#
 98 36.5 32.2 48.4





#30

Naphthalene

Concen: 20.487 ng/ul

RT: 11.267 min Scan# 11

Delta R.T. 0.003 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:128 Resp: 409239

Ion Ratio Lower Upper

128 100

129 11.5 7.4 11.0#

127 12.7 10.5 15.7



#32

4-Chloroaniline

Concen: 21.079 ng/ul

RT: 11.379 min Scan# 1360

Delta R.T. 0.003 min

Lab File: BG059516.D




Acq: 28 Oct 2023 11:06


Tgt Ion:127 Resp: 169889

Ion Ratio Lower Upper

127 100

129 29.8 30.4 45.6#









#33
Hexachlorobutadiene
Concen: 18.844 ng/ul
RT: 11.496 min Scan# 11
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:225 Resp: 56370
Ion Ratio Lower Upper
225 100
223 59.1 50.8 76.2
227 62.0 49.8 74.8



#34
Caprolactam
Concen: 11.841 ng/ul
RT: 12.143 min Scan# 1490
Delta R.T. -0.014 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:113 Resp: 21601
Ion Ratio Lower Upper
113 100
55 98.4 86.9 130.3
56 77.8 64.4 96.6









#35
4-Chloro-3-methylphenol
Concen: 21.386 ng/ul
RT: 12.436 min Scan# 11
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:107 Resp: 127243
Ion Ratio Lower Upper
107 100
144 21.2 21.0 31.6
142 80.6 68.2 102.4



#36
2-Methylnaphthalene
Concen: 20.399 ng/ul
RT: 12.842 min Scan# 1609
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:142 Resp: 280334
Ion Ratio Lower Upper
142 100
141 89.3 71.7 107.5





#37

1-Methylnaphthalene

Concen: 20.343 ng/ul

RT: 13.059 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:142 Resp: 278455

Ion Ratio Lower Upper

142 100

141 90.6 73.1 109.7

116 3.6 3.3 4.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.917 ng/ul

RT: 13.183 min Scan# 1667

Delta R.T. 0.003 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Tgt Ion:216 Resp: 114774




Ion Ratio Lower Upper


216 100

214 78.3 63.7 95.5

179 25.5 19.2 28.8

108 18.8 16.5 24.7









#40
Hexachlorocyclopentadiene
Concen: 17.748 ng/ul
RT: 13.141 min Scan# 1659
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:237 Resp: 59622
Ion Ratio Lower Upper
237 100
235 52.7 46.6 69.8
272 12.5 13.0 19.4#



#41
2,4,6-Trichlorophenol
Concen: 18.095 ng/ul
RT: 13.418 min Scan# 1707
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:196 Resp: 72204
Ion Ratio Lower Upper
196 100
198 100.1 69.6 104.4
200 33.7 23.8 35.8









#42
2,4,5-Trichlorophenol
Concen: 18.620 ng/ul
RT: 13.488 min Scan# 1718
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:196 Resp: 80441
Ion Ratio Lower Upper
196 100
198 91.6 79.8 119.6
200 31.8 28.0 42.0



#43
1,1'-Biphenyl
Concen: 19.981 ng/ul
RT: 13.829 min Scan# 1777
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:154 Resp: 361153
Ion Ratio Lower Upper
154 100
153 38.3 30.7 46.1
76 9.4 8.2 12.4









#44
2-Chloronaphthalene
Concen: 19.324 ng/ul
RT: 13.882 min Scan# 1786
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:162 Resp: 271469
Ion Ratio Lower Upper
162 100
164 36.3 23.9 35.9#
127 43.6 35.3 52.9



#45
2-Nitroaniline
Concen: 20.337 ng/ul
RT: 14.093 min Scan# 1822
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion: 65 Resp: 53159
Ion Ratio Lower Upper
65 100
92 67.7 59.6 89.4
138 134.7 109.0 163.6









#47
Dimethylphthalate
Concen: 19.964 ng/ul
RT: 14.434 min Scan# 1880
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:163 Resp: 357034
Ion Ratio Lower Upper
163 100
194 2.6 2.6 3.8
164 11.8 6.6 10.0#



#48
2,6-Dinitrotoluene
Concen: 20.477 ng/ul
RT: 14.569 min Scan# 1903
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:165 Resp: 55243
Ion Ratio Lower Upper
165 100
89 68.0 41.0 61.4#
121 21.9 12.3 18.5#









#50
Acenaphthylene
Concen: 19.813 ng/ul
RT: 14.722 min Scan# 1929
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:152 Resp: 484934
Ion Ratio Lower Upper
152 100
151 20.4 16.6 25.0
153 14.7 9.4 14.2#



#51
3-Nitroaniline
Concen: 20.781 ng/ul
RT: 14.910 min Scan# 1961
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:138 Resp: 69880
Ion Ratio Lower Upper
138 100
108 15.7 17.1 25.7#
92 104.7 101.8 152.8









#52
Acenaphthene
Concen: 19.193 ng/ul
RT: 15.057 min Scan# 1986
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:153 Resp: 306215
Ion Ratio Lower Upper
153 100
152 46.3 37.6 56.4
154 88.2 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 19.924 ng/ul
RT: 15.098 min Scan# 1993
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:184 Resp: 30540
Ion Ratio Lower Upper
184 100
63 40.8 35.0 52.6
154 58.3 49.8 74.8









#55
4-Nitrophenol
Concen: 19.375 ng/ul
RT: 15.169 min Scan# 2004
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:109 Resp: 66279
Ion Ratio Lower Upper
109 100
139 96.2 72.2 108.2
65 79.3 57.3 85.9



#56
Dibenzofuran
Concen: 20.199 ng/ul
RT: 15.386 min Scan# 2042
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:168 Resp: 417804
Ion Ratio Lower Upper
168 100
139 44.7 33.8 50.8





#57

2,4-Dinitrotoluene

Concen: 20.435 ng/ul

RT: 15.345 min Scan# 2035

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:165 Resp: 77545

Ion Ratio Lower Upper

165 100

63 36.6 30.9 46.3

182 2.3 1.8 2.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.340 ng/ul

RT: 15.592 min Scan# 2077

Delta R.T. 0.003 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Tgt Ion:232 Resp: 68965




Ion Ratio Lower Upper


232 100

131 61.8 44.2 66.2

130 3.1 2.3 3.5

166 36.6 32.6 48.8









#59
Diethylphthalate
Concen: 19.965 ng/ul
RT: 15.785 min Scan# 2109
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:149 Resp: 389487
Ion Ratio Lower Upper
149 100
177 19.5 16.3 24.5
150 13.9 9.1 13.7#



#61
Fluorene
Concen: 19.909 ng/ul
RT: 16.032 min Scan# 2152
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:166 Resp: 355491
Ion Ratio Lower Upper
166 100
165 91.4 74.4 111.6
167 12.2 9.1 13.7









#62
4-Chlorophenyl-phenylether
Concen: 19.469 ng/ul
RT: 16.020 min Scan# 2150
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:204 Resp: 148572
Ion Ratio Lower Upper
204 100
206 33.1 28.8 43.2
141 85.9 68.6 103.0



#63
4-Nitroaniline
Concen: 20.361 ng/ul
RT: 16.067 min Scan# 2158
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:138 Resp: 72659
Ion Ratio Lower Upper
138 100
92 48.1 49.8 74.8#
108 122.9 100.9 151.3









#66
4,6-Dinitro-2-methylphenol
Concen: 20.008 ng/ul
RT: 16.091 min Scan# 2162
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:198 Resp: 44589
Ion Ratio Lower Upper
198 100
182 4.6 1.4 2.0#
77 26.1 15.8 23.8#



#67
N-Nitrosodiphenylamine
Concen: 19.594 ng/ul
RT: 16.238 min Scan# 2187
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:169 Resp: 298950
Ion Ratio Lower Upper
169 100
168 67.3 54.2 81.2
167 36.6 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 12.356 ng/ul
RT: 16.919 min Scan# 21
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:248 Resp: 55208
Ion Ratio Lower Upper
248 100
250 95.5 74.2 111.2
141 108.6 88.5 132.7



#69
Hexachlorobenzene
Concen: 20.456 ng/ul
RT: 17.008 min Scan# 2318
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:284 Resp: 105633
Ion Ratio Lower Upper
284 100
142 41.8 39.6 59.4
249 28.4 28.2 42.2









#70
Atrazine
Concen: 21.854 ng/ul
RT: 17.172 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:200 Resp: 103322
Ion Ratio Lower Upper
200 100
173 26.3 25.7 38.5
215 39.8 32.1 48.1



#71
Pentachlorophenol
Concen: 18.160 ng/ul
RT: 17.360 min Scan# 2378
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:266 Resp: 53574
Ion Ratio Lower Upper
266 100
264 55.7 46.2 69.2
268 55.7 46.4 69.6





#72

Phenanthrene

Concen: 19.428 ng/ul

RT: 17.789 min Scan# 2450

Delta R.T. 0.003 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:178 Resp: 563959

Ion Ratio Lower Upper

178 100

179 14.4 11.3 16.9

176 18.8 15.4 23.0



#74

Anthracene

Concen: 20.216 ng/ul

RT: 17.877 min Scan# 2466

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06




Tgt Ion:178 Resp: 592524


Ion Ratio Lower Upper

178 100

179 14.5 13.1 19.7

176 17.1 13.9 20.9





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.535 ng/uL

RT: 13.788 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:216 Resp: 116157

Ion Ratio Lower Upper

216 100

214 82.6 70.2 105.2

218 47.2 39.9 59.9



#76

Pentachlorobenzene

Concen: 19.501 ng/uL

RT: 15.280 min Scan# 2024

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06




Tgt Ion:250 Resp: 110033


Ion Ratio Lower Upper

250 100

248 65.2 48.7 73.1

252 66.5 51.2 76.8









#77
 Carbazole
 Concen: 22.133 ng/ul
 RT: 18.147 min Scan# 2512
 Delta R.T. -0.002 min
 Lab File: BG059516.D
 Acq: 28 Oct 2023 11:06




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020459


Tgt Ion:167 Resp: 557265
 Ion Ratio Lower Upper
 167 100
 166 21.8 19.4 29.2
 139 13.9 11.7 17.5



#78
 Di-n-butylphthalate
 Concen: 20.532 ng/ul
 RT: 18.670 min Scan# 2601
 Delta R.T. 0.003 min
 Lab File: BG059516.D
 Acq: 28 Oct 2023 11:06

Tgt Ion:149 Resp: 710291
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.5 11.3
 104 5.6 5.3 7.9





#80

Fluoranthene

Concen: 19.424 ng/ul

RT: 19.781 min Scan# 21

Delta R.T. 0.003 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:202 Resp: 690776

Ion Ratio Lower Upper

202 100

101 7.6 7.0 10.6

100 7.4 5.2 7.8



#82

Pyrene

Concen: 19.587 ng/ul

RT: 20.139 min Scan# 2851

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06




Tgt Ion:202 Resp: 723508


Ion Ratio Lower Upper

202 100

101 10.3 7.9 11.9

100 9.3 7.9 11.9









#83
Butylbenzylphthalate
Concen: 19.709 ng/ul
RT: 21.015 min Scan# 3000
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:149 Resp: 328795
Ion Ratio Lower Upper
149 100
91 67.8 52.2 78.2
206 16.5 15.1 22.7



#84
3,3'-Dichlorobenzidine
Concen: 19.643 ng/ul
RT: 21.966 min Scan# 3162
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:252 Resp: 214394
Ion Ratio Lower Upper
252 100
254 67.0 48.9 73.3
126 9.9 11.5 17.3#









#85
Benzo(a)anthracene
Concen: 19.594 ng/ul
RT: 22.060 min Scan# 3177
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:228 Resp: 676752
Ion Ratio Lower Upper
228 100
229 19.6 16.7 25.1
226 27.0 20.3 30.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.062 ng/ul
RT: 21.914 min Scan# 3153
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:149 Resp: 474499
Ion Ratio Lower Upper
149 100
167 23.2 19.4 29.2
279 3.0 2.2 3.4









#87
Chrysene
Concen: 19.485 ng/ul
RT: 22.131 min Scan# 31
Delta R.T. -0.002 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:228 Resp: 626059
Ion Ratio Lower Upper
228 100
226 27.5 21.0 31.4
229 19.3 15.1 22.7



#89
Di-n-octyl phthalate
Concen: 21.506 ng/ul
RT: 23.277 min Scan# 3385
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:149 Resp: 829102









#90
Benzo(b)fluoranthene
Concen: 20.603 ng/ul
RT: 24.528 min Scan# 3597
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06




Instrument :
BNA_G
ClientSampleId :
SSTD020459


Tgt Ion:252 Resp: 657608
Ion Ratio Lower Upper
252 100
253 20.6 17.8 26.8
125 8.8 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 20.718 ng/ul
RT: 24.605 min Scan# 3611
Delta R.T. 0.003 min
Lab File: BG059516.D
Acq: 28 Oct 2023 11:06

Tgt Ion:252 Resp: 660529
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 8.0 7.9 11.9





#93

Benzo(a)pyrene

Concen: 19.980 ng/ul

RT: 25.527 min Scan# 31

Delta R.T. 0.009 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:252 Resp: 602750

Ion Ratio Lower Upper

252 100

253 20.8 18.5 27.7

125 9.0 7.6 11.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 14.888 ng/ul

RT: 29.887 min Scan# 4510

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06




Tgt Ion:276 Resp: 569264


Ion Ratio Lower Upper

276 100

138 15.6 15.3 22.9

277 23.0 16.4 24.6





#95

Dibenzo(a,h)anthracene

Concen: 19.606 ng/ul

RT: 29.981 min Scan# 4

Delta R.T. -0.002 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Instrument :

BNA_G

ClientSampleId :

SSTD020459





Tgt Ion:278 Resp: 619192

Ion Ratio Lower Upper

278 100

139 13.2 11.0 16.6

279 21.5 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 19.325 ng/ul

RT: 31.209 min Scan# 4735

Delta R.T. -0.008 min

Lab File: BG059516.D

Acq: 28 Oct 2023 11:06

Tgt Ion:276 Resp: 604804

Ion Ratio Lower Upper

276 100

138 21.9 16.2 24.2

277 24.5 18.2 27.4

