

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102823\
 Data File : BG059530.D
 Acq On : 28 Oct 2023 21:23
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020460

Quant Time: Oct 30 02:16:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.364	152	73828	20.000	ng/ul	0.00
20) Naphthalene-d8	11.208	136	365986	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.986	164	227308	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.742	188	490181	20.000	ng/ul	# 0.00
79) Chrysene-d12	22.078	240	405908	20.000	ng/ul	0.00
88) Perylene-d12	25.674	264	240438	20.000	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.740	96	366	0.185	ng/uL	0.09
4) Pyridine-d5	4.093	84	121680	22.067	ng/ul	0.00
7) Phenol-d5	7.465	99	173375	22.853	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.677	67	81638	23.428	ng/ul	0.00
11) 2-Chlorophenol-d4	7.877	132	141194	22.940	ng/ul	0.00
15) 4-Methylphenol-d8	9.040	113	145076	24.103	ng/ul	0.00
21) Nitrobenzene-d5	9.557	128	65166	25.042	ng/ul	0.00
24) 2-Nitrophenol-d4	10.280	143	72538	27.153	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.803	165	132795	22.562	ng/ul	0.00
31) 4-Chloroaniline-d4	11.349	131	210028	21.677	ng/ul	0.00
46) Dimethylphthalate-d6	14.387	166	410535	19.935	ng/ul	0.00
49) Acenaphthylene-d8	14.692	160	499955	20.687	ng/ul	0.00
54) 4-Nitrophenol-d4	15.151	143	80836	22.890	ng/ul	0.00
60) Fluorene-d10	15.973	176	362832	20.616	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.079	200	59818	25.769	ng/ul	0.00
73) Anthracene-d10	17.836	188	563248	21.210	ng/ul	0.00
81) Pyrene-d10	20.109	212	604804	21.436	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.433	264	565184	39.919	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.705	88	8805	3.987	ng/uL#	84
5) Pyridine	4.116	79	121378	22.316	ng/ul#	86
6) Benzaldehyde	7.507	77	91222	26.695	ng/ul	98
8) Phenol	7.495	94	164076	22.940	ng/ul#	90
10) Bis(2-Chloroethyl)ether	7.777	93	123348	23.517	ng/ul	96
12) 2-Chlorophenol	7.906	128	141647	23.214	ng/ul	95
13) 2-Methylphenol	8.776	108	138114	23.164	ng/ul	95
14) 2,2'-oxybis(1-Chloropr...	8.870	45	102622	20.791	ng/ul#	90
16) Acetophenone	9.205	105	219489	23.134	ng/ul	97
17) N-Nitroso-di-n-propyla...	9.169	70	96241	21.528	ng/ul#	91
18) 4-Methylphenol	9.105	108	147687	23.280	ng/ul	91
19) Hexachloroethane	9.451	117	56226	23.201	ng/ul	93
22) Nitrobenzene	9.598	77	144699	25.761	ng/ul	92
23) Isophorone	10.115	82	294829	20.843	ng/ul#	94
25) 2-Nitrophenol	10.315	139	76156	26.506	ng/ul	93
26) 2,4-Dimethylphenol	10.333	107	171438	22.685	ng/ul	89
27) Bis(2-Chloroethoxy)met...	10.597	93	165588	21.024	ng/ul	94
29) 2,4-Dichlorophenol	10.832	162	131791	23.144	ng/ul	87
30) Naphthalene	11.267	128	483723	21.206	ng/ul	98
32) 4-Chloroaniline	11.373	127	207383	22.533	ng/ul	89
33) Hexachlorobutadiene	11.496	225	73191	21.425	ng/ul	94
34) Caprolactam	12.142	113	31429	15.087	ng/ul	93
35) 4-Chloro-3-methylphenol	12.436	107	154404	22.725	ng/ul	91
36) 2-Methylnaphthalene	12.842	142	335559	21.382	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102823\
 Data File : BG059530.D
 Acq On : 28 Oct 2023 21:23
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020460

Quant Time: Oct 30 02:16:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.059	142	334927	21.427	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.182	216	137048	20.904	ng/ul	98
40) Hexachlorocyclopentadiene	13.135	237	84369	22.076	ng/ul#	97
41) 2,4,6-Trichlorophenol	13.417	196	96218	21.195	ng/ul	93
42) 2,4,5-Trichlorophenol	13.482	196	107348	21.841	ng/ul	98
43) 1,1'-Biphenyl	13.829	154	425748	20.704	ng/ul	94
44) 2-Chloronaphthalene	13.881	162	330542	20.682	ng/ul	98
45) 2-Nitroaniline	14.087	65	76726	25.801	ng/ul	95
47) Dimethylphthalate	14.434	163	413713	20.334	ng/ul#	93
48) 2,6-Dinitrotoluene	14.569	165	68689	22.380	ng/ul#	67
50) Acenaphthylene	14.722	152	558891	20.071	ng/ul	99
51) 3-Nitroaniline	14.904	138	95121	24.864	ng/ul#	75
52) Acenaphthene	15.051	153	356103	19.619	ng/ul	99
53) 2,4-Dinitrophenol	15.104	184	41007	23.515	ng/ul	95
55) 4-Nitrophenol	15.168	109	100450	25.810	ng/ul	98
56) Dibenzofuran	15.386	168	485873	20.647	ng/ul	95
57) 2,4-Dinitrotoluene	15.344	165	109036	25.256	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.585	232	90061	23.347	ng/ul#	89
59) Diethylphthalate	15.779	149	455819	20.538	ng/ul	98
61) Fluorene	16.032	166	409894	20.177	ng/ul	86
62) 4-Chlorophenyl-phenyle...	16.020	204	179320	20.654	ng/ul	92
63) 4-Nitroaniline	16.067	138	93065	22.923	ng/ul	85
66) 4,6-Dinitro-2-methylph...	16.091	198	61884	25.295	ng/ul#	83
67) N-Nitrosodiphenylamine	16.232	169	349795	20.884	ng/ul	100
68) 4-Bromophenyl-phenylether	16.913	248	99530	20.291	ng/ul	95
69) Hexachlorobenzene	17.001	284	119255	21.037	ng/ul	96
70) Atrazine	17.166	200	118889	22.907	ng/ul	90
71) Pentachlorophenol	17.354	266	79813	24.645	ng/ul	96
72) Phenanthrene	17.783	178	657832	20.643	ng/ul	98
74) Anthracene	17.877	178	664372	20.648	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.787	216	140876	21.581	ng/ul	87
76) Pentachlorobenzene	15.280	250	132153	21.335	ng/ul	97
77) Carbazole	18.147	167	638670	23.106	ng/ul	96
78) Di-n-butylphthalate	18.664	149	806357	21.232	ng/ul	99
80) Fluoranthene	19.775	202	735188	21.376	ng/ul	97
82) Pyrene	20.139	202	750563	21.010	ng/ul	98
83) Butylbenzylphthalate	21.008	149	378727	23.474	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.960	252	255595	24.214	ng/ul	96
85) Benzo(a)anthracene	22.054	228	723774	21.668	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.901	149	548055	22.766	ng/ul	98
87) Chrysene	22.125	228	670099	21.565	ng/ul	95
89) Di-n-octyl phthalate	23.259	149	926079	43.992	ng/ul	100
90) Benzo(b)fluoranthene	24.592	252	711124	40.802	ng/ul	100
91) Benzo(k)fluoranthene	24.592	252	713866	41.006	ng/ul	99
93) Benzo(a)pyrene	25.515	252	656381	39.847	ng/ul	95
94) Indeno(1,2,3-cd)pyrene	29.869	276	726519	34.796	ng/ul	95
95) Dibenzo(a,h)anthracene	29.963	278	696752	40.405	ng/ul	96
96) Benzo(g,h,i)perylene	31.179	276	314261	18.390	ng/ul#	91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 3.987 ng/uL
RT: 3.705 min Scan# 54
Delta R.T. 0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion: 88 Resp: 8805
Ion Ratio Lower Upper
88 100
43 12.9 15.7 23.5#
58 34.0 36.2 54.2#



#5
Pyridine
Concen: 22.316 ng/uL
RT: 4.116 min Scan# 124
Delta R.T. 0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion: 79 Resp: 121378
Ion Ratio Lower Upper
79 100
52 51.3 53.3 79.9#
51 29.7 22.4 33.6









#6
Benzaldehyde
Concen: 26.695 ng/ul
RT: 7.507 min Scan# 700
Delta R.T. 0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion: 77 Resp: 91222
Ion Ratio Lower Upper
77 100
105 101.3 83.6 125.4
106 98.1 78.6 118.0



#8
Phenol
Concen: 22.940 ng/ul
RT: 7.495 min Scan# 699
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion: 94 Resp: 164076
Ion Ratio Lower Upper
94 100
65 30.1 22.5 33.7
66 47.4 31.1 46.7#









#10
Bis(2-Chloroethyl)ether
Concen: 23.517 ng/ul
RT: 7.777 min Scan# 747
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion: 93 Resp: 123348
Ion Ratio Lower Upper
93 100
63 62.7 52.5 78.7
95 35.1 26.5 39.7



#12
2-Chlorophenol
Concen: 23.214 ng/ul
RT: 7.906 min Scan# 769
Delta R.T. -0.009 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:128 Resp: 141647
Ion Ratio Lower Upper
128 100
64 33.4 30.2 45.2
130 33.6 27.9 41.9





#13

2-Methylphenol

Concen: 23.164 ng/ul

RT: 8.776 min Scan# 917

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :





SSTD020460

Tgt Ion:108 Resp: 138114

Ion Ratio Lower Upper

108 100

107 81.2 68.6 102.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.791 ng/ul

RT: 8.870 min Scan# 933

Delta R.T. -0.009 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion: 45 Resp: 102622


Ion Ratio Lower Upper

45 100

77 25.6 15.0 22.6#

79 13.6 10.0 15.0









#16
Acetophenone
Concen: 23.134 ng/ul
RT: 9.205 min Scan# 990
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:105 Resp: 219489
Ion Ratio Lower Upper
105 100
77 72.4 59.6 89.4
51 19.0 16.2 24.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.528 ng/ul
RT: 9.169 min Scan# 984
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion: 70 Resp: 96241
Ion Ratio Lower Upper
70 100
42 46.4 33.8 50.8
101 11.9 5.4 8.0#
130 28.5 18.2 27.4#









#18
4-Methylphenol
Concen: 23.280 ng/ul
RT: 9.105 min Scan# 91
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:108 Resp: 147687
Ion Ratio Lower Upper
108 100
107 101.8 88.9 133.3



#19
Hexachloroethane
Concen: 23.201 ng/ul
RT: 9.451 min Scan# 1032
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:117 Resp: 56226
Ion Ratio Lower Upper
117 100
201 99.4 75.8 113.6
199 61.0 42.3 63.5









#22
Nitrobenzene
Concen: 25.761 ng/ul
RT: 9.598 min Scan# 1057
Delta R.T. -0.009 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion: 77 Resp: 144699
Ion Ratio Lower Upper
77 100
123 46.1 42.6 63.8
65 13.3 11.0 16.4



#23
Isophorone
Concen: 20.843 ng/ul
RT: 10.115 min Scan# 1145
Delta R.T. 0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion: 82 Resp: 294829
Ion Ratio Lower Upper
82 100
95 8.8 5.6 8.4#
138 18.3 12.3 18.5









#25
2-Nitrophenol
Concen: 26.506 ng/ul
RT: 10.315 min Scan# 1179
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:139 Resp: 76156
Ion Ratio Lower Upper
139 100
65 53.8 39.7 59.5
109 30.6 21.0 31.6



#26
2,4-Dimethylphenol
Concen: 22.685 ng/ul
RT: 10.333 min Scan# 1182
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:107 Resp: 171438
Ion Ratio Lower Upper
107 100
121 48.5 42.7 64.1
122 77.4 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.024 ng/ul
 RT: 10.597 min Scan# 11
 Delta R.T. -0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020460


Tgt Ion: 93 Resp: 165588
 Ion Ratio Lower Upper
 93 100
 95 30.2 27.1 40.7
 123 14.1 10.3 15.5



#29
 2,4-Dichlorophenol
 Concen: 23.144 ng/ul
 RT: 10.832 min Scan# 1267
 Delta R.T. 0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23

Tgt Ion:162 Resp: 131791
 Ion Ratio Lower Upper
 162 100
 164 65.9 45.3 67.9
 98 32.5 32.2 48.4









#30
Naphthalene
Concen: 21.206 ng/ul
RT: 11.267 min Scan# 11
Delta R.T. 0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:128 Resp: 483723
Ion Ratio Lower Upper
128 100
129 10.3 7.4 11.0
127 12.7 10.5 15.7



#32
4-Chloroaniline
Concen: 22.533 ng/ul
RT: 11.373 min Scan# 1359
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:127 Resp: 207383
Ion Ratio Lower Upper
127 100
129 31.2 30.4 45.6





#33

Hexachlorobutadiene

Concen: 21.425 ng/ul

RT: 11.496 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:225 Resp: 73191

Ion Ratio Lower Upper

225 100

223 60.0 50.8 76.2

227 68.8 49.8 74.8



#34

Caprolactam

Concen: 15.087 ng/ul

RT: 12.142 min Scan# 1490

Delta R.T. -0.014 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:113 Resp: 31429


Ion Ratio Lower Upper

113 100

55 100.1 86.9 130.3

56 75.4 64.4 96.6









#35
 4-Chloro-3-methylphenol
 Concen: 22.725 ng/ul
 RT: 12.436 min Scan# 11
 Delta R.T. -0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020460


Tgt Ion:107 Resp: 154404
 Ion Ratio Lower Upper
 107 100
 144 27.5 21.0 31.6
 142 74.9 68.2 102.4



#36
 2-Methylnaphthalene
 Concen: 21.382 ng/ul
 RT: 12.842 min Scan# 1609
 Delta R.T. -0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23

Tgt Ion:142 Resp: 335559
 Ion Ratio Lower Upper
 142 100
 141 88.4 71.7 107.5





#37

1-Methylnaphthalene

Concen: 21.427 ng/ul

RT: 13.059 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:142 Resp: 334927

Ion Ratio Lower Upper

142 100

141 93.5 73.1 109.7

116 4.6 3.3 4.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.904 ng/ul

RT: 13.182 min Scan# 1667

Delta R.T. 0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Tgt Ion:216 Resp: 137048




Ion Ratio Lower Upper


216 100

214 77.8 63.7 95.5

179 22.9 19.2 28.8

108 22.5 16.5 24.7









#40
Hexachlorocyclopentadiene
Concen: 22.076 ng/ul
RT: 13.135 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:237 Resp: 84369
Ion Ratio Lower Upper
237 100
235 56.8 46.6 69.8
272 12.9 13.0 19.4#



#41
2,4,6-Trichlorophenol
Concen: 21.195 ng/ul
RT: 13.417 min Scan# 1707
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:196 Resp: 96218
Ion Ratio Lower Upper
196 100
198 94.2 69.6 104.4
200 33.0 23.8 35.8





#42

2,4,5-Trichlorophenol

Concen: 21.841 ng/ul

RT: 13.482 min Scan# 1718

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:196 Resp: 107348

Ion Ratio Lower Upper

196 100

198 97.7 79.8 119.6

200 35.4 28.0 42.0



#43

1,1'-Biphenyl

Concen: 20.704 ng/ul

RT: 13.829 min Scan# 1777

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:154 Resp: 425748


Ion Ratio Lower Upper

154 100

153 43.1 30.7 46.1

76 10.1 8.2 12.4





#44

2-Chloronaphthalene

Concen: 20.682 ng/ul

RT: 13.881 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:162 Resp: 330542

Ion Ratio Lower Upper

162 100

164 30.9 23.9 35.9

127 45.9 35.3 52.9



#45

2-Nitroaniline

Concen: 25.801 ng/ul

RT: 14.087 min Scan# 1821

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion: 65 Resp: 76726


Ion Ratio Lower Upper

65 100

92 79.2 59.6 89.4

138 129.9 109.0 163.6









#47
Dimethylphthalate
Concen: 20.334 ng/ul
RT: 14.434 min Scan# 1880
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:163 Resp: 413713
Ion Ratio Lower Upper
163 100
194 2.5 2.6 3.8#
164 11.3 6.6 10.0#



#48
2,6-Dinitrotoluene
Concen: 22.380 ng/ul
RT: 14.569 min Scan# 1903
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:165 Resp: 68689
Ion Ratio Lower Upper
165 100
89 73.1 41.0 61.4#
121 32.0 12.3 18.5#









#50
Acenaphthylene
Concen: 20.071 ng/ul
RT: 14.722 min Scan# 1929
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument : BNA_G
ClientSampleId : SSTD020460


Tgt Ion:152 Resp: 558891
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0
153 12.8 9.4 14.2



#51
3-Nitroaniline
Concen: 24.864 ng/ul
RT: 14.904 min Scan# 1960
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:138 Resp: 95121
Ion Ratio Lower Upper
138 100
108 10.9 17.1 25.7#
92 97.8 101.8 152.8#









#52
Acenaphthene
Concen: 19.619 ng/ul
RT: 15.051 min Scan# 1985
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:153 Resp: 356103
Ion Ratio Lower Upper
153 100
152 48.9 37.6 56.4
154 86.3 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 23.515 ng/ul
RT: 15.104 min Scan# 1994
Delta R.T. 0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:184 Resp: 41007
Ion Ratio Lower Upper
184 100
63 40.7 35.0 52.6
154 65.7 49.8 74.8









#55
4-Nitrophenol
Concen: 25.810 ng/ul
RT: 15.168 min Scan# 2004
Delta R.T. 0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:109 Resp: 100450
Ion Ratio Lower Upper
109 100
139 88.5 72.2 108.2
65 69.2 57.3 85.9



#56
Dibenzofuran
Concen: 20.647 ng/ul
RT: 15.386 min Scan# 2042
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:168 Resp: 485873
Ion Ratio Lower Upper
168 100
139 45.3 33.8 50.8









#57
2,4-Dinitrotoluene
Concen: 25.256 ng/ul
RT: 15.344 min Scan# 2035
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:165 Resp: 109036
Ion Ratio Lower Upper
165 100
63 41.0 30.9 46.3
182 2.2 1.8 2.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.347 ng/ul
RT: 15.585 min Scan# 2076
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:232 Resp: 90061
Ion Ratio Lower Upper
232 100
131 64.2 44.2 66.2
130 4.1 2.3 3.5#
166 34.0 32.6 48.8









#59
Diethylphthalate
Concen: 20.538 ng/ul
RT: 15.779 min Scan# 2109
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:149 Resp: 455819
Ion Ratio Lower Upper
149 100
177 19.7 16.3 24.5
150 12.7 9.1 13.7



#61
Fluorene
Concen: 20.177 ng/ul
RT: 16.032 min Scan# 2152
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:166 Resp: 409894
Ion Ratio Lower Upper
166 100
165 107.3 74.4 111.6
167 13.0 9.1 13.7









#62
4-Chlorophenyl-phenylether
Concen: 20.654 ng/ul
RT: 16.020 min Scan# 2150
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:204 Resp: 179320
Ion Ratio Lower Upper
204 100
206 31.7 28.8 43.2
141 78.7 68.6 103.0



#63
4-Nitroaniline
Concen: 22.923 ng/ul
RT: 16.067 min Scan# 2158
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:138 Resp: 93065
Ion Ratio Lower Upper
138 100
92 55.9 49.8 74.8
108 147.7 100.9 151.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 25.295 ng/ul
 RT: 16.091 min Scan# 2162
 Delta R.T. -0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020460


Tgt Ion:198 Resp: 61884
 Ion Ratio Lower Upper
 198 100
 182 3.7 1.4 2.0#
 77 28.2 15.8 23.8#



#67
 N-Nitrosodiphenylamine
 Concen: 20.884 ng/ul
 RT: 16.232 min Scan# 2186
 Delta R.T. -0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23

Tgt Ion:169 Resp: 349795
 Ion Ratio Lower Upper
 169 100
 168 68.1 54.2 81.2
 167 39.3 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 20.291 ng/ul
RT: 16.913 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:248 Resp: 99530
Ion Ratio Lower Upper
248 100
250 89.2 74.2 111.2
141 117.0 88.5 132.7



#69
Hexachlorobenzene
Concen: 21.037 ng/ul
RT: 17.001 min Scan# 2317
Delta R.T. -0.003 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:284 Resp: 119255
Ion Ratio Lower Upper
284 100
142 49.7 39.6 59.4
249 29.6 28.2 42.2





#70

Atrazine

Concen: 22.907 ng/ul

RT: 17.166 min Scan# 21

Delta R.T. -0.009 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:200 Resp: 118889

Ion Ratio Lower Upper

200 100

173 26.8 25.7 38.5

215 47.0 32.1 48.1



#71

Pentachlorophenol

Concen: 24.645 ng/ul

RT: 17.354 min Scan# 2377

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:266 Resp: 79813


Ion Ratio Lower Upper

266 100

264 62.6 46.2 69.2

268 58.9 46.4 69.6





#72

Phenanthrene

Concen: 20.643 ng/ul

RT: 17.783 min Scan# 2450

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:178 Resp: 657832

Ion Ratio Lower Upper

178 100

179 15.3 11.3 16.9

176 20.1 15.4 23.0



#74

Anthracene

Concen: 20.648 ng/ul

RT: 17.877 min Scan# 2466

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:178 Resp: 664372


Ion Ratio Lower Upper

178 100

179 15.0 13.1 19.7

176 17.4 13.9 20.9





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.581 ng/uL

RT: 13.787 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:216 Resp: 140876

Ion Ratio Lower Upper

216 100

214 72.4 70.2 105.2

218 45.7 39.9 59.9



#76

Pentachlorobenzene

Concen: 21.335 ng/uL

RT: 15.280 min Scan# 2024

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:250 Resp: 132153


Ion Ratio Lower Upper

250 100

248 56.9 48.7 73.1

252 65.3 51.2 76.8









#77
 Carbazole
 Concen: 23.106 ng/ul
 RT: 18.147 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020460


Tgt Ion:167 Resp: 638670
 Ion Ratio Lower Upper
 167 100
 166 21.0 19.4 29.2
 139 14.9 11.7 17.5



#78
 Di-n-butylphthalate
 Concen: 21.232 ng/ul
 RT: 18.664 min Scan# 2600
 Delta R.T. -0.003 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23

Tgt Ion:149 Resp: 806357
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.5 11.3
 104 7.1 5.3 7.9





#80

Fluoranthene

Concen: 21.376 ng/ul

RT: 19.775 min Scan# 21

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:202 Resp: 735188

Ion Ratio Lower Upper

202 100

101 7.9 7.0 10.6

100 7.7 5.2 7.8



#82

Pyrene

Concen: 21.010 ng/ul

RT: 20.139 min Scan# 2851

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:202 Resp: 750563


Ion Ratio Lower Upper

202 100

101 9.2 7.9 11.9

100 9.4 7.9 11.9









#83
 Butylbenzylphthalate
 Concen: 23.474 ng/ul
 RT: 21.008 min Scan# 21
 Delta R.T. -0.009 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020460


Tgt Ion:149 Resp: 378727
 Ion Ratio Lower Upper
 149 100
 91 70.8 52.2 78.2
 206 16.7 15.1 22.7



#84
 3,3'-Dichlorobenzidine
 Concen: 24.214 ng/ul
 RT: 21.960 min Scan# 3161
 Delta R.T. -0.009 min
 Lab File: BG059530.D
 Acq: 28 Oct 2023 21:23

Tgt Ion:252 Resp: 255595
 Ion Ratio Lower Upper
 252 100
 254 64.5 48.9 73.3
 126 13.1 11.5 17.3





#85

Benzo(a)anthracene

Concen: 21.668 ng/ul

RT: 22.054 min Scan# 3177

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:228 Resp: 723774

Ion Ratio Lower Upper

228 100

229 18.7 16.7 25.1

226 23.9 20.3 30.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.766 ng/ul

RT: 21.901 min Scan# 3151

Delta R.T. -0.014 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:149 Resp: 548055


Ion Ratio Lower Upper

149 100

167 23.3 19.4 29.2

279 2.9 2.2 3.4





#87

Chrysene

Concen: 21.565 ng/ul

RT: 22.125 min Scan# 31

Delta R.T. -0.009 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:228 Resp: 670099

Ion Ratio Lower Upper

228 100

226 29.9 21.0 31.4

229 19.7 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 43.992 ng/ul




RT: 23.259 min Scan# 3382


Delta R.T. -0.014 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Tgt Ion:149 Resp: 926079









#90
Benzo(b)fluoranthene
Concen: 40.802 ng/ul
RT: 24.592 min Scan# 3609
Delta R.T. 0.068 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23




Instrument :
BNA_G
ClientSampleId :
SSTD020460


Tgt Ion:252 Resp: 711124
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 11.0 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 41.006 ng/ul
RT: 24.592 min Scan# 3609
Delta R.T. -0.009 min
Lab File: BG059530.D
Acq: 28 Oct 2023 21:23

Tgt Ion:252 Resp: 713866
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 11.0 7.9 11.9





#93

Benzo(a)pyrene

Concen: 39.847 ng/ul

RT: 25.515 min Scan# 31

Delta R.T. -0.003 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:252 Resp: 656381

Ion Ratio Lower Upper

252 100

253 19.8 18.5 27.7

125 10.2 7.6 11.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 34.796 ng/ul

RT: 29.869 min Scan# 4507

Delta R.T. -0.020 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23




Tgt Ion:276 Resp: 726519


Ion Ratio Lower Upper

276 100

138 16.1 15.3 22.9

277 22.5 16.4 24.6





#95

Dibenzo(a,h)anthracene

Concen: 40.405 ng/ul

RT: 29.963 min Scan# 4

Delta R.T. -0.020 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Instrument :

BNA_G

ClientSampleId :

SSTD020460





Tgt Ion:278 Resp: 696752

Ion Ratio Lower Upper

278 100

139 12.1 11.0 16.6

279 21.1 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 18.390 ng/ul

RT: 31.179 min Scan# 4730

Delta R.T. -0.038 min

Lab File: BG059530.D

Acq: 28 Oct 2023 21:23

Tgt Ion:276 Resp: 314261

Ion Ratio Lower Upper

276 100

138 13.7 16.2 24.2#

277 20.5 18.2 27.4

