

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102825\
 Data File : BG064614.D
 Acq On : 28 Oct 2025 23:20
 Operator : RC/JU
 Sample : Q3460-17 5X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-5

Quant Time: Oct 29 01:00:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 16:40:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.219	152	38929	20.000	ng	0.00
21) Naphthalene-d8	11.045	136	159793	20.000	ng	0.00
39) Acenaphthene-d10	14.841	164	141528	20.000	ng	0.00
64) Phenanthrene-d10	17.573	188	365617	20.000	ng	0.00
76) Chrysene-d12	21.850	240	316306	20.000	ng	-0.01
86) Perylene-d12	25.187	264	344008	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.745	112	41823	18.033	ng	0.00
7) Phenol-d6	7.349	99	56079	17.621	ng	0.00
23) Nitrobenzene-d5	9.376	82	32337	12.155	ng	-0.01
42) 2,4,6-Tribromophenol	16.315	330	35142	17.204	ng	0.00
45) 2-Fluorobiphenyl	13.466	172	96509	10.336	ng	-0.01
79) Terphenyl-d14	20.164	244	177616	10.597	ng	-0.01


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102825\
Data File : BG064614.D
Acq On : 28 Oct 2025 23:20
Operator : RC/JU
Sample : Q3460-17 5X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-5

Quant Time: Oct 29 01:00:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.219 min Scan# 81
Delta R.T. -0.002 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20




Instrument :
BNA_G
ClientSampleId :
WC-5


Tgt Ion:152 Resp: 38929
Ion Ratio Lower Upper
152 100
150 147.9 122.2 183.4
115 57.9 50.8 76.2



#5
2-Fluorophenol
Concen: 18.033 ng
RT: 5.745 min Scan# 452
Delta R.T. -0.002 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20

Tgt Ion:112 Resp: 41823
Ion Ratio Lower Upper
112 100
64 77.8 60.7 91.1
63 41.7 33.4 50.0









#7
Phenol-d6
Concen: 17.621 ng
RT: 7.349 min Scan# 71
Delta R.T. -0.007 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20




Instrument :
BNA_G
ClientSampleId :
WC-5


Tgt Ion: 99 Resp: 56079
Ion Ratio Lower Upper
99 100
42 24.2 18.3 27.5
71 39.3 28.6 43.0



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.045 min Scan# 1354
Delta R.T. -0.007 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20

Tgt Ion:136 Resp: 159793
Ion Ratio Lower Upper
136 100
137 10.9 9.2 13.8
54 16.7 12.6 19.0
68 6.1 5.7 8.5





#23

Nitrobenzene-d5

Concen: 12.155 ng

RT: 9.376 min Scan# 1072

Delta R.T. -0.014 min

Lab File: BG064614.D

Acq: 28 Oct 2025 23:20

Instrument :

BNA_G

ClientSampleId :

WC-5





Tgt Ion: 82 Resp: 32337

Ion Ratio Lower Upper

82 100

128 38.3 31.3 46.9

54 73.8 56.6 84.8



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.841 min Scan# 2000

Delta R.T. -0.002 min

Lab File: BG064614.D

Acq: 28 Oct 2025 23:20




Tgt Ion:164 Resp: 141528


Ion Ratio Lower Upper

164 100

162 100.7 77.0 115.6

160 45.5 34.4 51.6









#42
2,4,6-Tribromophenol
Concen: 17.204 ng
RT: 16.315 min Scan# 21
Delta R.T. -0.008 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20




Instrument :
BNA_G
ClientSampleId :
WC-5


Tgt Ion:330 Resp: 35142
Ion Ratio Lower Upper
330 100
332 96.4 78.0 117.0
141 27.1 21.8 32.6



#45
2-Fluorobiphenyl
Concen: 10.336 ng
RT: 13.466 min Scan# 1766
Delta R.T. -0.013 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20

Tgt Ion:172 Resp: 96509
Ion Ratio Lower Upper
172 100
171 35.1 29.4 44.0
170 22.5 19.1 28.7









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.573 min Scan# 2465
Delta R.T. -0.007 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20




Instrument :
BNA_G
ClientSampleId :
WC-5


Tgt Ion:188 Resp: 365617
Ion Ratio Lower Upper
188 100
94 7.4 5.7 8.5
80 8.4 6.8 10.2



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.850 min Scan# 3193
Delta R.T. -0.013 min
Lab File: BG064614.D
Acq: 28 Oct 2025 23:20

Tgt Ion:240 Resp: 316306
Ion Ratio Lower Upper
240 100
120 6.7 6.5 9.7
236 26.8 20.2 30.2









#79
 Terphenyl-d14
 Concen: 10.597 ng
 RT: 20.164 min Scan# 2906
 Delta R.T. -0.013 min
 Lab File: BG064614.D
 Acq: 28 Oct 2025 23:20

Instrument :
 BNA_G
 ClientSampleId :
 WC-5

Tgt Ion:244 Resp: 177616

Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.1	9.1
122	6.1	5.6	8.4



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.187 min Scan# 3761
 Delta R.T. -0.013 min
 Lab File: BG064614.D
 Acq: 28 Oct 2025 23:20

Tgt Ion:264 Resp: 344008

Ion	Ratio	Lower	Upper
264	100		
260	24.5	19.0	28.4
265	22.0	17.6	26.4

