

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG102918\
 Data File : BG037625.D
 Acq On : 29 Oct 2018 16:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 17:42:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.607	0.628	-3.5	112	0.00
3	Pyridine	1.529	1.506	1.5	105	0.00
4	n-Nitrosodimethylamine	0.545	0.567	-4.0	112	0.00
5 S	2-Fluorophenol	1.196	1.219	-1.9	108	0.00
6	Aniline	2.207	2.216	-0.4	107	0.00
7 S	Phenol-d6	1.809	1.833	-1.3	107	0.00
8	2-Chlorophenol	1.399	1.427	-2.0	108	0.00
9	Benzaldehyde	1.132	1.029	9.1	97	0.00
10 C	Phenol	1.909	1.941	-1.7	108	0.00
11	bis(2-Chloroethyl)ether	1.450	1.430	1.4	106	0.00
12	1,3-Dichlorobenzene	1.558	1.545	0.8	107	0.00
13 C	1,4-Dichlorobenzene	1.594	1.574	1.3	106	0.00
14	1,2-Dichlorobenzene	1.533	1.486	3.1	104	0.00
15	Benzyl Alcohol	1.337	1.359	-1.6	105	0.00
16	2,2'-oxybis(1-Chloropropane	2.594	2.592	0.1	107	0.00
17	2-Methylphenol	1.262	1.283	-1.7	106	0.00
18	Hexachloroethane	0.543	0.556	-2.4	109	0.00
19 P	n-Nitroso-di-n-propylamine	1.246	1.246	0.0	103	0.00
20	3+4-Methylphenols	1.804	1.827	-1.3	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
22	Acetophenone	0.502	0.488	2.8	103	0.00
23 S	Nitrobenzene-d5	0.357	0.363	-1.7	107	0.00
24	Nitrobenzene	0.371	0.376	-1.3	106	0.00
25	Isophorone	0.652	0.651	0.2	103	0.00
26 C	2-Nitrophenol	0.167	0.191	-14.4	117	0.00
27	2,4-Dimethylphenol	0.298	0.295	1.0	104	0.00
28	bis(2-Chloroethoxy)methane	0.427	0.420	1.6	103	0.00
29 C	2,4-Dichlorophenol	0.332	0.342	-3.0	108	0.00
30	1,2,4-Trichlorobenzene	0.361	0.358	0.8	107	0.00
31	Naphthalene	0.982	0.964	1.8	106	0.00
32	Benzoic acid	0.194	0.245	-26.3#	130	0.00
33	4-Chloroaniline	0.462	0.462	0.0	105	0.00
34 C	Hexachlorobutadiene	0.214	0.214	0.0	105	0.00
35	Caprolactam	0.133	0.137	-3.0	105	0.00
36 C	4-Chloro-3-methylphenol	0.375	0.380	-1.3	106	0.00
37	2-Methylnaphthalene	0.716	0.702	2.0	103	0.00
38	1-Methylnaphthalene	0.695	0.678	2.4	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	0.645	0.656	-1.7	106	0.00
41 P	Hexachlorocyclopentadiene	0.308	0.310	-0.6	102	0.00
42 S	2,4,6-Tribromophenol	0.218	0.227	-4.1	103	0.00
43 C	2,4,6-Trichlorophenol	0.431	0.453	-5.1	106	0.00
44	2,4,5-Trichlorophenol	0.469	0.494	-5.3	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG102918\
 Data File : BG037625.D
 Acq On : 29 Oct 2018 16:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 17:42:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.326	1.311	1.1	104	0.00
46	1,1'-Biphenyl	1.656	1.650	0.4	105	0.00
47	2-Chloronaphthalene	1.253	1.248	0.4	104	0.00
48	2-Nitroaniline	0.380	0.406	-6.8	109	0.00
49	Acenaphthylene	1.885	1.900	-0.8	104	0.00
50	Dimethylphthalate	1.547	1.549	-0.1	104	0.00
51	2,6-Dinitrotoluene	0.342	0.360	-5.3	106	0.00
52 C	Acenaphthene	1.161	1.149	1.0	105	0.00
53	3-Nitroaniline	0.380	0.396	-4.2	104	0.00
54 P	2,4-Dinitrophenol	0.100	0.122	-22.0	128	0.00
55	Dibenzofuran	1.923	1.898	1.3	104	0.00
56 P	4-Nitrophenol	0.281	0.288	-2.5	103	0.00
57	2,4-Dinitrotoluene	0.449	0.497	-10.7	109	0.00
58	Fluorene	1.440	1.419	1.5	104	0.00
59	2,3,4,6-Tetrachlorophenol	0.402	0.411	-2.2	105	0.00
60	Diethylphthalate	1.588	1.601	-0.8	105	0.00
61	4-Chlorophenyl-phenylether	0.840	0.825	1.8	103	0.00
62	4-Nitroaniline	0.378	0.385	-1.9	102	0.00
63	Azobenzene	1.516	1.486	2.0	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	0.093	0.110	-18.3	121	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.601	0.2	103	0.00
67	4-Bromophenyl-phenylether	0.220	0.221	-0.5	105	0.00
68	Hexachlorobenzene	0.228	0.223	2.2	102	0.00
69	Atrazine	0.218	0.214	1.8	99	0.00
70 C	Pentachlorophenol	0.152	0.154	-1.3	102	0.00
71	Phenanthrene	1.016	0.986	3.0	102	0.00
72	Anthracene	0.998	0.969	2.9	101	0.00
73	Carbazole	0.998	0.985	1.3	102	0.00
74	Di-n-butylphthalate	1.110	1.125	-1.4	103	0.00
75 C	Fluoranthene	1.153	1.104	4.2	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzidine	0.680	0.645	5.1	92	0.00
78	Pyrene	1.307	1.319	-0.9	101	0.00
79 S	Terphenyl-d14	0.936	0.930	0.6	100	0.00
80	Butylbenzylphthalate	0.535	0.586	-9.5	105	0.00
81	Benzo(a)anthracene	1.183	1.194	-0.9	101	0.00
82	3,3'-Dichlorobenzidine	0.459	0.484	-5.4	102	0.00
83	Chrysene	1.138	1.143	-0.4	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.750	0.822	-9.6	105	0.00
85 c	Di-n-octyl phthalate	1.223	1.342	-9.7	105	-0.01
86	Indeno(1,2,3-cd)pyrene	1.220	1.200	1.6	96	0.00
87 I	Perylene-d12	1.000	1.000	0.0	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG102918\
Data File : BG037625.D
Acq On : 29 Oct 2018 16:34
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 17:42:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 14:06:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.096	1.097	-0.1	101	0.00
89	Benzo(k)fluoranthene	1.074	1.073	0.1	101	0.00
90 C	Benzo(a)pyrene	1.042	1.047	-0.5	101	0.00
91	Dibenzo(a,h)anthracene	0.961	0.918	4.5	94	0.00
92	Benzo(q,h,i)perylene	0.972	0.925	4.8	96	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0