

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102920\
 Data File : BG047179.D
 Acq On : 31 Oct 2020 7:43
 Operator : CG/JU
 Sample : L4507-16
 Misc : GCMS CONFIRMATION
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0BN3

Quant Time: Nov 01 23:20:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG102220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 29 16:49:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	63168	20.00	ng/ul	0.00
18) Naphthalene-d8	10.86	136	234835	20.00	ng/ul	-0.02
36) Acenaphthene-d10	14.68	164	164456	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.43	188	397438	20.00	ng/ul	0.00
78) Chrysene-d12	21.69	240	414590	20.00	ng/ul	-0.02
86) Perylene-d12	24.93	264	423104	20.00	ng/ul	-0.02

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL
5) Phenol-d5	0.00	99	0	0.00	ng/ul
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul
44) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul
47) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
58) Fluorene-d10	0.00	176	0	0.00	ng/ul
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul
71) Anthracene-d10	0.00	188	0d	0.00	ng/ul
79) Pyrene-d10	0.00	212	0	0.00	ng/ul
90) Benzo(a)pyrene-d12	0.00	264	0	0.00	ng/ul

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102920\
Data File : BG047179.D
Acq On : 31 Oct 2020 7:43
Operator : CG/JU
Sample : L4507-16
Misc : GCMS CONFIRMATION
ALS Vial : 69 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0BN3

Quant Time: Nov 01 23:20:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG102220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 29 16:49:45 2020
Response via : Initial Calibration

