


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG102920\
 Data File : BG047110.D
 Acq On : 29 Oct 2020 8:15
 Operator : CG/JU
 Sample : DFTPP47
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP47

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG102220MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Oct 29 16:49:45 2020



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.5	51512	PASS
68	69	0.00	2	0.4	246	PASS
69	198	0.00	100	33.7	55146	PASS
70	69	0.00	2	0.4	218	PASS
127	198	10	80	40.2	65829	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	163565	PASS
199	198	5	9	7.1	11631	PASS
275	198	10	60	30.2	49394	PASS
365	198	1	100	4.0	6559	PASS
441	198	0.01	100	13.4	21880	PASS
442	198	50	100	81.6	133421	PASS
443	442	15	24	19.1	25458	PASS

Scan 2465 (17.824 min): BG047110.D
DFTPP47

m/z	Abundance
40.00	287.0
44.00	478.0
73.10	205.0
206.90	551.0
273.20	172.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

m/z	Abundance
37.10	249.0
38.10	1014.0
39.10	6069.0
40.10	378.0
41.20	390.0
44.10	954.0
45.20	221.0
49.00	492.0
50.10	17224.0
51.10	64440.0
52.10	3540.0
55.10	254.0
56.00	1985.0
57.00	4717.0
58.10	246.0
59.20	213.0
61.00	708.0
62.00	1386.0
63.10	3402.0
64.10	702.0
65.10	962.0
69.00	71560.0
69.90	356.0
73.10	830.0
74.10	7852.0
75.10	11547.0
76.10	3259.0
77.10	73744.0
78.10	4246.0
79.00	4995.0
80.10	4294.0
81.00	5868.0
82.00	918.0
83.10	1386.0
85.00	801.0
86.10	1886.0
87.00	1221.0
87.80	174.0
88.10	237.0
89.00	175.0
91.10	1795.0
92.00	1373.0
93.00	10256.0
94.10	536.0
94.90	192.0
96.10	401.0
96.90	239.0
98.00	6879.0
99.10	5321.0
100.00	596.0
101.00	2960.0
102.00	364.0
103.00	996.0
104.10	2186.0
105.10	1766.0
107.00	26344.0
108.00	4449.0
109.10	1289.0
110.00	40464.0
111.00	6825.0
112.00	691.0
113.00	154.0
116.00	1473.0
117.00	20920.0
118.00	1747.0
119.10	250.0
119.90	241.0
120.20	168.0
122.10	1763.0
122.90	2321.0
123.90	1260.0
125.00	1110.0
127.00	82136.0
128.00	6749.0
129.00	34488.0
130.00	3093.0
130.90	690.0
132.20	323.0
132.80	176.0
134.00	946.0
135.10	3509.0
136.10	1406.0
137.10	1051.0
137.80	307.0
139.90	402.0
141.00	5151.0
142.00	1880.0
143.10	1294.0
144.10	288.0
145.00	178.0
146.00	1124.0
147.00	2514.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

148.00	5295.0
149.10	1305.0
150.10	253.0
151.20	456.0
152.00	361.0
153.00	1935.0
154.00	987.0
155.00	3114.0
156.10	4167.0
157.00	582.0
157.80	1132.0
159.00	670.0
160.10	1371.0
161.00	2232.0
162.10	505.0
162.90	162.0
164.00	159.0
165.00	2107.0
166.00	1757.0
167.00	8098.0
168.10	4737.0
169.00	662.0
169.90	719.0
171.00	612.0
171.90	1268.0
173.10	1264.0
174.00	1911.0
175.10	4045.0
176.10	917.0
177.10	1767.0
178.00	687.0
179.00	7890.0
180.00	4760.0
181.00	2049.0
182.00	696.0
184.20	751.0
185.00	3083.0
186.10	25448.0
187.00	6791.0
188.10	818.0
189.00	1533.0
190.00	400.0
191.10	777.0
192.10	2648.0
193.00	2470.0
194.10	268.0
195.10	402.0
196.00	4961.0
198.00	183232.0
199.00	13542.0
200.00	1048.0
201.40	812.0
202.00	345.0
203.10	1624.0
204.10	7028.0
205.00	12968.0
206.00	52512.0
207.00	7817.0
208.00	2278.0
209.10	603.0
210.30	756.0
211.00	2225.0
213.20	173.0
213.90	229.0
215.00	698.0
216.00	1272.0
217.00	13685.0
218.00	1964.0
219.00	231.0
219.80	240.0
221.00	7019.0
221.80	2197.0
223.10	3209.0
224.10	32416.0
225.10	8664.0
226.00	607.0
227.00	14460.0
228.10	1820.0
229.00	2935.0
230.00	388.0
231.00	1271.0
232.00	218.0
232.90	602.0
233.90	910.0
235.00	926.0
236.10	640.0
237.00	959.0
238.00	260.0
239.10	325.0
239.90	538.0
241.00	710.0
241.90	1870.0
243.00	1577.0
244.00	23456.0
245.00	3335.0
246.00	5305.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

247.00	905.0
247.90	369.0
249.00	933.0
250.20	169.0
251.10	191.0
252.00	495.0
252.80	670.0
253.10	663.0
255.00	124384.0
256.00	18008.0
257.10	1147.0
258.00	7120.0
259.00	1413.0
259.90	222.0
261.30	258.0
265.00	2995.0
266.00	526.0
270.90	426.0
272.00	290.0
273.00	4063.0
274.00	9214.0
275.00	52224.0
276.00	6877.0
277.10	4407.0
278.00	889.0
279.10	314.0
281.90	262.0
283.20	471.0
284.10	462.0
285.10	691.0
289.00	366.0
290.10	226.0
292.00	212.0
293.00	868.0
293.80	239.0
296.00	15044.0
297.00	2238.0
298.10	237.0
301.00	455.0
302.10	427.0
303.10	1446.0
304.00	191.0
308.10	168.0
314.00	700.0
315.10	917.0
316.00	1065.0
321.10	268.0
322.00	334.0
323.00	4362.0
324.00	704.0
326.20	182.0
327.00	967.0
328.00	330.0
333.10	557.0
334.10	3321.0
335.10	994.0
335.90	252.0
341.10	617.0
346.10	957.0
347.00	315.0
352.00	1466.0
353.00	924.0
354.00	1440.0
355.00	157.0
357.00	160.0
365.00	6112.0
365.90	803.0
371.10	229.0
372.10	1834.0
373.10	857.0
374.00	239.0
382.90	527.0
384.20	186.0
389.90	185.0
390.90	338.0
402.10	904.0
403.00	903.0
404.10	428.0
421.10	780.0
422.00	999.0
423.00	6599.0
424.00	1556.0
438.40	158.0
441.10	19816.0
442.10	122056.0
443.10	21776.0
444.00	2131.0
444.90	309.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

m/z	Abundance
37.20	325.0
38.10	591.0
39.10	4527.0
40.00	529.0
41.10	284.0
43.00	240.0
44.00	786.0
50.10	14979.0
51.10	59136.0
52.10	3494.0
55.20	490.0
56.10	2083.0
57.00	4752.0
58.00	180.0
61.10	891.0
62.00	929.0
63.10	3303.0
64.10	383.0
65.10	1497.0
69.00	62216.0
70.00	300.0
71.20	156.0
73.10	747.0
74.10	6436.0
75.10	10536.0
76.10	3164.0
77.10	69984.0
78.10	5134.0
79.00	4794.0
80.00	3747.0
81.10	6350.0
82.00	1015.0
83.00	1079.0
85.10	1513.0
86.00	1285.0
87.10	596.0
87.90	217.0
91.10	1180.0
92.10	1902.0
93.00	9161.0
93.90	489.0
96.10	376.0
98.10	7439.0
99.10	5503.0
100.00	400.0
101.00	3542.0
101.90	234.0
103.00	866.0
104.00	1862.0
105.00	1724.0
107.10	23336.0
108.10	4568.0
110.00	38672.0
111.00	5865.0
112.10	626.0
113.70	158.0
114.20	157.0
116.00	1862.0
117.00	19368.0
118.10	1357.0
118.90	283.0
120.10	329.0
122.00	1800.0
123.10	2802.0
123.90	930.0
125.10	1133.0
127.00	73232.0
128.10	6266.0
129.00	31792.0
130.00	3149.0
131.00	796.0
132.00	209.0
134.00	912.0
135.00	2781.0
136.10	1042.0
137.00	1243.0
139.00	272.0
140.00	566.0
141.10	5188.0
142.10	1535.0
143.10	998.0
144.10	227.0
144.90	209.0
146.00	834.0
147.10	2639.0
148.00	4906.0
149.10	1020.0
149.90	258.0
151.10	735.0
152.10	489.0
152.90	1412.0
154.10	1327.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

155.10	2830.0
156.00	4602.0
157.10	669.0
157.90	851.0
159.00	696.0
160.00	1463.0
161.00	2407.0
162.00	789.0
163.20	264.0
164.10	212.0
164.90	2192.0
166.00	1635.0
167.00	8656.0
168.00	3614.0
169.10	613.0
170.00	318.0
170.90	640.0
171.90	1304.0
173.00	1285.0
174.00	2194.0
175.00	4681.0
176.10	979.0
177.00	2468.0
178.10	781.0
179.00	8118.0
180.00	4833.0
181.10	2386.0
181.90	528.0
182.60	418.0
184.00	740.0
185.00	4053.0
186.00	24568.0
187.10	8408.0
188.00	1003.0
189.00	2113.0
190.00	151.0
191.00	883.0
192.00	2193.0
193.00	1966.0
194.10	482.0
194.90	498.0
195.20	453.0
196.00	4928.0
198.00	198208.0
199.00	14326.0
200.00	1267.0
201.50	702.0
203.00	2269.0
204.00	7510.0
205.00	14125.0
206.00	57392.0
207.00	7279.0
207.90	2011.0
209.00	679.0
211.00	2127.0
212.10	413.0
213.00	254.0
215.00	609.0
216.10	1210.0
217.00	16960.0
218.00	1666.0
219.00	341.0
219.90	311.0
221.00	7219.0
222.10	1877.0
223.00	3321.0
224.00	32160.0
225.00	8866.0
226.00	1117.0
227.00	14994.0
228.00	2505.0
229.00	3106.0
230.00	612.0
231.00	1062.0
232.00	246.0
232.90	330.0
234.00	1288.0
235.00	1011.0
236.10	623.0
237.00	1107.0
238.20	156.0
239.00	417.0
239.90	235.0
241.00	969.0
242.00	1817.0
243.00	2159.0
244.00	25312.0
245.00	3483.0
246.00	5488.0
247.00	968.0
248.00	280.0
248.90	1203.0
250.10	256.0
251.10	227.0
252.10	383.0
253.00	712.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

255.00	139904.0
256.00	19040.0
257.00	1426.0
258.00	7558.0
259.00	1263.0
261.20	168.0
264.10	455.0
265.00	2874.0
266.00	412.0
268.00	210.0
270.20	271.0
271.10	204.0
273.00	3943.0
274.00	10737.0
275.00	58072.0
276.00	7430.0
277.00	5518.0
277.90	645.0
278.90	206.0
282.00	350.0
282.90	791.0
284.10	641.0
285.10	1134.0
286.00	163.0
289.00	166.0
290.90	281.0
292.30	172.0
293.00	1187.0
294.10	407.0
295.10	522.0
296.00	18264.0
297.00	2297.0
297.90	306.0
301.00	377.0
302.10	243.0
303.00	2244.0
304.10	469.0
308.00	305.0
309.00	235.0
309.30	223.0
310.10	535.0
314.20	847.0
315.00	1908.0
316.10	975.0
317.00	379.0
321.00	506.0
323.10	5540.0
324.00	1000.0
326.00	244.0
327.10	1343.0
328.00	642.0
328.80	208.0
330.20	227.0
332.10	408.0
332.90	494.0
334.00	3757.0
335.00	842.0
341.10	668.0
346.00	1119.0
351.10	199.0
352.00	1627.0
353.10	1058.0
354.00	1605.0
355.20	474.0
359.10	152.0
363.70	233.0
365.00	8229.0
366.00	1086.0
369.90	489.0
370.90	545.0
372.10	2513.0
372.90	785.0
383.00	776.0
384.10	203.0
390.00	222.0
391.10	236.0
392.00	310.0
401.10	250.0
402.10	810.0
403.10	1988.0
404.00	687.0
421.00	1108.0
422.10	1222.0
423.10	8695.0
424.10	1641.0
424.90	295.0
432.20	194.0
441.10	26776.0
442.10	160192.0
443.10	31968.0
444.00	2959.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

m/z	Abundance
38.00	372.0
39.10	2158.0
40.00	227.0
41.30	161.0
44.00	803.0
49.20	228.0
50.10	8467.0
51.10	30960.0
52.10	1456.0
56.00	1051.0
57.10	2153.0
58.00	178.0
61.10	269.0
62.10	420.0
63.10	1255.0
64.00	337.0
65.10	524.0
68.00	740.0
69.00	31664.0
74.10	4088.0
75.00	5067.0
76.10	1351.0
77.10	32592.0
78.10	1783.0
79.00	2616.0
80.00	1814.0
81.00	2860.0
82.00	908.0
83.00	813.0
84.20	192.0
85.00	705.0
86.00	599.0
87.20	381.0
88.00	307.0
91.00	693.0
92.00	583.0
93.00	4270.0
94.20	223.0
95.20	189.0
96.00	162.0
98.00	3482.0
99.00	3001.0
100.00	220.0
101.00	1872.0
103.00	349.0
104.20	886.0
105.00	1227.0
105.90	319.0
107.00	13208.0
108.00	1485.0
108.90	248.0
110.00	20264.0
111.10	2861.0
112.00	589.0
115.50	279.0
116.10	652.0
117.10	11391.0
117.90	359.0
118.20	325.0
120.00	184.0
122.00	931.0
123.00	910.0
124.00	779.0
125.10	704.0
125.90	323.0
127.00	42120.0
128.00	3066.0
129.00	17520.0
130.00	1624.0
130.90	182.0
133.10	180.0
134.00	665.0
135.00	1289.0
136.00	722.0
137.00	637.0
137.90	263.0
139.10	193.0
140.10	477.0
141.00	2045.0
142.10	786.0
143.00	565.0
143.90	182.0
145.00	235.0
146.00	433.0
147.00	1658.0
148.00	2982.0
149.00	617.0
149.80	297.0
151.20	266.0
152.00	249.0
153.00	1354.0
154.10	488.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

155.00	2275.0
156.00	2137.0
157.10	273.0
158.00	572.0
159.10	570.0
160.00	842.0
161.00	1407.0
162.00	434.0
165.00	1087.0
165.90	1216.0
167.00	5252.0
168.10	1942.0
168.90	506.0
170.20	174.0
170.90	221.0
172.00	656.0
173.00	736.0
174.00	1058.0
175.00	2000.0
176.10	451.0
177.00	1043.0
177.90	325.0
179.00	4096.0
180.10	2175.0
181.00	1268.0
182.00	257.0
182.60	322.0
184.00	391.0
185.00	2331.0
186.00	16264.0
187.10	4242.0
187.90	393.0
189.10	909.0
190.10	369.0
190.90	406.0
192.00	973.0
193.00	1563.0
194.10	274.0
194.80	325.0
196.00	3674.0
198.00	109256.0
199.00	7026.0
200.00	478.0
201.30	541.0
201.80	236.0
203.00	1040.0
204.00	4450.0
205.10	7779.0
206.00	33248.0
207.00	4420.0
208.00	1199.0
209.00	433.0
210.10	630.0
211.10	1331.0
212.10	309.0
214.90	370.0
216.00	610.0
217.00	9751.0
218.00	1372.0
221.10	4555.0
222.00	1798.0
223.00	1673.0
224.10	19384.0
225.10	5054.0
226.10	646.0
227.00	9191.0
228.00	1278.0
229.00	1942.0
229.90	385.0
231.00	1234.0
232.90	320.0
234.00	594.0
235.00	851.0
236.00	268.0
236.90	732.0
239.10	495.0
239.90	233.0
241.00	779.0
242.00	1052.0
243.00	1214.0
244.00	14728.0
245.00	1789.0
246.00	3506.0
246.90	460.0
249.00	379.0
250.90	266.0
252.00	378.0
253.00	785.0
253.80	1267.0
255.00	83696.0
256.00	10889.0
257.00	957.0
258.00	4308.0
259.00	900.0
264.20	344.0
265.00	1892.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47

266.10	569.0
271.10	168.0
272.20	259.0
273.00	2376.0
274.00	6451.0
275.00	37888.0
276.00	5076.0
277.00	2979.0
278.00	670.0
285.00	640.0
289.00	209.0
289.90	187.0
291.00	192.0
293.00	806.0
293.90	243.0
296.00	11404.0
297.10	1693.0
301.00	202.0
303.00	1594.0
304.10	353.0
313.00	157.0
314.10	734.0
315.00	1110.0
316.00	580.0
317.00	336.0
321.10	254.0
322.00	320.0
323.00	3875.0
324.00	523.0
327.00	577.0
327.90	498.0
332.00	304.0
333.10	448.0
334.00	2854.0
335.10	451.0
336.00	185.0
341.10	486.0
346.00	705.0
350.00	164.0
352.10	1290.0
353.00	658.0
354.00	1205.0
355.00	195.0
365.00	5336.0
366.10	1053.0
370.80	165.0
372.00	1584.0
373.10	504.0
378.40	152.0
383.10	440.0
390.10	237.0
390.80	174.0
391.90	195.0
402.10	604.0
403.00	1022.0
404.00	507.0
421.00	1155.0
422.10	991.0
423.00	6087.0
424.00	1091.0
439.80	203.0
441.00	19048.0
442.10	118016.0
443.10	22632.0
444.10	1584.0

Instrument :
BNA_G
ClientSampleId :
DFTPP47