

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102921\
 Data File : BG050785.D
 Acq On : 29 Oct 2021 20:03
 Operator : CG/JU
 Sample : M4372-15
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GW-1

Quant Time: Oct 30 01:38:35 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.244	152	31215	20.000	ng	-0.01
21) Naphthalene-d8	11.070	136	133466	20.000	ng	#-0.01
39) Acenaphthene-d10	14.866	164	88026	20.000	ng	0.00
64) Phenanthrene-d10	17.615	188	201476	20.000	ng	# 0.00
76) Chrysene-d12	21.916	240	189673	20.000	ng	# 0.00
86) Perylene-d12	25.342	264	191315	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.782	112	142700	68.369	ng	0.00
7) Phenol-d6	7.386	99	120657	39.927	ng	-0.01
23) Nitrobenzene-d5	9.419	82	297232	92.189	ng	0.00
42) 2,4,6-Tribromophenol	16.358	330	122710	146.151	ng	0.00
45) 2-Fluorobiphenyl	13.491	172	550699	94.156	ng	-0.01
79) Terphenyl-d14	20.206	244	863812	88.754	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102921\
Data File : BG050785.D
Acq On : 29 Oct 2021 20:03
Operator : CG/JU
Sample : M4372-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GW-1

Quant Time: Oct 30 01:38:35 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 10:56:22 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.244 min Scan# 809
Delta R.T. -0.014 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03




Instrument :
BNA_G
ClientSampled :
GW-1


Tgt Ion:152 Resp: 31215
Ion Ratio Lower Upper
152 100
150 167.6 119.4 179.0
115 64.5 60.2 90.4



#5
2-Fluorophenol
Concen: 68.369 ng
RT: 5.782 min Scan# 390
Delta R.T. -0.008 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03

Tgt Ion:112 Resp: 142700
Ion Ratio Lower Upper
112 100
64 70.2 46.7 70.1#
63 37.6 35.3 52.9









#7
Phenol-d6
Concen: 39.927 ng
RT: 7.386 min Scan# 663
Delta R.T. -0.014 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03




Instrument :
BNA_G
ClientSampled :
GW-1


Tgt Ion: 99 Resp: 120657
Ion Ratio Lower Upper
99 100
42 20.7 20.5 30.7
71 39.3 41.5 62.3#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.070 min Scan# 1290
Delta R.T. -0.014 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03

Tgt Ion:136 Resp: 133466
Ion Ratio Lower Upper
136 100
137 10.8 8.2 12.2
54 10.2 5.5 8.3#
68 7.9 2.3 3.5#









#23
Nitrobenzene-d5
Concen: 92.189 ng
RT: 9.419 min Scan# 1009
Delta R.T. -0.008 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03




Instrument :
BNA_G
ClientSampleId :
GW-1


Tgt Ion: 82 Resp: 297232
Ion Ratio Lower Upper
82 100
128 37.5 29.8 44.6
54 50.5 45.8 68.8



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.866 min Scan# 1936
Delta R.T. -0.008 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03

Tgt Ion: 164 Resp: 88026
Ion Ratio Lower Upper
164 100
162 100.5 81.4 122.0
160 42.1 34.9 52.3









#42
2,4,6-Tribromophenol
Concen: 146.151 ng
RT: 16.358 min Scan# 2190
Delta R.T. -0.003 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03




Instrument :
BNA_G
ClientSampleId :
GW-1


Tgt Ion:330 Resp: 122710
Ion Ratio Lower Upper
330 100
332 96.9 77.3 115.9
141 54.6 35.6 53.4#



#45
2-Fluorobiphenyl
Concen: 94.156 ng
RT: 13.491 min Scan# 1702
Delta R.T. -0.014 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03

Tgt Ion:172 Resp: 550699
Ion Ratio Lower Upper
172 100
171 37.1 28.6 43.0
170 25.0 18.6 27.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.615 min Scan# 2404
Delta R.T. -0.003 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03




Instrument :
BNA_G
ClientSampleId :
GW-1


Tgt Ion:188 Resp: 201476
Ion Ratio Lower Upper
188 100
94 9.5 5.4 8.0#
80 11.8 3.7 5.5#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.916 min Scan# 3136
Delta R.T. -0.008 min
Lab File: BG050785.D
Acq: 29 Oct 2021 20:03

Tgt Ion:240 Resp: 189673
Ion Ratio Lower Upper
240 100
120 10.4 4.2 6.2#
236 26.5 21.8 32.6









#79
 Terphenyl-d14
 Concen: 88.754 ng
 RT: 20.206 min Scan# 2845
 Delta R.T. -0.003 min
 Lab File: BG050785.D
 Acq: 29 Oct 2021 20:03

Instrument :
 BNA_G
 ClientSampleId :
 GW-1

Tgt Ion:244 Resp: 863812
 Ion Ratio Lower Upper
 244 100
 212 8.3 6.9 10.3
 122 11.0 5.0 7.6#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.342 min Scan# 3719
 Delta R.T. 0.009 min
 Lab File: BG050785.D
 Acq: 29 Oct 2021 20:03

Tgt Ion:264 Resp: 191315
 Ion Ratio Lower Upper
 264 100
 260 23.7 19.4 29.2
 265 20.9 16.9 25.3

