


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102922\
 Data File : BG055359.D
 Acq On : 29 Oct 2022 10:27
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Oct 29 04:52:38 2022






AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.4	15860	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.0	16153	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.5	22272	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	48920	PASS
199	198	5	9	6.8	3310	PASS
275	198	10	60	30.1	14704	PASS
365	198	1	100	4.2	2034	PASS
441	198	0.01	100	17.0	8323	PASS
442	442	50	100	100.0	50651	PASS
443	442	15	24	19.8	10024	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102922\
Data File : BG055359.D
Acq On : 29 Oct 2022 10:27
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 29 11:04:51 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 29 04:52:38 2022
Response via : Initial Calibration



TIC: BG055359.D\data.ms

(70) Pentachlorophenol (C)

17.259min (-0.007) 37945.18 ng




response 48396

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.50	63.76
264.00	64.40	59.85
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102922\
Data File : BG055359.D
Acq On : 29 Oct 2022 10:27
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 29 11:04:51 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 29 04:52:38 2022
Response via : Initial Calibration



TIC: BG055359.D\data.ms

(77) Benzidine

19.809min (-0.007) 0.00 ng

response 374555

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.80	14.35
183.00	12.50	12.34
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/29/2022	BNA_G	BG055359.D
Compound Name	Response	Retention Time
DDT	177338	21.031
DDD	3723	20.59
DDE	254	20.085
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3977	181315	2.19

Instrument :
BNA_G
ClientSampleId :
DFTPP