

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103123\
 Data File : BG059562.D
 Acq On : 31 Oct 2023 15:27
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465

Quant Time: Nov 01 03:18:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.355	152	77886	20.000	ng/ul	0.00
20) Naphthalene-d8	11.205	136	365122	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.989	164	246306	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.739	188	521946	20.000	ng/ul	0.00
79) Chrysene-d12	22.081	240	494676	20.000	ng/ul	0.00
88) Perylene-d12	25.682	264	516383	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.649	96	20535	9.865	ng/uL	0.00
4) Pyridine-d5	4.084	84	112325	19.309	ng/ul	0.00
7) Phenol-d5	7.462	99	151105	18.880	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.674	67	67307	18.309	ng/ul	0.00
11) 2-Chlorophenol-d4	7.874	132	119567	18.414	ng/ul	0.00
15) 4-Methylphenol-d8	9.037	113	121504	19.135	ng/ul	0.00
21) Nitrobenzene-d5	9.548	128	60479	23.296	ng/ul	-0.01
24) 2-Nitrophenol-d4	10.277	143	63811	23.942	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.805	165	118591	20.196	ng/ul	0.00
31) 4-Chloroaniline-d4	11.346	131	185736	19.215	ng/ul	0.00
46) Dimethylphthalate-d6	14.384	166	388983	17.432	ng/ul	0.00
49) Acenaphthylene-d8	14.689	160	466511	17.814	ng/ul	0.00
54) 4-Nitrophenol-d4	15.147	143	80248	20.971	ng/ul	0.00
60) Fluorene-d10	15.976	176	341024	17.882	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.076	200	52454	21.221	ng/ul	0.00
73) Anthracene-d10	17.838	188	539442	19.077	ng/ul	0.00
81) Pyrene-d10	20.106	212	666254	19.377	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.435	264	584026	19.207	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.690	88	14898	6.394	ng/uL#	79
5) Pyridine	4.107	79	109352	19.057	ng/ul	89
6) Benzaldehyde	7.498	77	82357	22.845	ng/ul	97
8) Phenol	7.492	94	144912	19.205	ng/ul#	89
10) Bis(2-Chloroethyl)ether	7.768	93	102363	18.500	ng/ul	98
12) 2-Chlorophenol	7.903	128	126838	19.704	ng/ul	96
13) 2-Methylphenol	8.773	108	115459	18.356	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.855	45	30852	5.925	ng/ul#	89
16) Acetophenone	9.201	105	183304	18.313	ng/ul	95
17) N-Nitroso-di-n-propyla...	9.160	70	75904	16.094	ng/ul#	94
18) 4-Methylphenol	9.102	108	123119	18.396	ng/ul	97
19) Hexachloroethane	9.448	117	47410	18.544	ng/ul#	79
22) Nitrobenzene	9.595	77	123955	22.121	ng/ul	98
23) Isophorone	10.106	82	239992	17.007	ng/ul#	91
25) 2-Nitrophenol	10.306	139	65540	22.865	ng/ul#	82
26) 2,4-Dimethylphenol	10.330	107	138593	18.383	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.588	93	144027	18.330	ng/ul	94
29) 2,4-Dichlorophenol	10.829	162	113996	20.066	ng/ul#	84
30) Naphthalene	11.258	128	427605	18.790	ng/ul#	97
32) 4-Chloroaniline	11.370	127	180482	19.656	ng/ul	91
33) Hexachlorobutadiene	11.487	225	64127	18.816	ng/ul	95
34) Caprolactam	12.139	113	19822	9.538	ng/ul#	88
35) 4-Chloro-3-methylphenol	12.433	107	140861	20.781	ng/ul	90
36) 2-Methylnaphthalene	12.838	142	311345	19.886	ng/ul	96


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103123\
 Data File : BG059562.D
 Acq On : 31 Oct 2023 15:27
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465

Quant Time: Nov 01 03:18:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG102723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 28 03:56:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.050	142	312634	20.048	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	13.179	216	132557	18.659	ng/ul	90
40) Hexachlorocyclopentadiene	13.132	237	65954	15.926	ng/ul	87
41) 2,4,6-Trichlorophenol	13.414	196	92502	18.804	ng/ul	89
42) 2,4,5-Trichlorophenol	13.485	196	102233	19.196	ng/ul	95
43) 1,1'-Biphenyl	13.825	154	396696	17.804	ng/ul	99
44) 2-Chloronaphthalene	13.878	162	305124	17.619	ng/ul	96
45) 2-Nitroaniline	14.090	65	78795	24.453	ng/ul	94
47) Dimethylphthalate	14.431	163	376952	17.098	ng/ul	100
48) 2,6-Dinitrotoluene	14.566	165	69022	20.754	ng/ul#	75
50) Acenaphthylene	14.719	152	538385	17.843	ng/ul	98
51) 3-Nitroaniline	14.907	138	92998	22.434	ng/ul#	86
52) Acenaphthene	15.053	153	342718	17.425	ng/ul	98
53) 2,4-Dinitrophenol	15.100	184	37567	19.880	ng/ul#	86
55) 4-Nitrophenol	15.171	109	92933	22.037	ng/ul	91
56) Dibenzofuran	15.382	168	458412	17.978	ng/ul	95
57) 2,4-Dinitrotoluene	15.341	165	109717	23.454	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.588	232	86544	20.705	ng/ul#	91
59) Diethylphthalate	15.776	149	420476	17.484	ng/ul	99
61) Fluorene	16.035	166	392199	17.817	ng/ul	100
62) 4-Chlorophenyl-phenyle...	16.017	204	169552	18.023	ng/ul	93
63) 4-Nitroaniline	16.070	138	101634	23.102	ng/ul	94
66) 4,6-Dinitro-2-methylph...	16.088	198	58464	22.443	ng/ul#	91
67) N-Nitrosodiphenylamine	16.234	169	332336	18.635	ng/ul	97
68) 4-Bromophenyl-phenylether	16.916	248	101328	19.400	ng/ul	97
69) Hexachlorobenzene	17.004	284	120720	19.999	ng/ul	90
70) Atrazine	17.169	200	108111	19.563	ng/ul#	87
71) Pentachlorophenol	17.357	266	69900	20.270	ng/ul	94
72) Phenanthrene	17.874	178	660632	19.470	ng/ul	97
74) Anthracene	17.874	178	660632	19.283	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.784	216	131757	18.956	ng/ul	85
76) Pentachlorobenzene	15.277	250	131853	19.991	ng/ul	96
77) Carbazole	18.144	167	593130	20.153	ng/ul	95
78) Di-n-butylphthalate	18.661	149	763265	18.874	ng/ul	98
80) Fluoranthene	19.771	202	718139	17.134	ng/ul	99
82) Pyrene	20.142	202	826046	18.974	ng/ul#	95
83) Butylbenzylphthalate	21.005	149	394801	20.080	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.963	252	275928	21.450	ng/ul	97
85) Benzo(a)anthracene	22.057	228	756682	18.588	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.898	149	569985	19.428	ng/ul	99
87) Chrysene	22.128	228	709740	18.742	ng/ul	95
89) Di-n-octyl phthalate	23.256	149	941064	20.815	ng/ul	100
90) Benzo(b)fluoranthene	24.519	252	706571	18.877	ng/ul	99
91) Benzo(k)fluoranthene	24.519	252	706571	18.898	ng/ul	99
93) Benzo(a)pyrene	25.518	252	671649	18.985	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.877	276	746190	16.641	ng/ul#	88
95) Dibenzo(a,h)anthracene	29.971	278	708579	19.133	ng/ul#	96
96) Benzo(g,h,i)perylene	31.205	276	675557	18.407	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 6.394 ng/uL
RT: 3.690 min Scan# 51
Delta R.T. -0.012 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 88 Resp: 14898
Ion Ratio Lower Upper
88 100
43 19.9 15.7 23.5
58 64.6 36.2 54.2#



#5
Pyridine
Concen: 19.057 ng/uL
RT: 4.107 min Scan# 122
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion: 79 Resp: 109352
Ion Ratio Lower Upper
79 100
52 54.4 53.3 79.9
51 28.8 22.4 33.6









#6
Benzaldehyde
Concen: 22.845 ng/ul
RT: 7.498 min Scan# 699
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 77 Resp: 82357
Ion Ratio Lower Upper
77 100
105 102.3 83.6 125.4
106 95.2 78.6 118.0



#8
Phenol
Concen: 19.205 ng/ul
RT: 7.492 min Scan# 698
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion: 94 Resp: 144912
Ion Ratio Lower Upper
94 100
65 31.2 22.5 33.7
66 48.2 31.1 46.7#









#10
Bis(2-Chloroethyl)ether
Concen: 18.500 ng/ul
RT: 7.768 min Scan# 746
Delta R.T. -0.012 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 93 Resp: 102363
Ion Ratio Lower Upper
93 100
63 65.0 52.5 78.7
95 35.3 26.5 39.7



#12
2-Chlorophenol
Concen: 19.704 ng/ul
RT: 7.903 min Scan# 768
Delta R.T. -0.012 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:128 Resp: 126838
Ion Ratio Lower Upper
128 100
64 35.1 30.2 45.2
130 32.4 27.9 41.9









#13
2-Methylphenol
Concen: 18.356 ng/ul
RT: 8.773 min Scan# 917
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:108 Resp: 115459
Ion Ratio Lower Upper
108 100
107 87.3 68.6 102.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 5.925 ng/ul
RT: 8.855 min Scan# 930
Delta R.T. -0.023 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion: 45 Resp: 30852
Ion Ratio Lower Upper
45 100
77 23.9 15.0 22.6#
79 16.7 10.0 15.0#









#16
Acetophenone
Concen: 18.313 ng/ul
RT: 9.201 min Scan# 91
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:105 Resp: 183304
Ion Ratio Lower Upper
105 100
77 69.4 59.6 89.4
51 21.0 16.2 24.4



#17
N-Nitroso-di-n-propylamine
Concen: 16.094 ng/ul
RT: 9.160 min Scan# 982
Delta R.T. -0.012 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion: 70 Resp: 75904
Ion Ratio Lower Upper
70 100
42 43.5 33.8 50.8
101 9.5 5.4 8.0#
130 28.4 18.2 27.4#









#18
4-Methylphenol
Concen: 18.396 ng/ul
RT: 9.102 min Scan# 91
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:108 Resp: 123119
Ion Ratio Lower Upper
108 100
107 107.8 88.9 133.3



#19
Hexachloroethane
Concen: 18.544 ng/ul
RT: 9.448 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:117 Resp: 47410
Ion Ratio Lower Upper
117 100
201 71.4 75.8 113.6#
199 64.8 42.3 63.5#









#22
Nitrobenzene
Concen: 22.121 ng/ul
RT: 9.595 min Scan# 1056
Delta R.T. -0.012 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 77 Resp: 123955
Ion Ratio Lower Upper
77 100
123 52.3 42.6 63.8
65 15.6 11.0 16.4



#23
Isophorone
Concen: 17.007 ng/ul
RT: 10.106 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion: 82 Resp: 239992
Ion Ratio Lower Upper
82 100
95 8.1 5.6 8.4
138 20.6 12.3 18.5#









#25
2-Nitrophenol
Concen: 22.865 ng/ul
RT: 10.306 min Scan# 1177
Delta R.T. -0.012 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:139 Resp: 65540
Ion Ratio Lower Upper
139 100
65 42.6 39.7 59.5
109 43.8 21.0 31.6#



#26
2,4-Dimethylphenol
Concen: 18.383 ng/ul
RT: 10.330 min Scan# 1181
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:107 Resp: 138593
Ion Ratio Lower Upper
107 100
121 44.9 42.7 64.1
122 88.9 72.6 108.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.330 ng/ul
 RT: 10.588 min Scan# 1225
 Delta R.T. -0.012 min
 Lab File: BG059562.D
 Acq: 31 Oct 2023 15:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465


Tgt Ion: 93 Resp: 144027
 Ion Ratio Lower Upper
 93 100
 95 30.0 27.1 40.7
 123 12.0 10.3 15.5



#29
 2,4-Dichlorophenol
 Concen: 20.066 ng/ul
 RT: 10.829 min Scan# 1266
 Delta R.T. 0.000 min
 Lab File: BG059562.D
 Acq: 31 Oct 2023 15:27

Tgt Ion:162 Resp: 113996
 Ion Ratio Lower Upper
 162 100
 164 69.5 45.3 67.9#
 98 32.3 32.2 48.4









#30
Naphthalene
Concen: 18.790 ng/ul
RT: 11.258 min Scan# 11
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:128 Resp: 427605
Ion Ratio Lower Upper
128 100
129 11.1 7.4 11.0#
127 13.9 10.5 15.7



#32
4-Chloroaniline
Concen: 19.656 ng/ul
RT: 11.370 min Scan# 1358
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:127 Resp: 180482
Ion Ratio Lower Upper
127 100
129 32.5 30.4 45.6





#33

Hexachlorobutadiene

Concen: 18.816 ng/ul

RT: 11.487 min Scan# 1

Delta R.T. -0.012 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:225 Resp: 64127

Ion Ratio Lower Upper

225 100

223 63.0 50.8 76.2

227 55.7 49.8 74.8



#34

Caprolactam

Concen: 9.538 ng/ul

RT: 12.139 min Scan# 1489

Delta R.T. -0.018 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:113 Resp: 19822


Ion Ratio Lower Upper

113 100

55 116.7 86.9 130.3

56 96.8 64.4 96.6#









#35
4-Chloro-3-methylphenol
Concen: 20.781 ng/ul
RT: 12.433 min Scan# 1539
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:107 Resp: 140861
Ion Ratio Lower Upper
107 100
144 23.2 21.0 31.6
142 75.4 68.2 102.4



#36
2-Methylnaphthalene
Concen: 19.886 ng/ul
RT: 12.838 min Scan# 1608
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:142 Resp: 311345
Ion Ratio Lower Upper
142 100
141 86.1 71.7 107.5





#37

1-Methylnaphthalene

Concen: 20.048 ng/ul

RT: 13.050 min Scan# 1

Delta R.T. -0.012 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:142 Resp: 312634

Ion Ratio Lower Upper

142 100

141 88.0 73.1 109.7

116 5.4 3.3 4.9#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.659 ng/ul

RT: 13.179 min Scan# 1666

Delta R.T. 0.000 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Tgt Ion:216 Resp: 132557




Ion Ratio Lower Upper


216 100

214 68.9 63.7 95.5

179 20.1 19.2 28.8

108 19.5 16.5 24.7





#40

Hexachlorocyclopentadiene

Concen: 15.926 ng/ul

RT: 13.132 min Scan# 1

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:237 Resp: 65954

Ion Ratio Lower Upper

237 100

235 69.5 46.6 69.8

272 14.2 13.0 19.4



#41

2,4,6-Trichlorophenol

Concen: 18.804 ng/ul

RT: 13.414 min Scan# 1706

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:196 Resp: 92502


Ion Ratio Lower Upper

196 100

198 97.6 69.6 104.4

200 34.6 23.8 35.8





#42

2,4,5-Trichlorophenol

Concen: 19.196 ng/ul

RT: 13.485 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:196 Resp: 102233

Ion Ratio Lower Upper

196 100

198 93.6 79.8 119.6

200 33.3 28.0 42.0



#43

1,1'-Biphenyl

Concen: 17.804 ng/ul

RT: 13.825 min Scan# 1776

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:154 Resp: 396696


Ion Ratio Lower Upper

154 100

153 38.4 30.7 46.1

76 8.5 8.2 12.4









#44
2-Chloronaphthalene
Concen: 17.619 ng/ul
RT: 13.878 min Scan# 11
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:162 Resp: 305124
Ion Ratio Lower Upper
162 100
164 34.1 23.9 35.9
127 44.6 35.3 52.9



#45
2-Nitroaniline
Concen: 24.453 ng/ul
RT: 14.090 min Scan# 1821
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion: 65 Resp: 78795
Ion Ratio Lower Upper
65 100
92 78.9 59.6 89.4
138 128.9 109.0 163.6









#47
Dimethylphthalate
Concen: 17.098 ng/ul
RT: 14.431 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:163 Resp: 376952
Ion Ratio Lower Upper
163 100
194 2.9 2.6 3.8
164 8.3 6.6 10.0



#48
2,6-Dinitrotoluene
Concen: 20.754 ng/ul
RT: 14.566 min Scan# 1902
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:165 Resp: 69022
Ion Ratio Lower Upper
165 100
89 67.2 41.0 61.4#
121 28.4 12.3 18.5#





#50

Acenaphthylene

Concen: 17.843 ng/ul

RT: 14.719 min Scan# 1928

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:152 Resp: 538385

Ion Ratio Lower Upper

152 100

151 19.8 16.6 25.0

153 12.8 9.4 14.2



#51

3-Nitroaniline

Concen: 22.434 ng/ul

RT: 14.907 min Scan# 1960

Delta R.T. 0.000 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:138 Resp: 92998


Ion Ratio Lower Upper

138 100

108 10.3 17.1 25.7#

92 113.1 101.8 152.8









#52
Acenaphthene
Concen: 17.425 ng/ul
RT: 15.053 min Scan# 1985
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:153 Resp: 342718
Ion Ratio Lower Upper
153 100
152 48.9 37.6 56.4
154 87.8 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 19.880 ng/ul
RT: 15.100 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:184 Resp: 37567
Ion Ratio Lower Upper
184 100
63 56.4 35.0 52.6#
154 70.6 49.8 74.8









#55
4-Nitrophenol
Concen: 22.037 ng/ul
RT: 15.171 min Scan# 2005
Delta R.T. 0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:109 Resp: 92933
Ion Ratio Lower Upper
109 100
139 92.8 72.2 108.2
65 84.9 57.3 85.9



#56
Dibenzofuran
Concen: 17.978 ng/ul
RT: 15.382 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:168 Resp: 458412
Ion Ratio Lower Upper
168 100
139 45.5 33.8 50.8





#57

2,4-Dinitrotoluene

Concen: 23.454 ng/ul

RT: 15.341 min Scan# 2035

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:165 Resp: 109717

Ion Ratio Lower Upper

165 100

63 41.9 30.9 46.3

182 2.2 1.8 2.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.705 ng/ul

RT: 15.588 min Scan# 2076

Delta R.T. 0.000 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Tgt Ion:232 Resp: 86544




Ion Ratio Lower Upper


232 100

131 61.8 44.2 66.2

130 4.6 2.3 3.5#

166 34.5 32.6 48.8









#59
Diethylphthalate
Concen: 17.484 ng/ul
RT: 15.776 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:149 Resp: 420476
Ion Ratio Lower Upper
149 100
177 20.3 16.3 24.5
150 12.6 9.1 13.7



#61
Fluorene
Concen: 17.817 ng/ul
RT: 16.035 min Scan# 2152
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:166 Resp: 392199
Ion Ratio Lower Upper
166 100
165 93.0 74.4 111.6
167 10.3 9.1 13.7









#62
4-Chlorophenyl-phenylether
Concen: 18.023 ng/ul
RT: 16.017 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:204 Resp: 169552
Ion Ratio Lower Upper
204 100
206 35.3 28.8 43.2
141 77.5 68.6 103.0



#63
4-Nitroaniline
Concen: 23.102 ng/ul
RT: 16.070 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:138 Resp: 101634
Ion Ratio Lower Upper
138 100
92 50.8 49.8 74.8
108 125.1 100.9 151.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 22.443 ng/ul
 RT: 16.088 min Scan# 2161
 Delta R.T. -0.006 min
 Lab File: BG059562.D
 Acq: 31 Oct 2023 15:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465


Tgt Ion:198 Resp: 58464
 Ion Ratio Lower Upper
 198 100
 182 4.6 1.4 2.0#
 77 24.1 15.8 23.8#



#67
 N-Nitrosodiphenylamine
 Concen: 18.635 ng/ul
 RT: 16.234 min Scan# 2186
 Delta R.T. 0.000 min
 Lab File: BG059562.D
 Acq: 31 Oct 2023 15:27

Tgt Ion:169 Resp: 332336
 Ion Ratio Lower Upper
 169 100
 168 66.0 54.2 81.2
 167 37.0 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 19.400 ng/ul
RT: 16.916 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:248 Resp: 101328
Ion Ratio Lower Upper
248 100
250 90.1 74.2 111.2
141 114.1 88.5 132.7



#69
Hexachlorobenzene
Concen: 19.999 ng/ul
RT: 17.004 min Scan# 2317
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:284 Resp: 120720
Ion Ratio Lower Upper
284 100
142 40.2 39.6 59.4
249 32.7 28.2 42.2









#70
Atrazine
Concen: 19.563 ng/ul
RT: 17.169 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:200 Resp: 108111
Ion Ratio Lower Upper
200 100
173 25.5 25.7 38.5#
215 48.5 32.1 48.1#



#71
Pentachlorophenol
Concen: 20.270 ng/ul
RT: 17.357 min Scan# 2377
Delta R.T. 0.000 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:266 Resp: 69900
Ion Ratio Lower Upper
266 100
264 59.9 46.2 69.2
268 65.1 46.4 69.6









#72
Phenanthrene
Concen: 19.470 ng/ul
RT: 17.874 min Scan# 2465
Delta R.T. 0.088 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:178 Resp: 660632
Ion Ratio Lower Upper
178 100
179 15.8 11.3 16.9
176 18.3 15.4 23.0



#74
Anthracene
Concen: 19.283 ng/ul
RT: 17.874 min Scan# 2465
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:178 Resp: 660632
Ion Ratio Lower Upper
178 100
179 15.8 13.1 19.7
176 18.3 13.9 20.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.956 ng/uL
RT: 13.784 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:216 Resp: 131757
Ion Ratio Lower Upper
216 100
214 72.1 70.2 105.2
218 42.5 39.9 59.9



#76
Pentachlorobenzene
Concen: 19.991 ng/uL
RT: 15.277 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BG059562.D
Acq: 31 Oct 2023 15:27

Tgt Ion:250 Resp: 131853
Ion Ratio Lower Upper
250 100
248 58.9 48.7 73.1
252 59.8 51.2 76.8









#77
 Carbazole
 Concen: 20.153 ng/ul
 RT: 18.144 min Scan# 2511
 Delta R.T. -0.006 min
 Lab File: BG059562.D
 Acq: 31 Oct 2023 15:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465


Tgt Ion:167 Resp: 593130
 Ion Ratio Lower Upper
 167 100
 166 20.6 19.4 29.2
 139 14.4 11.7 17.5



#78
 Di-n-butylphthalate
 Concen: 18.874 ng/ul
 RT: 18.661 min Scan# 2599
 Delta R.T. -0.006 min
 Lab File: BG059562.D
 Acq: 31 Oct 2023 15:27

Tgt Ion:149 Resp: 763265
 Ion Ratio Lower Upper
 149 100
 150 8.3 7.5 11.3
 104 6.1 5.3 7.9





#80

Fluoranthene

Concen: 17.134 ng/ul

RT: 19.771 min Scan# 2789

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:202 Resp: 718139

Ion Ratio Lower Upper

202 100

101 9.0 7.0 10.6

100 7.1 5.2 7.8



#82

Pyrene

Concen: 18.974 ng/ul

RT: 20.142 min Scan# 2851

Delta R.T. 0.000 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:202 Resp: 826046


Ion Ratio Lower Upper

202 100

101 9.2 7.9 11.9

100 6.6 7.9 11.9#





#83

Butylbenzylphthalate

Concen: 20.080 ng/ul

RT: 21.005 min Scan# 2999

Delta R.T. -0.012 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:149 Resp: 394801

Ion Ratio Lower Upper

149 100

91 71.4 52.2 78.2

206 17.8 15.1 22.7



#84

3,3'-Dichlorobenzidine

Concen: 21.450 ng/ul

RT: 21.963 min Scan# 3161

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:252 Resp: 275928


Ion Ratio Lower Upper

252 100

254 63.4 48.9 73.3

126 12.5 11.5 17.3





#85

Benzo(a)anthracene

Concen: 18.588 ng/ul

RT: 22.057 min Scan# 3177

Delta R.T. 0.000 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:228 Resp: 756682

Ion Ratio Lower Upper

228 100

229 22.6 16.7 25.1

226 25.8 20.3 30.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.428 ng/ul

RT: 21.898 min Scan# 3150

Delta R.T. -0.018 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:149 Resp: 569985


Ion Ratio Lower Upper

149 100

167 24.7 19.4 29.2

279 2.8 2.2 3.4





#87

Chrysene

Concen: 18.742 ng/ul

RT: 22.128 min Scan# 31

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:228 Resp: 709740

Ion Ratio Lower Upper

228 100

226 29.6 21.0 31.4

229 19.8 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 20.815 ng/ul




RT: 23.256 min Scan# 3381


Delta R.T. -0.018 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Tgt Ion:149 Resp: 941064





#90

Benzo(b)fluoranthene

Concen: 18.877 ng/ul

RT: 24.519 min Scan# 3

Delta R.T. -0.006 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:252 Resp: 706571

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8

125 10.2 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 18.898 ng/ul

RT: 24.519 min Scan# 3596

Delta R.T. -0.082 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:252 Resp: 706571


Ion Ratio Lower Upper

252 100

253 22.1 17.9 26.9

125 10.2 7.9 11.9





#93

Benzo(a)pyrene

Concen: 18.985 ng/ul

RT: 25.518 min Scan# 31

Delta R.T. 0.000 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:252 Resp: 671649

Ion Ratio Lower Upper

252 100

253 22.1 18.5 27.7

125 9.8 7.6 11.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.641 ng/ul

RT: 29.877 min Scan# 4508

Delta R.T. -0.012 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27




Tgt Ion:276 Resp: 746190


Ion Ratio Lower Upper

276 100

138 14.3 15.3 22.9#

277 27.0 16.4 24.6#





#95

Dibenzo(a,h)anthracene

Concen: 19.133 ng/ul

RT: 29.971 min Scan# 4

Delta R.T. -0.012 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:278 Resp: 708579

Ion Ratio Lower Upper

278 100

139 10.8 11.0 16.6#

279 22.1 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 18.407 ng/ul

RT: 31.205 min Scan# 4734

Delta R.T. -0.012 min

Lab File: BG059562.D

Acq: 31 Oct 2023 15:27

Tgt Ion:276 Resp: 675557

Ion Ratio Lower Upper

276 100

138 18.2 16.2 24.2

277 23.6 18.2 27.4

