

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG110119\  
 Data File : BG043466.D  
 Acq On : 1 Nov 2019 15:23  
 Operator : JU  
 Sample : K5602-04  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 20190941-AMENDED-COMPOSITE-

Quant Time: Nov 01 23:00:00 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Oct 21 16:37:43 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.01  | 152  | 41594    | 20.00 | ng    | -0.01    |
| 21) Naphthalene-d8          | 10.82 | 136  | 157687   | 20.00 | ng    | -0.01    |
| 39) Acenaphthene-d10        | 14.64 | 164  | 112833   | 20.00 | ng    | -0.01    |
| 64) Phenanthrene-d10        | 17.38 | 188  | 276683   | 20.00 | ng    | -0.02    |
| 76) Chrysene-d12            | 21.66 | 240  | 299993   | 20.00 | ng    | -0.01    |
| 87) Perylene-d12            | 24.84 | 264  | 365323   | 20.00 | ng    | -0.03    |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.60  | 112  | 11783    | 5.19  | ng    | 0.00     |
| 7) Phenol-d6                | 7.20  | 99   | 82231    | 25.18 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.18  | 82   | 131784   | 43.09 | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 16.14 | 330  | 5506     | 2.56  | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.27 | 172  | 341300   | 41.80 | ng    | -0.01    |
| 79) Terphenyl-d14           | 19.99 | 244  | 713660   | 44.63 | ng    | -0.02    |
| Target Compounds            |       |      |          |       |       |          |
| 50) Dimethylphthalate       | 14.09 | 163  | 65700    | 7.517 | ng    | 100      |
| 75) Fluoranthene            | 19.44 | 202  | 39291    | 2.127 | ng    | 97       |
| 78) Pyrene                  | 19.80 | 202  | 40021    | 2.155 | ng    | 96       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG110119\  
Data File : BG043466.D  
Acq On : 1 Nov 2019 15:23  
Operator : JU  
Sample : K5602-04  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
20190941-AMENDED-COMPOSITE-

Quant Time: Nov 01 23:00:00 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Oct 21 16:37:43 2019  
Response via : Initial Calibration

