

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110218\
 Data File : BG037775.D
 Acq On : 3 Nov 2018 4:31
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 03 04:06:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.607	0.566	6.8	103	0.00
3	Pyridine	1.529	1.441	5.8	103	0.00
4	n-Nitrosodimethylamine	0.545	0.522	4.2	105	0.00
5 S	2-Fluorophenol	1.196	1.171	2.1	106	0.00
6	Aniline	2.207	2.150	2.6	105	0.00
7 S	Phenol-d6	1.809	1.756	2.9	104	0.00
8	2-Chlorophenol	1.399	1.381	1.3	107	0.00
9	Benzaldehyde	1.132	0.991	12.5	95	0.00
10 C	Phenol	1.909	1.842	3.5	104	0.00
11	bis(2-Chloroethyl)ether	1.450	1.430	1.4	108	0.00
12	1,3-Dichlorobenzene	1.558	1.578	-1.3	111	-0.01
13 C	1,4-Dichlorobenzene	1.594	1.587	0.4	109	0.00
14	1,2-Dichlorobenzene	1.533	1.510	1.5	108	0.00
15	Benzyl Alcohol	1.337	1.277	4.5	101	0.00
16	2,2'-oxybis(1-Chloropropane	2.594	2.550	1.7	107	-0.01
17	2-Methylphenol	1.262	1.217	3.6	102	0.00
18	Hexachloroethane	0.543	0.576	-6.1	115	-0.01
19 P	n-Nitroso-di-n-propylamine	1.246	1.164	6.6	98	0.00
20	3+4-Methylphenols	1.804	1.750	3.0	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.502	0.479	4.6	101	0.00
23 S	Nitrobenzene-d5	0.357	0.366	-2.5	107	0.00
24	Nitrobenzene	0.371	0.376	-1.3	106	0.00
25	Isophorone	0.652	0.633	2.9	99	-0.01
26 C	2-Nitrophenol	0.167	0.192	-15.0	117	-0.01
27	2,4-Dimethylphenol	0.298	0.289	3.0	101	0.00
28	bis(2-Chloroethoxy)methane	0.427	0.418	2.1	102	-0.01
29 C	2,4-Dichlorophenol	0.332	0.341	-2.7	107	0.00
30	1,2,4-Trichlorobenzene	0.361	0.370	-2.5	109	-0.01
31	Naphthalene	0.982	0.977	0.5	106	0.00
32	Benzoic acid	0.194	0.214	-10.3	113	0.00
33	4-Chloroaniline	0.462	0.454	1.7	102	-0.01
34 C	Hexachlorobutadiene	0.214	0.233	-8.9	113	-0.01
35	Caprolactam	0.133	0.130	2.3	99	0.00
36 C	4-Chloro-3-methylphenol	0.375	0.364	2.9	101	0.00
37	2-Methylnaphthalene	0.716	0.708	1.1	103	-0.01
38	1-Methylnaphthalene	0.695	0.682	1.9	104	-0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.645	0.683	-5.9	109	-0.01
41 P	Hexachlorocyclopentadiene	0.308	0.350	-13.6	114	0.00
42 S	2,4,6-Tribromophenol	0.218	0.230	-5.5	103	0.00
43 C	2,4,6-Trichlorophenol	0.431	0.453	-5.1	105	0.00
44	2,4,5-Trichlorophenol	0.469	0.479	-2.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110218\
 Data File : BG037775.D
 Acq On : 3 Nov 2018 4:31
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 03 04:06:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.326	1.325	0.1	104	0.00
46	1,1'-Biphenyl	1.656	1.652	0.2	104	0.00
47	2-Chloronaphthalene	1.253	1.264	-0.9	105	0.00
48	2-Nitroaniline	0.380	0.402	-5.8	107	0.00
49	Acenaphthylene	1.885	1.891	-0.3	103	0.00
50	Dimethylphthalate	1.547	1.508	2.5	100	0.00
51	2,6-Dinitrotoluene	0.342	0.355	-3.8	103	0.00
52 C	Acenaphthene	1.161	1.143	1.6	103	0.00
53	3-Nitroaniline	0.380	0.385	-1.3	100	0.00
54 P	2,4-Dinitrophenol	0.100	0.113	-13.0	118	0.00
55	Dibenzofuran	1.923	1.873	2.6	102	0.00
56 P	4-Nitrophenol	0.281	0.267	5.0	94	0.00
57	2,4-Dinitrotoluene	0.449	0.492	-9.6	107	0.00
58	Fluorene	1.440	1.396	3.1	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.402	0.399	0.7	101	0.00
60	Diethylphthalate	1.588	1.558	1.9	101	0.00
61	4-Chlorophenyl-phenylether	0.840	0.821	2.3	101	0.00
62	4-Nitroaniline	0.378	0.383	-1.3	100	0.00
63	Azobenzene	1.516	1.445	4.7	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	0.093	0.114	-22.6	118	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.619	-2.8	100	0.00
67	4-Bromophenyl-phenylether	0.220	0.229	-4.1	102	0.00
68	Hexachlorobenzene	0.228	0.232	-1.8	99	0.00
69	Atrazine	0.218	0.215	1.4	93	0.00
70 C	Pentachlorophenol	0.152	0.157	-3.3	97	0.00
71	Phenanthrene	1.016	1.020	-0.4	99	0.00
72	Anthracene	0.998	0.998	0.0	98	0.00
73	Carbazole	0.998	1.008	-1.0	98	0.00
74	Di-n-butylphthalate	1.110	1.150	-3.6	99	0.00
75 C	Fluoranthene	1.153	1.147	0.5	97	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzidine	0.680	0.634	6.8	90	0.00
78	Pyrene	1.307	1.283	1.8	98	0.00
79 S	Terphenyl-d14	0.936	0.913	2.5	98	0.00
80	Butylbenzylphthalate	0.535	0.568	-6.2	102	0.00
81	Benzo(a)anthracene	1.183	1.188	-0.4	100	0.00
82	3,3'-Dichlorobenzidine	0.459	0.467	-1.7	98	0.00
83	Chrysene	1.138	1.137	0.1	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.750	0.797	-6.3	101	-0.01
85 c	Di-n-octyl phthalate	1.223	1.321	-8.0	103	-0.03
86	Indeno(1,2,3-cd)pyrene	1.220	1.105	9.4	88	-0.02
87 I	Perylene-d12	1.000	1.000	0.0	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110218\
Data File : BG037775.D
Acq On : 3 Nov 2018 4:31
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 03 04:06:53 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 14:06:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.096	1.089	0.6	98	0.00
89	Benzo(k)fluoranthene	1.074	1.103	-2.7	101	0.00
90 C	Benzo(a)pyrene	1.042	1.069	-2.6	101	-0.01
91	Dibenzo(a,h)anthracene	0.961	0.849	11.7	85	-0.02
92	Benzo(q,h,i)perylene	0.972	0.895	7.9	90	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0