

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110222\
 Data File : BG055408.D
 Acq On : 2 Nov 2022 9:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 03 06:02:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 02 05:05:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	87	0.01
2	1,4-Dioxane	40.000	38.883	2.8	86	0.01
3	Pyridine	40.000	39.048	2.4	86	0.01
4	n-Nitrosodimethylamine	40.000	39.847	0.4	87	0.01
5 S	2-Fluorophenol	80.000	79.169	1.0	88	0.01
6	Aniline	40.000	39.665	0.8	90	0.01
7 S	Phenol-d6	80.000	81.016	-1.3	90	0.00
8	2-Chlorophenol	40.000	39.259	1.9	89	0.01
9	Benzaldehyde	40.000	40.311	-0.8	90	0.00
10 C	Phenol	40.000	40.176	-0.4	91	0.00
11	bis(2-Chloroethyl)ether	40.000	39.496	1.3	91	0.00
12	1,3-Dichlorobenzene	40.000	38.477	3.8	88	0.01
13 C	1,4-Dichlorobenzene	40.000	39.100	2.2	89	0.01
14	1,2-Dichlorobenzene	40.000	38.974	2.6	89	0.01
15	Benzyl Alcohol	40.000	41.348	-3.4	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.585	3.5	89	0.00
17	2-Methylphenol	40.000	39.762	0.6	91	0.00
18	Hexachloroethane	40.000	39.234	1.9	91	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	38.268	4.3	90	0.00
20	3+4-Methylphenols	40.000	39.651	0.9	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	93	0.01
22	Acetophenone	40.000	37.955	5.1	91	0.01
23 S	Nitrobenzene-d5	80.000	76.388	4.5	90	0.01
24	Nitrobenzene	40.000	38.015	5.0	90	0.01
25	Isophorone	40.000	37.514	6.2	92	0.00
26 C	2-Nitrophenol	40.000	38.283	4.3	91	0.01
27	2,4-Dimethylphenol	40.000	37.985	5.0	93	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.796	5.5	93	0.00
29 C	2,4-Dichlorophenol	40.000	38.024	4.9	92	0.00
30	1,2,4-Trichlorobenzene	40.000	38.307	4.2	92	0.01
31	Naphthalene	40.000	38.156	4.6	92	0.01
32	Benzoic acid	40.000	38.107	4.7	97	-0.02
33	4-Chloroaniline	40.000	39.311	1.7	94	0.00
34 C	Hexachlorobutadiene	40.000	37.765	5.6	92	0.00
35	Caprolactam	40.000	37.837	5.4	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.263	1.8	94	0.00
37	2-Methylnaphthalene	40.000	37.837	5.4	93	0.01
38	1-Methylnaphthalene	40.000	37.858	5.4	93	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	38.120	4.7	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.117	9.7	91	0.01
42 S	2,4,6-Tribromophenol	80.000	82.003	-2.5	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.972	2.6	94	0.00
44	2,4,5-Trichlorophenol	40.000	39.388	1.5	95	0.00
45 S	2-Fluorobiphenyl	80.000	76.360	4.6	94	0.01
46	1,1'-Biphenyl	40.000	38.323	4.2	95	0.01
47	2-Chloronaphthalene	40.000	38.956	2.6	95	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110222\
 Data File : BG055408.D
 Acq On : 2 Nov 2022 9:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 03 06:02:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 02 05:05:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.929	0.2	95	0.00
49	Acenaphthylene	40.000	38.518	3.7	94	0.00
50	Dimethylphthalate	40.000	38.832	2.9	96	0.00
51	2,6-Dinitrotoluene	40.000	39.424	1.4	95	0.00
52 C	Acenaphthene	40.000	38.651	3.4	95	0.01
53	3-Nitroaniline	40.000	39.658	0.9	94	0.00
54 P	2,4-Dinitrophenol	40.000	36.827	7.9	92	0.00
55	Dibenzofuran	40.000	39.053	2.4	95	0.00
56 P	4-Nitrophenol	40.000	42.037	-5.1	95	0.00
57	2,4-Dinitrotoluene	40.000	41.186	-3.0	97	0.00
58	Fluorene	40.000	39.086	2.3	96	0.01
59	2,3,4,6-Tetrachlorophenol	40.000	39.345	1.6	96	0.00
60	Diethylphthalate	40.000	39.678	0.8	97	0.00
61	4-Chlorophenyl-phenylether	40.000	38.993	2.5	97	0.01
62	4-Nitroaniline	40.000	40.673	-1.7	96	0.00
63	Azobenzene	40.000	39.258	1.9	97	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.576	-1.4	95	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.108	4.7	98	0.00
67	4-Bromophenyl-phenylether	40.000	37.795	5.5	97	0.00
68	Hexachlorobenzene	40.000	38.239	4.4	97	0.00
69	Atrazine	40.000	38.829	2.9	98	0.00
70 C	Pentachlorophenol	40.000	37.002	7.5	95	0.00
71	Phenanthrene	40.000	38.602	3.5	97	0.00
72	Anthracene	40.000	38.446	3.9	97	0.01
73	Carbazole	40.000	39.002	2.5	97	0.00
74	Di-n-butylphthalate	40.000	39.740	0.6	98	0.01
75 C	Fluoranthene	40.000	38.802	3.0	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.01
77	Benzidine	40.000	45.377	-13.4	103	0.00
78	Pyrene	40.000	38.568	3.6	98	0.00
79 S	Terphenyl-d14	80.000	77.443	3.2	100	0.00
80	Butylbenzylphthalate	40.000	38.525	3.7	97	0.01
81	Benzo(a)anthracene	40.000	38.846	2.9	98	0.01
82	3,3'-Dichlorobenzidine	40.000	39.622	0.9	99	0.01
83	Chrysene	40.000	38.549	3.6	99	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	38.892	2.8	98	0.02
85 c	Di-n-octyl phthalate	40.000	38.740	3.1	98	0.02
86 I	Perylene-d12	20.000	20.000	0.0	97	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	38.359	4.1	98	0.03
88	Benzo(b)fluoranthene	40.000	37.997	5.0	95	0.02
89	Benzo(k)fluoranthene	40.000	38.980	2.6	100	0.02
90 C	Benzo(a)pyrene	40.000	38.231	4.4	98	0.02
91	Dibenzo(a,h)anthracene	40.000	38.788	3.0	99	0.03
92	Benzo(g,h,i)perylene	40.000	38.050	4.9	97	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110222\
Data File : BG055408.D
Acq On : 2 Nov 2022 9:45
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 03 06:02:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 02 05:05:12 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0