

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110323\
 Data File : BG059606.D
 Acq On : 3 Nov 2023 19:03
 Operator : MA/JU
 Sample : SSTD00567
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005441

Quant Time: Nov 04 01:57:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 01:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.353	152	16398	20.000	ng/ul	# 0.00
20) Naphthalene-d8	11.202	136	75845	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.986	164	60877	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.836	188	34931	20.000	ng/ul	0.10
79) Chrysene-d12	22.201	240	319	20.000	ng/ul	# 0.13
88) Perylene-d12	25.815	264	706	20.000	ng/ul	# 0.14
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.641	96	1217	2.887	ng/uL	0.00
4) Pyridine-d5	4.087	84	6260	5.238	ng/ul	0.00
7) Phenol-d5	7.466	99	7532	4.620	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.671	67	4165	5.330	ng/ul	0.00
11) 2-Chlorophenol-d4	7.871	132	4734	3.694	ng/ul	0.00
15) 4-Methylphenol-d8	9.040	113	5276	4.063	ng/ul	0.00
21) Nitrobenzene-d5	9.546	128	1398	2.584	ng/ul	0.00
24) 2-Nitrophenol-d4	10.274	143	2630	4.645	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.803	165	5313	4.425	ng/ul	0.00
31) 4-Chloroaniline-d4	11.343	131	9361	4.812	ng/ul	0.00
46) Dimethylphthalate-d6	14.381	166	22101	4.115	ng/ul	0.00
49) Acenaphthylene-d8	14.681	160	25990	4.105	ng/ul	0.00
54) 4-Nitrophenol-d4	15.163	143	1102	1.204	ng/ul	0.00
60) Fluorene-d10	15.967	176	20944	4.459	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.073	200	2806	16.039	ng/ul	0.00
73) Anthracene-d10	17.836	188	34931	18.635	ng/ul	0.00
81) Pyrene-d10	20.210	212	159	7.266	ng/ul	0.11
92) Benzo(a)pyrene-d12	25.544	264	261	6.317	ng/ul	0.12
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.688	88	1408	2.850	ng/uL#	80
5) Pyridine	4.105	79	5606	4.651	ng/ul#	90
6) Benzaldehyde	7.489	77	1843	2.335	ng/ul	99
8) Phenol	7.495	94	6874	4.481	ng/ul#	78
10) Bis(2-Chloroethyl)ether	7.759	93	4170	3.675	ng/ul	98
12) 2-Chlorophenol	7.900	128	4266	3.287	ng/ul	92
13) 2-Methylphenol	8.770	108	6338	4.889	ng/ul	84
14) 2,2'-oxybis(1-Chloropr...	8.864	45	2132	2.246	ng/ul	92
16) Acetophenone	9.199	105	11886	5.578	ng/ul	91
17) N-Nitroso-di-n-propyla...	9.152	70	6299	6.134	ng/ul#	84
18) 4-Methylphenol	9.105	108	4980	3.606	ng/ul	83
19) Hexachloroethane	9.440	117	2478	4.722	ng/ul#	62
22) Nitrobenzene	9.593	77	7102	5.690	ng/ul#	87
23) Isophorone	10.104	82	19801	6.421	ng/ul#	86
25) 2-Nitrophenol	10.309	139	3833	6.281	ng/ul#	87
26) 2,4-Dimethylphenol	10.321	107	9422	5.912	ng/ul#	72
27) Bis(2-Chloroethoxy)met...	10.591	93	8613	5.368	ng/ul#	89
29) 2,4-Dichlorophenol	10.826	162	5356	4.655	ng/ul#	85
30) Naphthalene	11.255	128	21885	4.729	ng/ul#	96
32) 4-Chloroaniline	11.373	127	8692	4.709	ng/ul#	74
33) Hexachlorobutadiene	11.484	225	4168	5.819	ng/ul	97
34) Caprolactam	12.125	113	788	1.930	ng/ul	96
35) 4-Chloro-3-methylphenol	12.436	107	8444	5.954	ng/ul#	60
36) 2-Methylnaphthalene	12.830	142	15788	4.892	ng/ul#	72


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110323\
 Data File : BG059606.D
 Acq On : 3 Nov 2023 19:03
 Operator : MA/JU
 Sample : SSTD00567
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005441

Quant Time: Nov 04 01:57:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 01:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.047	142	16181	4.989	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	13.171	216	7830	4.515	ng/ul#	71
40) Hexachlorocyclopentadiene	13.130	237	2288	2.232	ng/ul#	95
41) 2,4,6-Trichlorophenol	13.406	196	4626	3.961	ng/ul	98
42) 2,4,5-Trichlorophenol	13.406	196	4626	3.593	ng/ul	92
43) 1,1'-Biphenyl	13.823	154	23818	4.350	ng/ul	92
44) 2-Chloronaphthalene	13.876	162	16328	3.874	ng/ul	93
45) 2-Nitroaniline	14.081	65	2050	2.505	ng/ul	90
47) Dimethylphthalate	14.422	163	21060	3.915	ng/ul#	97
48) 2,6-Dinitrotoluene	14.563	165	4471	5.400	ng/ul#	61
50) Acenaphthylene	14.716	152	28035	3.845	ng/ul#	93
51) 3-Nitroaniline	14.910	138	1565	1.539	ng/ul#	41
52) Acenaphthene	15.051	153	19537	4.097	ng/ul	88
53) 2,4-Dinitrophenol	15.092	184	2169	4.364	ng/ul#	47
55) 4-Nitrophenol	15.163	109	4999	4.496	ng/ul#	72
56) Dibenzofuran	15.380	168	27726	4.466	ng/ul#	84
57) 2,4-Dinitrotoluene	15.339	165	5744	4.703	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	15.586	232	5378	5.186	ng/ul#	88
59) Diethylphthalate	15.768	149	23063	3.979	ng/ul#	92
61) Fluorene	16.026	166	23539	4.389	ng/ul	92
62) 4-Chlorophenyl-phenyle...	16.009	204	10764	4.525	ng/ul#	89
63) 4-Nitroaniline	16.056	138	3404	3.243	ng/ul#	83
66) 4,6-Dinitro-2-methylph...	16.114	198	124	0.673	ng/ul#	28
67) N-Nitrosodiphenylamine	16.296	169	147	0.124	ng/ul#	48
68) 4-Bromophenyl-phenylether	16.978	248	245	0.689	ng/ul#	1
69) Hexachlorobenzene	17.002	284	4990	12.517	ng/ul	88
70) Atrazine	17.154	200	7039	18.941	ng/ul	94
71) Pentachlorophenol	17.354	266	3607	15.591	ng/ul#	81
72) Phenanthrene	17.777	178	37067	16.525	ng/ul	95
74) Anthracene	17.871	178	43035	19.025	ng/ul#	95
75) 1,2,3,4-Tetrachloroben...	13.776	216	6865	15.179	ng/uL	89
76) Pentachlorobenzene	15.268	250	7071	16.050	ng/uL#	89
77) Carbazole	18.147	167	33544	17.522	ng/ul	96
78) Di-n-butylphthalate	18.835	149	98	0.038	ng/ul#	75
80) Fluoranthene	19.857	202	202	7.548	ng/ul#	1
82) Pyrene	20.133	202	44396	1598.790	ng/ul#	93
83) Butylbenzylphthalate	20.997	149	16612	1368.588	ng/ul	93
84) 3,3'-Dichlorobenzidine	22.125	252	216	26.474	ng/ul#	1
85) Benzo(a)anthracene	22.043	228	46856	1780.981	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	22.049	149	586	32.612	ng/ul#	58
87) Chrysene	22.119	228	41713	1724.343	ng/ul	93
89) Di-n-octyl phthalate	23.359	149	78	1.389	ng/ul	100
90) Benzo(b)fluoranthene	24.640	252	229	4.649	ng/ul#	1
91) Benzo(k)fluoranthene	24.593	252	48591	976.796	ng/ul	97
93) Benzo(a)pyrene	25.638	252	56	1.162	ng/ul#	64
94) Indeno(1,2,3-cd)pyrene	30.010	276	887	14.631	ng/ul#	41
95) Dibenzo(a,h)anthracene	30.104	278	150	2.977	ng/ul#	1
96) Benzo(g,h,i)perylene	31.367	276	246	4.942	ng/ul#	30





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 2.850 ng/uL
RT: 3.688 min Scan# 51
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion: 88 Resp: 1408
Ion Ratio Lower Upper
88 100
43 27.1 21.1 31.7
58 47.9 22.7 34.1#



#5
Pyridine
Concen: 4.651 ng/uL
RT: 4.105 min Scan# 122
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion: 79 Resp: 5606
Ion Ratio Lower Upper
79 100
52 31.4 27.8 41.6
51 16.1 19.8 29.6#









#6
Benzaldehyde
Concen: 2.335 ng/ul
RT: 7.489 min Scan# 698
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion: 77 Resp: 1843
Ion Ratio Lower Upper
77 100
105 73.0 59.9 89.9
106 74.4 59.7 89.5



#8
Phenol
Concen: 4.481 ng/ul
RT: 7.495 min Scan# 699
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion: 94 Resp: 6874
Ion Ratio Lower Upper
94 100
65 25.9 32.2 48.4#
66 72.6 45.5 68.3#









#10
Bis(2-Chloroethyl)ether
Concen: 3.675 ng/ul
RT: 7.759 min Scan# 744
Delta R.T. -0.011 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion: 93 Resp: 4170
Ion Ratio Lower Upper
93 100
63 56.2 44.2 66.2
95 34.7 26.3 39.5



#12
2-Chlorophenol
Concen: 3.287 ng/ul
RT: 7.900 min Scan# 768
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:128 Resp: 4266
Ion Ratio Lower Upper
128 100
64 52.1 45.6 68.4
130 27.2 27.0 40.6





#13

2-Methylphenol

Concen: 4.889 ng/ul

RT: 8.770 min Scan# 917

Delta R.T. -0.005 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :





SSTD005441

Tgt Ion:108 Resp: 6338

Ion Ratio Lower Upper

108 100

107 74.1 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 2.246 ng/ul

RT: 8.864 min Scan# 932

Delta R.T. 0.013 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03




Tgt Ion: 45 Resp: 2132


Ion Ratio Lower Upper

45 100

77 50.6 35.8 53.8

79 28.7 25.4 38.2









#16
Acetophenone
Concen: 5.578 ng/ul
RT: 9.199 min Scan# 91
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:105 Resp: 11886
Ion Ratio Lower Upper
105 100
77 72.6 66.2 99.2
51 18.7 15.3 22.9



#17
N-Nitroso-di-n-propylamine
Concen: 6.134 ng/ul
RT: 9.152 min Scan# 981
Delta R.T. -0.011 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion: 70 Resp: 6299
Ion Ratio Lower Upper
70 100
42 38.4 24.2 36.2#
101 7.9 5.1 7.7#
130 21.0 8.6 13.0#









#18
4-Methylphenol
Concen: 3.606 ng/ul
RT: 9.105 min Scan# 91
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:108 Resp: 4980
Ion Ratio Lower Upper
108 100
107 90.0 86.2 129.2



#19
Hexachloroethane
Concen: 4.722 ng/ul
RT: 9.440 min Scan# 1030
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:117 Resp: 2478
Ion Ratio Lower Upper
117 100
201 112.6 68.2 102.4#
199 0.0 26.3 39.5#









#22
Nitrobenzene
Concen: 5.690 ng/ul
RT: 9.593 min Scan# 1056
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion: 77 Resp: 7102
Ion Ratio Lower Upper
77 100
123 37.4 32.5 48.7
65 26.7 8.0 12.0#



#23
Isophorone
Concen: 6.421 ng/ul
RT: 10.104 min Scan# 1143
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion: 82 Resp: 19801
Ion Ratio Lower Upper
82 100
95 5.0 8.6 13.0#
138 17.7 10.2 15.2#









#25
2-Nitrophenol
Concen: 6.281 ng/ul
RT: 10.309 min Scan# 1178
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:139 Resp: 3833
Ion Ratio Lower Upper
139 100
65 36.9 28.4 42.6
109 55.5 33.3 49.9#



#26
2,4-Dimethylphenol
Concen: 5.912 ng/ul
RT: 10.321 min Scan# 1180
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:107 Resp: 9422
Ion Ratio Lower Upper
107 100
121 28.2 28.0 42.0
122 41.5 57.4 86.2#





#27

Bis(2-Chloroethoxy)methane

Concen: 5.368 ng/ul

RT: 10.591 min Scan# 1226

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion: 93 Resp: 8613

Ion Ratio Lower Upper

93 100

95 35.0 24.6 37.0

123 3.1 8.3 12.5#



#29

2,4-Dichlorophenol

Concen: 4.655 ng/ul

RT: 10.826 min Scan# 1266

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03




Tgt Ion:162 Resp: 5356


Ion Ratio Lower Upper

162 100

164 54.5 57.0 85.4#

98 31.3 23.6 35.4









#30
Naphthalene
Concen: 4.729 ng/ul
RT: 11.255 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:128 Resp: 21885
Ion Ratio Lower Upper
128 100
129 9.1 7.3 10.9
127 15.5 10.1 15.1#



#32
4-Chloroaniline
Concen: 4.709 ng/ul
RT: 11.373 min Scan# 1359
Delta R.T. 0.007 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:127 Resp: 8692
Ion Ratio Lower Upper
127 100
129 21.0 29.0 43.6#





#33

Hexachlorobutadiene

Concen: 5.819 ng/ul

RT: 11.484 min Scan# 11

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion:225 Resp: 4168

Ion Ratio Lower Upper

225 100

223 55.9 43.7 65.5

227 68.3 52.1 78.1



#34

Caprolactam

Concen: 1.930 ng/ul

RT: 12.125 min Scan# 1487

Delta R.T. -0.022 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03




Tgt Ion:113 Resp: 788


Ion Ratio Lower Upper

113 100

55 56.8 47.4 71.2

56 79.3 66.6 99.8









#35
4-Chloro-3-methylphenol
Concen: 5.954 ng/ul
RT: 12.436 min Scan# 11
Delta R.T. 0.007 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:107 Resp: 8444
Ion Ratio Lower Upper
107 100
144 22.7 17.8 26.6
142 41.2 69.5 104.3#



#36
2-Methylnaphthalene
Concen: 4.892 ng/ul
RT: 12.830 min Scan# 1607
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:142 Resp: 15788
Ion Ratio Lower Upper
142 100
141 120.8 75.1 112.7#









#37
1-Methylnaphthalene
Concen: 4.989 ng/ul
RT: 13.047 min Scan# 1644
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:142 Resp: 16181
Ion Ratio Lower Upper
142 100
141 88.2 70.9 106.3
116 7.7 5.0 7.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.515 ng/ul
RT: 13.171 min Scan# 1665
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:216 Resp: 7830
Ion Ratio Lower Upper
216 100
214 43.8 60.4 90.6#
179 17.7 16.3 24.5
108 34.1 18.9 28.3#









#40
Hexachlorocyclopentadiene
Concen: 2.232 ng/ul
RT: 13.130 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:237 Resp: 2288
Ion Ratio Lower Upper
237 100
235 55.1 44.9 67.3
272 0.0 7.2 10.8#



#41
2,4,6-Trichlorophenol
Concen: 3.961 ng/ul
RT: 13.406 min Scan# 1705
Delta R.T. -0.011 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:196 Resp: 4626
Ion Ratio Lower Upper
196 100
198 91.7 74.0 111.0
200 24.6 22.0 33.0





#42

2,4,5-Trichlorophenol

Concen: 3.593 ng/ul

RT: 13.406 min Scan# 1

Delta R.T. -0.075 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion:196 Resp: 4626

Ion Ratio Lower Upper

196 100

198 91.7 65.8 98.6

200 24.6 19.5 29.3



#43

1,1'-Biphenyl

Concen: 4.350 ng/ul

RT: 13.823 min Scan# 1776

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03




Tgt Ion:154 Resp: 23818


Ion Ratio Lower Upper

154 100

153 40.6 27.8 41.8

76 12.2 9.2 13.8





#44

2-Chloronaphthalene

Concen: 3.874 ng/ul

RT: 13.876 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion:162 Resp: 16328

Ion Ratio Lower Upper

162 100

164 24.8 22.2 33.4

127 34.7 31.4 47.2



#45

2-Nitroaniline

Concen: 2.505 ng/ul

RT: 14.081 min Scan# 1820

Delta R.T. -0.005 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03




Tgt Ion: 65 Resp: 2050


Ion Ratio Lower Upper

65 100

92 58.3 56.5 84.7

138 125.9 94.6 141.8









#47
Dimethylphthalate
Concen: 3.915 ng/ul
RT: 14.422 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:163 Resp: 21060
Ion Ratio Lower Upper
163 100
194 4.5 3.0 4.4#
164 7.8 7.0 10.6



#48
2,6-Dinitrotoluene
Concen: 5.400 ng/ul
RT: 14.563 min Scan# 1902
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:165 Resp: 4471
Ion Ratio Lower Upper
165 100
89 53.0 79.2 118.8#
121 18.2 17.9 26.9









#50
Acenaphthylene
Concen: 3.845 ng/ul
RT: 14.716 min Scan# 1928
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:152 Resp: 28035
Ion Ratio Lower Upper
152 100
151 24.0 15.1 22.7#
153 12.5 10.2 15.2



#51
3-Nitroaniline
Concen: 1.539 ng/ul
RT: 14.910 min Scan# 1961
Delta R.T. 0.007 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:138 Resp: 1565
Ion Ratio Lower Upper
138 100
108 25.8 8.2 12.2#
92 65.4 110.6 165.8#









#52
 Acenaphthene
 Concen: 4.097 ng/ul
 RT: 15.051 min Scan# 1985
 Delta R.T. 0.007 min
 Lab File: BG059606.D
 Acq: 3 Nov 2023 19:03




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005441


Tgt Ion:	153	Resp:	19537
Ion Ratio	Lower	Upper	
153	100		
152	45.0	39.4	59.0
154	101.9	70.7	106.1



#53
 2,4-Dinitrophenol
 Concen: 4.364 ng/ul
 RT: 15.092 min Scan# 1992
 Delta R.T. -0.011 min
 Lab File: BG059606.D
 Acq: 3 Nov 2023 19:03

Tgt Ion:	184	Resp:	2169
Ion Ratio	Lower	Upper	
184	100		
63	14.9	33.8	50.8#
154	30.2	65.1	97.7#









#55
4-Nitrophenol
Concen: 4.496 ng/ul
RT: 15.163 min Scan# 2004
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:109 Resp: 4999
Ion Ratio Lower Upper
109 100
139 68.0 48.8 73.2
65 42.2 63.3 94.9#



#56
Dibenzofuran
Concen: 4.466 ng/ul
RT: 15.380 min Scan# 2041
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:168 Resp: 27726
Ion Ratio Lower Upper
168 100
139 41.7 42.3 63.5#





#57

2,4-Dinitrotoluene

Concen: 4.703 ng/ul

RT: 15.339 min Scan# 2034

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion:165 Resp: 5744

Ion Ratio Lower Upper

165 100

63 31.3 34.3 51.5#

182 0.0 1.0 1.4#



#58

2,3,4,6-Tetrachlorophenol

Concen: 5.186 ng/ul

RT: 15.586 min Scan# 2076

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Tgt Ion:232 Resp: 5378




Ion Ratio Lower Upper


232 100

131 47.5 40.7 61.1

130 15.3 3.0 4.4#

166 14.8 21.5 32.3#









#59
Diethylphthalate
Concen: 3.979 ng/ul
RT: 15.768 min Scan# 2107
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:149 Resp: 23063
Ion Ratio Lower Upper
149 100
177 25.3 16.2 24.4#
150 14.2 9.9 14.9



#61
Fluorene
Concen: 4.389 ng/ul
RT: 16.026 min Scan# 2151
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:166 Resp: 23539
Ion Ratio Lower Upper
166 100
165 98.8 73.0 109.4
167 10.1 9.3 13.9









#62
4-Chlorophenyl-phenylether
Concen: 4.525 ng/ul
RT: 16.009 min Scan# 2148
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:204 Resp: 10764
Ion Ratio Lower Upper
204 100
206 10.8 24.1 36.1#
141 77.4 62.5 93.7



#63
4-Nitroaniline
Concen: 3.243 ng/ul
RT: 16.056 min Scan# 2156
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:138 Resp: 3404
Ion Ratio Lower Upper
138 100
92 26.0 43.8 65.6#
108 172.9 150.1 225.1









#66
4,6-Dinitro-2-methylphenol
Concen: 0.673 ng/ul
RT: 16.114 min Scan# 2166
Delta R.T. 0.025 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:198 Resp: 124
Ion Ratio Lower Upper
198 100
182 0.0 4.3 6.5#
77 83.8 29.2 43.8#



#67
N-Nitrosodiphenylamine
Concen: 0.124 ng/ul
RT: 16.296 min Scan# 2197
Delta R.T. 0.066 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:169 Resp: 147
Ion Ratio Lower Upper
169 100
168 100.0 51.2 76.8#
167 0.0 30.2 45.4#









#68
4-Bromophenyl-phenylether
Concen: 0.689 ng/ul
RT: 16.978 min Scan# 211
Delta R.T. 0.072 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:248 Resp: 245
Ion Ratio Lower Upper
248 100
250 0.0 75.6 113.4#
141 0.0 85.4 128.0#



#69
Hexachlorobenzene
Concen: 12.517 ng/ul
RT: 17.002 min Scan# 2317
Delta R.T. 0.007 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:284 Resp: 4990
Ion Ratio Lower Upper
284 100
142 50.1 49.8 74.6
249 34.1 25.3 37.9





#70

Atrazine

Concen: 18.941 ng/ul

RT: 17.154 min Scan# 21

Delta R.T. -0.005 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion:200 Resp: 7039

Ion Ratio Lower Upper

200 100

173 26.4 21.8 32.6

215 33.0 30.6 46.0



#71

Pentachlorophenol

Concen: 15.591 ng/ul

RT: 17.354 min Scan# 2377

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03




Tgt Ion:266 Resp: 3607


Ion Ratio Lower Upper

266 100

264 61.7 46.6 70.0

268 38.1 50.3 75.5#









#72
Phenanthrene
Concen: 16.525 ng/ul
RT: 17.777 min Scan# 2449
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:178 Resp: 37067
Ion Ratio Lower Upper
178 100
179 14.4 11.8 17.6
176 16.8 16.4 24.6



#74
Anthracene
Concen: 19.025 ng/ul
RT: 17.871 min Scan# 2465
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:178 Resp: 43035
Ion Ratio Lower Upper
178 100
179 11.4 12.0 18.0#
176 18.7 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 15.179 ng/uL
RT: 13.776 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:216 Resp: 6865
Ion Ratio Lower Upper
216 100
214 89.8 81.8 122.6
218 63.6 46.3 69.5



#76
Pentachlorobenzene
Concen: 16.050 ng/uL
RT: 15.268 min Scan# 2022
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:250 Resp: 7071
Ion Ratio Lower Upper
250 100
248 51.8 43.6 65.4
252 46.0 47.2 70.8#









#77
 Carbazole
 Concen: 17.522 ng/ul
 RT: 18.147 min Scan# 21
 Delta R.T. 0.007 min
 Lab File: BG059606.D
 Acq: 3 Nov 2023 19:03




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005441


Tgt Ion:	167	Resp:	33544
Ion Ratio	Lower	Upper	
167	100		
166	23.8	17.9	26.9
139	12.8	11.8	17.8



#78
 Di-n-butylphthalate
 Concen: 0.038 ng/ul
 RT: 18.835 min Scan# 2629
 Delta R.T. 0.183 min
 Lab File: BG059606.D
 Acq: 3 Nov 2023 19:03

Tgt Ion:	149	Resp:	98
Ion Ratio	Lower	Upper	
149	100		
150	0.0	7.6	11.4#
104	0.0	6.6	9.8#









#80
Fluoranthene
Concen: 7.548 ng/ul
RT: 19.857 min Scan# 2803
Delta R.T. 0.089 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:202 Resp: 202
Ion Ratio Lower Upper
202 100
101 67.5 4.6 7.0#
100 0.0 4.8 7.2#



#82
Pyrene
Concen: 1598.790 ng/ul
RT: 20.133 min Scan# 2850
Delta R.T. 0.001 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:202 Resp: 44396
Ion Ratio Lower Upper
202 100
101 5.4 4.6 6.8
100 10.5 5.1 7.7#





#83

Butylbenzylphthalate

Concen: 1368.588 ng/ul

RT: 20.997 min Scan# 2997

Delta R.T. 0.001 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion:149 Resp: 16612

Ion Ratio Lower Upper

149 100

91 73.1 64.6 96.8

206 17.3 13.4 20.2



#84

3,3'-Dichlorobenzidine

Concen: 26.474 ng/ul

RT: 22.125 min Scan# 3189

Delta R.T. 0.166 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03




Tgt Ion:252 Resp: 216


Ion Ratio Lower Upper

252 100

254 0.0 54.9 82.3#

126 108.7 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 1780.981 ng/ul
RT: 22.043 min Scan# 3175
Delta R.T. -0.005 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:228 Resp: 46856
Ion Ratio Lower Upper
228 100
229 25.6 17.7 26.5
226 27.0 21.1 31.7



#86
Bis(2-ethylhexyl)phthalate
Concen: 32.612 ng/ul
RT: 22.049 min Scan# 3176
Delta R.T. 0.160 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:149 Resp: 586
Ion Ratio Lower Upper
149 100
167 0.0 18.0 27.0#
279 0.0 2.6 4.0#









#87
 Chrysene
 Concen: 1724.343 ng/ul
 RT: 22.119 min Scan# 3188
 Delta R.T. -0.005 min
 Lab File: BG059606.D
 Acq: 3 Nov 2023 19:03

Instrument :
 BNA_G
 ClientSampleId :
 SST005441




Tgt Ion:228 Resp: 41713


Ion	Ratio	Lower	Upper
228	100		
226	30.5	21.7	32.5
229	16.7	15.7	23.5



#89
 Di-n-octyl phthalate
 Concen: 1.389 ng/ul
 RT: 23.359 min Scan# 3399
 Delta R.T. 0.125 min
 Lab File: BG059606.D
 Acq: 3 Nov 2023 19:03

Tgt Ion:149 Resp: 78









#90
Benzo(b)fluoranthene
Concen: 4.649 ng/ul
RT: 24.640 min Scan# 301
Delta R.T. 0.136 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:252 Resp: 229
Ion Ratio Lower Upper
252 100
253 221.2 15.1 22.7#
125 69.1 2.4 3.6#



#91
Benzo(k)fluoranthene
Concen: 976.796 ng/ul
RT: 24.593 min Scan# 3609
Delta R.T. 0.007 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:252 Resp: 48591
Ion Ratio Lower Upper
252 100
253 20.5 17.9 26.9
125 5.4 4.1 6.1









#93
Benzo(a)pyrene
Concen: 1.162 ng/ul
RT: 25.638 min Scan# 31
Delta R.T. 0.130 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03




Instrument :
BNA_G
ClientSampleId :
SSTD005441


Tgt Ion:252 Resp: 56
Ion Ratio Lower Upper
252 100
253 0.0 14.9 22.3#
125 0.0 3.8 5.8#



#94
Indeno(1,2,3-cd)pyrene
Concen: 14.631 ng/ul
RT: 30.010 min Scan# 4531
Delta R.T. 0.154 min
Lab File: BG059606.D
Acq: 3 Nov 2023 19:03

Tgt Ion:276 Resp: 887
Ion Ratio Lower Upper
276 100
138 39.7 5.4 8.2#
277 0.0 18.5 27.7#





#95

Dibenzo(a,h)anthracene

Concen: 2.977 ng/ul

RT: 30.104 min Scan# 4

Delta R.T. 0.154 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Instrument :

BNA_G

ClientSampleId :

SSTD005441





Tgt Ion:278 Resp: 150

Ion Ratio Lower Upper

278 100

139 0.0 4.3 6.5#

279 125.5 17.8 26.8#



#96

Benzo(g,h,i)perylene

Concen: 4.942 ng/ul

RT: 31.367 min Scan# 4762

Delta R.T. 0.171 min

Lab File: BG059606.D

Acq: 3 Nov 2023 19:03

Tgt Ion:276 Resp: 246

Ion Ratio Lower Upper

276 100

138 60.9 4.6 6.8#

277 0.0 17.5 26.3#

