

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110323\
 Data File : BG059607.D
 Acq On : 3 Nov 2023 19:43
 Operator : MA/JU
 Sample : SST01068
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST010442

Quant Time: Nov 04 01:58:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 01:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.353	152	22137	20.000	ng/ul	0.00
20) Naphthalene-d8	11.203	136	101621	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.981	164	69656	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.736	188	169558	20.000	ng/ul	0.00
79) Chrysene-d12	22.073	240	145053	20.000	ng/ul	# 0.00
88) Perylene-d12	25.674	264	211569	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.647	96	2230	3.918	ng/uL	0.00
4) Pyridine-d5	4.076	84	17818	11.045	ng/ul	0.00
7) Phenol-d5	7.460	99	19799	8.996	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.666	67	12088	11.459	ng/ul	0.00
11) 2-Chlorophenol-d4	7.872	132	14942	8.637	ng/ul	0.00
15) 4-Methylphenol-d8	9.041	113	16931	9.657	ng/ul	0.00
21) Nitrobenzene-d5	9.546	128	6182	8.527	ng/ul	0.00
24) 2-Nitrophenol-d4	10.263	143	8328	10.979	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.803	165	13615	8.463	ng/ul	0.00
31) 4-Chloroaniline-d4	11.338	131	24335	9.336	ng/ul	0.00
46) Dimethylphthalate-d6	14.376	166	55539	9.038	ng/ul	0.00
49) Acenaphthylene-d8	14.687	160	65881	9.094	ng/ul	0.00
54) 4-Nitrophenol-d4	15.151	143	9363	8.940	ng/ul	0.00
60) Fluorene-d10	15.968	176	55965	10.412	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.068	200	10783	12.698	ng/ul	0.00
73) Anthracene-d10	17.736	188	169558	18.635	ng/ul	-0.10
81) Pyrene-d10	20.104	212	92943	9.340	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.422	264	126981	10.255	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.688	88	2236	3.352	ng/uL#	84
5) Pyridine	4.105	79	18144	11.150	ng/ul	94
6) Benzaldehyde	7.490	77	17226	16.165	ng/ul#	75
8) Phenol	7.495	94	20647	9.969	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.766	93	14622	9.546	ng/ul	96
12) 2-Chlorophenol	7.895	128	14116	8.058	ng/ul#	80
13) 2-Methylphenol	8.770	108	16743	9.567	ng/ul#	76
14) 2,2'-oxybis(1-Chloropr...	8.864	45	5299	4.135	ng/ul#	84
16) Acetophenone	9.199	105	33866	11.773	ng/ul	97
17) N-Nitroso-di-n-propyla...	9.158	70	14920	10.762	ng/ul#	81
18) 4-Methylphenol	9.105	108	18370	9.853	ng/ul	88
19) Hexachloroethane	9.434	117	6725	9.494	ng/ul#	66
22) Nitrobenzene	9.599	77	20291	12.134	ng/ul	93
23) Isophorone	10.104	82	48381	11.709	ng/ul	98
25) 2-Nitrophenol	10.304	139	9351	11.437	ng/ul	94
26) 2,4-Dimethylphenol	10.333	107	23914	11.200	ng/ul#	87
27) Bis(2-Chloroethoxy)met...	10.580	93	22246	10.347	ng/ul	92
29) 2,4-Dichlorophenol	10.827	162	14664	9.513	ng/ul#	85
30) Naphthalene	11.250	128	61367	9.897	ng/ul#	92
32) 4-Chloroaniline	11.367	127	23839	9.640	ng/ul#	87
33) Hexachlorobutadiene	11.479	225	10741	11.193	ng/ul#	91
34) Caprolactam	12.137	113	4662	8.520	ng/ul#	76
35) 4-Chloro-3-methylphenol	12.437	107	18101	9.526	ng/ul#	76
36) 2-Methylnaphthalene	12.830	142	42722	9.880	ng/ul	91


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110323\
 Data File : BG059607.D
 Acq On : 3 Nov 2023 19:43
 Operator : MA/JU
 Sample : SSTD01068
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD010442

Quant Time: Nov 04 01:58:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 01:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.048	142	45326	10.431	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	13.177	216	19122	9.637	ng/ul	90
40) Hexachlorocyclopentadiene	13.124	237	13312	11.347	ng/ul#	90
41) 2,4,6-Trichlorophenol	13.406	196	8806	6.591	ng/ul#	64
42) 2,4,5-Trichlorophenol	13.488	196	8111	5.506	ng/ul	93
43) 1,1'-Biphenyl	13.817	154	56866	9.078	ng/ul	94
44) 2-Chloronaphthalene	13.870	162	44903	9.310	ng/ul	92
45) 2-Nitroaniline	14.088	65	10390	11.096	ng/ul	94
47) Dimethylphthalate	14.423	163	58949	9.577	ng/ul#	97
48) 2,6-Dinitrotoluene	14.564	165	9900	10.451	ng/ul#	70
50) Acenaphthylene	14.711	152	78263	9.380	ng/ul	97
51) 3-Nitroaniline	14.904	138	9509	8.174	ng/ul#	94
52) Acenaphthene	15.045	153	51610	9.459	ng/ul	92
53) 2,4-Dinitrophenol	15.104	184	2724	4.790	ng/ul	89
55) 4-Nitrophenol	15.163	109	13935	10.953	ng/ul	97
56) Dibenzofuran	15.380	168	66407	9.349	ng/ul	96
57) 2,4-Dinitrotoluene	15.339	165	14598	10.446	ng/ul#	96
58) 2,3,4,6-Tetrachlorophenol	15.586	232	10743	9.055	ng/ul#	75
59) Diethylphthalate	15.768	149	62176	9.374	ng/ul	99
61) Fluorene	16.027	166	58802	9.583	ng/ul#	73
62) 4-Chlorophenyl-phenyle...	16.009	204	30436	11.183	ng/ul	92
63) 4-Nitroaniline	16.062	138	9997	8.323	ng/ul	83
66) 4,6-Dinitro-2-methylph...	16.085	198	8950	10.008	ng/ul#	84
67) N-Nitrosodiphenylamine	16.226	169	52416	9.139	ng/ul	98
68) 4-Bromophenyl-phenylether	16.914	248	15898	9.205	ng/ul#	87
69) Hexachlorobenzene	16.996	284	16096	8.318	ng/ul#	87
70) Atrazine	17.161	200	18317	10.154	ng/ul#	94
71) Pentachlorophenol	17.355	266	9519	8.476	ng/ul	94
72) Phenanthrene	17.778	178	96932	8.903	ng/ul#	93
74) Anthracene	17.872	178	101154	9.212	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.776	216	20359	9.274	ng/ul#	80
76) Pentachlorobenzene	15.269	250	20009	9.356	ng/ul	96
77) Carbazole	18.142	167	91682	9.866	ng/ul#	89
78) Di-n-butylphthalate	18.653	149	97023	7.749	ng/ul	99
80) Fluoranthene	19.769	202	110556	9.085	ng/ul#	95
82) Pyrene	20.134	202	112089	8.877	ng/ul#	97
83) Butylbenzylphthalate	20.997	149	40220	7.287	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.955	252	35960	9.693	ng/ul#	90
85) Benzo(a)anthracene	22.049	228	114495	9.571	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.884	149	60059	7.350	ng/ul#	96
87) Chrysene	22.049	228	114495	10.409	ng/ul	95
89) Di-n-octyl phthalate	23.236	149	101062	6.005	ng/ul	100
90) Benzo(b)fluoranthene	24.499	252	137885	9.342	ng/ul#	93
91) Benzo(k)fluoranthene	24.581	252	141598	9.499	ng/ul#	96
93) Benzo(a)pyrene	25.498	252	134609	9.323	ng/ul#	89
94) Indeno(1,2,3-cd)pyrene	29.846	276	55065	3.031	ng/ul#	97
95) Dibenzo(a,h)anthracene	29.928	278	56181	3.720	ng/ul#	91
96) Benzo(g,h,i)perylene	31.179	276	19334	1.296	ng/ul#	91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 3.352 ng/uL
RT: 3.688 min Scan# 51
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion: 88 Resp: 2236
Ion Ratio Lower Upper
88 100
43 15.4 21.1 31.7#
58 34.4 22.7 34.1#



#5
Pyridine
Concen: 11.150 ng/uL
RT: 4.105 min Scan# 122
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion: 79 Resp: 18144
Ion Ratio Lower Upper
79 100
52 37.6 27.8 41.6
51 20.4 19.8 29.6









#6
Benzaldehyde
Concen: 16.165 ng/ul
RT: 7.490 min Scan# 699
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion: 77 Resp: 17226
Ion Ratio Lower Upper
77 100
105 57.2 59.9 89.9#
106 49.1 59.7 89.5#



#8
Phenol
Concen: 9.969 ng/ul
RT: 7.495 min Scan# 699
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion: 94 Resp: 20647
Ion Ratio Lower Upper
94 100
65 41.6 32.2 48.4
66 74.6 45.5 68.3#









#10
Bis(2-Chloroethyl)ether
Concen: 9.546 ng/ul
RT: 7.766 min Scan# 746
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion: 93 Resp: 14622
Ion Ratio Lower Upper
93 100
63 59.6 44.2 66.2
95 32.1 26.3 39.5



#12
2-Chlorophenol
Concen: 8.058 ng/ul
RT: 7.895 min Scan# 767
Delta R.T. -0.010 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:128 Resp: 14116
Ion Ratio Lower Upper
128 100
64 37.4 45.6 68.4#
130 39.4 27.0 40.6





#13

2-Methylphenol

Concen: 9.567 ng/ul

RT: 8.770 min Scan# 917

Delta R.T. -0.004 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :





SSTD010442

Tgt Ion:108 Resp: 16743

Ion Ratio Lower Upper

108 100

107 66.9 71.4 107.0#



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.135 ng/ul

RT: 8.864 min Scan# 932

Delta R.T. 0.013 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion: 45 Resp: 5299


Ion Ratio Lower Upper

45 100

77 59.9 35.8 53.8#

79 28.6 25.4 38.2









#16
Acetophenone
Concen: 11.773 ng/ul
RT: 9.199 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:105 Resp: 33866
Ion Ratio Lower Upper
105 100
77 85.0 66.2 99.2
51 20.5 15.3 22.9



#17
N-Nitroso-di-n-propylamine
Concen: 10.762 ng/ul
RT: 9.158 min Scan# 982
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion: 70 Resp: 14920
Ion Ratio Lower Upper
70 100
42 39.9 24.2 36.2#
101 11.8 5.1 7.7#
130 20.2 8.6 13.0#









#18
4-Methylphenol
Concen: 9.853 ng/ul
RT: 9.105 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:108 Resp: 18370
Ion Ratio Lower Upper
108 100
107 95.1 86.2 129.2



#19
Hexachloroethane
Concen: 9.494 ng/ul
RT: 9.434 min Scan# 1029
Delta R.T. -0.010 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:117 Resp: 6725
Ion Ratio Lower Upper
117 100
201 47.3 68.2 102.4#
199 41.1 26.3 39.5#









#22
Nitrobenzene
Concen: 12.134 ng/ul
RT: 9.599 min Scan# 1057
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion: 77 Resp: 20291
Ion Ratio Lower Upper
77 100
123 36.1 32.5 48.7
65 11.8 8.0 12.0



#23
Isophorone
Concen: 11.709 ng/ul
RT: 10.104 min Scan# 1143
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion: 82 Resp: 48381
Ion Ratio Lower Upper
82 100
95 10.8 8.6 13.0
138 14.3 10.2 15.2









#25
2-Nitrophenol
Concen: 11.437 ng/ul
RT: 10.304 min Scan# 1178
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:139 Resp: 9351
Ion Ratio Lower Upper
139 100
65 38.6 28.4 42.6
109 37.7 33.3 49.9



#26
2,4-Dimethylphenol
Concen: 11.200 ng/ul
RT: 10.333 min Scan# 1182
Delta R.T. 0.007 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:107 Resp: 23914
Ion Ratio Lower Upper
107 100
121 36.7 28.0 42.0
122 56.5 57.4 86.2#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.347 ng/ul
 RT: 10.580 min Scan# 1224
 Delta R.T. -0.010 min
 Lab File: BG059607.D
 Acq: 3 Nov 2023 19:43




Instrument :
 BNA_G
 ClientSampleId :
 SSTD010442


Tgt Ion: 93 Resp: 22246
 Ion Ratio Lower Upper
 93 100
 95 25.3 24.6 37.0
 123 9.3 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 9.513 ng/ul
 RT: 10.827 min Scan# 1266
 Delta R.T. 0.002 min
 Lab File: BG059607.D
 Acq: 3 Nov 2023 19:43

Tgt Ion:162 Resp: 14664
 Ion Ratio Lower Upper
 162 100
 164 62.4 57.0 85.4
 98 15.1 23.6 35.4#









#30
Naphthalene
Concen: 9.897 ng/ul
RT: 11.250 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:128 Resp: 61367
Ion Ratio Lower Upper
128 100
129 8.8 7.3 10.9
127 17.6 10.1 15.1#



#32
4-Chloroaniline
Concen: 9.640 ng/ul
RT: 11.367 min Scan# 1358
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:127 Resp: 23839
Ion Ratio Lower Upper
127 100
129 28.9 29.0 43.6#





#33

Hexachlorobutadiene

Concen: 11.193 ng/ul

RT: 11.479 min Scan# 11

Delta R.T. -0.004 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:225 Resp: 10741

Ion Ratio Lower Upper

225 100

223 40.0 43.7 65.5#

227 65.0 52.1 78.1



#34

Caprolactam

Concen: 8.520 ng/ul

RT: 12.137 min Scan# 1489

Delta R.T. -0.010 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:113 Resp: 4662


Ion Ratio Lower Upper

113 100

55 55.8 47.4 71.2

56 49.0 66.6 99.8#









#35
4-Chloro-3-methylphenol
Concen: 9.526 ng/ul
RT: 12.437 min Scan# 11
Delta R.T. 0.007 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:107 Resp: 18101
Ion Ratio Lower Upper
107 100
144 18.0 17.8 26.6
142 61.5 69.5 104.3#



#36
2-Methylnaphthalene
Concen: 9.880 ng/ul
RT: 12.830 min Scan# 1607
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:142 Resp: 42722
Ion Ratio Lower Upper
142 100
141 102.6 75.1 112.7









#37
1-Methylnaphthalene
Concen: 10.431 ng/ul
RT: 13.048 min Scan# 1644
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:142 Resp: 45326
Ion Ratio Lower Upper
142 100
141 91.4 70.9 106.3
116 4.6 5.0 7.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.637 ng/ul
RT: 13.177 min Scan# 1666
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:216 Resp: 19122
Ion Ratio Lower Upper
216 100
214 64.3 60.4 90.6
179 16.6 16.3 24.5
108 21.9 18.9 28.3









#40
Hexachlorocyclopentadiene
Concen: 11.347 ng/ul
RT: 13.124 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:237 Resp: 13312
Ion Ratio Lower Upper
237 100
235 49.1 44.9 67.3
272 13.0 7.2 10.8#



#41
2,4,6-Trichlorophenol
Concen: 6.591 ng/ul
RT: 13.406 min Scan# 1705
Delta R.T. -0.010 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:196 Resp: 8806
Ion Ratio Lower Upper
196 100
198 50.6 74.0 111.0#
200 33.6 22.0 33.0#





#42

2,4,5-Trichlorophenol

Concen: 5.506 ng/ul

RT: 13.488 min Scan# 1718

Delta R.T. 0.007 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:196 Resp: 8111

Ion Ratio Lower Upper

196 100

198 87.6 65.8 98.6

200 29.1 19.5 29.3



#43

1,1'-Biphenyl

Concen: 9.078 ng/ul

RT: 13.817 min Scan# 1775

Delta R.T. -0.004 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:154 Resp: 56866


Ion Ratio Lower Upper

154 100

153 39.4 27.8 41.8

76 12.0 9.2 13.8









#44
2-Chloronaphthalene
Concen: 9.310 ng/ul
RT: 13.870 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:162 Resp: 44903
Ion Ratio Lower Upper
162 100
164 29.2 22.2 33.4
127 46.8 31.4 47.2



#45
2-Nitroaniline
Concen: 11.096 ng/ul
RT: 14.088 min Scan# 1821
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion: 65 Resp: 10390
Ion Ratio Lower Upper
65 100
92 77.2 56.5 84.7
138 123.1 94.6 141.8









#47
Dimethylphthalate
Concen: 9.577 ng/ul
RT: 14.423 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:163 Resp: 58949
Ion Ratio Lower Upper
163 100
194 2.9 3.0 4.4#
164 9.8 7.0 10.6



#48
2,6-Dinitrotoluene
Concen: 10.451 ng/ul
RT: 14.564 min Scan# 1902
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:165 Resp: 9900
Ion Ratio Lower Upper
165 100
89 65.2 79.2 118.8#
121 28.9 17.9 26.9#









#50
Acenaphthylene
Concen: 9.380 ng/ul
RT: 14.711 min Scan# 1928
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:152 Resp: 78263
Ion Ratio Lower Upper
152 100
151 17.2 15.1 22.7
153 11.9 10.2 15.2



#51
3-Nitroaniline
Concen: 8.174 ng/ul
RT: 14.904 min Scan# 1960
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:138 Resp: 9509
Ion Ratio Lower Upper
138 100
108 5.1 8.2 12.2#
92 132.1 110.6 165.8









#52
Acenaphthene
Concen: 9.459 ng/ul
RT: 15.045 min Scan# 1984
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:153 Resp: 51610
Ion Ratio Lower Upper
153 100
152 49.3 39.4 59.0
154 100.0 70.7 106.1



#53
2,4-Dinitrophenol
Concen: 4.790 ng/ul
RT: 15.104 min Scan# 1994
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:184 Resp: 2724
Ion Ratio Lower Upper
184 100
63 37.1 33.8 50.8
154 70.8 65.1 97.7









#55
4-Nitrophenol
Concen: 10.953 ng/ul
RT: 15.163 min Scan# 2005
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:109 Resp: 13935
Ion Ratio Lower Upper
109 100
139 62.5 48.8 73.2
65 76.4 63.3 94.9



#56
Dibenzofuran
Concen: 9.349 ng/ul
RT: 15.380 min Scan# 2041
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:168 Resp: 66407
Ion Ratio Lower Upper
168 100
139 56.0 42.3 63.5





#57

2,4-Dinitrotoluene

Concen: 10.446 ng/ul

RT: 15.339 min Scan# 2034

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:165 Resp: 14598

Ion Ratio Lower Upper

165 100

63 45.3 34.3 51.5

182 2.4 1.0 1.4#



#58

2,3,4,6-Tetrachlorophenol

Concen: 9.055 ng/ul

RT: 15.586 min Scan# 2076

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Tgt Ion:232 Resp: 10743




Ion Ratio Lower Upper


232 100

131 69.3 40.7 61.1#

130 0.0 3.0 4.4#

166 40.1 21.5 32.3#









#59
Diethylphthalate
Concen: 9.374 ng/ul
RT: 15.768 min Scan# 2107
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:149 Resp: 62176
Ion Ratio Lower Upper
149 100
177 19.8 16.2 24.4
150 12.9 9.9 14.9



#61
Fluorene
Concen: 9.583 ng/ul
RT: 16.027 min Scan# 2151
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:166 Resp: 58802
Ion Ratio Lower Upper
166 100
165 119.3 73.0 109.4#
167 15.8 9.3 13.9#









#62
4-Chlorophenyl-phenylether
Concen: 11.183 ng/ul
RT: 16.009 min Scan# 2148
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:204 Resp: 30436
Ion Ratio Lower Upper
204 100
206 28.0 24.1 36.1
141 69.8 62.5 93.7



#63
4-Nitroaniline
Concen: 8.323 ng/ul
RT: 16.062 min Scan# 2157
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:138 Resp: 9997
Ion Ratio Lower Upper
138 100
92 59.1 43.8 65.6
108 217.6 150.1 225.1









#66
4,6-Dinitro-2-methylphenol
Concen: 10.008 ng/ul
RT: 16.085 min Scan# 2161
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:198 Resp: 8950
Ion Ratio Lower Upper
198 100
182 0.0 4.3 6.5#
77 26.9 29.2 43.8#



#67
N-Nitrosodiphenylamine
Concen: 9.139 ng/ul
RT: 16.226 min Scan# 2185
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:169 Resp: 52416
Ion Ratio Lower Upper
169 100
168 61.7 51.2 76.8
167 38.1 30.2 45.4









#68
4-Bromophenyl-phenylether
Concen: 9.205 ng/ul
RT: 16.914 min Scan# 21
Delta R.T. 0.007 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:248 Resp: 15898
Ion Ratio Lower Upper
248 100
250 92.6 75.6 113.4
141 129.5 85.4 128.0#



#69
Hexachlorobenzene
Concen: 8.318 ng/ul
RT: 16.996 min Scan# 2316
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:284 Resp: 16096
Ion Ratio Lower Upper
284 100
142 59.7 49.8 74.6
249 48.7 25.3 37.9#





#70

Atrazine

Concen: 10.154 ng/ul

RT: 17.161 min Scan# 21

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:200 Resp: 18317

Ion Ratio Lower Upper

200 100

173 21.2 21.8 32.6#

215 39.1 30.6 46.0



#71

Pentachlorophenol

Concen: 8.476 ng/ul

RT: 17.355 min Scan# 2377

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:266 Resp: 9519


Ion Ratio Lower Upper

266 100

264 51.2 46.6 70.0

268 60.2 50.3 75.5





#72

Phenanthrene

Concen: 8.903 ng/ul

RT: 17.778 min Scan# 2449

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:178 Resp: 96932

Ion Ratio Lower Upper

178 100

179 17.8 11.8 17.6#

176 17.4 16.4 24.6



#74

Anthracene

Concen: 9.212 ng/ul

RT: 17.872 min Scan# 2465

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:178 Resp: 101154


Ion Ratio Lower Upper

178 100

179 15.7 12.0 18.0

176 15.7 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.274 ng/uL
RT: 13.776 min Scan# 1768
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:216 Resp: 20359
Ion Ratio Lower Upper
216 100
214 79.6 81.8 122.6#
218 45.4 46.3 69.5#



#76
Pentachlorobenzene
Concen: 9.356 ng/uL
RT: 15.269 min Scan# 2022
Delta R.T. -0.004 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:250 Resp: 20009
Ion Ratio Lower Upper
250 100
248 55.9 43.6 65.4
252 54.2 47.2 70.8





#77

Carbazole

Concen: 9.866 ng/ul

RT: 18.142 min Scan# 2511

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:167 Resp: 91682

Ion Ratio Lower Upper

167 100

166 15.8 17.9 26.9#

139 11.8 11.8 17.8#



#78

Di-n-butylphthalate

Concen: 7.749 ng/ul

RT: 18.653 min Scan# 2598

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:149 Resp: 97023


Ion Ratio Lower Upper

149 100

150 10.4 7.6 11.4

104 8.3 6.6 9.8





#80

Fluoranthene

Concen: 9.085 ng/ul

RT: 19.769 min Scan# 21

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:202 Resp: 110556

Ion Ratio Lower Upper

202 100

101 6.2 4.6 7.0

100 3.0 4.8 7.2#



#82

Pyrene

Concen: 8.877 ng/ul

RT: 20.134 min Scan# 2850

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:202 Resp: 112089


Ion Ratio Lower Upper

202 100

101 4.1 4.6 6.8#

100 7.2 5.1 7.7





#83

Butylbenzylphthalate

Concen: 7.287 ng/ul

RT: 20.997 min Scan# 2997

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:149 Resp: 40220

Ion Ratio Lower Upper

149 100

91 73.6 64.6 96.8

206 18.4 13.4 20.2



#84

3,3'-Dichlorobenzidine

Concen: 9.693 ng/ul

RT: 21.955 min Scan# 3160

Delta R.T. -0.004 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:252 Resp: 35960


Ion Ratio Lower Upper

252 100

254 60.7 54.9 82.3

126 5.9 8.2 12.4#





#85

Benzo(a)anthracene

Concen: 9.571 ng/ul

RT: 22.049 min Scan# 3176

Delta R.T. 0.002 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:228 Resp: 114495

Ion Ratio Lower Upper

228 100

229 19.6 17.7 26.5

226 22.8 21.1 31.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 7.350 ng/ul

RT: 21.884 min Scan# 3148

Delta R.T. -0.004 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:149 Resp: 60059


Ion Ratio Lower Upper

149 100

167 24.2 18.0 27.0

279 5.7 2.6 4.0#









#87
Chrysene
Concen: 10.409 ng/ul
RT: 22.049 min Scan# 3189
Delta R.T. -0.075 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43




Instrument :
BNA_G
ClientSampleId :
SSTD010442


Tgt Ion:228 Resp: 114495
Ion Ratio Lower Upper
228 100
226 22.8 21.7 32.5
229 19.6 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 6.005 ng/ul
RT: 23.236 min Scan# 3378
Delta R.T. 0.002 min
Lab File: BG059607.D
Acq: 3 Nov 2023 19:43

Tgt Ion:149 Resp: 101062





#90

Benzo(b)fluoranthene

Concen: 9.342 ng/ul

RT: 24.499 min Scan# 3593

Delta R.T. -0.004 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:252 Resp: 137885

Ion Ratio Lower Upper

252 100

253 16.6 15.1 22.7

125 8.7 2.4 3.6#



#91

Benzo(k)fluoranthene

Concen: 9.499 ng/ul

RT: 24.581 min Scan# 3607

Delta R.T. -0.004 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:252 Resp: 141598


Ion Ratio Lower Upper

252 100

253 24.1 17.9 26.9

125 3.3 4.1 6.1#





#93

Benzo(a)pyrene

Concen: 9.323 ng/ul

RT: 25.498 min Scan# 31

Delta R.T. -0.010 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:252 Resp: 134609

Ion Ratio Lower Upper

252 100

253 24.3 14.9 22.3#

125 3.6 3.8 5.8#



#94

Indeno(1,2,3-cd)pyrene

Concen: 3.031 ng/ul

RT: 29.846 min Scan# 4503

Delta R.T. -0.010 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43




Tgt Ion:276 Resp: 55065


Ion Ratio Lower Upper

276 100

138 8.2 5.4 8.2#

277 21.5 18.5 27.7





#95

Dibenzo(a,h)anthracene

Concen: 3.720 ng/ul

RT: 29.928 min Scan# 4

Delta R.T. -0.022 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Instrument :

BNA_G

ClientSampleId :

SSTD010442





Tgt Ion:278 Resp: 56181

Ion Ratio Lower Upper

278 100

139 11.4 4.3 6.5#

279 18.8 17.8 26.8



#96

Benzo(g,h,i)perylene

Concen: 1.296 ng/ul

RT: 31.179 min Scan# 4730

Delta R.T. -0.016 min

Lab File: BG059607.D

Acq: 3 Nov 2023 19:43

Tgt Ion:276 Resp: 19334

Ion Ratio Lower Upper

276 100

138 7.5 4.6 6.8#

277 17.1 17.5 26.3#

