

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110323\
 Data File : BG059608.D
 Acq On : 3 Nov 2023 20:24
 Operator : MA/JU
 Sample : SST02069
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020443

Quant Time: Nov 04 01:58:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 01:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.346	152	19908	20.000	ng/u1	0.00
20) Naphthalene-d8	11.201	136	130276	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.985	164	88740	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.735	188	206026	20.000	ng/u1	0.00
79) Chrysene-d12	22.071	240	176853	20.000	ng/u1	0.00
88) Perylene-d12	25.678	264	250172	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.645	96	5439	10.626	ng/uL	0.00
4) Pyridine-d5	4.080	84	41675	28.725	ng/u1	0.00
7) Phenol-d5	7.465	99	55441	28.012	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.670	67	29903	31.520	ng/u1	0.00
11) 2-Chlorophenol-d4	7.870	132	36695	23.586	ng/u1	0.00
15) 4-Methylphenol-d8	9.039	113	43306	27.467	ng/u1	0.00
21) Nitrobenzene-d5	9.550	128	18732	20.155	ng/u1	0.00
24) 2-Nitrophenol-d4	10.267	143	20871	21.462	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.796	165	40560	19.666	ng/u1	0.00
31) 4-Chloroaniline-d4	11.342	131	62846	18.808	ng/u1	0.00
46) Dimethylphthalate-d6	14.374	166	144140	18.412	ng/u1	0.00
49) Acenaphthylene-d8	14.685	160	173370	18.784	ng/u1	0.00
54) 4-Nitrophenol-d4	15.156	143	23411	17.547	ng/u1	0.00
60) Fluorene-d10	15.972	176	137866	20.134	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.072	200	23303	22.584	ng/u1	0.00
73) Anthracene-d10	17.835	188	213675	19.327	ng/u1	0.00
81) Pyrene-d10	20.103	212	225082	18.552	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.420	264	283971	19.395	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.687	88	6774	11.292	ng/uL	100
5) Pyridine	4.104	79	45483	31.079	ng/u1	100
6) Benzaldehyde	7.494	77	38584	40.261	ng/u1	100
8) Phenol	7.494	94	52735	28.313	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.770	93	38615	28.031	ng/u1	100
12) 2-Chlorophenol	7.905	128	37440	23.764	ng/u1	100
13) 2-Methylphenol	8.775	108	45698	29.037	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.851	45	9552	8.288	ng/u1	100
16) Acetophenone	9.198	105	83007	32.088	ng/u1	100
17) N-Nitroso-di-n-propyla...	9.163	70	41291	33.118	ng/u1	100
18) 4-Methylphenol	9.104	108	47037	28.055	ng/u1	100
19) Hexachloroethane	9.445	117	16955	26.615	ng/u1	100
22) Nitrobenzene	9.597	77	55610	25.939	ng/u1	100
23) Isophorone	10.103	82	130174	24.574	ng/u1	100
25) 2-Nitrophenol	10.308	139	23788	22.695	ng/u1	100
26) 2,4-Dimethylphenol	10.326	107	59928	21.893	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.590	93	53351	19.357	ng/u1	100
29) 2,4-Dichlorophenol	10.825	162	38087	19.273	ng/u1	100
30) Naphthalene	11.254	128	150199	18.896	ng/u1	100
32) 4-Chloroaniline	11.366	127	58661	18.504	ng/u1	100
33) Hexachlorobutadiene	11.483	225	25791	20.964	ng/u1	100
34) Caprolactam	12.147	113	3593	5.122	ng/u1	100
35) 4-Chloro-3-methylphenol	12.429	107	51813	21.270	ng/u1	100
36) 2-Methylnaphthalene	12.835	142	110294	19.897	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110323\
 Data File : BG059608.D
 Acq On : 3 Nov 2023 20:24
 Operator : MA/JU
 Sample : SSTD02069
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020443

Quant Time: Nov 04 01:58:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 01:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.046	142	113650	20.402	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.176	216	48008	18.992	ng/ul	100
40) Hexachlorocyclopentadiene	13.128	237	31311	20.950	ng/ul	100
41) 2,4,6-Trichlorophenol	13.416	196	30462	17.895	ng/ul	100
42) 2,4,5-Trichlorophenol	13.481	196	33828	18.025	ng/ul	100
43) 1,1'-Biphenyl	13.822	154	157676	19.757	ng/ul	100
44) 2-Chloronaphthalene	13.875	162	117532	19.129	ng/ul	100
45) 2-Nitroaniline	14.086	65	26051	21.839	ng/ul	100
47) Dimethylphthalate	14.427	163	147780	18.845	ng/ul	100
48) 2,6-Dinitrotoluene	14.568	165	24373	20.195	ng/ul	100
50) Acenaphthylene	14.715	152	195517	18.394	ng/ul	100
51) 3-Nitroaniline	14.903	138	29329	19.789	ng/ul	100
52) Acenaphthene	15.044	153	128640	18.506	ng/ul	100
53) 2,4-Dinitrophenol	15.103	184	17010	23.477	ng/ul	100
55) 4-Nitrophenol	15.167	109	40070	24.721	ng/ul	100
56) Dibenzofuran	15.379	168	173927	19.220	ng/ul	100
57) 2,4-Dinitrotoluene	15.338	165	42792	24.035	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.584	232	30184	19.969	ng/ul	100
59) Diethylphthalate	15.772	149	152755	18.077	ng/ul	100
61) Fluorene	16.025	166	149213	19.087	ng/ul	100
62) 4-Chlorophenyl-phenyle...	16.007	204	73578	21.221	ng/ul	100
63) 4-Nitroaniline	16.060	138	28566	18.669	ng/ul	100
66) 4,6-Dinitro-2-methylph...	16.090	198	24440	22.491	ng/ul	100
67) N-Nitrosodiphenylamine	16.231	169	132375	18.994	ng/ul	100
68) 4-Bromophenyl-phenylether	16.906	248	45004	21.445	ng/ul	100
69) Hexachlorobenzene	16.995	284	43269	18.403	ng/ul	100
70) Atrazine	17.159	200	46010	20.991	ng/ul	100
71) Pentachlorophenol	17.353	266	26579	19.479	ng/ul	100
72) Phenanthrene	17.776	178	243875	18.434	ng/ul	100
74) Anthracene	17.870	178	252291	18.910	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.781	216	46410	17.398	ng/ul	100
76) Pentachlorobenzene	15.273	250	51550	19.839	ng/ul	100
77) Carbazole	18.140	167	219026	19.398	ng/ul	100
78) Di-n-butylphthalate	18.651	149	247683	16.280	ng/ul	100
80) Fluoranthene	19.768	202	278548	18.774	ng/ul	100
82) Pyrene	20.132	202	288658	18.750	ng/ul	100
83) Butylbenzylphthalate	20.996	149	108418	16.111	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.959	252	91745	20.282	ng/ul	100
85) Benzo(a)anthracene	22.047	228	279737	19.179	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.889	149	156962	15.756	ng/ul	100
87) Chrysene	22.124	228	256602	19.133	ng/ul	100
89) Di-n-octyl phthalate	23.234	149	265399	13.337	ng/ul	100
90) Benzo(b)fluoranthene	24.503	252	298665	17.113	ng/ul	100
91) Benzo(k)fluoranthene	24.586	252	311084	17.648	ng/ul	100
93) Benzo(a)pyrene	25.508	252	325590	19.070	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	29.856	276	334753	15.583	ng/ul	100
95) Dibenzo(a,h)anthracene	29.950	278	334967	18.758	ng/ul	100
96) Benzo(g,h,i)perylene	31.172	276	156091	8.849	ng/ul#	92





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.292 ng/uL
RT: 3.687 min Scan# 51
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion: 88 Resp: 6774
Ion Ratio Lower Upper
88 100
43 26.4 21.1 31.7
58 28.4 22.7 34.1



#5
Pyridine
Concen: 31.079 ng/uL
RT: 4.104 min Scan# 122
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion: 79 Resp: 45483
Ion Ratio Lower Upper
79 100
52 34.7 27.8 41.6
51 24.7 19.8 29.6









#6
Benzaldehyde
Concen: 40.261 ng/ul
RT: 7.494 min Scan# 699
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion: 77 Resp: 38584
Ion Ratio Lower Upper
77 100
105 74.9 59.9 89.9
106 74.6 59.7 89.5



#8
Phenol
Concen: 28.313 ng/ul
RT: 7.494 min Scan# 699
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion: 94 Resp: 52735
Ion Ratio Lower Upper
94 100
65 40.3 32.2 48.4
66 56.9 45.5 68.3









#10
 Bis(2-Chloroethyl)ether
 Concen: 28.031 ng/ul
 RT: 7.770 min Scan# 746
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020443

Tgt Ion: 93 Resp: 38615




Ion	Ratio	Lower	Upper
93	100		
63	55.2	44.2	66.2
95	32.9	26.3	39.5




#12
 2-Chlorophenol
 Concen: 23.764 ng/ul
 RT: 7.905 min Scan# 769
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Tgt Ion: 128 Resp: 37440

Ion	Ratio	Lower	Upper
128	100		
64	57.0	45.6	68.4
130	33.8	27.0	40.6





#13

2-Methylphenol

Concen: 29.037 ng/ul

RT: 8.775 min Scan# 917

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :





SSTD020443

Tgt Ion:108 Resp: 45698

Ion Ratio Lower Upper

108 100

107 89.2 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 8.288 ng/ul

RT: 8.851 min Scan# 930

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion: 45 Resp: 9552


Ion Ratio Lower Upper

45 100

77 44.8 35.8 53.8

79 31.8 25.4 38.2









#16
Acetophenone
Concen: 32.088 ng/ul
RT: 9.198 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:105 Resp: 83007
Ion Ratio Lower Upper
105 100
77 82.7 66.2 99.2
51 19.1 15.3 22.9



#17
N-Nitroso-di-n-propylamine
Concen: 33.118 ng/ul
RT: 9.163 min Scan# 983
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion: 70 Resp: 41291
Ion Ratio Lower Upper
70 100
42 30.2 24.2 36.2
101 6.4 5.1 7.7
130 10.8 8.6 13.0









#18
4-Methylphenol
Concen: 28.055 ng/ul
RT: 9.104 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:108 Resp: 47037
Ion Ratio Lower Upper
108 100
107 107.7 86.2 129.2



#19
Hexachloroethane
Concen: 26.615 ng/ul
RT: 9.445 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:117 Resp: 16955
Ion Ratio Lower Upper
117 100
201 85.3 68.2 102.4
199 32.9 26.3 39.5









#22
Nitrobenzene
Concen: 25.939 ng/ul
RT: 9.597 min Scan# 1057
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion: 77 Resp: 55610
Ion Ratio Lower Upper
77 100
123 40.6 32.5 48.7
65 10.0 8.0 12.0



#23
Isophorone
Concen: 24.574 ng/ul
RT: 10.103 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion: 82 Resp: 130174
Ion Ratio Lower Upper
82 100
95 10.8 8.6 13.0
138 12.7 10.2 15.2









#25
2-Nitrophenol
Concen: 22.695 ng/ul
RT: 10.308 min Scan# 1178
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:139 Resp: 23788
Ion Ratio Lower Upper
139 100
65 35.5 28.4 42.6
109 41.6 33.3 49.9



#26
2,4-Dimethylphenol
Concen: 21.893 ng/ul
RT: 10.326 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:107 Resp: 59928
Ion Ratio Lower Upper
107 100
121 35.0 28.0 42.0
122 71.8 57.4 86.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.357 ng/ul
 RT: 10.590 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020443


Tgt Ion: 93 Resp: 53351
 Ion Ratio Lower Upper
 93 100
 95 30.8 24.6 37.0
 123 10.4 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 19.273 ng/ul
 RT: 10.825 min Scan# 1266
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Tgt Ion: 162 Resp: 38087
 Ion Ratio Lower Upper
 162 100
 164 71.2 57.0 85.4
 98 29.5 23.6 35.4









#30
Naphthalene
Concen: 18.896 ng/ul
RT: 11.254 min Scan# 1339
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:128 Resp: 150199
Ion Ratio Lower Upper
128 100
129 9.1 7.3 10.9
127 12.6 10.1 15.1



#32
4-Chloroaniline
Concen: 18.504 ng/ul
RT: 11.366 min Scan# 1358
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:127 Resp: 58661
Ion Ratio Lower Upper
127 100
129 36.3 29.0 43.6





#33

Hexachlorobutadiene

Concen: 20.964 ng/ul

RT: 11.483 min Scan# 11

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:225 Resp: 25791

Ion Ratio Lower Upper

225 100

223 54.6 43.7 65.5

227 65.1 52.1 78.1



#34

Caprolactam

Concen: 5.122 ng/ul

RT: 12.147 min Scan# 1491

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:113 Resp: 3593


Ion Ratio Lower Upper

113 100

55 59.3 47.4 71.2

56 83.2 66.6 99.8









#35
4-Chloro-3-methylphenol
Concen: 21.270 ng/ul
RT: 12.429 min Scan# 1539
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:107 Resp: 51813
Ion Ratio Lower Upper
107 100
144 22.2 17.8 26.6
142 86.9 69.5 104.3



#36
2-Methylnaphthalene
Concen: 19.897 ng/ul
RT: 12.835 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:142 Resp: 110294
Ion Ratio Lower Upper
142 100
141 93.9 75.1 112.7









#37
1-Methylnaphthalene
Concen: 20.402 ng/ul
RT: 13.046 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:142 Resp: 113650
Ion Ratio Lower Upper
142 100
141 88.6 70.9 106.3
116 6.2 5.0 7.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.992 ng/ul
RT: 13.176 min Scan# 1666
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:216 Resp: 48008
Ion Ratio Lower Upper
216 100
214 75.5 60.4 90.6
179 20.4 16.3 24.5
108 23.6 18.9 28.3





#40

Hexachlorocyclopentadiene

Concen: 20.950 ng/ul

RT: 13.128 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:237 Resp: 31311

Ion Ratio Lower Upper

237 100

235 56.1 44.9 67.3

272 9.0 7.2 10.8



#41

2,4,6-Trichlorophenol

Concen: 17.895 ng/ul

RT: 13.416 min Scan# 1707

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:196 Resp: 30462


Ion Ratio Lower Upper

196 100

198 92.5 74.0 111.0

200 27.5 22.0 33.0





#42

2,4,5-Trichlorophenol

Concen: 18.025 ng/ul

RT: 13.481 min Scan# 1718

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:196 Resp: 33828

Ion Ratio Lower Upper

196 100

198 82.2 65.8 98.6

200 24.4 19.5 29.3



#43

1,1'-Biphenyl

Concen: 19.757 ng/ul

RT: 13.822 min Scan# 1776

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:154 Resp: 157676


Ion Ratio Lower Upper

154 100

153 34.8 27.8 41.8

76 11.5 9.2 13.8









#44
2-Chloronaphthalene
Concen: 19.129 ng/ul
RT: 13.875 min Scan# 117532
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion	Ratio	Lower	Upper
162	100		
164	27.8	22.2	33.4
127	39.3	31.4	47.2



#45
2-Nitroaniline
Concen: 21.839 ng/ul
RT: 14.086 min Scan# 1821
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion	Ratio	Lower	Upper
65	100		
92	70.6	56.5	84.7
138	118.2	94.6	141.8









#47
Dimethylphthalate
Concen: 18.845 ng/ul
RT: 14.427 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:163 Resp: 147780
Ion Ratio Lower Upper
163 100
194 3.7 3.0 4.4
164 8.8 7.0 10.6



#48
2,6-Dinitrotoluene
Concen: 20.195 ng/ul
RT: 14.568 min Scan# 1903
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:165 Resp: 24373
Ion Ratio Lower Upper
165 100
89 99.0 79.2 118.8
121 22.4 17.9 26.9









#50
Acenaphthylene
Concen: 18.394 ng/ul
RT: 14.715 min Scan# 1928
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:152 Resp: 195517
Ion Ratio Lower Upper
152 100
151 18.9 15.1 22.7
153 12.7 10.2 15.2



#51
3-Nitroaniline
Concen: 19.789 ng/ul
RT: 14.903 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:138 Resp: 29329
Ion Ratio Lower Upper
138 100
108 10.2 8.2 12.2
92 138.2 110.6 165.8









#52
Acenaphthene
Concen: 18.506 ng/ul
RT: 15.044 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:153 Resp: 128640
Ion Ratio Lower Upper
153 100
152 49.2 39.4 59.0
154 88.4 70.7 106.1



#53
2,4-Dinitrophenol
Concen: 23.477 ng/ul
RT: 15.103 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:184 Resp: 17010
Ion Ratio Lower Upper
184 100
63 42.3 33.8 50.8
154 81.4 65.1 97.7









#55
4-Nitrophenol
Concen: 24.721 ng/ul
RT: 15.167 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:109 Resp: 40070
Ion Ratio Lower Upper
109 100
139 61.0 48.8 73.2
65 79.1 63.3 94.9



#56
Dibenzofuran
Concen: 19.220 ng/ul
RT: 15.379 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:168 Resp: 173927
Ion Ratio Lower Upper
168 100
139 52.9 42.3 63.5









#57
2,4-Dinitrotoluene
Concen: 24.035 ng/ul
RT: 15.338 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:165 Resp: 42792
Ion Ratio Lower Upper
165 100
63 42.9 34.3 51.5
182 1.2 1.0 1.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.969 ng/ul
RT: 15.584 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:232 Resp: 30184
Ion Ratio Lower Upper
232 100
131 50.9 40.7 61.1
130 3.7 3.0 4.4
166 26.9 21.5 32.3









#59
Diethylphthalate
Concen: 18.077 ng/ul
RT: 15.772 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:149 Resp: 152755
Ion Ratio Lower Upper
149 100
177 20.3 16.2 24.4
150 12.4 9.9 14.9



#61
Fluorene
Concen: 19.087 ng/ul
RT: 16.025 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:166 Resp: 149213
Ion Ratio Lower Upper
166 100
165 91.2 73.0 109.4
167 11.6 9.3 13.9









#62
4-Chlorophenyl-phenylether
Concen: 21.221 ng/ul
RT: 16.007 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:204 Resp: 73578
Ion Ratio Lower Upper
204 100
206 30.1 24.1 36.1
141 78.1 62.5 93.7



#63
4-Nitroaniline
Concen: 18.669 ng/ul
RT: 16.060 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:138 Resp: 28566
Ion Ratio Lower Upper
138 100
92 54.7 43.8 65.6
108 187.6 150.1 225.1









#66
 4,6-Dinitro-2-methylphenol
 Concen: 22.491 ng/ul
 RT: 16.090 min Scan# 2162
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020443


Tgt Ion:198 Resp: 24440
 Ion Ratio Lower Upper
 198 100
 182 5.4 4.3 6.5
 77 36.5 29.2 43.8



#67
 N-Nitrosodiphenylamine
 Concen: 18.994 ng/ul
 RT: 16.231 min Scan# 2186
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Tgt Ion:169 Resp: 132375
 Ion Ratio Lower Upper
 169 100
 168 64.0 51.2 76.8
 167 37.8 30.2 45.4









#68
4-Bromophenyl-phenylether
Concen: 21.445 ng/ul
RT: 16.906 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:248 Resp: 45004
Ion Ratio Lower Upper
248 100
250 94.5 75.6 113.4
141 106.7 85.4 128.0



#69
Hexachlorobenzene
Concen: 18.403 ng/ul
RT: 16.995 min Scan# 2316
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:284 Resp: 43269
Ion Ratio Lower Upper
284 100
142 62.2 49.8 74.6
249 31.6 25.3 37.9





#70

Atrazine

Concen: 20.991 ng/ul

RT: 17.159 min Scan# 2344

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:200 Resp: 46010

Ion Ratio Lower Upper

200 100

173 27.2 21.8 32.6

215 38.3 30.6 46.0



#71

Pentachlorophenol

Concen: 19.479 ng/ul

RT: 17.353 min Scan# 2377

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:266 Resp: 26579


Ion Ratio Lower Upper

266 100

264 58.3 46.6 70.0

268 62.9 50.3 75.5





#72

Phenanthrene

Concen: 18.434 ng/ul

RT: 17.776 min Scan# 2449

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:178 Resp: 243875

Ion Ratio Lower Upper

178 100

179 14.7 11.8 17.6

176 20.5 16.4 24.6



#74

Anthracene

Concen: 18.910 ng/ul

RT: 17.870 min Scan# 2465

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:178 Resp: 252291


Ion Ratio Lower Upper

178 100

179 15.0 12.0 18.0

176 19.5 15.6 23.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 17.398 ng/uL

RT: 13.781 min Scan# 1769

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:216 Resp: 46410

Ion Ratio Lower Upper

216 100

214 102.2 81.8 122.6

218 57.9 46.3 69.5



#76

Pentachlorobenzene

Concen: 19.839 ng/uL

RT: 15.273 min Scan# 2023

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:250 Resp: 51550


Ion Ratio Lower Upper

250 100

248 54.5 43.6 65.4

252 59.0 47.2 70.8





#77

Carbazole

Concen: 19.398 ng/ul

RT: 18.140 min Scan# 2511

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:167 Resp: 219026

Ion Ratio Lower Upper

167 100

166 22.4 17.9 26.9

139 14.8 11.8 17.8



#78

Di-n-butylphthalate

Concen: 16.280 ng/ul

RT: 18.651 min Scan# 2598

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:149 Resp: 247683


Ion Ratio Lower Upper

149 100

150 9.5 7.6 11.4

104 8.2 6.6 9.8





#80

Fluoranthene

Concen: 18.774 ng/ul

RT: 19.768 min Scan# 2788

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:202 Resp: 278548

Ion Ratio Lower Upper

202 100

101 5.8 4.6 7.0

100 6.0 4.8 7.2



#82

Pyrene

Concen: 18.750 ng/ul

RT: 20.132 min Scan# 2850

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:202 Resp: 288658


Ion Ratio Lower Upper

202 100

101 5.7 4.6 6.8

100 6.4 5.1 7.7





#83

Butylbenzylphthalate

Concen: 16.111 ng/ul

RT: 20.996 min Scan# 2997

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:149 Resp: 108418

Ion Ratio Lower Upper

149 100

91 80.7 64.6 96.8

206 16.8 13.4 20.2



#84

3,3'-Dichlorobenzidine

Concen: 20.282 ng/ul

RT: 21.959 min Scan# 3161

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:252 Resp: 91745


Ion Ratio Lower Upper

252 100

254 68.6 54.9 82.3

126 10.3 8.2 12.4









#85
 Benzo(a)anthracene
 Concen: 19.179 ng/ul
 RT: 22.047 min Scan# 31176
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020443

Tgt Ion:228 Resp: 279737




Ion	Ratio	Lower	Upper
228	100		
229	22.1	17.7	26.5
226	26.4	21.1	31.7




#86
 Bis(2-ethylhexyl)phthalate
 Concen: 15.756 ng/ul
 RT: 21.889 min Scan# 3149
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Tgt Ion:149 Resp: 156962

Ion	Ratio	Lower	Upper
149	100		
167	22.5	18.0	27.0
279	3.3	2.6	4.0









#87
Chrysene
Concen: 19.133 ng/ul
RT: 22.124 min Scan# 3189
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:228 Resp: 256602
Ion Ratio Lower Upper
228 100
226 27.1 21.7 32.5
229 19.6 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 13.337 ng/ul
RT: 23.234 min Scan# 3378
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:149 Resp: 265399









#90
Benzo(b)fluoranthene
Concen: 17.113 ng/ul
RT: 24.503 min Scan# 3594
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24




Instrument :
BNA_G
ClientSampleId :
SSTD020443


Tgt Ion:252 Resp: 298665
Ion Ratio Lower Upper
252 100
253 18.9 15.1 22.7
125 3.0 2.4 3.6



#91
Benzo(k)fluoranthene
Concen: 17.648 ng/ul
RT: 24.586 min Scan# 3608
Delta R.T. 0.000 min
Lab File: BG059608.D
Acq: 3 Nov 2023 20:24

Tgt Ion:252 Resp: 311084
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 5.1 4.1 6.1





#93

Benzo(a)pyrene

Concen: 19.070 ng/ul

RT: 25.508 min Scan# 3765

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24

Instrument :

BNA_G

ClientSampleId :

SSTD020443





Tgt Ion:252 Resp: 325590

Ion Ratio Lower Upper

252 100

253 18.6 14.9 22.3

125 4.8 3.8 5.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 15.583 ng/ul

RT: 29.856 min Scan# 4505

Delta R.T. 0.000 min

Lab File: BG059608.D

Acq: 3 Nov 2023 20:24




Tgt Ion:276 Resp: 334753


Ion Ratio Lower Upper

276 100

138 6.8 5.4 8.2

277 23.1 18.5 27.7









#95
 Dibenzo(a,h)anthracene
 Concen: 18.758 ng/ul
 RT: 29.950 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020443

Tgt Ion:278 Resp: 334967
 Ion Ratio Lower Upper
 278 100
 139 5.4 4.3 6.5
 279 22.3 17.8 26.8



#96
 Benzo(g,h,i)perylene
 Concen: 8.849 ng/ul
 RT: 31.172 min Scan# 4729
 Delta R.T. -0.024 min
 Lab File: BG059608.D
 Acq: 3 Nov 2023 20:24

Tgt Ion:276 Resp: 156091
 Ion Ratio Lower Upper
 276 100
 138 12.0 4.6 6.8#
 277 19.3 17.5 26.3

