

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG110518\  
 Data File : BG037822.D  
 Acq On : 6 Nov 2018 3:32  
 Operator : JU/SJ  
 Sample : J5812-13  
 Misc :  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 TP2-B

Quant Time: Nov 06 06:02:20 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Oct 18 14:06:42 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.10  | 152  | 60145    | 20.00  | ng    | -0.01    |
| 21) Naphthalene-d8          | 10.92 | 136  | 278162   | 20.00  | ng    | -0.01    |
| 39) Acenaphthene-d10        | 14.73 | 164  | 181917   | 20.00  | ng    | -0.01    |
| 64) Phenanthrene-d10        | 17.47 | 188  | 394504   | 20.00  | ng    | -0.01    |
| 76) Chrysene-d12            | 21.76 | 240  | 354933   | 20.00  | ng    | -0.02    |
| 87) Perylene-d12            | 25.09 | 264  | 330678   | 20.00  | ng    | -0.02    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.64  | 112  | 369187   | 102.62 | ng    | 0.00     |
| 7) Phenol-d6                | 7.24  | 99   | 548290   | 100.79 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 9.27  | 82   | 332587   | 66.98  | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 16.22 | 330  | 181877   | 91.66  | ng    | -0.01    |
| 45) 2-Fluorobiphenyl        | 13.35 | 172  | 802586   | 66.53  | ng    | -0.02    |
| 79) Terphenyl-d14           | 20.08 | 244  | 1095482  | 65.95  | ng    | -0.01    |
| Target Compounds            |       |      |          |        |       |          |
| 10) Phenol                  | 7.27  | 94   | 23889    | 4.162  | ng    | 97       |
| 50) Dimethylphthalate       | 14.17 | 163  | 101046   | 7.180  | ng    | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG110518\  
Data File : BG037822.D  
Acq On : 6 Nov 2018 3:32  
Operator : JU/SJ  
Sample : J5812-13  
Misc :  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
TP2-B

Quant Time: Nov 06 06:02:20 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Oct 18 14:06:42 2018  
Response via : Initial Calibration

