

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110518\
 Data File : BG037797.D
 Acq On : 5 Nov 2018 11:17
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 13:48:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 05 13:45:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	-0.01
2	1,4-Dioxane	0.607	0.573	5.6	89	0.00
3	Pyridine	1.529	1.463	4.3	89	0.00
4	n-Nitrosodimethylamine	0.545	0.491	9.9	85	0.00
5 S	2-Fluorophenol	1.196	1.165	2.6	90	-0.01
6	Aniline	2.207	2.169	1.7	91	-0.01
7 S	Phenol-d6	1.809	1.763	2.5	89	-0.01
8	2-Chlorophenol	1.399	1.381	1.3	91	-0.01
9	Benzaldehyde	1.132	0.973	14.0	80	-0.01
10 C	Phenol	1.909	1.857	2.7	90	0.00
11	bis(2-Chloroethyl)ether	1.450	1.398	3.6	90	-0.01
12	1,3-Dichlorobenzene	1.558	1.603	-2.9	96	-0.01
13 C	1,4-Dichlorobenzene	1.594	1.594	0.0	94	-0.01
14	1,2-Dichlorobenzene	1.533	1.543	-0.7	94	-0.01
15	Benzyl Alcohol	1.337	1.265	5.4	85	-0.01
16	2,2'-oxybis(1-Chloropropane	2.594	2.485	4.2	89	-0.01
17	2-Methylphenol	1.262	1.240	1.7	89	-0.01
18	Hexachloroethane	0.543	0.562	-3.5	96	-0.02
19 P	n-Nitroso-di-n-propylamine	1.246	1.181	5.2	85	-0.01
20	3+4-Methylphenols	1.804	1.754	2.8	89	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	94	-0.01
22	Acetophenone	0.502	0.476	5.2	90	-0.01
23 S	Nitrobenzene-d5	0.357	0.346	3.1	91	-0.01
24	Nitrobenzene	0.371	0.361	2.7	91	-0.01
25	Isophorone	0.652	0.617	5.4	87	-0.02
26 C	2-Nitrophenol	0.167	0.175	-4.8	95	-0.02
27	2,4-Dimethylphenol	0.298	0.282	5.4	88	-0.01
28	bis(2-Chloroethoxy)methane	0.427	0.402	5.9	87	-0.02
29 C	2,4-Dichlorophenol	0.332	0.329	0.9	92	-0.01
30	1,2,4-Trichlorobenzene	0.361	0.362	-0.3	96	-0.02
31	Naphthalene	0.982	0.968	1.4	94	-0.01
32	Benzoic acid	0.194	0.160	17.5	76	-0.02
33	4-Chloroaniline	0.462	0.449	2.8	90	-0.02
34 C	Hexachlorobutadiene	0.214	0.221	-3.3	96	-0.02
35	Caprolactam	0.133	0.125	6.0	85	-0.01
36 C	4-Chloro-3-methylphenol	0.375	0.356	5.1	88	-0.01
37	2-Methylnaphthalene	0.716	0.712	0.6	93	-0.02
38	1-Methylnaphthalene	0.695	0.689	0.9	94	-0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	92	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.645	0.685	-6.2	98	-0.01
41 P	Hexachlorocyclopentadiene	0.308	0.308	0.0	89	-0.01
42 S	2,4,6-Tribromophenol	0.218	0.217	0.5	87	-0.01
43 C	2,4,6-Trichlorophenol	0.431	0.437	-1.4	90	-0.01
44	2,4,5-Trichlorophenol	0.469	0.465	0.9	88	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110518\
 Data File : BG037797.D
 Acq On : 5 Nov 2018 11:17
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 13:48:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 05 13:45:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.326	1.348	-1.7	95	-0.01
46	1,1'-Biphenyl	1.656	1.655	0.1	93	-0.01
47	2-Chloronaphthalene	1.253	1.275	-1.8	94	-0.01
48	2-Nitroaniline	0.380	0.375	1.3	89	-0.01
49	Acenaphthylene	1.885	1.905	-1.1	92	-0.01
50	Dimethylphthalate	1.547	1.496	3.3	88	-0.01
51	2,6-Dinitrotoluene	0.342	0.345	-0.9	90	0.00
52 C	Acenaphthene	1.161	1.149	1.0	92	-0.01
53	3-Nitroaniline	0.380	0.374	1.6	87	-0.01
54 P	2,4-Dinitrophenol	0.100	0.077	23.0	71	-0.01
55	Dibenzofuran	1.923	1.894	1.5	92	-0.01
56 P	4-Nitrophenol	0.281	0.239	14.9	75	0.00
57	2,4-Dinitrotoluene	0.449	0.461	-2.7	89	0.00
58	Fluorene	1.440	1.403	2.6	91	-0.01
59	2,3,4,6-Tetrachlorophenol	0.402	0.386	4.0	87	-0.01
60	Diethylphthalate	1.588	1.544	2.8	89	-0.01
61	4-Chlorophenyl-phenylether	0.840	0.836	0.5	92	-0.01
62	4-Nitroaniline	0.378	0.367	2.9	85	-0.01
63	Azobenzene	1.516	1.432	5.5	88	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	87	-0.01
65	4,6-Dinitro-2-methylphenol	0.093	0.091	2.2	84	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.617	-2.5	89	-0.01
67	4-Bromophenyl-phenylether	0.220	0.231	-5.0	91	-0.01
68	Hexachlorobenzene	0.228	0.231	-1.3	88	-0.01
69	Atrazine	0.218	0.207	5.0	80	-0.01
70 C	Pentachlorophenol	0.152	0.134	11.8	74	-0.01
71	Phenanthrene	1.016	1.016	0.0	88	-0.01
72	Anthracene	0.998	1.004	-0.6	88	-0.01
73	Carbazole	0.998	0.996	0.2	86	0.00
74	Di-n-butylphthalate	1.110	1.115	-0.5	85	-0.02
75 C	Fluoranthene	1.153	1.145	0.7	86	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	89	-0.01
77	Benzidine	0.680	0.487	28.4#	61	-0.01
78	Pyrene	1.307	1.298	0.7	88	0.00
79 S	Terphenyl-d14	0.936	0.930	0.6	88	-0.01
80	Butylbenzylphthalate	0.535	0.545	-1.9	86	-0.01
81	Benzo(a)anthracene	1.183	1.180	0.3	88	-0.01
82	3,3'-Dichlorobenzidine	0.459	0.450	2.0	84	-0.01
83	Chrysene	1.138	1.147	-0.8	89	-0.01
84	Bis(2-ethylhexyl)phthalate	0.750	0.779	-3.9	88	-0.02
85 c	Di-n-octyl phthalate	1.223	1.278	-4.5	88	-0.04
86	Indeno(1,2,3-cd)pyrene	1.220	1.046	14.3	74	-0.05
87 I	Perylene-d12	1.000	1.000	0.0	85	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110518\
Data File : BG037797.D
Acq On : 5 Nov 2018 11:17
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 05 13:48:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 05 13:45:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.096	1.086	0.9	83	-0.02
89	Benzo(k)fluoranthene	1.074	1.172	-9.1	92	-0.02
90 C	Benzo(a)pyrene	1.042	1.061	-1.8	85	-0.02
91	Dibenzo(a,h)anthracene	0.961	0.819	14.8	70	-0.05
92	Benzo(q,h,i)perylene	0.972	0.826	15.0	71	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0