


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110518\
 Data File : BG037824.D
 Acq On : 6 Nov 2018 4:50
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG101818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 18 14:06:42 2018



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	21802	PASS
68	69	0.00	2	0.5	101	PASS
69	198	0.00	100	32.6	18858	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.4	25136	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	57920	PASS
199	198	5	9	7.0	4067	PASS
275	198	10	60	26.1	15118	PASS
365	198	1	100	3.1	1777	PASS
441	198	0.01	100	8.7	5043	PASS
442	198	50	100	55.5	32136	PASS
443	442	15	24	18.4	5911	PASS