

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110821\
 Data File : BG050944.D
 Acq On : 10 Nov 2021 15:34
 Operator : CG/JU
 Sample : PB140576BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS576

Quant Time: Nov 10 16:18:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 14:49:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.231	152	36676	20.000	ng/u1	0.00
20) Naphthalene-d8	11.052	136	161443	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.853	164	105860	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.597	188	235403	20.000	ng/u1	0.00
79) Chrysene-d12	21.892	240	202027	20.000	ng/u1	-0.01
88) Perylene-d12	25.288	264	204008	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.590	96	6101	5.369	ng/uL	0.00
4) Pyridine-d5	4.013	84	89089	26.207	ng/u1	0.00
7) Phenol-d5	7.368	99	102712	26.251	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.544	67	68219	26.991	ng/u1	0.00
11) 2-Chlorophenol-d4	7.755	132	72993	26.919	ng/u1	0.00
15) 4-Methylphenol-d8	8.925	113	79503	25.810	ng/u1	0.00
21) Nitrobenzene-d5	9.401	128	38273	27.897	ng/u1	0.00
24) 2-Nitrophenol-d4	10.123	143	42321	27.742	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.664	165	72482	28.206	ng/u1	0.00
31) 4-Chloroaniline-d4	11.187	131	90724	23.313	ng/u1	0.00
46) Dimethylphthalate-d6	14.242	166	234586	28.965	ng/u1	0.00
49) Acenaphthylene-d8	14.547	160	289473	28.688	ng/u1	0.00
54) 4-Nitrophenol-d4	15.035	143	40355	27.481	ng/u1	0.00
60) Fluorene-d10	15.840	176	204301	28.476	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.958	200	38420	26.920	ng/u1	0.00
73) Anthracene-d10	17.697	188	321013	28.843	ng/u1	0.00
81) Pyrene-d10	19.970	212	373676	28.637	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.059	264	316758	28.087	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.625	88	13756	11.021	ng/uL	98
5) Pyridine	4.030	79	93585	26.595	ng/u1	95
6) Benzaldehyde	7.362	77	71034	28.784	ng/u1	99
8) Phenol	7.397	94	110281	27.247	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.638	93	83273	27.487	ng/u1	98
12) 2-Chlorophenol	7.791	128	76050	27.623	ng/u1	96
13) 2-Methylphenol	8.660	108	82038	27.422	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.748	45	129542	27.147	ng/u1	100
16) Acetophenone	9.054	105	129160	26.992	ng/u1	97
17) N-Nitroso-di-n-propyla...	9.030	70	75961	26.311	ng/u1	98
18) 4-Methylphenol	8.989	108	86751	27.234	ng/u1	94
19) Hexachloroethane	9.318	117	32389	28.137	ng/u1	93
22) Nitrobenzene	9.442	77	109124	28.519	ng/u1	98
23) Isophorone	9.965	82	209855	28.259	ng/u1	100
25) 2-Nitrophenol	10.158	139	45002	29.404	ng/u1	97
26) 2,4-Dimethylphenol	10.200	107	97328	28.896	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.440	93	116194	29.039	ng/u1	96
29) 2,4-Dichlorophenol	10.693	162	72953	29.105	ng/u1	96
30) Naphthalene	11.104	128	253891	28.759	ng/u1	98
32) 4-Chloroaniline	11.210	127	93353	24.159	ng/u1	98
33) Hexachlorobutadiene	11.375	225	46974	28.553	ng/u1	98
34) Caprolactam	11.968	113	28677	26.949	ng/u1	93
35) 4-Chloro-3-methylphenol	12.309	107	91945	28.715	ng/u1	98
36) 2-Methylnaphthalene	12.691	142	170757	28.392	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110821\
 Data File : BG050944.D
 Acq On : 10 Nov 2021 15:34
 Operator : CG/JU
 Sample : PB140576BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SLCS576

Quant Time: Nov 10 16:18:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 14:49:05 2021
 Response via : Initial Calibration




Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.908	142	173819	28.523	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.055	216	92654	30.038	ng/ul	98
40) Hexachlorocyclopentadiene	13.026	237	37839	25.546	ng/ul	96
41) 2,4,6-Trichlorophenol	13.290	196	60643	30.046	ng/ul	95
42) 2,4,5-Trichlorophenol	13.366	196	65151	30.066	ng/ul	98
43) 1,1'-Biphenyl	13.684	154	231267	29.888	ng/ul	99
44) 2-Chloronaphthalene	13.737	162	180700	29.801	ng/ul	99
45) 2-Nitroaniline	13.936	65	70879	29.422	ng/ul	95
47) Dimethylphthalate	14.289	163	242645	29.963	ng/ul	100
48) 2,6-Dinitrotoluene	14.424	165	49719	29.334	ng/ul	98
50) Acenaphthylene	14.577	152	298207	29.488	ng/ul	98
51) 3-Nitroaniline	14.753	138	48429	27.627	ng/ul	88
52) Acenaphthene	14.918	153	195243	29.361	ng/ul	96
53) 2,4-Dinitrophenol	14.965	184	21170	22.641	ng/ul	90
55) 4-Nitrophenol	15.053	109	37893	28.136	ng/ul	93
56) Dibenzofuran	15.247	168	281688	29.595	ng/ul	98
57) 2,4-Dinitrotoluene	15.205	165	72747	30.075	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.470	232	50251	29.509	ng/ul	95
59) Diethylphthalate	15.646	149	256335	29.572	ng/ul	100
61) Fluorene	15.893	166	221154	29.355	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.881	204	116330	29.654	ng/ul	97
63) 4-Nitroaniline	15.916	138	51892	29.839	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.969	198	39296	28.234	ng/ul#	96
67) N-Nitrosodiphenylamine	16.093	169	198067	30.101	ng/ul	98
68) 4-Bromophenyl-phenylether	16.774	248	69951	29.873	ng/ul	96
69) Hexachlorobenzene	16.898	284	72769	30.229	ng/ul	95
70) Atrazine	17.033	200	80148	28.723	ng/ul	98
71) Pentachlorophenol	17.244	266	30863	27.925	ng/ul	95
72) Phenanthrene	17.638	178	377963	30.079	ng/ul	99
74) Anthracene	17.732	178	376698	29.878	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.654	216	94965	29.628	ng/uL	98
76) Pentachlorobenzene	15.164	250	86001	28.958	ng/uL	99
77) Carbazole	17.996	167	353009	31.238	ng/ul	99
78) Di-n-butylphthalate	18.531	149	449102	30.245	ng/ul	100
80) Fluoranthene	19.636	202	461591	29.476	ng/ul	99
82) Pyrene	20.000	202	450396	29.435	ng/ul	100
83) Butylbenzylphthalate	20.864	149	195205	29.667	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.774	252	125962	25.602	ng/ul	98
85) Benzo(a)anthracene	21.868	228	417142	29.825	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.739	149	280040	29.650	ng/ul	98
87) Chrysene	21.939	228	401989	30.088	ng/ul	98
89) Di-n-octyl phthalate	23.008	149	478563	28.814	ng/ul	100
90) Benzo(b)fluoranthene	24.201	252	422188	29.049	ng/ul	99
91) Benzo(k)fluoranthene	24.271	252	395601	29.008	ng/ul	98
93) Benzo(a)pyrene	25.129	252	396139	28.619	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.189	276	446091	28.951	ng/ul	99
95) Dibenzo(a,h)anthracene	29.260	278	377628	28.964	ng/ul	98
96) Benzo(g,h,i)perylene	30.417	276	375098	29.081	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 10 16:18:24 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 02 14:49:05 2021
Response via : Initial Calibration







Abundance Scan 23 (3.625 min): BG050942.D\data.ms (-17)



#2
1,4-Dioxane
Concen: 11.021 ng/uL
RT: 3.625 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34


Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:	88	Resp:	13756
Ion	Ratio	Lower	Upper
88	100		
43	31.4	23.0	34.4
58	78.8	62.4	93.6



#5
Pyridine
Concen: 26.595 ng/uL
RT: 4.030 min Scan# 92
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:	79	Resp:	93585
Ion	Ratio	Lower	Upper
79	100		
52	80.4	60.8	91.2
51	35.2	30.5	45.7









#6
Benzaldehyde
Concen: 28.784 ng/ul
RT: 7.362 min Scan# 61
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion: 77 Resp: 71034
Ion Ratio Lower Upper
77 100
105 88.3 70.4 105.6
106 78.5 61.2 91.8



#8
Phenol
Concen: 27.247 ng/ul
RT: 7.397 min Scan# 665
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion: 94 Resp: 110281
Ion Ratio Lower Upper
94 100
65 32.4 25.1 37.7
66 40.9 31.6 47.4









#10
Bis(2-Chloroethyl)ether
Concen: 27.487 ng/ul
RT: 7.638 min Scan# 706
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Instrument :
BNA_G
ClientSampleId :
SLCS576

Tgt Ion: 93 Resp: 83273




Ion	Ratio	Lower	Upper
93	100		
63	83.6	68.6	103.0
95	32.7	26.1	39.1




#12
2-Chlorophenol
Concen: 27.623 ng/ul
RT: 7.791 min Scan# 732
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion: 128 Resp: 76050

Ion	Ratio	Lower	Upper
128	100		
64	58.7	45.3	67.9
130	31.1	27.1	40.7









#13
2-Methylphenol
Concen: 27.422 ng/ul
RT: 8.660 min Scan# 880
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:108 Resp: 82038
Ion Ratio Lower Upper
108 100
107 88.5 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 27.147 ng/ul
RT: 8.748 min Scan# 895
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion: 45 Resp: 129542
Ion Ratio Lower Upper
45 100
77 13.4 10.7 16.1
79 10.8 8.5 12.7









#16
Acetophenone
Concen: 26.992 ng/ul
RT: 9.054 min Scan# 947
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:105 Resp: 129160
Ion Ratio Lower Upper
105 100
77 87.7 67.3 100.9
51 30.0 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 26.311 ng/ul
RT: 9.030 min Scan# 943
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion: 70 Resp: 75961
Ion Ratio Lower Upper
70 100
42 55.6 46.0 69.0
101 8.8 7.0 10.4
130 16.8 13.4 20.2





#18

4-Methylphenol

Concen: 27.234 ng/ul

RT: 8.989 min Scan# 91

Delta R.T. -0.009 min

Lab File: BG050944.D

Acq: 10 Nov 2021 15:34

Instrument :

BNA_G

ClientSampleId :





SLCS576

Tgt Ion:108 Resp: 86751

Ion Ratio Lower Upper

108 100

107 112.0 95.0 142.4



#19

Hexachloroethane

Concen: 28.137 ng/ul

RT: 9.318 min Scan# 992

Delta R.T. -0.009 min

Lab File: BG050944.D

Acq: 10 Nov 2021 15:34




Tgt Ion:117 Resp: 32389


Ion Ratio Lower Upper

117 100

201 91.3 77.2 115.8

199 55.1 49.9 74.9









#22
Nitrobenzene
Concen: 28.519 ng/ul
RT: 9.442 min Scan# 1014
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion: 77 Resp: 109124
Ion Ratio Lower Upper
77 100
123 36.1 30.1 45.1
65 12.8 9.6 14.4



#23
Isophorone
Concen: 28.259 ng/ul
RT: 9.965 min Scan# 1102
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion: 82 Resp: 209855
Ion Ratio Lower Upper
82 100
95 6.8 5.6 8.4
138 14.9 11.8 17.8









#25
2-Nitrophenol
Concen: 29.404 ng/ul
RT: 10.158 min Scan# 1135
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:139 Resp: 45002
Ion Ratio Lower Upper
139 100
65 61.8 48.0 72.0
109 32.8 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 28.896 ng/ul
RT: 10.200 min Scan# 1142
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:107 Resp: 97328
Ion Ratio Lower Upper
107 100
121 47.2 39.3 58.9
122 76.1 63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 29.039 ng/ul
 RT: 10.440 min Scan# 1183
 Delta R.T. -0.009 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34

Instrument :
 BNA_G
 ClientSampleId :
 SLCS576

Tgt Ion: 93 Resp: 116194




Ion	Ratio	Lower	Upper
93	100		
95	30.6	26.5	39.7
123	9.9	8.1	12.1




#29
 2,4-Dichlorophenol
 Concen: 29.105 ng/ul
 RT: 10.693 min Scan# 1226
 Delta R.T. -0.009 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34

Tgt Ion: 162 Resp: 72953

Ion	Ratio	Lower	Upper
162	100		
164	63.8	49.4	74.0
98	41.1	30.4	45.6









#30
Naphthalene
Concen: 28.759 ng/ul
RT: 11.104 min Scan# 11
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:128 Resp: 253891
Ion Ratio Lower Upper
128 100
129 10.7 8.5 12.7
127 13.6 11.8 17.6



#32
4-Chloroaniline
Concen: 24.159 ng/ul
RT: 11.210 min Scan# 1314
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:127 Resp: 93353
Ion Ratio Lower Upper
127 100
129 31.4 26.2 39.2









#33
Hexachlorobutadiene
Concen: 28.553 ng/ul
RT: 11.375 min Scan# 11
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:225 Resp: 46974
Ion Ratio Lower Upper
225 100
223 64.1 50.6 75.8
227 65.1 49.8 74.8



#34
Caprolactam
Concen: 26.949 ng/ul
RT: 11.968 min Scan# 1443
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:113 Resp: 28677
Ion Ratio Lower Upper
113 100
55 199.8 147.0 220.6
56 137.0 109.2 163.8









#35
 4-Chloro-3-methylphenol
 Concen: 28.715 ng/ul
 RT: 12.309 min Scan# 1501
 Delta R.T. -0.009 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34




Instrument :
 BNA_G
 ClientSampleId :
 SLCS576


Tgt Ion:	107	Resp:	91945
Ion	Ratio	Lower	Upper
107	100		
144	22.9	19.8	29.6
142	71.0	56.2	84.4



#36
 2-Methylnaphthalene
 Concen: 28.392 ng/ul
 RT: 12.691 min Scan# 1566
 Delta R.T. -0.009 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34

Tgt Ion:	142	Resp:	170757
Ion	Ratio	Lower	Upper
142	100		
141	91.2	73.3	109.9









#37
1-Methylnaphthalene
Concen: 28.523 ng/ul
RT: 12.908 min Scan# 1603
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:142 Resp: 173819
Ion Ratio Lower Upper
142 100
141 95.5 77.8 116.8
116 4.2 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.038 ng/ul
RT: 13.055 min Scan# 1628
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:216 Resp: 92654
Ion Ratio Lower Upper
216 100
214 74.6 61.0 91.4
179 20.2 16.2 24.2
108 17.4 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 25.546 ng/ul
RT: 13.026 min Scan# 1623
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:237 Resp: 37839
Ion Ratio Lower Upper
237 100
235 63.9 49.4 74.0
272 11.2 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 30.046 ng/ul
RT: 13.290 min Scan# 1668
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:196 Resp: 60643
Ion Ratio Lower Upper
196 100
198 99.0 74.3 111.5
200 30.9 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 30.066 ng/ul
RT: 13.366 min Scan# 1
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:196 Resp: 65151
Ion Ratio Lower Upper
196 100
198 94.2 77.0 115.4
200 30.4 25.4 38.2



#43
1,1'-Biphenyl
Concen: 29.888 ng/ul
RT: 13.684 min Scan# 1735
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:154 Resp: 231267
Ion Ratio Lower Upper
154 100
153 43.3 34.5 51.7
76 15.6 13.1 19.7









#44
2-Chloronaphthalene
Concen: 29.801 ng/ul
RT: 13.737 min Scan# 1744
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:162 Resp: 180700
Ion Ratio Lower Upper
162 100
164 31.9 26.1 39.1
127 38.4 31.0 46.4



#45
2-Nitroaniline
Concen: 29.422 ng/ul
RT: 13.936 min Scan# 1778
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion: 65 Resp: 70879
Ion Ratio Lower Upper
65 100
92 67.0 53.0 79.4
138 92.4 67.4 101.0









#47
Dimethylphthalate
Concen: 29.963 ng/ul
RT: 14.289 min Scan# 1838
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:163 Resp: 242645
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.4
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 29.334 ng/ul
RT: 14.424 min Scan# 1861
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:165 Resp: 49719
Ion Ratio Lower Upper
165 100
89 72.5 57.4 86.0
121 25.9 18.8 28.2









#50
Acenaphthylene
Concen: 29.488 ng/ul
RT: 14.577 min Scan# 1887
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:152 Resp: 298207
Ion Ratio Lower Upper
152 100
151 20.5 17.4 26.0
153 13.5 10.3 15.5



#51
3-Nitroaniline
Concen: 27.627 ng/ul
RT: 14.753 min Scan# 1917
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:138 Resp: 48429
Ion Ratio Lower Upper
138 100
108 12.1 8.8 13.2
92 136.3 98.0 147.0









#52
Acenaphthene
Concen: 29.361 ng/ul
RT: 14.918 min Scan# 1945
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:153 Resp: 195243
Ion Ratio Lower Upper
153 100
152 48.4 36.5 54.7
154 85.8 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 22.641 ng/ul
RT: 14.965 min Scan# 1953
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:184 Resp: 21170
Ion Ratio Lower Upper
184 100
63 71.1 66.2 99.2
154 62.8 53.6 80.4









#55
4-Nitrophenol
Concen: 28.136 ng/ul
RT: 15.053 min Scan# 1968
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:109 Resp: 37893
Ion Ratio Lower Upper
109 100
139 113.7 88.7 133.1
65 128.9 113.6 170.4



#56
Dibenzofuran
Concen: 29.595 ng/ul
RT: 15.247 min Scan# 2001
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:168 Resp: 281688
Ion Ratio Lower Upper
168 100
139 40.2 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 30.075 ng/ul
RT: 15.205 min Scan# 1994
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:165 Resp: 72747
Ion Ratio Lower Upper
165 100
63 52.4 42.1 63.1
182 2.5 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 29.509 ng/ul
RT: 15.470 min Scan# 2039
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:232 Resp: 50251
Ion Ratio Lower Upper
232 100
131 54.2 38.9 58.3
130 3.2 2.4 3.6
166 33.2 25.8 38.8









#59
Diethylphthalate
Concen: 29.572 ng/ul
RT: 15.646 min Scan# 2069
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:149 Resp: 256335
Ion Ratio Lower Upper
149 100
177 21.1 16.8 25.2
150 12.7 10.1 15.1



#61
Fluorene
Concen: 29.355 ng/ul
RT: 15.893 min Scan# 2111
Delta R.T. -0.015 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:166 Resp: 221154
Ion Ratio Lower Upper
166 100
165 102.5 81.0 121.6
167 14.0 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 29.654 ng/ul
RT: 15.881 min Scan# 2109
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:204 Resp: 116330
Ion Ratio Lower Upper
204 100
206 32.7 27.4 41.2
141 58.3 48.6 73.0



#63
4-Nitroaniline
Concen: 29.839 ng/ul
RT: 15.916 min Scan# 2115
Delta R.T. -0.003 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:138 Resp: 51892
Ion Ratio Lower Upper
138 100
92 60.4 49.3 73.9
108 98.0 72.6 108.8









#66
4,6-Dinitro-2-methylphenol
Concen: 28.234 ng/ul
RT: 15.969 min Scan# 2124
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:198 Resp: 39296
Ion Ratio Lower Upper
198 100
182 3.8 3.8 5.8#
77 31.5 23.2 34.8



#67
N-Nitrosodiphenylamine
Concen: 30.101 ng/ul
RT: 16.093 min Scan# 2145
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:169 Resp: 198067
Ion Ratio Lower Upper
169 100
168 71.3 58.5 87.7
167 39.5 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 29.873 ng/ul
RT: 16.774 min Scan# 211
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:248 Resp: 69951
Ion Ratio Lower Upper
248 100
250 97.8 75.5 113.3
141 80.4 67.6 101.4



#69
Hexachlorobenzene
Concen: 30.229 ng/ul
RT: 16.898 min Scan# 2282
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:284 Resp: 72769
Ion Ratio Lower Upper
284 100
142 42.8 31.9 47.9
249 34.8 25.8 38.6









#70
Atrazine
Concen: 28.723 ng/ul
RT: 17.033 min Scan# 21
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:200 Resp: 80148
Ion Ratio Lower Upper
200 100
173 26.1 20.2 30.2
215 49.3 38.1 57.1



#71
Pentachlorophenol
Concen: 27.925 ng/ul
RT: 17.244 min Scan# 2341
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:266 Resp: 30863
Ion Ratio Lower Upper
266 100
264 70.7 54.3 81.5
268 68.4 51.0 76.6









#72
Phenanthrene
Concen: 30.079 ng/u1
RT: 17.638 min Scan# 2408
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:178 Resp: 377963
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 19.7 15.5 23.3



#74
Anthracene
Concen: 29.878 ng/u1
RT: 17.732 min Scan# 2424
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:178 Resp: 376698
Ion Ratio Lower Upper
178 100
179 15.4 13.0 19.6
176 19.0 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.628 ng/uL
RT: 13.654 min Scan# 11
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:216 Resp: 94965
Ion Ratio Lower Upper
216 100
214 78.6 61.3 91.9
218 45.2 37.0 55.6



#76
Pentachlorobenzene
Concen: 28.958 ng/uL
RT: 15.164 min Scan# 1987
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:250 Resp: 86001
Ion Ratio Lower Upper
250 100
248 63.5 50.5 75.7
252 62.9 51.2 76.8









#77
 Carbazole
 Concen: 31.238 ng/ul
 RT: 17.996 min Scan# 2470
 Delta R.T. -0.009 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34




Instrument :
 BNA_G
 ClientSampleId :
 SLCS576


Tgt Ion:167 Resp: 353009
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.3 25.9
 139 14.2 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 30.245 ng/ul
 RT: 18.531 min Scan# 2560
 Delta R.T. -0.009 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34

Tgt Ion:149 Resp: 449102
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 6.3 5.0 7.6









#80
Fluoranthene
Concen: 29.476 ng/ul
RT: 19.636 min Scan# 21
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:202 Resp: 461591
Ion Ratio Lower Upper
202 100
101 11.6 9.3 13.9
100 9.3 7.0 10.6



#82
Pyrene
Concen: 29.435 ng/ul
RT: 20.000 min Scan# 2810
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:202 Resp: 450396
Ion Ratio Lower Upper
202 100
101 13.2 10.7 16.1
100 11.3 9.0 13.6









#83
Butylbenzylphthalate
Concen: 29.667 ng/ul
RT: 20.864 min Scan# 2957
Delta R.T. -0.009 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:	149	Resp:	195205
Ion	Ratio	Lower	Upper
149	100		
91	80.4	65.2	97.8
206	20.1	16.2	24.2



#84
3,3'-Dichlorobenzidine
Concen: 25.602 ng/ul
RT: 21.774 min Scan# 3112
Delta R.T. -0.015 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:	252	Resp:	125962
Ion	Ratio	Lower	Upper
252	100		
254	62.4	50.8	76.2
126	12.7	9.0	13.4









#85
Benzo(a)anthracene
Concen: 29.825 ng/ul
RT: 21.868 min Scan# 3128
Delta R.T. -0.015 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:228 Resp: 417142
Ion Ratio Lower Upper
228 100
229 19.9 15.7 23.5
226 28.1 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 29.650 ng/ul
RT: 21.739 min Scan# 3106
Delta R.T. -0.015 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:149 Resp: 280040
Ion Ratio Lower Upper
149 100
167 25.6 21.1 31.7
279 3.0 2.7 4.1









#87
Chrysene
Concen: 30.088 ng/ul
RT: 21.939 min Scan# 3141
Delta R.T. -0.015 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:228 Resp: 401989
Ion Ratio Lower Upper
228 100
226 30.0 24.8 37.2
229 20.4 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 28.814 ng/ul
RT: 23.008 min Scan# 3322
Delta R.T. -0.015 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:149 Resp: 478563









#90
Benzo(b)fluoranthene
Concen: 29.049 ng/ul
RT: 24.201 min Scan# 31
Delta R.T. -0.021 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:252 Resp: 422188
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 10.2 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 29.008 ng/ul
RT: 24.271 min Scan# 3537
Delta R.T. -0.021 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:252 Resp: 395601
Ion Ratio Lower Upper
252 100
253 22.2 17.3 25.9
125 10.5 7.8 11.6









#93
Benzo(a)pyrene
Concen: 28.619 ng/ul
RT: 25.129 min Scan# 30
Delta R.T. -0.021 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34




Instrument :
BNA_G
ClientSampleId :
SLCS576


Tgt Ion:252 Resp: 396139
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 10.9 9.1 13.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 28.951 ng/ul
RT: 29.189 min Scan# 4374
Delta R.T. -0.032 min
Lab File: BG050944.D
Acq: 10 Nov 2021 15:34

Tgt Ion:276 Resp: 446091
Ion Ratio Lower Upper
276 100
138 19.7 15.5 23.3
277 25.1 20.5 30.7









#95
 Dibenzo(a,h)anthracene
 Concen: 28.964 ng/ul
 RT: 29.260 min Scan# 41
 Delta R.T. -0.021 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34

Instrument :
 BNA_G
 ClientSampleId :
 SLCS576

Tgt Ion:278 Resp: 377628
 Ion Ratio Lower Upper
 278 100
 139 16.6 12.2 18.4
 279 24.9 19.5 29.3



#96
 Benzo(g,h,i)perylene
 Concen: 29.081 ng/ul
 RT: 30.417 min Scan# 4583
 Delta R.T. -0.027 min
 Lab File: BG050944.D
 Acq: 10 Nov 2021 15:34

Tgt Ion:276 Resp: 375098
 Ion Ratio Lower Upper
 276 100
 138 20.7 16.6 24.8
 277 23.1 17.6 26.4

