

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110821\
 Data File : BG050961.D
 Acq On : 11 Nov 2021 3:54
 Operator : CG/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020547

Quant Time: Nov 11 04:26:49 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 14:49:05 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/11/2021
 Supervised By :mohammad ahmed 11/11/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.226	152	50649	20.000	ng/u1	0.00
20) Naphthalene-d8	11.052	136	239241	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.848	164	160057	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.591	188	355032	20.000	ng/u1	-0.01
79) Chrysene-d12	21.892	240	303063	20.000	ng/u1	-0.01
88) Perylene-d12	25.288	264	304376	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.584	96	11101	7.074	ng/uL	0.00
4) Pyridine-d5	4.007	84	83369	17.758	ng/u1	0.00
7) Phenol-d5	7.368	99	102918	19.047	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.539	67	66203	18.967	ng/u1	0.00
11) 2-Chlorophenol-d4	7.750	132	71958	19.216	ng/u1	-0.01
15) 4-Methylphenol-d8	8.925	113	79586	18.709	ng/u1	0.00
21) Nitrobenzene-d5	9.401	128	38568	18.970	ng/u1	0.00
24) 2-Nitrophenol-d4	10.124	143	42507	18.803	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.664	165	73289	19.246	ng/u1	0.00
31) 4-Chloroaniline-d4	11.181	131	107023	18.559	ng/u1	0.00
46) Dimethylphthalate-d6	14.242	166	233843	19.096	ng/u1	0.00
49) Acenaphthylene-d8	14.548	160	302345	19.818	ng/u1	0.00
54) 4-Nitrophenol-d4	15.036	143	39925	17.982	ng/u1	0.00
60) Fluorene-d10	15.835	176	210114	19.369	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.952	200	38040	17.673	ng/u1	0.00
73) Anthracene-d10	17.691	188	327053	19.484	ng/u1	-0.01
81) Pyrene-d10	19.971	212	375163	19.166	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.053	264	314603	18.697	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.625	88	12056	6.995	ng/uL	90
5) Pyridine	4.031	79	89605	18.439	ng/u1	99
6) Benzaldehyde	7.356	77	79628	23.365	ng/u1	96
8) Phenol	7.397	94	106709	19.091	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.633	93	79673	19.043	ng/u1	99
12) 2-Chlorophenol	7.785	128	73228	19.260	ng/u1	97
13) 2-Methylphenol	8.661	108	77332	18.718	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.749	45	120087	18.223	ng/u1	98
16) Acetophenone	9.054	105	125953	19.060	ng/u1	99
17) N-Nitroso-di-n-propyla...	9.031	70	75831	19.020	ng/u1	97
18) 4-Methylphenol	8.990	108	84486	19.206	ng/u1	96
19) Hexachloroethane	9.319	117	30925	19.454	ng/u1	95
22) Nitrobenzene	9.442	77	106520	18.786	ng/u1	96
23) Isophorone	9.959	82	203101	18.456	ng/u1	100
25) 2-Nitrophenol	10.159	139	43685	19.262	ng/u1	97
26) 2,4-Dimethylphenol	10.200	107	95733	19.180	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.441	93	110452	18.628	ng/u1	99
29) 2,4-Dichlorophenol	10.694	162	72449	19.505	ng/u1	97
30) Naphthalene	11.105	128	251436	19.220	ng/u1	99
32) 4-Chloroaniline	11.205	127	107185	18.718	ng/u1	96
33) Hexachlorobutadiene	11.369	225	48222	19.780	ng/u1	97
34) Caprolactam	11.963	113	28168m	17.863	ng/u1	
35) 4-Chloro-3-methylphenol	12.309	107	89049	18.767	ng/u1	97
36) 2-Methylnaphthalene	12.691	142	168677	18.926	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110821\
 Data File : BG050961.D
 Acq On : 11 Nov 2021 3:54
 Operator : CG/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020547

Quant Time: Nov 11 04:26:49 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 14:49:05 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/11/2021
 Supervised By :mohammad ahmed 11/11/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.909	142	174740	19.350	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.050	216	94425	20.247	ng/ul	95
40) Hexachlorocyclopentadiene	13.026	237	37091	16.562	ng/ul	98
41) 2,4,6-Trichlorophenol	13.291	196	60345	19.774	ng/ul	96
42) 2,4,5-Trichlorophenol	13.361	196	62258	19.003	ng/ul	100
43) 1,1'-Biphenyl	13.684	154	227506	19.446	ng/ul	97
44) 2-Chloronaphthalene	13.731	162	182197	19.873	ng/ul	98
45) 2-Nitroaniline	13.931	65	68510	18.809	ng/ul	98
47) Dimethylphthalate	14.289	163	234148	19.123	ng/ul	99
48) 2,6-Dinitrotoluene	14.419	165	49445	19.294	ng/ul	98
50) Acenaphthylene	14.577	152	298053	19.493	ng/ul	99
51) 3-Nitroaniline	14.754	138	53821	20.306	ng/ul	90
52) Acenaphthene	14.912	153	197248	19.619	ng/ul	98
53) 2,4-Dinitrophenol	14.965	184	18524	13.103	ng/ul	89
55) 4-Nitrophenol	15.053	109	35839	17.600	ng/ul	96
56) Dibenzofuran	15.247	168	279504	19.422	ng/ul	98
57) 2,4-Dinitrotoluene	15.206	165	70262	19.212	ng/ul#	100
58) 2,3,4,6-Tetrachlorophenol	15.470	232	50411	19.579	ng/ul#	99
59) Diethylphthalate	15.647	149	249744	19.056	ng/ul	99
61) Fluorene	15.893	166	219242	19.248	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.876	204	117648	19.835	ng/ul	96
63) 4-Nitroaniline	15.911	138	54678	20.795	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.970	198	37557	17.892	ng/ul	99
67) N-Nitrosodiphenylamine	16.093	169	197218	19.873	ng/ul	98
68) 4-Bromophenyl-phenylether	16.775	248	71072	20.125	ng/ul	96
69) Hexachlorobenzene	16.892	284	74056	20.398	ng/ul	98
70) Atrazine	17.033	200	81700	19.414	ng/ul	97
71) Pentachlorophenol	17.239	266	28315	16.987	ng/ul	100
72) Phenanthrene	17.638	178	370082	19.528	ng/ul	100
74) Anthracene	17.727	178	372813	19.606	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.655	216	100006	20.688	ng/uL	99
76) Pentachlorobenzene	15.165	250	91221	20.366	ng/uL	99
77) Carbazole	17.997	167	340645	19.987	ng/ul	99
78) Di-n-butylphthalate	18.531	149	432297	19.303	ng/ul	99
80) Fluoranthene	19.636	202	453770	19.316	ng/ul	99
82) Pyrene	20.000	202	443268	19.311	ng/ul	98
83) Butylbenzylphthalate	20.864	149	188508	19.098	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.775	252	144873	19.629	ng/ul	98
85) Benzo(a)anthracene	21.869	228	401008	19.113	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.740	149	268313	18.937	ng/ul	99
87) Chrysene	21.939	228	382649	19.092	ng/ul	99
89) Di-n-octyl phthalate	23.009	149	457896	18.479	ng/ul	100
90) Benzo(b)fluoranthene	24.201	252	410629	18.937	ng/ul	99
91) Benzo(k)fluoranthene	24.272	252	371091	18.238	ng/ul	98
93) Benzo(a)pyrene	25.130	252	382275	18.510	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.190	276	430673	18.733	ng/ul	96
95) Dibenzo(a,h)anthracene	29.248	278	364575	18.742	ng/ul	97
96) Benzo(g,h,i)perylene	30.412	276	358044	18.606	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

