


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110821\
 Data File : BG050875.D
 Acq On : 8 Nov 2021 10:42
 Operator : CG/JU
 Sample : DFTPP429
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP429

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Title : SVOA CALIBRATION
 Last Update : Tue Nov 02 14:49:05 2021



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.1	32163	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	47.9	33461	PASS
70	69	0.00	2	0.4	126	PASS
127	198	10	80	54.5	38035	PASS
197	198	0.00	2	0.3	210	PASS
198	198	100	100	100.0	69840	PASS
199	198	5	9	7.0	4865	PASS
275	198	10	60	25.8	18030	PASS
365	198	1	100	3.4	2362	PASS
441	198	0.01	100	8.4	5897	PASS
442	198	50	100	51.9	36219	PASS
443	442	15	24	18.7	6781	PASS

Scan 2463 (17.958 min): BG050875.D\data.ms
DFTPP429

m/z	Abundance
40.00	576.0
44.00	959.0
207.10	167.0

Instrument :

BNA_G

ClientSampleId :

DFTPP429

m/z	Abundance
38.10	285.0
39.00	2422.0
40.00	499.0
44.00	842.0
50.00	8409.0
51.00	31472.0
52.10	2009.0
55.00	168.0
55.90	957.0
57.00	2122.0
58.00	177.0
61.00	381.0
62.00	557.0
63.00	1339.0
65.00	591.0
69.00	32936.0
70.00	203.0
73.00	378.0
74.00	3287.0
75.00	5526.0
76.00	2144.0
77.00	39968.0
78.00	2520.0
79.00	2117.0
80.00	2031.0
81.00	3104.0
82.00	819.0
83.00	616.0
84.90	655.0
85.90	562.0
87.00	264.0
91.00	542.0
92.10	617.0
93.00	4302.0
94.00	413.0
96.00	197.0
98.00	3168.0
99.00	2614.0
100.00	223.0
100.90	1691.0
103.00	562.0
104.00	918.0
105.00	973.0
107.00	11493.0
108.00	1736.0
109.00	307.0
110.00	17608.0
111.00	2992.0
112.00	314.0
116.00	752.0
117.00	9889.0
117.90	725.0
120.10	159.0
121.90	871.0
123.00	1092.0
123.90	591.0
125.00	404.0
127.00	36008.0
128.00	3019.0
129.00	13967.0
129.90	1193.0
131.00	345.0
133.90	613.0
134.90	957.0
135.90	343.0
137.00	612.0
139.90	197.0
140.90	2075.0
141.90	516.0
143.00	383.0
145.90	452.0
147.00	1136.0
148.00	2281.0
149.00	513.0
151.10	329.0
152.20	190.0
153.00	620.0
153.90	531.0
155.00	1077.0
156.00	1512.0
156.90	333.0
158.00	340.0
159.00	276.0
160.00	867.0
161.00	893.0
162.00	356.0
164.90	880.0
166.00	589.0
167.00	3441.0
168.00	2009.0
169.00	293.0
169.90	217.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

170.80	151.0
171.90	285.0
173.00	411.0
174.00	1053.0
175.00	1781.0
176.00	321.0
177.00	725.0
178.00	381.0
179.00	2776.0
180.00	2041.0
181.00	803.0
182.00	207.0
183.90	296.0
185.00	1495.0
186.00	9920.0
187.00	2841.0
187.90	280.0
189.00	786.0
191.00	320.0
192.00	863.0
193.00	1014.0
194.00	168.0
196.00	1884.0
198.00	61392.0
198.90	4491.0
199.90	382.0
201.40	351.0
203.00	700.0
204.00	2468.0
205.00	4377.0
206.00	18160.0
207.00	2451.0
208.00	635.0
209.00	191.0
210.10	232.0
211.10	647.0
215.00	300.0
216.00	412.0
217.00	5437.0
217.90	695.0
221.00	2179.0
221.90	737.0
223.00	977.0
224.00	9926.0
225.00	2721.0
226.00	299.0
227.00	4294.0
228.00	644.0
229.00	880.0
231.10	257.0
232.90	153.0
233.80	197.0
235.00	276.0
235.90	200.0
236.90	409.0
241.00	247.0
242.00	524.0
243.00	566.0
244.00	7209.0
245.00	955.0
246.00	1453.0
247.00	401.0
249.10	251.0
253.00	284.0
255.00	36416.0
256.00	5624.0
257.10	503.0
258.00	2171.0
259.00	377.0
265.10	992.0
266.00	170.0
273.00	1136.0
274.00	2826.0
275.00	15379.0
276.00	2083.0
277.00	1352.0
278.00	237.0
283.00	199.0
284.10	250.0
285.20	180.0
293.10	234.0
296.00	4644.0
297.00	700.0
303.00	501.0
304.00	184.0
314.00	188.0
315.00	538.0
316.00	193.0
323.10	1434.0
324.10	169.0
327.00	283.0
328.10	207.0
334.00	810.0
345.90	211.0
352.00	313.0
353.10	165.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

353.90	294.0
365.00	1860.0
366.00	276.0
372.00	456.0
403.00	209.0
420.90	170.0
422.10	302.0
423.10	1228.0
424.00	376.0
441.10	4275.0
442.10	25264.0
443.10	4716.0
444.10	474.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

m/z	Abundance
37.00	196.0
38.00	521.0
39.00	2806.0
40.00	679.0
44.00	942.0
49.00	287.0
50.00	9900.0
51.00	38224.0
52.00	1904.0
55.00	190.0
56.00	1219.0
57.00	2398.0
60.90	374.0
62.00	749.0
63.00	1542.0
64.00	235.0
65.00	761.0
69.00	39032.0
73.00	260.0
74.00	4622.0
75.00	6682.0
76.00	2712.0
77.00	45736.0
78.00	2839.0
79.00	3049.0
79.90	2482.0
81.00	3434.0
82.00	667.0
83.00	625.0
84.90	618.0
86.00	828.0
87.00	329.0
88.00	213.0
90.90	763.0
92.00	843.0
93.00	5719.0
94.00	435.0
95.90	311.0
98.00	4250.0
99.00	3054.0
100.00	343.0
101.00	1729.0
102.90	538.0
104.00	1161.0
105.00	1224.0
106.00	327.0
107.00	14174.0
108.00	2312.0
110.00	23528.0
111.00	3564.0
111.90	519.0
116.10	655.0
117.00	12315.0
118.00	749.0
122.00	958.0
123.00	1170.0
124.00	445.0
124.90	520.0
127.00	43648.0
128.00	3405.0
129.00	18952.0
130.00	1592.0
131.00	303.0
132.10	225.0
133.90	555.0
135.00	1631.0
136.00	627.0
137.00	561.0
138.00	236.0
140.10	290.0
140.90	2269.0
141.90	736.0
143.00	579.0
144.00	292.0
146.00	440.0
147.00	1203.0
148.00	3007.0
148.90	650.0
149.90	299.0
151.00	379.0
151.80	223.0
153.00	1070.0
154.00	600.0
155.00	1259.0
156.00	1555.0
157.10	341.0
157.90	472.0
158.90	475.0
159.90	780.0
161.00	1188.0
162.00	278.0
163.00	238.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

165.00	803.0
166.00	740.0
167.00	4442.0
168.00	1904.0
168.90	400.0
170.00	241.0
171.00	214.0
171.90	580.0
173.00	519.0
174.00	1086.0
175.00	1770.0
176.00	691.0
177.00	728.0
178.00	376.0
178.90	3468.0
180.00	2691.0
180.90	1163.0
181.90	275.0
184.00	442.0
185.00	1782.0
186.00	12640.0
187.00	3531.0
188.00	303.0
189.00	913.0
190.00	193.0
191.20	361.0
192.00	1108.0
193.00	1540.0
194.00	272.0
196.00	2226.0
196.70	631.0
197.90	80664.0
199.00	5596.0
199.90	401.0
201.50	331.0
203.00	670.0
204.00	3859.0
205.00	6084.0
206.00	23672.0
207.00	3627.0
208.00	860.0
208.90	291.0
210.00	583.0
211.00	848.0
215.10	237.0
216.00	964.0
217.00	6205.0
217.90	1027.0
221.00	2590.0
221.90	945.0
223.00	1354.0
224.00	13758.0
225.10	3683.0
226.10	331.0
227.00	6406.0
228.00	983.0
229.00	1424.0
231.00	522.0
232.00	154.0
234.00	385.0
235.00	458.0
235.80	159.0
237.00	366.0
239.00	347.0
240.00	218.0
241.00	340.0
242.00	757.0
243.00	579.0
244.00	10030.0
245.00	1171.0
246.00	2253.0
247.10	448.0
248.90	420.0
252.90	230.0
255.00	50392.0
256.00	7566.0
256.90	716.0
258.00	2646.0
259.00	478.0
260.00	162.0
264.00	163.0
265.00	1347.0
266.10	265.0
271.00	210.0
273.00	1303.0
274.00	3883.0
275.00	20344.0
276.00	3307.0
277.00	1793.0
278.00	288.0
283.00	240.0
284.00	161.0
285.10	299.0
293.00	387.0
296.00	6589.0
297.00	989.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

303.10	620.0
304.00	225.0
307.90	172.0
314.00	296.0
315.00	824.0
315.90	300.0
321.10	217.0
323.10	1970.0
324.00	346.0
327.00	393.0
331.90	205.0
334.00	1243.0
335.00	327.0
341.00	168.0
346.00	329.0
352.10	379.0
353.10	285.0
354.10	370.0
365.00	2637.0
366.00	342.0
372.10	695.0
373.10	194.0
383.00	285.0
402.00	312.0
403.00	329.0
421.00	297.0
422.00	304.0
423.10	2201.0
424.00	306.0
441.10	6620.0
442.10	40936.0
443.10	7553.0
444.00	650.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

m/z	Abundance
37.10	154.0
38.10	390.0
39.00	1973.0
39.90	472.0
41.00	194.0
44.00	823.0
49.00	263.0
50.00	7485.0
51.00	26792.0
52.00	1236.0
55.90	742.0
57.00	1916.0
58.00	243.0
60.90	398.0
62.00	348.0
63.00	1435.0
63.90	288.0
64.90	573.0
69.00	28416.0
70.00	176.0
73.00	302.0
74.00	3130.0
75.00	4684.0
76.00	1560.0
77.00	34304.0
78.00	2155.0
79.00	2486.0
80.00	1707.0
81.00	2953.0
82.00	635.0
82.90	507.0
85.00	306.0
86.00	629.0
87.00	440.0
90.90	523.0
92.00	585.0
93.00	3740.0
94.00	316.0
95.90	178.0
98.00	2944.0
99.00	2214.0
100.20	188.0
101.00	1440.0
103.00	374.0
103.90	752.0
105.00	983.0
106.00	221.0
107.00	9902.0
108.00	1393.0
110.00	17368.0
110.90	2287.0
112.00	348.0
116.00	541.0
117.00	8743.0
117.90	709.0
118.80	199.0
120.00	154.0
121.90	766.0
123.00	1259.0
124.00	411.0
124.90	485.0
127.00	34448.0
128.00	2665.0
129.00	13171.0
130.00	1066.0
131.00	378.0
134.00	397.0
135.00	938.0
135.90	302.0
137.00	630.0
139.90	281.0
141.00	2014.0
142.00	749.0
142.90	380.0
145.00	195.0
146.00	409.0
146.90	1061.0
148.00	2386.0
149.00	534.0
151.20	189.0
152.90	521.0
154.00	499.0
155.00	1164.0
156.00	1481.0
156.90	413.0
157.90	298.0
158.90	370.0
160.00	624.0
161.00	1010.0
161.90	265.0
164.00	170.0
164.90	786.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

166.00	506.0
167.00	3288.0
168.00	1616.0
169.00	277.0
171.00	178.0
172.00	301.0
173.10	392.0
174.00	851.0
175.00	1603.0
175.90	328.0
176.90	627.0
179.00	2937.0
180.00	1965.0
180.90	978.0
182.00	183.0
183.90	266.0
185.00	1488.0
186.00	10059.0
187.00	3172.0
187.90	258.0
188.90	858.0
190.90	314.0
192.00	925.0
193.00	1140.0
194.00	187.0
196.00	2049.0
198.00	67464.0
199.00	4509.0
200.00	320.0
201.30	335.0
203.00	653.0
204.00	2709.0
205.00	4973.0
206.00	21464.0
207.00	2954.0
208.00	634.0
209.00	299.0
210.00	388.0
211.00	723.0
215.00	168.0
216.00	632.0
217.00	5247.0
218.00	677.0
221.00	2591.0
221.80	952.0
223.00	1140.0
224.00	11040.0
225.00	2930.0
226.10	283.0
227.00	5228.0
228.00	703.0
229.00	832.0
231.00	406.0
234.00	382.0
235.10	460.0
236.00	237.0
237.00	423.0
238.90	172.0
240.00	188.0
241.00	430.0
242.00	749.0
243.00	770.0
244.00	8379.0
245.00	980.0
246.00	2128.0
247.00	465.0
249.00	315.0
252.90	309.0
255.00	44104.0
256.00	6065.0
257.00	543.0
258.00	2811.0
259.00	488.0
265.00	1399.0
273.00	1474.0
274.00	3427.0
275.00	18368.0
276.00	2427.0
277.00	1978.0
278.10	316.0
284.90	371.0
293.00	401.0
296.00	6486.0
297.00	866.0
303.00	554.0
304.00	208.0
314.00	296.0
315.00	627.0
315.90	264.0
321.00	241.0
323.10	1634.0
324.10	391.0
327.00	350.0
334.00	1157.0
335.00	320.0
346.00	455.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429

352.00	388.0
353.00	339.0
354.00	420.0
355.10	209.0
365.00	2588.0
366.00	310.0
372.00	879.0
373.10	264.0
383.00	169.0
390.10	180.0
391.00	176.0
402.00	351.0
403.10	471.0
420.90	305.0
422.00	345.0
423.10	1972.0
424.00	537.0
441.10	6795.0
442.10	42456.0
443.10	8075.0
444.10	793.0

Instrument :
BNA_G
ClientSampleId :
DFTPP429