


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG110918\  
 Data File : BG037919.D  
 Acq On : 9 Nov 2018 14:06  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG101818.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Oct 18 14:06:42 2018



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2474

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.2      | 29178   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.7       | 168     | PASS             |
| 69          | 198          | 0.00         | 100          | 33.3      | 25448   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.2       | 51      | PASS             |
| 127         | 198          | 10           | 80           | 45.0      | 34341   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 76352   | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 5345    | PASS             |
| 275         | 198          | 10           | 60           | 26.7      | 20408   | PASS             |
| 365         | 198          | 1            | 100          | 3.1       | 2335    | PASS             |
| 441         | 198          | 0.01         | 100          | 9.3       | 7106    | PASS             |
| 442         | 198          | 50           | 100          | 57.9      | 44178   | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 8402    | PASS             |