


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111119\  
 Data File : BG043602.D  
 Acq On : 12 Nov 2019 3:44  
 Operator : JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG102119.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Nov 04 15:22:15 2019



AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2498

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 32.4      | 10697   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.5       | 158     | PASS             |
| 69          | 198          | 0.00         | 100          | 32.5      | 10743   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 43.6      | 14403   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 33050   | PASS             |
| 199         | 198          | 5            | 9            | 7.7       | 2553    | PASS             |
| 275         | 198          | 10           | 60           | 38.1      | 12578   | PASS             |
| 365         | 198          | 1            | 100          | 5.3       | 1747    | PASS             |
| 441         | 198          | 0.01         | 100          | 16.2      | 5369    | PASS             |
| 442         | 198          | 50           | 100          | 91.8      | 30325   | PASS             |
| 443         | 442          | 15           | 24           | 20.3      | 6159    | PASS             |