

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111121\
 Data File : BG050970.D
 Acq On : 11 Nov 2021 16:17
 Operator : CG/JU
 Sample : M4522-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BG393

Quant Time: Nov 11 17:02:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 11 12:40:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.230	152	30664	20.000	ng/u1	0.00
20) Naphthalene-d8	11.056	136	134345	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.851	164	89261	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.595	188	196619	20.000	ng/u1	0.00
79) Chrysene-d12	21.890	240	180232	20.000	ng/u1	0.00
88) Perylene-d12	25.292	264	178512	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.594	96	4384	4.614	ng/uL	0.00
4) Pyridine-d5	4.029	84	18740	6.593	ng/u1	0.02
7) Phenol-d5	7.372	99	22417	6.853	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.542	67	61321	29.018	ng/u1	0.00
11) 2-Chlorophenol-d4	7.754	132	54341	23.969	ng/u1	0.00
15) 4-Methylphenol-d8	8.923	113	39680	15.408	ng/u1	0.00
21) Nitrobenzene-d5	9.399	128	37165	32.553	ng/u1	0.00
24) 2-Nitrophenol-d4	10.127	143	39623	31.213	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.668	165	62529	29.241	ng/u1	0.00
31) 4-Chloroaniline-d4	11.191	131	77181	23.834	ng/u1	0.00
46) Dimethylphthalate-d6	14.252	166	248775	36.429	ng/u1	0.01
49) Acenaphthylene-d8	14.552	160	290827	34.182	ng/u1	0.00
54) 4-Nitrophenol-d4	15.045	143	9794	7.910	ng/u1	0.01
60) Fluorene-d10	15.844	176	216496	35.787	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.962	200	40019	33.571	ng/u1	0.00
73) Anthracene-d10	17.695	188	365010	39.266	ng/u1	0.00
81) Pyrene-d10	19.975	212	431040	37.028	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.063	264	376904	38.193	ng/u1	0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111121\
Data File : BG050970.D
Acq On : 11 Nov 2021 16:17
Operator : CG/JU
Sample : M4522-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BG393

Quant Time: Nov 11 17:02:02 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 11 12:40:48 2021
Response via : Initial Calibration

