

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111121\
 Data File : BG051006.D
 Acq On : 12 Nov 2021 20:03
 Operator : CG/JU
 Sample : PB140666BS
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS666

Quant Time: Nov 15 00:46:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 15 00:27:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.241	152	34306	20.000	ng/u1	0.00
20) Naphthalene-d8	11.067	136	150992	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.862	164	97995	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.612	188	201756	20.000	ng/u1	0.00
79) Chrysene-d12	21.913	240	152285	20.000	ng/u1	0.00
88) Perylene-d12	25.332	264	149449	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.599	96	5292	4.979	ng/uL	0.00
4) Pyridine-d5	4.046	84	78562	24.706	ng/u1	-0.02
7) Phenol-d5	7.383	99	99927	27.304	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.553	67	61173	25.875	ng/u1	0.00
11) 2-Chlorophenol-d4	7.765	132	73973	29.165	ng/u1	0.00
15) 4-Methylphenol-d8	8.940	113	77807	27.005	ng/u1	0.00
21) Nitrobenzene-d5	9.416	128	38612	30.092	ng/u1	0.00
24) 2-Nitrophenol-d4	10.138	143	43694	30.625	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.679	165	73991	30.787	ng/u1	0.00
31) 4-Chloroaniline-d4	11.202	131	100880	27.717	ng/u1	0.00
46) Dimethylphthalate-d6	14.263	166	230810	30.786	ng/u1	0.00
49) Acenaphthylene-d8	14.563	160	290204	31.069	ng/u1	0.00
54) 4-Nitrophenol-d4	15.056	143	42811	31.493	ng/u1	0.00
60) Fluorene-d10	15.855	176	205739	30.978	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.973	200	39851	32.579	ng/u1	0.00
73) Anthracene-d10	17.712	188	297951	31.236	ng/u1	0.00
81) Pyrene-d10	19.986	212	315503	32.077	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.091	264	253395	30.671	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.634	88	10644	9.117	ng/uL#	95
5) Pyridine	4.069	79	80374	24.418	ng/u1	98
6) Benzaldehyde	7.377	77	62842	27.224	ng/u1	96
8) Phenol	7.412	94	100946	26.663	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.647	93	72853	25.708	ng/u1	99
12) 2-Chlorophenol	7.800	128	70620	27.423	ng/u1	96
13) 2-Methylphenol	8.675	108	76159	27.216	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.758	45	110304	24.712	ng/u1	98
16) Acetophenone	9.075	105	118372	26.447	ng/u1	99
17) N-Nitroso-di-n-propyla...	9.051	70	70215	26.001	ng/u1	97
18) 4-Methylphenol	9.010	108	79788	26.779	ng/u1	92
19) Hexachloroethane	9.328	117	29820	27.695	ng/u1	98
22) Nitrobenzene	9.457	77	98754	27.595	ng/u1	96
23) Isophorone	9.997	82	192537	27.722	ng/u1	99
25) 2-Nitrophenol	10.174	139	43126	30.129	ng/u1	97
26) 2,4-Dimethylphenol	10.215	107	83382	26.469	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.456	93	102560	27.406	ng/u1	99
29) 2,4-Dichlorophenol	10.708	162	70761	30.184	ng/u1	99
30) Naphthalene	11.120	128	238578	28.895	ng/u1	98
32) 4-Chloroaniline	11.225	127	96096	26.590	ng/u1	100
33) Hexachlorobutadiene	11.384	225	45735	29.724	ng/u1	97
34) Caprolactam	12.083	113	26054	26.178	ng/u1	99
35) 4-Chloro-3-methylphenol	12.330	107	86016	28.723	ng/u1	97
36) 2-Methylnaphthalene	12.706	142	158670	28.209	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111121\
 Data File : BG051006.D
 Acq On : 12 Nov 2021 20:03
 Operator : CG/JU
 Sample : PB140666BS
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS666

Quant Time: Nov 15 00:46:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 15 00:27:19 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.923	142	160423	28.147	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.064	216	89037	31.183	ng/ul	96
40) Hexachlorocyclopentadiene	13.035	237	23825	17.376	ng/ul	88
41) 2,4,6-Trichlorophenol	13.305	196	62491	33.446	ng/ul	98
42) 2,4,5-Trichlorophenol	13.376	196	66417	33.111	ng/ul	97
43) 1,1'-Biphenyl	13.699	154	216907	30.283	ng/ul	98
44) 2-Chloronaphthalene	13.752	162	170699	30.411	ng/ul	98
45) 2-Nitroaniline	13.952	65	63697	28.563	ng/ul	95
47) Dimethylphthalate	14.310	163	221222	29.510	ng/ul	99
48) 2,6-Dinitrotoluene	14.439	165	47547	30.304	ng/ul	90
50) Acenaphthylene	14.592	152	279256	29.830	ng/ul	97
51) 3-Nitroaniline	14.768	138	46862	28.878	ng/ul	96
52) Acenaphthene	14.927	153	185464	30.129	ng/ul	96
53) 2,4-Dinitrophenol	14.980	184	25490	29.450	ng/ul	91
55) 4-Nitrophenol	15.074	109	37078	29.741	ng/ul	97
56) Dibenzofuran	15.262	168	259448	29.446	ng/ul	99
57) 2,4-Dinitrotoluene	15.221	165	65236	29.134	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.485	232	54588	34.629	ng/ul	95
59) Diethylphthalate	15.661	149	228909	28.527	ng/ul	98
61) Fluorene	15.908	166	203430	29.170	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.890	204	109430	30.134	ng/ul	96
63) 4-Nitroaniline	15.932	138	48191	29.935	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.985	198	37618	31.536	ng/ul	94
67) N-Nitrosodiphenylamine	16.108	169	181699	32.219	ng/ul	98
68) 4-Bromophenyl-phenylether	16.789	248	66123	32.948	ng/ul	93
69) Hexachlorobenzene	16.913	284	68648	33.273	ng/ul	95
70) Atrazine	17.054	200	70358	29.420	ng/ul	99
71) Pentachlorophenol	17.259	266	41318	43.620	ng/ul	94
72) Phenanthrene	17.653	178	334528	31.062	ng/ul	99
74) Anthracene	17.747	178	325924	30.162	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.670	216	93044	33.870	ng/uL	97
76) Pentachlorobenzene	15.180	250	82993	32.605	ng/uL	100
77) Carbazole	18.012	167	298440	30.814	ng/ul	100
78) Di-n-butylphthalate	18.546	149	377112	29.632	ng/ul	99
80) Fluoranthene	19.651	202	366012	31.007	ng/ul	98
82) Pyrene	20.015	202	358845	31.112	ng/ul	98
83) Butylbenzylphthalate	20.879	149	146863	29.611	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.795	252	110220	29.720	ng/ul	99
85) Benzo(a)anthracene	21.889	228	317734	30.138	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.754	149	206061	28.943	ng/ul	99
87) Chrysene	21.960	228	306959	30.479	ng/ul	99
89) Di-n-octyl phthalate	23.029	149	349355	28.713	ng/ul	100
90) Benzo(b)fluoranthene	24.234	252	320841	30.135	ng/ul	99
91) Benzo(k)fluoranthene	24.304	252	292088	29.236	ng/ul	98
93) Benzo(a)pyrene	25.168	252	302709	29.853	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.263	276	340231	30.141	ng/ul	94
95) Dibenzo(a,h)anthracene	29.328	278	285668	29.910	ng/ul	97
96) Benzo(g,h,i)perylene	30.503	276	281904	29.835	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111121\
 Data File : BG051006.D
 Acq On : 12 Nov 2021 20:03
 Operator : CG/JU
 Sample : PB140666BS
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS666

Quant Time: Nov 15 00:46:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 15 00:27:19 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.117 ng/uL
RT: 3.634 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion: 88 Resp: 10644
Ion Ratio Lower Upper
88 100
43 38.6 23.0 34.4#
58 78.5 62.4 93.6



#5
Pyridine
Concen: 24.418 ng/uL
RT: 4.069 min Scan# 99
Delta R.T. -0.018 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion: 79 Resp: 80374
Ion Ratio Lower Upper
79 100
52 76.9 60.8 91.2
51 35.2 30.5 45.7









#6
Benzaldehyde
Concen: 27.224 ng/ul
RT: 7.377 min Scan# 662
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion: 77 Resp: 62842
Ion Ratio Lower Upper
77 100
105 89.0 70.4 105.6
106 83.0 61.2 91.8



#8
Phenol
Concen: 26.663 ng/ul
RT: 7.412 min Scan# 668
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion: 94 Resp: 100946
Ion Ratio Lower Upper
94 100
65 29.8 25.1 37.7
66 41.4 31.6 47.4









#10
Bis(2-Chloroethyl)ether
Concen: 25.708 ng/ul
RT: 7.647 min Scan# 709
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Instrument :
BNA_G
ClientSampleId :
SLCS666

Tgt Ion: 93 Resp: 72853




Ion	Ratio	Lower	Upper
93	100		
63	84.5	68.6	103.0
95	31.8	26.1	39.1




#12
2-Chlorophenol
Concen: 27.423 ng/ul
RT: 7.800 min Scan# 734
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion: 128 Resp: 70620

Ion	Ratio	Lower	Upper
128	100		
64	54.2	45.3	67.9
130	31.6	27.1	40.7





#13

2-Methylphenol

Concen: 27.216 ng/ul

RT: 8.675 min Scan# 884

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Instrument :

BNA_G

ClientSampleId :





SLCS666

Tgt Ion:108 Resp: 76159

Ion Ratio Lower Upper

108 100

107 94.5 71.8 107.6



#14

2,2'-oxybis(1-chloropropane)

Concen: 24.712 ng/ul

RT: 8.758 min Scan# 897

Delta R.T. -0.000 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03




Tgt Ion: 45 Resp: 110304


Ion Ratio Lower Upper

45 100

77 14.2 10.7 16.1

79 10.1 8.5 12.7









#16
Acetophenone
Concen: 26.447 ng/ul
RT: 9.075 min Scan# 91
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:105 Resp: 118372
Ion Ratio Lower Upper
105 100
77 83.6 67.3 100.9
51 29.5 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 26.001 ng/ul
RT: 9.051 min Scan# 947
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion: 70 Resp: 70215
Ion Ratio Lower Upper
70 100
42 59.9 46.0 69.0
101 9.0 7.0 10.4
130 18.2 13.4 20.2









#18
4-Methylphenol
Concen: 26.779 ng/ul
RT: 9.010 min Scan# 940
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:108 Resp: 79788
Ion Ratio Lower Upper
108 100
107 109.9 95.0 142.4



#19
Hexachloroethane
Concen: 27.695 ng/ul
RT: 9.328 min Scan# 994
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:117 Resp: 29820
Ion Ratio Lower Upper
117 100
201 99.7 77.2 115.8
199 62.8 49.9 74.9









#22
Nitrobenzene
Concen: 27.595 ng/ul
RT: 9.457 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion: 77 Resp: 98754
Ion Ratio Lower Upper
77 100
123 40.4 30.1 45.1
65 13.5 9.6 14.4



#23
Isophorone
Concen: 27.722 ng/ul
RT: 9.997 min Scan# 1108
Delta R.T. -0.018 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion: 82 Resp: 192537
Ion Ratio Lower Upper
82 100
95 7.4 5.6 8.4
138 15.4 11.8 17.8









#25
2-Nitrophenol
Concen: 30.129 ng/ul
RT: 10.174 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:139	Resp:	43126
Ion Ratio	Lower	Upper
139	100	
65	59.0	48.0 72.0
109	32.9	28.6 42.8



#26
2,4-Dimethylphenol
Concen: 26.469 ng/ul
RT: 10.215 min Scan# 1145
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:107	Resp:	83382
Ion Ratio	Lower	Upper
107	100	
121	47.9	39.3 58.9
122	81.9	63.7 95.5





#27

Bis(2-Chloroethoxy)methane

Concen: 27.406 ng/ul

RT: 10.456 min Scan# 1187

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Instrument :

BNA_G

ClientSampleId :

SLCS666





Tgt Ion: 93 Resp: 102560

Ion Ratio Lower Upper

93 100

95 32.6 26.5 39.7

123 9.3 8.1 12.1



#29

2,4-Dichlorophenol

Concen: 30.184 ng/ul

RT: 10.708 min Scan# 1229

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03




Tgt Ion:162 Resp: 70761


Ion Ratio Lower Upper

162 100

164 62.4 49.4 74.0

98 38.3 30.4 45.6









#30
Naphthalene
Concen: 28.895 ng/ul
RT: 11.120 min Scan# 1117
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:128 Resp: 238578
Ion Ratio Lower Upper
128 100
129 11.3 8.5 12.7
127 13.8 11.8 17.6



#32
4-Chloroaniline
Concen: 26.590 ng/ul
RT: 11.225 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:127 Resp: 96096
Ion Ratio Lower Upper
127 100
129 32.5 26.2 39.2









#33
Hexachlorobutadiene
Concen: 29.724 ng/u1
RT: 11.384 min Scan# 11
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:225 Resp: 45735
Ion Ratio Lower Upper
225 100
223 62.1 50.6 75.8
227 66.2 49.8 74.8



#34
Caprolactam
Concen: 26.178 ng/u1
RT: 12.083 min Scan# 1463
Delta R.T. -0.053 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:113 Resp: 26054
Ion Ratio Lower Upper
113 100
55 181.6 147.0 220.6
56 135.5 109.2 163.8









#35
 4-Chloro-3-methylphenol
 Concen: 28.723 ng/ul
 RT: 12.330 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03




Instrument :
 BNA_G
 ClientSampleId :
 SLCS666


Tgt Ion:	107	Resp:	86016
Ion	Ratio	Lower	Upper
107	100		
144	25.3	19.8	29.6
142	73.6	56.2	84.4



#36
 2-Methylnaphthalene
 Concen: 28.209 ng/ul
 RT: 12.706 min Scan# 1569
 Delta R.T. -0.006 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03

Tgt Ion:	142	Resp:	158670
Ion	Ratio	Lower	Upper
142	100		
141	93.6	73.3	109.9





#37

1-Methylnaphthalene

Concen: 28.147 ng/ul

RT: 12.923 min Scan# 1

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Instrument :

BNA_G

ClientSampleId :

SLCS666





Tgt Ion:142 Resp: 160423

Ion Ratio Lower Upper

142 100

141 95.6 77.8 116.8

116 3.9 3.3 4.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 31.183 ng/ul

RT: 13.064 min Scan# 1630

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Tgt Ion:216 Resp: 89037




Ion Ratio Lower Upper


216 100

214 79.5 61.0 91.4

179 21.5 16.2 24.2

108 19.2 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 17.376 ng/ul
RT: 13.035 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:237	Resp:	23825
Ion Ratio	Lower	Upper
237	100	
235	72.2	49.4 74.0
272	12.6	11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 33.446 ng/ul
RT: 13.305 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:196	Resp:	62491
Ion Ratio	Lower	Upper
196	100	
198	95.0	74.3 111.5
200	30.0	23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 33.111 ng/ul
RT: 13.376 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:196 Resp: 66417
Ion Ratio Lower Upper
196 100
198 92.5 77.0 115.4
200 30.7 25.4 38.2



#43
1,1'-Biphenyl
Concen: 30.283 ng/ul
RT: 13.699 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:154 Resp: 216907
Ion Ratio Lower Upper
154 100
153 42.2 34.5 51.7
76 14.4 13.1 19.7









#44
2-Chloronaphthalene
Concen: 30.411 ng/ul
RT: 13.752 min Scan# 1747
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:162 Resp: 170699
Ion Ratio Lower Upper
162 100
164 33.7 26.1 39.1
127 40.3 31.0 46.4



#45
2-Nitroaniline
Concen: 28.563 ng/ul
RT: 13.952 min Scan# 1781
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion: 65 Resp: 63697
Ion Ratio Lower Upper
65 100
92 66.8 53.0 79.4
138 91.7 67.4 101.0





#47

Dimethylphthalate

Concen: 29.510 ng/ul

RT: 14.310 min Scan# 1843

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Instrument :

BNA_G

ClientSampleId :

SLCS666





Tgt Ion:163 Resp: 221222

Ion Ratio Lower Upper

163 100

194 3.6 3.0 4.4

164 9.9 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 30.304 ng/ul

RT: 14.439 min Scan# 1864

Delta R.T. -0.000 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03




Tgt Ion:165 Resp: 47547


Ion Ratio Lower Upper

165 100

89 62.3 57.4 86.0

121 21.2 18.8 28.2









#50
Acenaphthylene
Concen: 29.830 ng/ul
RT: 14.592 min Scan# 1890
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:152 Resp: 279256
Ion Ratio Lower Upper
152 100
151 20.3 17.4 26.0
153 13.6 10.3 15.5



#51
3-Nitroaniline
Concen: 28.878 ng/ul
RT: 14.768 min Scan# 1920
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:138 Resp: 46862
Ion Ratio Lower Upper
138 100
108 10.5 8.8 13.2
92 127.3 98.0 147.0









#52
Acenaphthene
Concen: 30.129 ng/ul
RT: 14.927 min Scan# 1947
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:153 Resp: 185464
Ion Ratio Lower Upper
153 100
152 49.5 36.5 54.7
154 84.7 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 29.450 ng/ul
RT: 14.980 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:184 Resp: 25490
Ion Ratio Lower Upper
184 100
63 71.6 66.2 99.2
154 63.3 53.6 80.4









#55
4-Nitrophenol
Concen: 29.741 ng/ul
RT: 15.074 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:109 Resp: 37078
Ion Ratio Lower Upper
109 100
139 117.5 88.7 133.1
65 142.6 113.6 170.4



#56
Dibenzofuran
Concen: 29.446 ng/ul
RT: 15.262 min Scan# 2004
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:168 Resp: 259448
Ion Ratio Lower Upper
168 100
139 38.2 31.2 46.8





#57

2,4-Dinitrotoluene

Concen: 29.134 ng/ul

RT: 15.221 min Scan# 1998

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Instrument :

BNA_G

ClientSampleId :

SLCS666





Tgt Ion:165 Resp: 65236

Ion Ratio Lower Upper

165 100

63 49.6 42.1 63.1

182 2.3 1.8 2.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 34.629 ng/ul

RT: 15.485 min Scan# 2042

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Tgt Ion:232 Resp: 54588




Ion Ratio Lower Upper


232 100

131 45.4 38.9 58.3

130 2.4 2.4 3.6

166 29.1 25.8 38.8





#59

Diethylphthalate

Concen: 28.527 ng/ul

RT: 15.661 min Scan# 2073

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Instrument :

BNA_G

ClientSampleId :

SLCS666





Tgt Ion:149 Resp: 228909

Ion Ratio Lower Upper

149 100

177 20.4 16.8 25.2

150 11.8 10.1 15.1



#61

Fluorene

Concen: 29.170 ng/ul

RT: 15.908 min Scan# 2114

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03




Tgt Ion:166 Resp: 203430


Ion Ratio Lower Upper

166 100

165 104.1 81.0 121.6

167 14.3 11.0 16.4





#62

4-Chlorophenyl-phenylether

Concen: 30.134 ng/ul

RT: 15.890 min Scan# 2112

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03

Instrument :

BNA_G

ClientSampleId :

SLCS666





Tgt Ion:204 Resp: 109430

Ion Ratio Lower Upper

204 100

206 32.8 27.4 41.2

141 57.3 48.6 73.0



#63

4-Nitroaniline

Concen: 29.935 ng/ul

RT: 15.932 min Scan# 2118

Delta R.T. -0.006 min

Lab File: BG051006.D

Acq: 12 Nov 2021 20:03




Tgt Ion:138 Resp: 48191


Ion Ratio Lower Upper

138 100

92 57.6 49.3 73.9

108 90.2 72.6 108.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 31.536 ng/ul
 RT: 15.985 min Scan# 2127
 Delta R.T. -0.006 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03




Instrument :
 BNA_G
 ClientSampleId :
 SLCS666


Tgt Ion:198 Resp: 37618
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.8 5.8
 77 32.4 23.2 34.8



#67
 N-Nitrosodiphenylamine
 Concen: 32.219 ng/ul
 RT: 16.108 min Scan# 2148
 Delta R.T. -0.006 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03

Tgt Ion:169 Resp: 181699
 Ion Ratio Lower Upper
 169 100
 168 70.9 58.5 87.7
 167 39.5 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 32.948 ng/ul
RT: 16.789 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:248 Resp: 66123
Ion Ratio Lower Upper
248 100
250 99.0 75.5 113.3
141 75.8 67.6 101.4



#69
Hexachlorobenzene
Concen: 33.273 ng/ul
RT: 16.913 min Scan# 2285
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:284 Resp: 68648
Ion Ratio Lower Upper
284 100
142 35.7 31.9 47.9
249 33.3 25.8 38.6









#70
Atrazine
Concen: 29.420 ng/ul
RT: 17.054 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:200 Resp: 70358
Ion Ratio Lower Upper
200 100
173 24.9 20.2 30.2
215 48.2 38.1 57.1



#71
Pentachlorophenol
Concen: 43.620 ng/ul
RT: 17.259 min Scan# 2344
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:266 Resp: 41318
Ion Ratio Lower Upper
266 100
264 61.9 54.3 81.5
268 60.1 51.0 76.6









#72
Phenanthrene
Concen: 31.062 ng/ul
RT: 17.653 min Scan# 2411
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:178 Resp: 334528
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 18.9 15.5 23.3



#74
Anthracene
Concen: 30.162 ng/ul
RT: 17.747 min Scan# 2427
Delta R.T. -0.000 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:178 Resp: 325924
Ion Ratio Lower Upper
178 100
179 15.5 13.0 19.6
176 19.6 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 33.870 ng/uL
RT: 13.670 min Scan# 11
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:216 Resp: 93044
Ion Ratio Lower Upper
216 100
214 77.6 61.3 91.9
218 50.3 37.0 55.6



#76
Pentachlorobenzene
Concen: 32.605 ng/uL
RT: 15.180 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:250 Resp: 82993
Ion Ratio Lower Upper
250 100
248 63.4 50.5 75.7
252 64.4 51.2 76.8









#77
 Carbazole
 Concen: 30.814 ng/ul
 RT: 18.012 min Scan# 2473
 Delta R.T. -0.006 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03




Instrument :
 BNA_G
 ClientSampleId :
 SLCS666


Tgt Ion:167 Resp: 298440
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 13.6 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 29.632 ng/ul
 RT: 18.546 min Scan# 2563
 Delta R.T. -0.006 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03

Tgt Ion:149 Resp: 377112
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 5.7 5.0 7.6









#80
Fluoranthene
Concen: 31.007 ng/u1
RT: 19.651 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:202 Resp: 366012
Ion Ratio Lower Upper
202 100
101 10.5 9.3 13.9
100 8.5 7.0 10.6



#82
Pyrene
Concen: 31.112 ng/u1
RT: 20.015 min Scan# 2813
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:202 Resp: 358845
Ion Ratio Lower Upper
202 100
101 12.2 10.7 16.1
100 10.6 9.0 13.6









#83
Butylbenzylphthalate
Concen: 29.611 ng/ul
RT: 20.879 min Scan# 2960
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:149 Resp: 146863
Ion Ratio Lower Upper
149 100
91 77.9 65.2 97.8
206 21.0 16.2 24.2



#84
3,3'-Dichlorobenzidine
Concen: 29.720 ng/ul
RT: 21.795 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:252 Resp: 110220
Ion Ratio Lower Upper
252 100
254 63.1 50.8 76.2
126 10.3 9.0 13.4









#85
Benzo(a)anthracene
Concen: 30.138 ng/ul
RT: 21.889 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:228 Resp: 317734
Ion Ratio Lower Upper
228 100
229 20.0 15.7 23.5
226 27.8 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 28.943 ng/ul
RT: 21.754 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:149 Resp: 206061
Ion Ratio Lower Upper
149 100
167 25.6 21.1 31.7
279 3.5 2.7 4.1









#87
Chrysene
Concen: 30.479 ng/ul
RT: 21.960 min Scan# 3145
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:228 Resp: 306959
Ion Ratio Lower Upper
228 100
226 30.9 24.8 37.2
229 20.3 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 28.713 ng/ul
RT: 23.029 min Scan# 3326
Delta R.T. -0.006 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:149 Resp: 349355









#90
Benzo(b)fluoranthene
Concen: 30.135 ng/ul
RT: 24.234 min Scan# 3533
Delta R.T. -0.012 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:252 Resp: 320841
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 9.6 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 29.236 ng/ul
RT: 24.304 min Scan# 3543
Delta R.T. -0.012 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:252 Resp: 292088
Ion Ratio Lower Upper
252 100
253 22.7 17.3 25.9
125 9.4 7.8 11.6









#93
Benzo(a)pyrene
Concen: 29.853 ng/u1
RT: 25.168 min Scan# 30
Delta R.T. -0.012 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03




Instrument :
BNA_G
ClientSampleId :
SLCS666


Tgt Ion:252 Resp: 302709
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.2
125 10.6 9.1 13.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.141 ng/u1
RT: 29.263 min Scan# 4387
Delta R.T. -0.018 min
Lab File: BG051006.D
Acq: 12 Nov 2021 20:03

Tgt Ion:276 Resp: 340231
Ion Ratio Lower Upper
276 100
138 16.2 15.5 23.3
277 23.2 20.5 30.7









#95
 Dibenzo(a,h)anthracene
 Concen: 29.910 ng/ul
 RT: 29.328 min Scan# 41
 Delta R.T. -0.018 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03

Instrument :
 BNA_G
 ClientSampleId :
 SLCS666

Tgt Ion:278 Resp: 285668
 Ion Ratio Lower Upper
 278 100
 139 13.1 12.2 18.4
 279 23.3 19.5 29.3



#96
 Benzo(g,h,i)perylene
 Concen: 29.835 ng/ul
 RT: 30.503 min Scan# 4598
 Delta R.T. -0.018 min
 Lab File: BG051006.D
 Acq: 12 Nov 2021 20:03

Tgt Ion:276 Resp: 281904
 Ion Ratio Lower Upper
 276 100
 138 19.2 16.6 24.8
 277 24.6 17.6 26.4

