


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111121\
 Data File : BG050962.D
 Acq On : 11 Nov 2021 10:36
 Operator : CG/JU
 Sample : DFTPP430
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP430

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Title : SVOA CALIBRATION
 Last Update : Mon Nov 15 12:03:08 2021



AutoFind: Scans 2468, 2469, 2470; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.1	28469	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	44.1	29157	PASS
70	69	0.00	2	0.2	58	PASS
127	198	10	80	51.3	33909	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	66061	PASS
199	198	5	9	6.8	4524	PASS
275	198	10	60	26.6	17598	PASS
365	198	1	100	3.3	2212	PASS
441	198	0.01	100	8.8	5784	PASS
442	198	50	100	56.1	37035	PASS
443	442	15	24	19.4	7187	PASS

Scan 2462 (17.957 min): BG050962.D\data.ms
DFTPP430

m/z	Abundance
40.00	588.0
43.90	566.0
206.90	164.0

Instrument :

BNA_G

ClientSampleId :

DFTPP430

m/z	Abundance
38.00	582.0
39.00	2774.0
40.00	516.0
41.10	181.0
44.00	446.0
48.90	391.0
50.00	8864.0
51.00	35968.0
52.00	2026.0
55.00	195.0
56.00	953.0
57.00	2522.0
61.00	406.0
62.00	558.0
63.00	1530.0
64.00	214.0
65.10	819.0
69.00	35832.0
72.90	341.0
74.00	4035.0
75.00	6029.0
76.00	2029.0
77.00	44744.0
78.00	2658.0
79.00	2412.0
80.00	2301.0
81.00	3304.0
82.00	1105.0
82.90	772.0
84.90	476.0
86.00	888.0
87.00	329.0
87.90	281.0
91.00	636.0
92.00	712.0
93.00	5143.0
94.00	344.0
96.10	196.0
98.00	3616.0
99.00	2907.0
100.00	300.0
100.90	1604.0
102.90	711.0
104.00	1113.0
104.90	891.0
105.90	356.0
107.00	13165.0
108.00	1881.0
109.00	403.0
110.00	20656.0
111.00	3193.0
112.00	341.0
113.00	231.0
116.10	791.0
117.00	10013.0
117.90	864.0
121.90	836.0
123.00	1322.0
124.00	617.0
125.00	631.0
127.00	40184.0
128.00	2950.0
129.00	16488.0
130.00	1322.0
131.00	209.0
131.90	160.0
134.00	495.0
135.00	1474.0
135.90	657.0
137.00	724.0
140.00	194.0
141.00	2264.0
142.00	700.0
143.00	422.0
143.90	181.0
146.00	499.0
147.00	1200.0
148.00	2569.0
149.00	556.0
149.80	181.0
151.00	273.0
152.00	205.0
152.90	769.0
154.00	596.0
155.00	1213.0
156.00	1690.0
157.00	333.0
157.90	410.0
159.00	197.0
160.00	656.0
161.00	992.0
162.00	345.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

165.00	869.0
166.00	799.0
167.00	4170.0
168.00	1845.0
169.00	391.0
169.90	159.0
172.00	240.0
173.00	700.0
174.00	879.0
175.00	1925.0
176.00	460.0
177.00	825.0
178.00	289.0
179.00	3330.0
180.00	2078.0
180.90	1241.0
182.00	250.0
184.00	302.0
185.00	1471.0
186.00	11431.0
187.00	3969.0
188.00	359.0
188.90	1053.0
190.00	176.0
191.00	531.0
192.00	971.0
192.90	1128.0
194.00	278.0
196.00	2344.0
197.90	73864.0
198.90	5061.0
199.90	391.0
201.40	258.0
203.00	550.0
204.00	2858.0
205.00	5548.0
206.00	22584.0
207.00	3048.0
208.00	641.0
208.90	413.0
210.00	378.0
210.90	847.0
211.80	176.0
215.00	249.0
216.00	491.0
217.00	5851.0
217.90	739.0
221.00	2610.0
221.90	625.0
223.00	1587.0
224.00	12263.0
225.00	3078.0
226.00	352.0
227.00	5028.0
228.00	603.0
229.00	1055.0
231.00	447.0
234.00	420.0
235.00	700.0
236.00	286.0
237.00	269.0
239.00	317.0
240.00	167.0
240.90	289.0
242.00	593.0
243.00	689.0
244.00	8693.0
245.00	1320.0
246.00	1740.0
247.00	493.0
248.90	327.0
253.00	214.0
255.00	46464.0
256.00	6589.0
257.00	557.0
258.00	2979.0
259.00	413.0
265.00	1175.0
273.00	1572.0
274.00	3725.0
275.00	18856.0
276.00	2828.0
277.00	2010.0
277.90	283.0
283.10	215.0
283.90	210.0
285.00	239.0
293.00	403.0
293.90	171.0
295.00	211.0
296.00	5953.0
297.00	844.0
302.10	160.0
303.10	571.0
304.00	193.0
314.10	296.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

315.00	693.0
316.00	422.0
320.90	171.0
323.10	1672.0
323.90	320.0
326.90	399.0
327.90	183.0
333.00	172.0
334.00	1076.0
335.00	238.0
341.10	169.0
346.00	307.0
351.90	365.0
353.00	246.0
354.10	440.0
365.00	2282.0
365.90	241.0
372.00	712.0
373.00	194.0
383.00	160.0
402.10	352.0
403.00	336.0
421.10	191.0
422.10	323.0
423.00	1624.0
424.10	423.0
441.10	5283.0
442.10	34280.0
443.10	7384.0
444.00	609.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

m/z	Abundance
37.00	179.0
37.90	394.0
39.00	2134.0
40.00	570.0
44.00	619.0
49.00	309.0
50.00	8721.0
51.00	31800.0
52.00	1726.0
54.90	270.0
56.00	1062.0
57.00	2082.0
60.90	524.0
61.90	494.0
63.00	1407.0
63.90	177.0
65.00	833.0
69.00	33536.0
69.90	170.0
73.00	440.0
74.00	3461.0
75.00	5295.0
76.00	1856.0
77.00	38784.0
78.00	2653.0
79.00	2510.0
79.90	1997.0
81.00	2470.0
82.00	825.0
83.00	719.0
83.90	171.0
85.00	336.0
86.00	849.0
87.00	317.0
87.90	150.0
91.00	586.0
92.00	743.0
93.00	4284.0
94.00	224.0
96.10	296.0
98.00	3675.0
99.00	2572.0
100.00	161.0
101.00	1759.0
102.90	474.0
103.90	1194.0
105.00	1044.0
107.00	11561.0
108.00	1669.0
110.00	19992.0
111.00	3197.0
112.00	300.0
113.00	166.0
116.00	802.0
117.00	10471.0
117.90	613.0
121.90	642.0
123.00	1247.0
124.00	619.0
124.90	630.0
127.00	38992.0
128.00	3052.0
129.00	15748.0
130.00	1294.0
130.90	215.0
132.00	167.0
134.00	340.0
135.00	1446.0
135.90	506.0
136.90	618.0
138.10	159.0
140.00	169.0
141.00	1996.0
141.90	445.0
142.90	441.0
143.80	170.0
146.00	274.0
147.00	1460.0
148.00	2697.0
148.90	595.0
150.90	310.0
152.90	821.0
153.90	625.0
155.00	1256.0
156.00	2066.0
156.90	210.0
157.90	463.0
158.90	426.0
159.90	629.0
161.00	945.0
161.70	221.0
162.10	213.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

163.90	187.0
165.00	876.0
166.10	677.0
167.00	4345.0
168.00	2022.0
168.90	346.0
170.10	173.0
171.00	270.0
171.90	353.0
173.00	558.0
174.00	888.0
175.00	1675.0
175.90	481.0
177.00	969.0
177.90	344.0
178.90	3312.0
180.00	2262.0
181.00	1094.0
183.90	312.0
185.00	1459.0
186.00	11737.0
187.00	3339.0
188.10	391.0
188.90	954.0
189.90	246.0
190.90	409.0
192.00	1080.0
193.00	1387.0
193.90	259.0
194.80	162.0
196.00	2407.0
197.90	76792.0
198.90	5422.0
200.00	422.0
201.40	359.0
203.00	759.0
204.00	3582.0
205.00	5189.0
206.00	23056.0
207.00	2976.0
208.00	944.0
209.00	336.0
210.00	444.0
211.00	815.0
215.00	229.0
216.00	550.0
216.90	6126.0
217.90	719.0
221.00	2952.0
222.00	824.0
223.00	1437.0
224.00	13420.0
225.00	3416.0
226.00	379.0
227.00	5684.0
228.00	819.0
229.00	1066.0
230.90	358.0
233.90	434.0
235.00	387.0
235.90	382.0
237.00	344.0
239.10	182.0
240.00	187.0
241.00	299.0
242.00	692.0
243.00	748.0
244.00	9064.0
245.00	1272.0
246.00	2182.0
247.00	444.0
247.90	150.0
249.00	434.0
253.00	245.0
255.00	49432.0
256.00	6618.0
257.00	579.0
258.00	3268.0
259.00	493.0
264.00	152.0
265.00	1134.0
266.00	198.0
272.10	166.0
273.00	1452.0
274.00	3550.0
275.00	21136.0
276.00	2881.0
277.00	2163.0
278.00	297.0
282.90	246.0
284.10	195.0
285.00	376.0
292.90	354.0
295.00	157.0
296.00	6693.0
296.90	865.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

302.10	240.0
303.00	734.0
303.80	171.0
304.00	172.0
307.90	151.0
314.00	325.0
315.00	806.0
316.00	275.0
321.00	268.0
323.00	1634.0
324.10	442.0
327.00	329.0
327.90	176.0
332.00	153.0
333.00	189.0
334.00	1227.0
335.00	312.0
341.00	230.0
346.10	535.0
352.00	773.0
353.00	419.0
354.10	416.0
365.00	2664.0
366.00	372.0
371.20	157.0
372.00	1255.0
373.00	176.0
402.00	313.0
403.00	390.0
404.10	192.0
421.10	282.0
422.00	266.0
423.10	2259.0
424.10	492.0
441.10	6959.0
442.10	42872.0
443.10	8069.0
444.00	948.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

m/z	Abundance
38.10	247.0
39.00	1127.0
39.90	363.0
43.90	713.0
50.00	4758.0
51.00	17640.0
52.00	845.0
55.90	428.0
57.00	1008.0
61.00	220.0
61.90	253.0
63.00	640.0
65.00	276.0
69.00	18104.0
70.10	175.0
73.10	219.0
74.00	1825.0
75.00	3301.0
76.00	1048.0
77.00	22240.0
78.00	1535.0
79.00	1331.0
80.00	1014.0
81.00	1652.0
82.00	323.0
83.00	373.0
85.00	374.0
85.90	439.0
86.80	169.0
91.00	217.0
91.90	430.0
93.00	2386.0
93.90	170.0
96.10	269.0
97.30	206.0
98.00	1651.0
98.90	1671.0
99.90	187.0
100.90	991.0
102.90	340.0
103.90	692.0
105.00	569.0
107.00	6849.0
108.00	1013.0
108.90	217.0
110.00	10537.0
111.00	2076.0
112.00	230.0
116.00	412.0
117.00	5773.0
118.00	499.0
122.00	482.0
123.00	722.0
124.00	397.0
124.90	351.0
127.00	22552.0
127.90	1446.0
129.00	9271.0
130.00	958.0
133.90	260.0
135.00	613.0
135.90	298.0
137.00	394.0
139.90	166.0
141.00	1108.0
141.80	405.0
142.90	355.0
146.90	714.0
148.00	1499.0
149.00	281.0
151.10	293.0
153.00	368.0
154.00	307.0
155.00	711.0
156.00	940.0
156.90	233.0
158.90	152.0
159.90	359.0
160.90	608.0
162.00	194.0
165.00	680.0
166.00	431.0
167.00	2545.0
168.00	1158.0
168.90	183.0
172.00	419.0
173.00	319.0
173.90	381.0
175.00	1031.0
175.90	320.0
176.90	454.0
179.00	1801.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

180.00	1233.0
181.00	712.0
184.00	217.0
185.00	829.0
186.00	6794.0
187.00	2003.0
187.90	243.0
188.90	408.0
191.00	506.0
192.00	543.0
193.10	693.0
194.90	195.0
196.00	1384.0
197.90	47528.0
198.90	3088.0
199.90	181.0
201.40	163.0
203.00	546.0
204.00	2371.0
205.00	3458.0
206.00	14147.0
207.00	1739.0
208.00	698.0
209.80	172.0
211.00	660.0
214.90	181.0
216.00	240.0
217.00	4003.0
217.90	472.0
221.00	1887.0
221.90	608.0
223.00	759.0
224.00	8258.0
225.00	2006.0
227.00	4040.0
228.00	560.0
229.00	516.0
231.00	167.0
234.00	259.0
235.00	274.0
236.10	207.0
236.90	277.0
240.00	276.0
242.10	426.0
243.00	479.0
244.00	6136.0
245.00	785.0
246.00	1562.0
247.00	267.0
249.00	343.0
252.90	183.0
253.90	311.0
255.00	30776.0
256.00	4250.0
257.00	315.0
258.00	2053.0
259.00	313.0
264.90	832.0
272.00	192.0
272.90	904.0
274.00	2805.0
275.00	12801.0
276.00	1715.0
277.00	1338.0
278.00	236.0
285.00	284.0
293.00	294.0
296.00	4480.0
297.00	510.0
303.00	521.0
314.10	172.0
315.00	604.0
316.00	265.0
323.10	1292.0
323.90	213.0
327.00	267.0
328.00	195.0
334.00	997.0
335.00	282.0
340.90	154.0
346.00	365.0
352.00	392.0
352.90	294.0
354.10	348.0
365.00	1691.0
366.00	317.0
372.00	576.0
373.10	159.0
383.10	208.0
402.00	249.0
403.10	298.0
404.00	197.0
421.10	317.0
422.00	309.0
423.10	1787.0
424.00	443.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430

441.10	5110.0
442.10	33952.0
443.10	6108.0
444.10	707.0

Instrument :
BNA_G
ClientSampleId :
DFTPP430