

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG111222\  
 Data File : BG055587.D  
 Acq On : 12 Nov 2022 12:55  
 Operator : CG/JU  
 Sample : PB148983BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 PB148983BL

Quant Time: Nov 14 03:25:54 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG110922.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 10 03:39:46 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.270  | 152  | 37719    | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 11.097 | 136  | 153970   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.892 | 164  | 118987   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.630 | 188  | 317883   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.925 | 240  | 314514   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 25.356 | 264  | 325563   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.797  | 112  | 314992   | 159.064 | ng    | 0.00     |
| 7) Phenol-d6                | 7.407  | 99   | 412863   | 150.932 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.440  | 82   | 246083   | 108.348 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.373 | 330  | 242830   | 209.163 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.517 | 172  | 664995   | 84.301  | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.221 | 244  | 1420744  | 90.732  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG111222\  
Data File : BG055587.D  
Acq On : 12 Nov 2022 12:55  
Operator : CG/JU  
Sample : PB148983BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
PB148983BL

Quant Time: Nov 14 03:25:54 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG110922.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 10 03:39:46 2022  
Response via : Initial Calibration

