

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111222\
 Data File : BG055588.D
 Acq On : 12 Nov 2022 13:45
 Operator : CG/JU
 Sample : N5509-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-7

Quant Time: Nov 14 03:26:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 03:39:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.271	152	38085	20.000	ng	0.00
21) Naphthalene-d8	11.097	136	141390	20.000	ng	0.00
39) Acenaphthene-d10	14.886	164	98872	20.000	ng	0.00
64) Phenanthrene-d10	17.630	188	262494	20.000	ng	0.00
76) Chrysene-d12	21.925	240	287186	20.000	ng	0.00
86) Perylene-d12	25.351	264	299661	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.797	112	371049	185.571	ng	0.00
7) Phenol-d6	7.407	99	444050	160.773	ng	0.00
23) Nitrobenzene-d5	9.440	82	300851	144.248	ng	0.00
42) 2,4,6-Tribromophenol	16.367	330	274536	284.582	ng	0.00
45) 2-Fluorobiphenyl	13.517	172	736167	112.310	ng	0.00
79) Terphenyl-d14	20.215	244	1603257	112.131	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111222\
Data File : BG055588.D
Acq On : 12 Nov 2022 13:45
Operator : CG/JU
Sample : N5509-02
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-7

Quant Time: Nov 14 03:26:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 03:39:46 2022
Response via : Initial Calibration

