


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111319\
 Data File : BG043613.D
 Acq On : 13 Nov 2019 14:43
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG102119.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 04 15:22:15 2019



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.4	11066	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	32.8	12362	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.6	15292	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	37672	PASS
199	198	5	9	6.3	2370	PASS
275	198	10	60	32.7	12304	PASS
365	198	1	100	5.0	1898	PASS
441	198	0.01	100	14.5	5456	PASS
442	198	50	100	91.7	34554	PASS
443	442	15	24	20.3	6998	PASS