


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111419\
 Data File : BG043647.D
 Acq On : 15 Nov 2019 4:41
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG102119.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 04 15:22:15 2019



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.8	17199	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.0	18454	PASS
70	69	0.00	2	0.5	84	PASS
127	198	10	80	41.7	23304	PASS
197	198	0.00	2	0.3	187	PASS
198	198	100	100	100.0	55845	PASS
199	198	5	9	6.4	3592	PASS
275	198	10	60	32.0	17880	PASS
365	198	1	100	4.2	2323	PASS
441	198	0.01	100	13.3	7443	PASS
442	198	50	100	85.4	47672	PASS
443	442	15	24	19.2	9173	PASS