

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111425\
 Data File : BG064775.D
 Acq On : 14 Nov 2025 10:59
 Operator : RC/JU
 Sample : PB170517TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB170517TB

Quant Time: Nov 14 11:41:19 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 12 17:10:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.164	152	36175	20.000	ng	0.00
21) Naphthalene-d8	10.979	136	138484	20.000	ng	0.00
39) Acenaphthene-d10	14.774	164	113390	20.000	ng	0.00
64) Phenanthrene-d10	17.500	188	294619	20.000	ng	0.00
76) Chrysene-d12	21.748	240	423867	20.000	ng	0.00
86) Perylene-d12	25.003	264	416040	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.703	112	270690	130.666	ng	0.00
7) Phenol-d6	7.307	99	351133	127.437	ng	0.00
23) Nitrobenzene-d5	9.322	82	220889	78.864	ng	0.00
42) 2,4,6-Tribromophenol	16.249	330	271602	122.816	ng	0.00
45) 2-Fluorobiphenyl	13.405	172	650168	80.008	ng	0.00
79) Terphenyl-d14	20.092	244	1828266	84.847	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111425\
Data File : BG064775.D
Acq On : 14 Nov 2025 10:59
Operator : RC/JU
Sample : PB170517TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170517TB


Quant Time: Nov 14 11:41:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 12 17:10:56 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.164 min Scan# 80
Delta R.T. -0.003 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59

Instrument :
BNA_G
ClientSampleId :
PB170517TB







Tgt Ion:152 Resp: 36175
Ion Ratio Lower Upper
152 100
150 155.0 129.2 193.8
115 59.3 53.7 80.5



#5
2-Fluorophenol
Concen: 130.666 ng
RT: 5.703 min Scan# 445
Delta R.T. 0.003 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59

Tgt Ion:112 Resp: 270690
Ion Ratio Lower Upper
112 100
64 72.5 61.0 91.4
63 40.8 33.3 49.9









#7
Phenol-d6
Concen: 127.437 ng
RT: 7.307 min Scan# 718
Delta R.T. -0.003 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59




Instrument :
BNA_G
ClientSampleId :
PB170517TB


Tgt Ion: 99 Resp: 351133
Ion Ratio Lower Upper
99 100
42 21.4 16.8 25.2
71 38.7 31.1 46.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.979 min Scan# 1343
Delta R.T. -0.008 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59

Tgt Ion: 136 Resp: 138484
Ion Ratio Lower Upper
136 100
137 11.3 9.3 13.9
54 11.9 9.3 13.9
68 6.0 5.0 7.6





#23

Nitrobenzene-d5

Concen: 78.864 ng

RT: 9.322 min Scan# 1061

Delta R.T. -0.009 min

Lab File: BG064775.D

Acq: 14 Nov 2025 10:59

Instrument :

BNA_G

ClientSampleId :

PB170517TB





Tgt Ion: 82 Resp: 220889

Ion Ratio Lower Upper

82 100

128 41.3 34.1 51.1

54 63.6 54.3 81.5



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.774 min Scan# 1989

Delta R.T. -0.003 min

Lab File: BG064775.D

Acq: 14 Nov 2025 10:59




Tgt Ion:164 Resp: 113390


Ion Ratio Lower Upper

164 100

162 98.7 79.8 119.6


160 47.2 34.9 52.3








#42
2,4,6-Tribromophenol
Concen: 122.816 ng
RT: 16.249 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59


Instrument :
BNA_G
ClientSampleId :
PB170517TB





Tgt Ion:330 Resp: 271602
Ion Ratio Lower Upper
330 100
332 97.3 77.0 115.6
141 23.0 18.2 27.2




#45
2-Fluorobiphenyl
Concen: 80.008 ng
RT: 13.405 min Scan# 1756
Delta R.T. -0.003 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59



Tgt Ion:172 Resp: 650168
Ion Ratio Lower Upper
172 100
171 36.9 29.9 44.9
170 24.4 19.8 29.6









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.500 min Scan# 2453
Delta R.T. -0.003 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59




Instrument :
BNA_G
ClientSampleId :
PB170517TB


Tgt Ion:188 Resp: 294619
Ion Ratio Lower Upper
188 100
94 6.6 4.7 7.1
80 6.9 5.4 8.2



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.748 min Scan# 3176
Delta R.T. -0.009 min
Lab File: BG064775.D
Acq: 14 Nov 2025 10:59

Tgt Ion:240 Resp: 423867
Ion Ratio Lower Upper
240 100
120 5.4 4.2 6.4
236 25.1 20.8 31.2









#79
 Terphenyl-d14
 Concen: 84.847 ng
 RT: 20.092 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: BG064775.D
 Acq: 14 Nov 2025 10:59

Instrument :
 BNA_G
 ClientSampleId :
 PB170517TB

Tgt Ion:244 Resp: 1828266
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.8 8.6
 122 5.4 4.3 6.5



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.003 min Scan# 3730
 Delta R.T. -0.020 min
 Lab File: BG064775.D
 Acq: 14 Nov 2025 10:59

Tgt Ion:264 Resp: 416040
 Ion Ratio Lower Upper
 264 100
 260 24.2 17.2 25.8
 265 21.2 16.6 25.0

