

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111519\
 Data File : BG043664.D
 Acq On : 15 Nov 2019 23:46
 Operator : JU
 Sample : K5804-01DL 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL

Quant Time: Nov 16 01:14:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 04 15:22:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.99	152	27272	20.00	ng	-0.01
21) Naphthalene-d8	10.80	136	114514	20.00	ng	-0.02
39) Acenaphthene-d10	14.62	164	87315	20.00	ng	-0.02
64) Phenanthrene-d10	17.37	188	225700	20.00	ng	-0.02
76) Chrysene-d12	21.64	240	257080	20.00	ng	-0.02
87) Perylene-d12	24.82	264	298613	20.00	ng	-0.04

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.59	112	7303	4.91	ng	0.00
7) Phenol-d6	7.21	99	5100	2.38	ng	0.01
23) Nitrobenzene-d5	9.16	82	11562	5.21	ng	-0.01
42) 2,4,6-Tribromophenol	16.14	330	8206	4.94	ng	0.00
45) 2-Fluorobiphenyl	13.25	172	42504	6.73	ng	-0.01
79) Terphenyl-d14	19.98	244	101738	7.42	ng	-0.02

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.50	88	67	0.116	ng	# 1
3) Pyridine	3.84	79	66	0.045	ng	# 1
4) n-Nitrosodimethylamine	3.79	42	53	0.072	ng	# 1
6) Aniline	7.21	93	800	0.324	ng	# 1
9) Benzaldehyde	7.12	77	1184	1.125	ng	# 1
10) Phenol	7.21	94	123	0.063	ng	# 1
11) bis(2-Chloroethyl)ether	7.21	93	800	0.529	ng	# 1
12) 1,3-Dichlorobenzene	8.03	146	57	0.028	ng	# 1
13) 1,4-Dichlorobenzene	8.03	146	57	0.027	ng	# 25
14) 1,2-Dichlorobenzene	8.36	146	60	0.030	ng	# 23
15) Benzyl Alcohol	8.36	79	248	0.165	ng	# 1
16) 2,2'-oxybis(1-Chloropropan	8.34	45	55	0.025	ng	# 1
17) 2-Methylphenol	8.63	107	64	0.045	ng	# 1
18) Hexachloroethane	9.10	117	13357	17.777	ng	# 1
19) n-Nitroso-di-n-propylamine	9.16	70	1195	0.864	ng	# 67
20) 3+4-Methylphenols	8.63	107	64	0.033	ng	# 1
22) Acetophenone	8.81	105	693	0.236	ng	# 52
24) Nitrobenzene	9.18	77	1992	0.941	ng	# 6
25) Isophorone	9.49	82	83	0.020	ng	# 64
26) 2-Nitrophenol	9.42	139	57	0.056	ng	# 31
27) 2,4-Dimethylphenol	9.98	122	63	0.043	ng	# 1
28) bis(2-Chloroethoxy)methane	10.19	93	161	0.072	ng	# 49
31) Naphthalene	10.85	128	174556	30.030	ng	# 99
32) Benzoic acid	10.14	122	62	9.094	ng	# 38
33) 4-Chloroaniline	10.85	127	23438	9.837	ng	# 36
36) 4-Chloro-3-methylphenol	11.70	107	77	0.038	ng	# 22
37) 2-Methylnaphthalene	12.46	142	28468	6.383	ng	# 89
38) 1-Methylnaphthalene	12.46	142	28468	6.709	ng	# 86
40) 1,2,4,5-Tetrachlorobenzene	12.44	216	80	0.025	ng	# 26
46) 1,1'-Biphenyl	13.47	154	457	0.069	ng	# 43
49) Acenaphthylene	14.19	152	74	0.009	ng	# 60
50) Dimethylphthalate	14.10	163	2542	0.376	ng	# 75
51) 2,6-Dinitrotoluene	14.62	165	11808	8.037	ng	# 22
52) Acenaphthene	14.70	154	490	0.102	ng	# 36

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111519\
 Data File : BG043664.D
 Acq On : 15 Nov 2019 23:46
 Operator : JU
 Sample : K5804-01DL 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL

Quant Time: Nov 16 01:14:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 04 15:22:15 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
55) Dibenzofuran	15.06	168	78	0.010	nq	# 44
57) 2,4-Dinitrotoluene	14.62	165	11809	5.624	nq	# 25
58) Fluorene	15.71	166	374	0.057	nq	# 70
66) n-Nitrosodiphenylamine	16.13	169	69	0.011	nq	# 24
67) 4-Bromophenyl-phenylether	16.24	248	53	0.017	nq	# 18
71) Phenanthrene	17.41	178	573	0.050	nq	# 61
72) Anthracene	17.41	178	573	0.050	nq	# 61
74) Di-n-butylphthalate	18.33	149	67	0.005	nq	# 77
75) Fluoranthene	19.80	202	60	0.004	nq	# 65
77) Benzidine	19.98	184	1597	0.309	nq	# 67
78) Pyrene	19.80	202	60	0.004	nq	# 57
80) Butylbenzylphthalate	20.45	149	54	0.009	nq	# 27
81) Benzo(a)anthracene	21.63	228	817	0.051	nq	# 50
83) Chrysene	21.67	228	61	0.004	nq	# 48
84) Bis(2-ethylhexyl)phthalate	21.36	149	102738	11.997	nq	# 51
85) Di-n-octyl phthalate	22.72	149	62	0.004	nq	# 1
88) Benzo(b)fluoranthene	23.81	252	122	0.007	nq	# 60
89) Benzo(k)fluoranthene	23.93	252	58	0.003	nq	# 60
90) Benzo(a)pyrene	24.80	252	1233	0.076	nq	# 78


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111519\
Data File : BG043664.D
Acq On : 15 Nov 2019 23:46
Operator : JU
Sample : K5804-01DL 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL

Quant Time: Nov 16 01:14:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 04 15:22:15 2019
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.99 min Scan# 835
Delta R.T. -0.01 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL




Tot Ion:152 Resp: 27272
Ion Ratio Lower Upper
152 100
150 138.6 129.4 194.0
115 57.8 48.8 73.2






#2

1,4-Dioxane
Concen: 0.116 ng
RT: 3.50 min Scan# 70
Delta R.T. 0.03 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion: 88 Resp: 67
Ion Ratio Lower Upper
88 100
58 0.0 62.7 94.1#
43 655.2 23.7 35.5#





#3

Pvridine

Concen: 0.045 ng

RT: 3.84 min Scan# 128

Delta R.T. -0.04 min





Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Tot Ion: 79 Resp: 66

Ion	Ratio	Lower	Upper
79	100		
52	0.0	56.5	84.7#
51	239.4	33.2	49.8#

Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL



#4

n-Nitrosodimethylamine

Concen: 0.072 ng

RT: 3.79 min Scan# 119


Delta R.T. -0.00 min


Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Tgt Ion: 42 Resp: 53

Ion	Ratio	Lower	Upper
42	100		
74	0.0	97.3	145.9#
44	0.0	4.4	6.6#





#5

2-Fluorophenol

Concen: 4.907 ng

RT: 5.59 min Scan# 425

Delta R.T. -0.01 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:112 Resp: 7303

Ion Ratio Lower Upper

112 100

64 52.2 50.0 75.0




63 38.0 26.6 40.0



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 0.324 ng

RT: 7.21 min Scan# 702

Delta R.T. -0.12 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion: 93 Resp: 800

Ion Ratio Lower Upper

93 100

66 115.4 32.3 48.5#



65 556.3 16.1 24.1#




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 2.382 ng

RT: 7.21 min Scan# 701

Delta R.T. 0.01 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion: 99 Resp: 5100

Ion Ratio Lower Upper




99 100

42 31.7 14.6 21.8#

71 39.8 32.5 48.7



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#9

Benzaldehyde

Concen: 1.125 ng

RT: 7.12 min Scan# 687

Delta R.T. -0.02 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion: 77 Resp: 1184

Ion Ratio Lower Upper



77 100


105 456.2 68.6 108.6#

106 105.4 69.9 109.9



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 106.00 (105.70 to 106.70): BGC





#10

Phenol

Concen: 0.063 ng

RT: 7.21 min Scan# 702

Delta R.T. -0.01 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion: 94 Resp: 123

Ion Ratio Lower Upper




94 100

65 1365.3 12.6 52.6#

66 283.2 30.7 70.7#



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#11

bis(2-Chloroethyl)ether

Concen: 0.529 ng

RT: 7.21 min Scan# 702

Delta R.T. -0.22 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion: 93 Resp: 800

Ion Ratio Lower Upper



93 100


63 567.4 56.3 96.3#

95 0.0 12.2 52.2#



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC









#12
 1,3-Dichlorobenzene
 Concen: 0.028 ng
 RT: 8.03 min Scan# 841
 Delta R.T. 0.13 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46




Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL


Tot Ion	Ratio	Lower	Upper
146	100		
148	0.0	50.3	75.5#
75	121.7	24.6	37.0#



#13
 1,4-Dichlorobenzene
 Concen: 0.027 ng
 RT: 8.03 min Scan# 841
 Delta R.T. -0.01 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Tgt Ion	Ratio	Lower	Upper
146	100		
148	0.0	50.7	76.1#
111	0.0	33.4	50.2#









#14
 1,2-Dichlorobenzene
 Concen: 0.030 ng
 RT: 8.36 min Scan# 897
 Delta R.T. -0.01 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL

Tot Ion:146 Resp: 60
 Ion Ratio Lower Upper
 146 100
 148 0.0 53.2 79.8#
 111 0.0 35.8 53.6#




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#15
 Benzyl Alcohol
 Concen: 0.165 ng
 RT: 8.36 min Scan# 897
 Delta R.T. 0.10 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Tgt Ion: 79 Resp: 248
 Ion Ratio Lower Upper
 79 100
 108 0.0 53.6 80.4#
 77 481.0 49.6 74.4#



Abundance Ion 79.00 (78.70 to 79.70): BGC
 Ion 108.00 (107.70 to 108.70): E
 Ion 77.00 (76.70 to 77.70): BGC









#16
2,2'-oxvbis(1-Chloropropane)
Concen: 0.025 ng
RT: 8.34 min Scan# 894
Delta R.T. -0.20 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46




Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL


Tot Ion: 45 Resp: 55
Ion Ratio Lower Upper
45 100
77 214.6 0.0 34.9#
79 0.0 0.0 33.0



#17
2-Methylphenol
Concen: 0.045 ng
RT: 8.63 min Scan# 943
Delta R.T. 0.16 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion: 107 Resp: 64
Ion Ratio Lower Upper
107 100
108 0.0 90.2 135.4#
77 1565.2 40.2 60.2#
79 833.7 44.0 66.0#





#18

Hexachloroethane

Concen: 17.777 ng

RT: 9.10 min Scan# 1023

Delta R.T. 0.00 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:117 Resp: 13357

Ion Ratio Lower Upper




117 100

119 382.6 79.0 118.4#

201 0.0 105.4 158.2#



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E



#19

n-Nitroso-di-n-propylamine

Concen: 0.864 ng

RT: 9.16 min Scan# 1034

Delta R.T. 0.35 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Tgt Ion: 70 Resp: 1195


Ion Ratio Lower Upper

70 100



42 33.1 41.4 62.2#


101 0.0 7.9 11.9#

130 0.0 19.0 28.4#



Abundance Ion 70.00 (69.70 to 70.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 101.00 (100.70 to 101.70): BGC
Ion 130.00 (129.70 to 130.70): BGC





#20

3+4-Methylphenols

Concen: 0.033 ng

RT: 8.63 min Scan# 943

Delta R.T. -0.17 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL

Tot Ion:107 Resp: 64



Ion Ratio Lower Upper

107 100

108 0.0 66.6 106.6#

77 1565.2 20.9 60.9#

79 833.7 11.3 51.3#





Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.80 min Scan# 1312

Delta R.T. -0.02 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Tgt Ion:136 Resp: 114514



Ion Ratio Lower Upper

136 100

137 10.4 8.7 13.1

54 8.0 5.4 8.2

68 4.1 3.0 4.4





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC









#22
Acetophenone
Concen: 0.236 ng
RT: 8.81 min Scan# 973
Delta R.T. -0.03 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46




Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL


Tot Ion:105 Resp: 693
Ion Ratio Lower Upper
105 100
71 0.0 0.8 1.2#
51 0.0 21.0 31.4#
120 0.0 18.0 27.0#



#23
Nitrobenzene-d5
Concen: 5.205 ng
RT: 9.16 min Scan# 1034
Delta R.T. -0.01 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion: 82 Resp: 11562
Ion Ratio Lower Upper
82 100
128 39.0 35.0 52.6
54 42.9 40.2 60.2





#24

Nitrobenzene

Concen: 0.941 ng

RT: 9.18 min Scan# 1037

Delta R.T. -0.04 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL





Tot Ion: 77 Resp: 1992

Ion Ratio Lower Upper

77 100

123 0.0 35.0 52.6#

65 86.0 9.5 14.3#



#25

Isophorone

Concen: 0.020 ng

RT: 9.49 min Scan# 1090

Delta R.T. -0.24 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46




Tgt Ion: 82 Resp: 83


Ion Ratio Lower Upper

82 100

95 0.0 6.6 9.8#

138 0.0 15.2 22.8#









#26
2-Nitrophenol
Concen: 0.056 ng
RT: 9.42 min Scan# 1077
Delta R.T. -0.51 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL

Tot Ion:139 Resp: 57
Ion Ratio Lower Upper
139 100
109 0.0 26.5 39.7#
65 0.0 43.8 65.8#




Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BGC






#27
2,4-Dimethylphenol
Concen: 0.043 ng
RT: 9.98 min Scan# 1173
Delta R.T. -0.02 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion:122 Resp: 63
Ion Ratio Lower Upper
122 100
107 0.0 104.4 156.6#
121 0.0 47.2 70.8#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E









#28
bis(2-Chloroethoxy)methane
Concen: 0.072 ng
RT: 10.19 min Scan# 1209
Delta R.T. -0.02 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46




Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL


Tot Ion: 93 Resp: 161
Ion Ratio Lower Upper
93 100
95 0.0 26.0 39.0#
123 0.0 10.6 15.8#



#31
Naphthalene
Concen: 30.030 ng
RT: 10.85 min Scan# 1321
Delta R.T. -0.02 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion: 128 Resp: 174556
Ion Ratio Lower Upper
128 100
129 10.9 9.4 14.0
127 13.6 11.1 16.7





#32

Benzoic acid

Concen: 9.094 ng

RT: 10.14 min Scan# 1201

Delta R.T. -0.02 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:122 Resp: 62

Ion Ratio Lower Upper

122 100

105 0.0 103.3 143.3#




77 112.6 84.9 124.9



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BGC



#33

4-Chloroaniline

Concen: 9.837 ng

RT: 10.85 min Scan# 1321

Delta R.T. -0.13 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Tgt Ion:127 Resp: 23438


Ion Ratio Lower Upper

127 100

129 79.7 26.5 39.7#

65 0.0 26.4 39.6#

92 0.0 16.2 24.2#






Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BGC

Ion 92.00 (91.70 to 92.70): BGC






#36
4-Chloro-3-methylphenol
Concen: 0.038 ng
RT: 11.70 min Scan# 1466
Delta R.T. -0.41 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL

Tot Ion	Ratio	Lower	Upper
107	100		
144	0.0	20.9	31.3#
142	0.0	59.6	89.4#






Abundance

Ion 107.00 (106.70 to 107.70): E


Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E



#37
2-Methylnaphthalene
Concen: 6.383 ng
RT: 12.46 min Scan# 1595
Delta R.T. -0.01 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion	Ratio	Lower	Upper
142	100		
141	77.7	72.2	108.2
115	38.0	33.2	49.8






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E









#38
1-Methylnaphthalene
Concen: 6.709 ng
RT: 12.46 min Scan# 1595
Delta R.T. -0.23 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46




Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL


Tot Ion:142 Resp: 28468
Ion Ratio Lower Upper
142 100
141 77.7 73.2 109.8
116 4.3 3.6 5.4



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.62 min Scan# 1963
Delta R.T. -0.02 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion:164 Resp: 87315
Ion Ratio Lower Upper
164 100
162 101.0 83.5 125.3
160 44.0 36.5 54.7









#40
 1,2,4,5-Tetrachlorobenzene
 Concen: 0.025 ng
 RT: 12.44 min Scan# 1591
 Delta R.T. -0.40 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46




Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL


Tot Ion:216	Resp: 80
Ion Ratio	Lower Upper
216 100	
214 0.0	61.7 92.5#
179 0.0	15.8 23.8#
108 0.0	11.5 17.3#



#42
 2,4,6-Tribromophenol
 Concen: 4.939 ng
 RT: 16.14 min Scan# 2221
 Delta R.T. 0.00 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Tgt	Ion:330	Resp:	8206
Ion	Ratio	Lower	Upper
330	100		
332	97.3	77.4	116.0
141	26.4	22.8	34.2









#45
2-Fluorobiphenyl
Concen: 6.726 ng
RT: 13.25 min Scan# 1730
Delta R.T. -0.01 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46




Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL


Tot Ion:172 Resp: 42504
Ion Ratio Lower Upper
172 100
171 34.1 28.9 43.3
170 21.2 18.5 27.7



#46
1,1'-Biphenyl
Concen: 0.069 ng
RT: 13.47 min Scan# 1767
Delta R.T. -0.01 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion:154 Resp: 457
Ion Ratio Lower Upper
154 100
153 0.0 19.9 59.9#
76 0.0 0.0 29.4





#49

Acenaphthylene

Concen: 0.009 ng

RT: 14.19 min Scan# 1890

Delta R.T. -0.17 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:152 Resp: 74

Ion Ratio Lower Upper

152 100

151 0.0 16.2 24.4#




153 0.0 10.6 15.8#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 0.376 ng

RT: 14.10 min Scan# 1874

Delta R.T. 0.00 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion:163 Resp: 2542

Ion Ratio Lower Upper

163 100

194 0.0 3.8 5.6#



164 0.0 8.9 13.3#




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 8.037 ng
 RT: 14.62 min Scan# 1963
 Delta R.T. 0.41 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46




Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL


Tot Ion:165 Resp: 11808
 Ion Ratio Lower Upper
 165 100
 63 0.0 46.6 69.8#
 89 0.0 45.1 67.7#



#52
 Acenaphthene
 Concen: 0.102 ng
 RT: 14.70 min Scan# 1976
 Delta R.T. -0.01 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Tgt Ion:154 Resp: 490
 Ion Ratio Lower Upper
 154 100
 153 46.4 88.2 132.2#
 152 0.0 40.0 60.0#





#55

Dibenzenofuran

Concen: 0.010 ng

RT: 15.06 min Scan# 2037

Delta R.T. 0.02 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:168 Resp: 78

Ion Ratio Lower Upper

168 100

139 0.0 31.0 46.6#




169 0.0 11.7 17.5#



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#57

2,4-Dinitrotoluene

Concen: 5.624 ng

RT: 14.62 min Scan# 1963

Delta R.T. -0.38 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Tgt Ion:165 Resp: 11809


Ion Ratio Lower Upper

165 100

63 0.0 33.2 49.8#

89 0.0 52.2 78.4#

182 0.0 2.0 3.0#






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 89.00 (88.70 to 89.70): BGC

Ion 182.00 (181.70 to 182.70): E









#58
 Fluorene
 Concen: 0.057 ng
 RT: 15.71 min Scan# 2148
 Delta R.T. 0.02 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46




Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL


Tot Ion:166 Resp: 374
 Ion Ratio Lower Upper
 166 100
 165 69.5 79.0 118.6#
 167 0.0 10.5 15.7#



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.37 min Scan# 2431
 Delta R.T. -0.02 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Tgt Ion:188 Resp: 225700
 Ion Ratio Lower Upper
 188 100
 94 5.8 4.2 6.4
 80 6.0 4.5 6.7









#66
 n-Nitrosodiphenylamine
 Concen: 0.011 ng
 RT: 16.13 min Scan# 2220
 Delta R.T. 0.24 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Instrument :
 BNA_G
 ClientSampleId :
 GE-MW-58-1SDL

Tot Ion	Ratio	Lower	Upper
169	100		
168	0.0	53.8	80.6#
167	0.0	27.5	41.3#




Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E






#67
 4-Bromophenyl-phenylether
 Concen: 0.017 ng
 RT: 16.24 min Scan# 2239
 Delta R.T. -0.33 min
 Lab File: BG043664.D
 Acq: 15 Nov 2019 23:46

Tgt Ion	Ratio	Lower	Upper
248	100		
250	0.0	77.0	115.4#
141	102.0	47.5	71.3#



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E









#71
Phenanthrene
Concen: 0.050 ng
RT: 17.41 min Scan# 2438
Delta R.T. -0.02 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46




Instrument :
BNA_G
ClientSampled :
GE-MW-58-1SDL


Tot Ion:178 Resp: 573
Ion Ratio Lower Upper
178 100
176 37.2 15.4 23.0#
179 0.0 13.0 19.6#



#72
Anthracene
Concen: 0.050 ng
RT: 17.41 min Scan# 2438
Delta R.T. -0.11 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion:178 Resp: 573
Ion Ratio Lower Upper
178 100
176 37.2 15.5 23.3#
179 0.0 13.7 20.5#









#74
Di-n-butylphthalate
Concen: 0.005 ng
RT: 18.33 min Scan# 2594
Delta R.T. -0.01 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46




Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL


Tot Ion:149 Resp: 67
Ion Ratio Lower Upper
149 100
150 0.0 8.0 12.0#
104 0.0 4.4 6.6#



#75
Fluoranthene
Concen: 0.004 ng
RT: 19.80 min Scan# 2845
Delta R.T. 0.36 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion:202 Resp: 60
Ion Ratio Lower Upper
202 100
101 0.0 0.0 26.3
203 0.0 0.0 37.9





#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.64 min Scan# 3157

Delta R.T. -0.02 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:240 Resp: 257080

Ion Ratio Lower Upper

240 100

120 6.2 4.6 6.8




236 25.7 19.8 29.8



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Benzidine

Concen: 0.309 ng

RT: 19.98 min Scan# 2875

Delta R.T. 0.36 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion:184 Resp: 1597

Ion Ratio Lower Upper

184 100

185 0.0 11.0 16.4#



183 0.0 10.0 15.0#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#78

Pvrene

Concen: 0.004 ng

RT: 19.80 min Scan# 2845

Delta R.T. 0.00 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:202 Resp: 60

Ion Ratio Lower Upper


202 100

200 0.0 17.0 25.4#



203 0.0 14.7 22.1#



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E



Time-->



#79

Terphenyl-d14

Concen: 7.424 ng

RT: 19.98 min Scan# 2875

Delta R.T. -0.02 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion:244 Resp: 101738

Ion Ratio Lower Upper


244 100

212 7.1 5.8 8.6


122 6.5 5.4 8.2




Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E



Time-->









#80
Butylbenzylphthalate
Concen: 0.009 ng
RT: 20.45 min Scan# 2955
Delta R.T. -0.24 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Instrument :
BNA_G
ClientSampleId :
GE-MW-58-1SDL

Tot Ion	Ratio	Lower	Upper
149	100		
91	0.0	51.2	76.8#
206	0.0	22.5	33.7#




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 91.00 (90.70 to 91.70): BG
Ion 206.00 (205.70 to 206.70): E






#81
Benzo(a)anthracene
Concen: 0.051 ng
RT: 21.63 min Scan# 3156
Delta R.T. -0.01 min
Lab File: BG043664.D
Acq: 15 Nov 2019 23:46

Tgt Ion	Ratio	Lower	Upper
228	100		
226	0.0	22.2	33.2#
229	0.0	17.0	25.4#



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E





#83

Chrysene

Concen: 0.004 ng

RT: 21.67 min Scan# 3163

Delta R.T. -0.03 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :


BNA_G

ClientSampleId :

GE-MW-58-1SDL

Tot Ion:228 Resp: 61

Ion	Ratio	Lower	Upper
228	100		
226	0.0	25.1	37.7#
229	0.0	17.0	25.4#






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E

21.67



#84

Bis(2-ethylhexyl)phthalate

Concen: 11.997 ng

RT: 21.36 min Scan# 3109


Delta R.T. -0.18 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Tgt Ion:149 Resp: 102738

Ion	Ratio	Lower	Upper
149	100		
167	0.0	24.0	36.0#
279	0.0	5.4	8.2#





Abundance


Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E

21.36





#85

Di-n-octyl phthalate

Concen: 0.004 ng

RT: 22.72 min Scan# 3342

Delta R.T. -0.01 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:149 Resp: 62

Ion Ratio Lower Upper

149 100

167 0.0 1.4 2.2#

43 754.8 7.0 10.6#






Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BG0

Time--> 22.71 22.72 22.73 22.74



#87

Perylene-d12

Concen: 20.000 ng

RT: 24.82 min Scan# 3698

Delta R.T. -0.04 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion:264 Resp: 298613

Ion Ratio Lower Upper

264 100

260 22.8 17.5 26.3

265 21.9 17.4 26.0






Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

Time--> 24.70 24.80 24.90





#88

Benzo(b)fluoranthene

Concen: 0.007 ng

RT: 23.81 min Scan# 3526

Delta R.T. -0.04 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL


Tot Ion:252 Resp: 122

Ion Ratio Lower Upper

252 100

253 0.0 17.7 26.5#




125 0.0 4.5 6.7#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Benzo(k)fluoranthene

Concen: 0.003 ng

RT: 23.93 min Scan# 3547

Delta R.T. 0.02 min

Lab File: BG043664.D

Acq: 15 Nov 2019 23:46


Tgt Ion:252 Resp: 58

Ion Ratio Lower Upper

252 100

253 0.0 17.3 25.9#



125 0.0 4.6 6.8#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#90

Benzo(a)pyrene

Concen: 0.076 ng

RT: 24.80 min Scan# 3696

Delta R.T. 0.10 min

Lab File: BG043664.D


Acq: 15 Nov 2019 23:46

Instrument :

BNA_G

ClientSampleId :

GE-MW-58-1SDL



Tot Ion:	252	Resp:	1233
Ion	Ratio	Lower	Upper
252	100		
253	33.0	17.8	26.8#
125	0.0	5.4	8.2#

