


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG111519\
 Data File : BG043653.D
 Acq On : 15 Nov 2019 16:31
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG102119.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 04 15:22:15 2019



AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.2	8134	PASS
68	69	0.00	2	1.0	90	PASS
69	198	0.00	100	32.3	8703	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.9	11294	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	26946	PASS
199	198	5	9	7.1	1904	PASS
275	198	10	60	31.5	8499	PASS
365	198	1	100	4.7	1263	PASS
441	198	0.01	100	12.4	3351	PASS
442	198	50	100	88.3	23796	PASS
443	442	15	24	17.3	4115	PASS