

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111523\
 Data File : BG059709.D
 Acq On : 15 Nov 2023 21:07
 Operator : MA/JU
 Sample : 05102-01RE
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 67S-IDW-01RE

Quant Time: Nov 16 02:52:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 15 08:14:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.347	152	22264	20.000	ng	# 0.00
21) Naphthalene-d8	11.191	136	100093	20.000	ng	# 0.00
39) Acenaphthene-d10	14.975	164	68170	20.000	ng	0.00
64) Phenanthrene-d10	17.725	188	158383	20.000	ng	0.00
76) Chrysene-d12	22.055	240	156701	20.000	ng	0.00
86) Perylene-d12	25.645	264	184139	20.000	ng	# 0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.838	112	205961	159.708	ng	0.00
7) Phenol-d6	7.466	99	267823	137.423	ng	0.00
23) Nitrobenzene-d5	9.546	82	269888	115.915	ng	0.00
42) 2,4,6-Tribromophenol	16.461	330	131588	171.041	ng	0.00
45) 2-Fluorobiphenyl	13.600	172	529798	104.779	ng	0.00
79) Terphenyl-d14	20.304	244	1174270	154.611	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	11.238	128	23967	4.743	ng	# 94
37) 2-Methylnaphthalene	12.824	142	34872	8.262	ng	# 79
38) 1-Methylnaphthalene	13.042	142	32152	8.293	ng	# 84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111523\
Data File : BG059709.D
Acq On : 15 Nov 2023 21:07
Operator : MA/JU
Sample : 05102-01RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
67S-IDW-01RE

Quant Time: Nov 16 02:52:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 15 08:14:53 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.347 min Scan# 845
Delta R.T. -0.007 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07




Instrument :
BNA_G
ClientSampleId :
67S-IDW-01RE


Tgt Ion:152 Resp: 22264
Ion Ratio Lower Upper
152 100
150 154.7 113.9 170.9
115 100.0 59.4 89.2#



#5
2-Fluorophenol
Concen: 159.708 ng
RT: 5.838 min Scan# 417
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07

Tgt Ion:112 Resp: 205961
Ion Ratio Lower Upper
112 100
64 53.9 39.0 58.6
63 47.6 29.8 44.6#









#7
Phenol-d6
Concen: 137.423 ng
RT: 7.466 min Scan# 694
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07




Instrument :
BNA_G
ClientSampleId :
67S-IDW-01RE


Tgt Ion: 99 Resp: 267823
Ion Ratio Lower Upper
99 100
42 22.6 13.2 19.8#
71 75.2 51.3 76.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.191 min Scan# 1328
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07

Tgt Ion:136 Resp: 100093
Ion Ratio Lower Upper
136 100
137 9.4 7.7 11.5
54 6.5 4.2 6.2#
68 1.9 1.4 2.0









#23
Nitrobenzene-d5
Concen: 115.915 ng
RT: 9.546 min Scan# 1048
Delta R.T. -0.007 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07




Instrument :
BNA_G
ClientSampleId :
67S-IDW-01RE


Tgt Ion: 82 Resp: 269888
Ion Ratio Lower Upper
82 100
128 27.6 26.3 39.5
54 42.3 26.6 39.8#



#31
Naphthalene
Concen: 4.743 ng
RT: 11.238 min Scan# 1336
Delta R.T. -0.013 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07

Tgt Ion:128 Resp: 23967
Ion Ratio Lower Upper
128 100
129 13.3 8.6 12.8#
127 10.1 9.8 14.6









#37
2-Methylnaphthalene
Concen: 8.262 ng
RT: 12.824 min Scan# 1606
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07




Instrument :
BNA_G
ClientSampleId :
67S-IDW-01RE


Tgt Ion:142 Resp: 34872
Ion Ratio Lower Upper
142 100
141 62.4 70.9 106.3#
115 51.1 47.9 71.9



#38
1-Methylnaphthalene
Concen: 8.293 ng
RT: 13.042 min Scan# 1643
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07

Tgt Ion:142 Resp: 32152
Ion Ratio Lower Upper
142 100
141 70.2 68.3 102.5
116 7.8 4.4 6.6#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.975 min Scan# 1972
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07




Instrument :
BNA_G
ClientSampleId :
67S-IDW-01RE


Tgt Ion:164 Resp: 68170
Ion Ratio Lower Upper
164 100
162 101.5 77.8 116.6
160 47.0 38.0 57.0



#42
2,4,6-Tribromophenol
Concen: 171.041 ng
RT: 16.461 min Scan# 2225
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07

Tgt Ion:330 Resp: 131588
Ion Ratio Lower Upper
330 100
332 92.5 80.7 121.1
141 66.6 51.0 76.6









#45
2-Fluorobiphenyl
Concen: 104.779 ng
RT: 13.600 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07




Instrument :
BNA_G
ClientSampleId :
67S-IDW-01RE


Tgt Ion:172 Resp: 529798
Ion Ratio Lower Upper
172 100
171 33.1 27.9 41.9
170 20.9 19.0 28.4



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.725 min Scan# 2440
Delta R.T. -0.001 min
Lab File: BG059709.D
Acq: 15 Nov 2023 21:07

Tgt Ion:188 Resp: 158383
Ion Ratio Lower Upper
188 100
94 4.0 3.9 5.9
80 6.6 4.5 6.7





#76

Chrysene-d12

Concen: 20.000 ng

RT: 22.055 min Scan# 3177

Delta R.T. -0.007 min

Lab File: BG059709.D

Acq: 15 Nov 2023 21:07

Instrument :

BNA_G

ClientSampleId :

67S-IDW-01RE





Tgt Ion:240 Resp: 156701

Ion Ratio Lower Upper

240 100

120 3.1 2.6 3.8

236 28.2 19.4 29.0



#79

Terphenyl-d14

Concen: 154.611 ng

RT: 20.304 min Scan# 2879

Delta R.T. -0.001 min

Lab File: BG059709.D

Acq: 15 Nov 2023 21:07




Tgt Ion:244 Resp: 1174270


Ion Ratio Lower Upper

244 100

212 9.3 6.6 9.8

122 4.8 4.6 6.8





#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.645 min Scan# 31
 Delta R.T. -0.001 min
 Lab File: BG059709.D
 Acq: 15 Nov 2023 21:07

Instrument :
 BNA_G
 ClientSampleId :
 67S-IDW-01RE

Tgt Ion:264 Resp: 184139

Ion	Ratio	Lower	Upper
264	100		
260	21.9	15.6	23.4
265	20.6	13.4	20.2

