


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG111620\
Data File : BG047324.D
Acq On : 16 Nov 2020 10:40
Operator : CG/JU
Sample : DFTPP53
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP53

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG102220MA.M
Title : SVOA CALIBRATION
Last Update : Mon Nov 16 12:10:55 2020



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.2	37746	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	35.2	40026	PASS
70	69	0.00	2	0.6	254	PASS
127	198	10	80	40.8	46429	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	113832	PASS
199	198	5	9	7.2	8145	PASS
275	198	10	60	31.1	35421	PASS
365	198	1	100	3.7	4252	PASS
441	198	0.01	100	12.7	14420	PASS
442	198	50	100	80.0	91120	PASS
443	442	15	24	20.2	18404	PASS

Scan 2458 (17.782 min): BG047324.D
DFTPP53

m/z	Abundance
40.00	321.0
44.10	1112.0
57.10	237.0
74.90	205.0
207.00	567.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

m/z	Abundance
38.10	569.0
39.10	3291.0
40.10	281.0
44.00	1408.0
49.10	316.0
50.10	10111.0
51.10	41376.0
52.20	1751.0
52.80	226.0
55.10	473.0
56.10	1193.0
57.00	3340.0
58.00	298.0
61.00	458.0
62.20	460.0
63.00	1846.0
64.10	434.0
65.10	691.0
67.10	205.0
69.00	41640.0
70.10	231.0
71.00	164.0
73.10	420.0
74.10	4116.0
75.00	6987.0
76.10	2534.0
77.10	40992.0
78.10	2995.0
79.00	3116.0
80.00	2760.0
81.00	3527.0
82.00	829.0
83.00	943.0
85.00	586.0
86.00	1254.0
87.00	383.0
91.00	784.0
92.00	1117.0
93.00	5514.0
94.10	504.0
95.00	239.0
96.00	323.0
98.00	4759.0
99.00	3207.0
100.10	339.0
101.00	2101.0
102.90	476.0
104.00	1546.0
105.00	1855.0
107.00	14819.0
108.10	2042.0
110.00	22312.0
111.00	3893.0
112.10	656.0
113.00	418.0
116.10	1279.0
117.00	12056.0
118.00	856.0
119.90	253.0
122.00	965.0
123.00	1675.0
124.10	842.0
125.00	888.0
127.00	45160.0
128.10	3303.0
129.00	19224.0
130.00	1539.0
130.80	201.0
132.00	452.0
134.00	1028.0
134.90	1182.0
136.00	528.0
137.00	753.0
138.90	246.0
139.60	150.0
140.10	248.0
141.00	3249.0
142.10	576.0
143.00	813.0
144.00	255.0
146.10	334.0
146.90	1678.0
148.00	3000.0
149.00	739.0
151.00	333.0
152.10	327.0
153.00	663.0
153.80	380.0
155.00	1464.0
156.00	2281.0
157.10	571.0
158.00	524.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

159.00	739.0
160.00	535.0
161.00	1223.0
161.80	368.0
164.10	252.0
165.00	1807.0
166.00	640.0
167.00	4228.0
168.10	2500.0
169.10	435.0
170.00	267.0
170.80	248.0
172.00	651.0
173.00	776.0
174.00	1190.0
175.00	2963.0
176.00	486.0
177.00	1184.0
178.00	325.0
179.00	4990.0
180.00	3251.0
181.10	954.0
182.20	229.0
182.80	243.0
183.90	284.0
185.00	1881.0
186.00	14729.0
187.00	5359.0
187.90	487.0
189.00	1156.0
190.00	314.0
191.20	388.0
192.10	1369.0
193.00	1692.0
194.00	533.0
196.00	2972.0
198.00	106344.0
199.00	7720.0
200.00	640.0
200.90	430.0
201.40	615.0
203.00	945.0
204.00	3941.0
205.10	7481.0
206.00	29344.0
207.10	4629.0
208.00	1125.0
209.00	390.0
210.00	417.0
211.20	1055.0
213.20	175.0
215.00	718.0
216.10	884.0
217.00	8754.0
217.80	763.0
219.10	197.0
219.70	193.0
221.00	4122.0
221.80	1252.0
223.00	1872.0
224.00	16736.0
225.00	4459.0
227.00	7953.0
228.00	1412.0
229.00	1633.0
230.90	605.0
232.00	188.0
234.00	697.0
235.00	621.0
236.00	155.0
237.00	623.0
239.10	322.0
240.00	203.0
241.00	337.0
242.00	604.0
243.10	934.0
244.00	11593.0
245.10	2009.0
246.00	2430.0
247.00	268.0
248.10	158.0
249.00	567.0
251.10	195.0
251.30	179.0
252.70	233.0
253.10	261.0
255.00	67600.0
256.00	10175.0
256.90	676.0
258.00	3656.0
258.90	514.0
260.00	181.0
261.00	150.0
265.00	1684.0
266.00	265.0
270.00	223.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

272.20	216.0
273.00	1602.0
274.00	5387.0
275.00	28312.0
276.00	3346.0
277.00	2475.0
278.00	296.0
281.10	217.0
283.00	262.0
284.90	512.0
285.90	172.0
293.00	474.0
294.00	159.0
295.00	256.0
296.00	7923.0
297.00	1033.0
300.80	228.0
303.10	1139.0
303.80	377.0
305.50	201.0
312.90	164.0
313.70	299.0
313.90	286.0
315.00	950.0
316.00	590.0
317.00	156.0
321.00	266.0
323.00	2338.0
324.00	537.0
326.00	188.0
327.10	661.0
328.10	431.0
333.10	403.0
334.00	1366.0
335.10	531.0
339.90	212.0
346.10	343.0
351.40	188.0
352.10	576.0
353.00	786.0
354.10	870.0
359.00	156.0
365.00	3429.0
366.00	721.0
371.20	199.0
372.10	932.0
382.90	354.0
384.00	186.0
392.10	203.0
402.10	456.0
403.00	916.0
404.00	189.0
421.00	452.0
422.00	564.0
423.00	2884.0
424.10	620.0
441.10	9774.0
442.10	63736.0
443.00	13885.0
444.00	1497.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

m/z	Abundance
37.20	155.0
38.10	477.0
39.10	3104.0
40.10	333.0
44.00	912.0
49.00	586.0
50.10	10872.0
51.10	43496.0
52.10	2238.0
55.00	447.0
56.00	1245.0
57.10	2469.0
58.00	167.0
61.00	615.0
62.10	539.0
63.10	2162.0
64.10	389.0
65.10	783.0
67.10	286.0
69.00	48496.0
70.20	244.0
73.20	697.0
74.10	4528.0
75.00	7924.0
76.10	2262.0
77.10	50696.0
78.10	2932.0
79.00	4088.0
80.00	2420.0
81.10	3671.0
82.00	911.0
83.00	593.0
84.10	265.0
85.10	674.0
86.00	1120.0
87.00	615.0
88.10	240.0
89.00	167.0
91.00	977.0
92.00	1306.0
93.00	6131.0
93.80	278.0
96.00	263.0
98.00	5534.0
99.00	4077.0
100.00	245.0
101.00	2148.0
101.70	181.0
103.00	869.0
104.00	1419.0
105.10	1318.0
107.00	17232.0
108.10	2660.0
110.00	26840.0
111.10	3532.0
112.00	792.0
115.00	220.0
116.00	977.0
117.00	14747.0
118.00	931.0
119.10	215.0
120.00	179.0
122.00	1657.0
122.90	1377.0
124.00	746.0
125.00	782.0
127.00	55968.0
128.10	4500.0
129.00	26152.0
130.00	1711.0
131.10	371.0
132.10	243.0
134.00	1017.0
135.00	2384.0
136.00	758.0
137.00	907.0
140.20	282.0
141.00	3456.0
142.00	845.0
143.10	902.0
143.80	206.0
145.20	263.0
145.90	711.0
147.00	1995.0
148.00	2905.0
148.90	727.0
149.90	151.0
151.30	518.0
151.90	228.0
153.00	832.0
154.10	782.0
155.00	1726.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

156.00	2477.0
156.80	298.0
157.90	1140.0
159.10	518.0
159.90	1083.0
161.00	1637.0
162.10	338.0
165.00	1320.0
166.00	1395.0
167.00	6167.0
168.00	3169.0
168.90	347.0
170.00	321.0
171.00	352.0
171.90	261.0
173.10	980.0
174.10	1115.0
175.10	2908.0
176.10	819.0
177.00	1856.0
178.00	574.0
179.00	6124.0
180.00	3579.0
181.00	1463.0
182.00	336.0
182.80	192.0
184.30	382.0
185.10	3086.0
186.10	20312.0
187.00	6148.0
188.00	508.0
189.10	1483.0
189.90	295.0
190.90	780.0
192.00	2019.0
193.00	1789.0
194.10	491.0
195.00	303.0
196.00	3857.0
198.00	135744.0
199.00	9484.0
200.00	963.0
201.60	418.0
203.00	1208.0
204.00	5303.0
205.00	9998.0
206.00	41736.0
207.10	6429.0
208.10	1552.0
209.00	712.0
210.10	911.0
211.10	1764.0
215.00	634.0
215.90	959.0
217.00	11795.0
218.00	1180.0
219.00	249.0
221.00	4999.0
221.90	1384.0
223.00	2834.0
224.00	25352.0
225.00	5947.0
227.00	11129.0
228.10	1547.0
229.00	1978.0
230.00	329.0
231.00	721.0
233.10	289.0
233.90	643.0
235.00	535.0
235.90	575.0
236.90	512.0
237.90	340.0
239.00	563.0
239.80	229.0
240.20	216.0
240.90	247.0
242.00	1427.0
243.00	1090.0
244.00	17400.0
245.10	2548.0
246.00	4080.0
246.90	681.0
248.00	210.0
248.90	746.0
251.90	231.0
253.10	369.0
255.00	90568.0
256.00	13772.0
257.10	999.0
258.00	6077.0
258.90	1014.0
260.00	347.0
261.00	194.0
265.00	2359.0
266.10	242.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

267.90	203.0
269.90	261.0
270.80	365.0
271.90	368.0
273.00	2779.0
274.00	7184.0
275.00	43080.0
276.00	5816.0
277.00	3000.0
278.00	465.0
279.00	150.0
281.00	242.0
283.00	597.0
284.00	296.0
285.00	490.0
289.10	215.0
292.30	236.0
292.90	830.0
294.00	367.0
295.10	560.0
296.00	12744.0
297.00	1650.0
301.90	222.0
303.00	1863.0
303.90	257.0
309.00	192.0
309.90	195.0
314.10	523.0
314.90	1270.0
316.10	506.0
317.00	242.0
321.00	319.0
322.10	235.0
323.00	2973.0
324.00	596.0
325.90	189.0
327.00	729.0
328.00	418.0
329.00	234.0
332.00	312.0
333.00	210.0
334.00	2179.0
335.00	638.0
336.00	167.0
341.10	382.0
345.90	841.0
352.00	839.0
353.10	811.0
354.10	961.0
355.10	210.0
357.80	224.0
365.00	4626.0
366.00	589.0
370.90	324.0
372.10	1403.0
373.10	374.0
383.10	698.0
384.00	216.0
385.00	173.0
401.10	261.0
402.00	598.0
403.00	897.0
403.90	205.0
421.00	1317.0
422.00	1198.0
423.10	5825.0
424.00	1431.0
441.00	17072.0
442.10	106040.0
443.10	21488.0
444.10	1725.0
445.10	263.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

m/z	Abundance
38.00	343.0
39.10	1827.0
40.00	207.0
44.10	531.0
49.10	274.0
50.10	7413.0
51.10	28368.0
52.10	1574.0
55.10	382.0
56.00	694.0
57.00	1726.0
61.00	385.0
62.10	531.0
63.10	982.0
64.30	196.0
65.10	631.0
66.80	222.0
69.00	29944.0
70.00	287.0
74.10	3016.0
75.10	5063.0
76.10	1627.0
77.10	31720.0
78.00	1941.0
79.00	2235.0
80.10	1940.0
81.10	2328.0
82.00	716.0
83.00	789.0
84.00	166.0
85.00	582.0
86.00	634.0
87.10	369.0
88.10	170.0
90.90	376.0
92.00	831.0
93.00	5209.0
94.10	358.0
95.90	182.0
98.00	3020.0
99.00	2156.0
100.00	384.0
101.00	1475.0
102.90	386.0
104.00	1313.0
105.00	889.0
106.00	413.0
107.10	11309.0
108.00	1598.0
109.00	410.0
110.00	18424.0
111.00	2510.0
112.10	352.0
116.00	681.0
117.10	10032.0
118.00	438.0
118.90	313.0
120.00	263.0
122.00	718.0
123.10	1156.0
124.10	710.0
125.10	498.0
127.00	38160.0
128.00	2787.0
129.00	15895.0
130.00	2007.0
131.00	225.0
132.00	258.0
133.90	640.0
135.00	1591.0
136.10	449.0
137.00	868.0
138.10	284.0
139.00	255.0
140.20	406.0
141.00	2668.0
141.90	543.0
142.90	514.0
143.80	247.0
145.10	221.0
146.00	506.0
147.00	1015.0
148.00	2698.0
149.00	526.0
151.00	470.0
152.00	184.0
152.90	787.0
154.00	468.0
155.00	1133.0
156.10	2116.0
157.10	366.0
158.10	473.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

159.00	450.0
159.90	739.0
161.00	1106.0
161.90	351.0
165.00	1204.0
166.00	602.0
167.00	4627.0
168.00	1762.0
169.10	524.0
171.00	276.0
172.10	590.0
172.90	661.0
174.00	1069.0
175.00	2023.0
176.10	609.0
177.00	1091.0
179.00	4087.0
180.00	3085.0
181.00	1203.0
183.30	159.0
184.40	357.0
185.00	1939.0
186.00	13574.0
187.00	4248.0
188.00	511.0
189.10	1013.0
190.00	248.0
190.90	433.0
192.00	1068.0
193.00	1422.0
194.00	250.0
195.00	244.0
196.00	3222.0
198.00	99408.0
199.00	7231.0
199.90	642.0
201.40	644.0
202.30	213.0
203.00	696.0
204.00	4191.0
205.00	7432.0
206.00	29280.0
207.00	4420.0
208.00	1385.0
209.00	562.0
210.00	689.0
211.10	1254.0
211.60	466.0
212.70	179.0
215.00	585.0
216.10	477.0
217.00	8445.0
217.90	917.0
221.00	4562.0
222.10	1240.0
223.00	2238.0
224.00	17192.0
225.00	4667.0
226.00	645.0
227.00	7833.0
228.00	1328.0
229.00	1392.0
229.90	285.0
231.00	464.0
232.80	157.0
233.80	442.0
234.10	364.0
234.80	726.0
236.00	375.0
237.00	625.0
239.00	287.0
240.00	615.0
240.90	482.0
242.00	827.0
243.00	1002.0
244.00	12355.0
245.00	1635.0
246.00	3258.0
247.00	601.0
249.00	589.0
251.00	195.0
252.80	590.0
255.00	75552.0
256.00	10659.0
256.90	489.0
258.00	3549.0
259.00	705.0
264.20	362.0
265.00	1788.0
270.00	171.0
272.00	199.0
273.00	2276.0
274.00	5076.0
275.00	34872.0
276.00	4457.0
277.00	3668.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

277.90	619.0
282.00	220.0
283.00	304.0
284.00	208.0
285.00	501.0
289.20	178.0
293.00	674.0
293.90	199.0
295.00	207.0
296.00	10036.0
297.10	1024.0
302.00	164.0
303.10	1023.0
308.00	152.0
308.90	161.0
313.20	160.0
314.00	357.0
315.00	910.0
315.90	681.0
320.90	304.0
321.90	273.0
323.10	2839.0
324.00	669.0
327.00	532.0
328.10	542.0
332.00	301.0
333.10	555.0
334.00	2298.0
335.00	723.0
335.90	157.0
341.10	519.0
342.10	278.0
345.80	713.0
352.10	872.0
352.90	525.0
354.10	1041.0
354.90	294.0
365.00	4702.0
366.10	515.0
371.20	282.0
372.00	1873.0
372.90	565.0
383.00	585.0
392.10	271.0
402.00	804.0
403.00	679.0
420.90	654.0
422.00	852.0
423.00	4934.0
424.00	1571.0
425.10	164.0
441.00	16416.0
442.10	103584.0
443.00	19840.0
444.00	1734.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53