

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111623\
 Data File : BG059720.D
 Acq On : 16 Nov 2023 16:07
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG111623

Quant Time: Nov 17 01:53:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	28.413	29.0#	99	0.00
3	Pyridine	40.000	33.393	16.5	96	0.00
4	n-Nitrosodimethylamine	40.000	31.635	20.9	94	0.00
5 S	2-Fluorophenol	80.000	71.704	10.4	97	0.00
6	Aniline	40.000	35.740	10.6	95	0.00
7 S	Phenol-d6	80.000	72.492	9.4	95	0.00
8	2-Chlorophenol	40.000	34.728	13.2	86	0.00
9	Benzaldehyde	40.000	37.303	6.7	103	0.00
10 C	Phenol	40.000	35.758	10.6	93	0.00
11	bis(2-Chloroethyl)ether	40.000	34.928	12.7	90	0.00
12	1,3-Dichlorobenzene	40.000	35.810	10.5	95	0.00
13 C	1,4-Dichlorobenzene	40.000	36.530	8.7	98	0.00
14	1,2-Dichlorobenzene	40.000	36.889	7.8	101	0.00
15	Benzyl Alcohol	40.000	36.543	8.6	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.088	2.3	114	-0.01
17	2-Methylphenol	40.000	39.467	1.3	103	0.00
18	Hexachloroethane	40.000	37.263	6.8	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.010	2.5	99	0.00
20	3+4-Methylphenols	40.000	38.947	2.6	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22	Acetophenone	40.000	45.765	-14.4	100	0.00
23 S	Nitrobenzene-d5	80.000	93.496	-16.9	99	0.00
24	Nitrobenzene	40.000	47.202	-18.0	100	0.00
25	Isophorone	40.000	45.617	-14.0	99	0.00
26 C	2-Nitrophenol	40.000	45.864	-14.7	104	0.00
27	2,4-Dimethylphenol	40.000	41.633	-4.1	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	44.985	-12.5	100	0.00
29 C	2,4-Dichlorophenol	40.000	42.941	-7.4	89	0.00
30	1,2,4-Trichlorobenzene	40.000	43.397	-8.5	93	0.00
31	Naphthalene	40.000	42.795	-7.0	90	0.00
32	Benzoic acid	40.000	42.382	-6.0	96	-0.01
33	4-Chloroaniline	40.000	43.597	-9.0	93	0.00
34 C	Hexachlorobutadiene	40.000	43.773	-9.4	98	0.00
35	Caprolactam	40.000	41.097	-2.7	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	45.251	-13.1	99	0.00
37	2-Methylnaphthalene	40.000	40.881	-2.2	86	0.00
38	1-Methylnaphthalene	40.000	42.547	-6.4	91	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.623	-1.6	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.486	-11.2	98	0.00
42 S	2,4,6-Tribromophenol	80.000	86.433	-8.0	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.101	-2.8	90	0.00
44	2,4,5-Trichlorophenol	40.000	40.353	-0.9	90	0.00
45 S	2-Fluorobiphenyl	80.000	79.934	0.1	88	0.00
46	1,1'-Biphenyl	40.000	41.691	-4.2	93	0.00
47	2-Chloronaphthalene	40.000	41.432	-3.6	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111623\
 Data File : BG059720.D
 Acq On : 16 Nov 2023 16:07
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG111623

Quant Time: Nov 17 01:53:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.328	-13.3	99	0.00
49	Acenaphthylene	40.000	42.249	-5.6	96	0.00
50	Dimethylphthalate	40.000	41.712	-4.3	95	0.00
51	2,6-Dinitrotoluene	40.000	43.953	-9.9	102	0.00
52 C	Acenaphthene	40.000	42.410	-6.0	94	0.00
53	3-Nitroaniline	40.000	40.955	-2.4	91	0.00
54 P	2,4-Dinitrophenol	40.000	38.539	3.7	90	0.00
55	Dibenzofuran	40.000	42.209	-5.5	96	0.00
56 P	4-Nitrophenol	40.000	40.326	-0.8	95	0.00
57	2,4-Dinitrotoluene	40.000	43.201	-8.0	100	0.00
58	Fluorene	40.000	42.607	-6.5	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.454	-1.1	95	0.00
60	Diethylphthalate	40.000	43.794	-9.5	101	0.00
61	4-Chlorophenyl-phenylether	40.000	43.575	-8.9	98	0.00
62	4-Nitroaniline	40.000	41.506	-3.8	102	0.00
63	Azobenzene	40.000	44.658	-11.6	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.569	-11.4	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	44.615	-11.5	94	0.00
67	4-Bromophenyl-phenylether	40.000	43.697	-9.2	96	0.00
68	Hexachlorobenzene	40.000	42.685	-6.7	92	0.00
69	Atrazine	40.000	40.310	-0.8	94	0.00
70 C	Pentachlorophenol	40.000	44.228	-10.6	94	0.00
71	Phenanthrene	40.000	44.037	-10.1	95	0.00
72	Anthracene	40.000	43.185	-8.0	92	0.00
73	Carbazole	40.000	42.714	-6.8	92	0.00
74	Di-n-butylphthalate	40.000	44.774	-11.9	97	0.00
75 C	Fluoranthene	40.000	42.153	-5.4	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	37.802	5.5	89	0.00
78	Pyrene	40.000	41.800	-4.5	94	0.00
79 S	Terphenyl-d14	80.000	84.441	-5.6	96	0.00
80	Butylbenzylphthalate	40.000	43.362	-8.4	99	0.00
81	Benzo(a)anthracene	40.000	41.295	-3.2	94	0.00
82	3,3'-Dichlorobenzidine	40.000	41.862	-4.7	95	0.00
83	Chrysene	40.000	40.984	-2.5	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.002	-7.5	97	0.00
85 c	Di-n-octyl phthalate	40.000	38.366	4.1	75	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.667	-1.7	101	0.00
88	Benzo(b)fluoranthene	40.000	39.169	2.1	84	0.00
89	Benzo(k)fluoranthene	40.000	38.950	2.6	85	0.00
90 C	Benzo(a)pyrene	40.000	41.286	-3.2	100	0.00
91	Dibenzo(a,h)anthracene	40.000	39.698	0.8	100	0.00
92	Benzo(g,h,i)perylene	40.000	39.084	2.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111623\
Data File : BG059720.D
Acq On : 16 Nov 2023 16:07
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG111623

Quant Time: Nov 17 01:53:15 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 01:51:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0