

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
 Data File : BG059765.D
 Acq On : 18 Nov 2023 15:13
 Operator : MA/JU
 Sample : 05385-03 5X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 HR-03-11102023

Quant Time: Nov 20 04:43:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.342	152	31695	20.000	ng	0.00
21) Naphthalene-d8	11.191	136	137377	20.000	ng	0.00
39) Acenaphthene-d10	14.969	164	121683	20.000	ng	0.00
64) Phenanthrene-d10	17.725	188	237980	20.000	ng	# 0.00
76) Chrysene-d12	22.055	240	173676	20.000	ng	# 0.00
86) Perylene-d12	25.657	264	158368	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.839	112	30892	16.405	ng	0.00
7) Phenol-d6	7.460	99	43902	15.732	ng	0.00
23) Nitrobenzene-d5	9.546	82	44327	12.323	ng	0.00
42) 2,4,6-Tribromophenol	16.456	330	26309	19.056	ng	0.00
45) 2-Fluorobiphenyl	13.594	172	129873	14.366	ng	0.00
79) Terphenyl-d14	20.298	244	151341	12.864	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
Data File : BG059765.D
Acq On : 18 Nov 2023 15:13
Operator : MA/JU
Sample : 05385-03 5X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
HR-03-11102023

Quant Time: Nov 20 04:43:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 01:51:38 2023
Response via : Initial Calibration

