

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
 Data File : BG059728.D
 Acq On : 17 Nov 2023 12:04
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 17 12:38:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	129	0.00
2	1,4-Dioxane	40.000	36.293	9.3	157	0.00
3	Pyridine	40.000	39.053	2.4	140	0.00
4	n-Nitrosodimethylamine	40.000	40.765	-1.9	151	0.00
5 S	2-Fluorophenol	80.000	81.262	-1.6	137	0.00
6	Aniline	40.000	40.497	-1.2	134	0.00
7 S	Phenol-d6	80.000	84.634	-5.8	138	0.00
8	2-Chlorophenol	40.000	39.705	0.7	122	0.00
9	Benzaldehyde	40.000	42.698	-6.7	146	0.00
10 C	Phenol	40.000	43.601	-9.0	140	0.00
11	bis(2-Chloroethyl)ether	40.000	39.354	1.6	126	0.00
12	1,3-Dichlorobenzene	40.000	39.160	2.1	128	0.00
13 C	1,4-Dichlorobenzene	40.000	40.768	-1.9	136	0.00
14	1,2-Dichlorobenzene	40.000	41.035	-2.6	140	-0.32
15	Benzyl Alcohol	40.000	42.436	-6.1	134	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.128	-5.3	153	-0.02
17	2-Methylphenol	40.000	41.205	-3.0	134	0.00
18	Hexachloroethane	40.000	41.519	-3.8	137	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.542	-3.9	131	0.00
20	3+4-Methylphenols	40.000	42.661	-6.7	137	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	140	0.00
22	Acetophenone	40.000	39.042	2.4	132	0.00
23 S	Nitrobenzene-d5	80.000	83.709	-4.6	138	0.00
24	Nitrobenzene	40.000	41.687	-4.2	136	0.00
25	Isophorone	40.000	40.087	-0.2	135	0.00
26 C	2-Nitrophenol	40.000	40.502	-1.3	142	0.00
27	2,4-Dimethylphenol	40.000	38.832	2.9	136	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.316	1.7	136	0.00
29 C	2,4-Dichlorophenol	40.000	40.832	-2.1	132	0.00
30	1,2,4-Trichlorobenzene	40.000	42.364	-5.9	141	0.00
31	Naphthalene	40.000	40.500	-1.3	132	0.00
32	Benzoic acid	40.000	38.259	4.4	134	0.00
33	4-Chloroaniline	40.000	40.124	-0.3	132	0.00
34 C	Hexachlorobutadiene	40.000	39.842	0.4	138	0.00
35	Caprolactam	40.000	39.760	0.6	141	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.633	-9.1	148	0.00
37	2-Methylnaphthalene	40.000	39.820	0.4	130	0.00
38	1-Methylnaphthalene	40.000	40.665	-1.7	135	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.845	-4.6	126	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.938	-12.3	134	0.00
42 S	2,4,6-Tribromophenol	80.000	88.028	-10.0	133	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.607	-16.5	138	0.00
44	2,4,5-Trichlorophenol	40.000	43.796	-9.5	132	0.00
45 S	2-Fluorobiphenyl	80.000	88.138	-10.2	131	0.00
46	1,1'-Biphenyl	40.000	43.408	-8.5	132	0.00
47	2-Chloronaphthalene	40.000	45.162	-12.9	141	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
 Data File : BG059728.D
 Acq On : 17 Nov 2023 12:04
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 17 12:38:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	49.049	-22.6	145	0.00
49	Acenaphthylene	40.000	43.098	-7.7	133	0.00
50	Dimethylphthalate	40.000	43.393	-8.5	133	0.00
51	2,6-Dinitrotoluene	40.000	42.022	-5.1	131	0.00
52 C	Acenaphthene	40.000	44.113	-10.3	133	0.00
53	3-Nitroaniline	40.000	41.521	-3.8	125	0.00
54 P	2,4-Dinitrophenol	40.000	38.588	3.5	122	0.00
55	Dibenzofuran	40.000	43.996	-10.0	135	0.00
56 P	4-Nitrophenol	40.000	40.785	-2.0	130	0.00
57	2,4-Dinitrotoluene	40.000	43.745	-9.4	137	0.00
58	Fluorene	40.000	43.101	-7.8	129	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.613	-6.5	136	0.00
60	Diethylphthalate	40.000	44.617	-11.5	140	0.00
61	4-Chlorophenyl-phenylether	40.000	43.676	-9.2	133	0.00
62	4-Nitroaniline	40.000	38.938	2.7	129	0.00
63	Azobenzene	40.000	45.367	-13.4	137	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	131	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.267	-3.2	132	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.418	-6.0	129	0.00
67	4-Bromophenyl-phenylether	40.000	42.548	-6.4	135	0.00
68	Hexachlorobenzene	40.000	42.613	-6.5	133	0.00
69	Atrazine	40.000	39.919	0.2	134	0.00
70 C	Pentachlorophenol	40.000	43.075	-7.7	132	0.00
71	Phenanthrene	40.000	42.601	-6.5	133	0.00
72	Anthracene	40.000	40.636	-1.6	126	0.00
73	Carbazole	40.000	40.064	-0.2	125	0.00
74	Di-n-butylphthalate	40.000	41.300	-3.2	129	0.00
75 C	Fluoranthene	40.000	39.298	1.8	129	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	136	0.00
77	Benzydine	40.000	37.311	6.7	128	0.00
78	Pyrene	40.000	39.454	1.4	131	0.00
79 S	Terphenyl-d14	80.000	79.427	0.7	133	0.00
80	Butylbenzylphthalate	40.000	40.395	-1.0	136	0.00
81	Benzo(a)anthracene	40.000	39.592	1.0	132	0.00
82	3,3'-Dichlorobenzidine	40.000	38.734	3.2	129	0.00
83	Chrysene	40.000	42.054	-5.1	139	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.164	-2.9	136	0.00
85 c	Di-n-octyl phthalate	40.000	37.611	6.0	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.546	-3.9	102	0.00
88	Benzo(b)fluoranthene	40.000	45.653	-14.1	98	0.00
89	Benzo(k)fluoranthene	40.000	44.943	-12.4	98	0.00
90 C	Benzo(a)pyrene	40.000	41.327	-3.3	99	0.00
91	Dibenzo(a,h)anthracene	40.000	40.689	-1.7	102	0.00
92	Benzo(g,h,i)perylene	40.000	40.123	-0.3	99	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
Data File : BG059728.D
Acq On : 17 Nov 2023 12:04
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 17 12:38:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 01:51:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0