

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
 Data File : BG059735.D
 Acq On : 17 Nov 2023 18:10
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 18 04:17:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	88	0.00
2	1,4-Dioxane	40.000	39.050	2.4	116	0.00
3	Pyridine	40.000	33.263	16.8	81	0.00
4	n-Nitrosodimethylamine	40.000	37.755	5.6	96	0.00
5 S	2-Fluorophenol	80.000	77.992	2.5	90	0.00
6	Aniline	40.000	42.337	-5.8	95	0.00
7 S	Phenol-d6	80.000	88.692	-10.9	99	0.00
8	2-Chlorophenol	40.000	40.652	-1.6	85	0.00
9	Benzaldehyde	40.000	42.292	-5.7	99	0.00
10 C	Phenol	40.000	43.702	-9.3	96	0.00
11	bis(2-Chloroethyl)ether	40.000	39.509	1.2	86	0.00
12	1,3-Dichlorobenzene	40.000	39.979	0.1	90	0.00
13 C	1,4-Dichlorobenzene	40.000	40.155	-0.4	91	0.00
14	1,2-Dichlorobenzene	40.000	39.266	1.8	91	0.00
15	Benzyl Alcohol	40.000	44.146	-10.4	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.753	3.1	96	-0.02
17	2-Methylphenol	40.000	41.537	-3.8	92	0.00
18	Hexachloroethane	40.000	44.221	-10.6	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.251	-3.1	89	0.00
20	3+4-Methylphenols	40.000	45.018	-12.5	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	41.825	-4.6	97	0.00
23 S	Nitrobenzene-d5	80.000	84.817	-6.0	96	0.00
24	Nitrobenzene	40.000	41.858	-4.6	94	0.00
25	Isophorone	40.000	40.942	-2.4	95	0.00
26 C	2-Nitrophenol	40.000	41.511	-3.8	100	0.00
27	2,4-Dimethylphenol	40.000	42.470	-6.2	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.006	-7.5	102	0.00
29 C	2,4-Dichlorophenol	40.000	44.024	-10.1	98	0.00
30	1,2,4-Trichlorobenzene	40.000	41.566	-3.9	95	0.00
31	Naphthalene	40.000	40.290	-0.7	90	0.00
32	Benzoic acid	40.000	43.628	-9.1	105	0.00
33	4-Chloroaniline	40.000	42.030	-5.1	95	0.00
34 C	Hexachlorobutadiene	40.000	38.477	3.8	91	0.00
35	Caprolactam	40.000	39.528	1.2	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	45.408	-13.5	106	0.00
37	2-Methylnaphthalene	40.000	41.171	-2.9	93	0.00
38	1-Methylnaphthalene	40.000	41.819	-4.5	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.415	4.0	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.806	-9.5	105	0.00
42 S	2,4,6-Tribromophenol	80.000	84.176	-5.2	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.362	-3.4	98	0.00
44	2,4,5-Trichlorophenol	40.000	40.514	-1.3	98	0.00
45 S	2-Fluorobiphenyl	80.000	79.286	0.9	94	0.00
46	1,1'-Biphenyl	40.000	39.515	1.2	96	0.00
47	2-Chloronaphthalene	40.000	40.251	-0.6	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
 Data File : BG059735.D
 Acq On : 17 Nov 2023 18:10
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 18 04:17:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.600	-16.5	110	0.00
49	Acenaphthylene	40.000	40.746	-1.9	100	0.00
50	Dimethylphthalate	40.000	40.867	-2.2	101	0.00
51	2,6-Dinitrotoluene	40.000	41.689	-4.2	104	0.00
52 C	Acenaphthene	40.000	41.094	-2.7	99	0.00
53	3-Nitroaniline	40.000	38.785	3.0	94	0.00
54 P	2,4-Dinitrophenol	40.000	40.681	-1.7	103	0.00
55	Dibenzofuran	40.000	39.926	0.2	98	0.00
56 P	4-Nitrophenol	40.000	40.546	-1.4	103	0.00
57	2,4-Dinitrotoluene	40.000	41.754	-4.4	105	0.00
58	Fluorene	40.000	41.752	-4.4	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.485	-3.7	106	0.00
60	Diethylphthalate	40.000	40.855	-2.1	102	0.00
61	4-Chlorophenyl-phenylether	40.000	41.666	-4.2	101	0.00
62	4-Nitroaniline	40.000	38.946	2.6	104	0.00
63	Azobenzene	40.000	39.633	0.9	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.909	-19.8	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	46.465	-16.2	99	0.00
67	4-Bromophenyl-phenylether	40.000	44.804	-12.0	99	0.00
68	Hexachlorobenzene	40.000	45.567	-13.9	99	0.00
69	Atrazine	40.000	40.958	-2.4	96	0.00
70 C	Pentachlorophenol	40.000	44.884	-12.2	96	0.00
71	Phenanthrene	40.000	42.339	-5.8	92	0.00
72	Anthracene	40.000	42.525	-6.3	92	0.00
73	Carbazole	40.000	42.279	-5.7	92	0.00
74	Di-n-butylphthalate	40.000	42.132	-5.3	92	0.00
75 C	Fluoranthene	40.000	42.255	-5.6	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzydine	40.000	40.525	-1.3	94	0.00
78	Pyrene	40.000	43.226	-8.1	97	0.00
79 S	Terphenyl-d14	80.000	88.778	-11.0	100	0.00
80	Butylbenzylphthalate	40.000	42.246	-5.6	96	0.00
81	Benzo(a)anthracene	40.000	42.284	-5.7	95	0.00
82	3,3'-Dichlorobenzidine	40.000	43.383	-8.5	97	0.00
83	Chrysene	40.000	42.188	-5.5	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.965	-9.9	98	0.00
85 c	Di-n-octyl phthalate	40.000	37.349	6.6	72	0.00
86 I	Perylene-d12	20.000	20.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	46.295	-15.7	98	-0.02
88	Benzo(b)fluoranthene	40.000	34.907	12.7	65	0.00
89	Benzo(k)fluoranthene	40.000	35.806	10.5	67	0.00
90 C	Benzo(a)pyrene	40.000	40.538	-1.3	84	0.00
91	Dibenzo(a,h)anthracene	40.000	46.270	-15.7	100	0.00
92	Benzo(g,h,i)perylene	40.000	46.468	-16.2	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111723\
Data File : BG059735.D
Acq On : 17 Nov 2023 18:10
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 18 04:17:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 01:51:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0