





#39  
 2,4,6-Trichlorophenol  
 Concen: 20.099 ng/ul  
 RT: 13.32 min Scan# 1699  
 Delta R.T. 0.00 min  
 Lab File: BG047372.D  
 Acq: 20 Nov 2020 4:08

Instrument :  
 BNA\_G  
 ClientSampled :

Tot Ion:196 Resp: 70092  
 Ion Ratio Lower Upper  
 196 100  
 198 103.8 76.6 114.8  
 200 27.8 28.1 42.1#



#40  
 2,4,5-Trichlorophenol  
 Concen: 20.259 ng/ul  
 RT: 13.39 min Scan# 1711  
 Delta R.T. 0.00 min  
 Lab File: BG047372.D  
 Acq: 20 Nov 2020 4:08

Tgt Ion:196 Resp: 73351  
 Ion Ratio Lower Upper  
 196 100  
 198 95.2 77.7 116.5  
 200 29.4 27.2 40.8

