

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111921\
 Data File : BG051142.D
 Acq On : 20 Nov 2021 4:10
 Operator : CG/JU
 Sample : M4758-19
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP10

Quant Time: Nov 20 04:51:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 19 16:26:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.208	152	24520	20.000	ng	0.00
21) Naphthalene-d8	11.029	136	105570	20.000	ng	# 0.00
39) Acenaphthene-d10	14.830	164	76144	20.000	ng	0.00
64) Phenanthrene-d10	17.580	188	186191	20.000	ng	# 0.00
76) Chrysene-d12	21.875	240	171499	20.000	ng	# 0.00
86) Perylene-d12	25.277	264	169026	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.747	112	152493	96.138	ng	0.00
7) Phenol-d6	7.356	99	204375	91.481	ng	0.00
23) Nitrobenzene-d5	9.384	82	144449	62.460	ng	0.00
42) 2,4,6-Tribromophenol	16.322	330	80872	99.927	ng	0.00
45) 2-Fluorobiphenyl	13.455	172	296352	58.729	ng	0.00
79) Terphenyl-d14	20.171	244	507769	54.126	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111921\
Data File : BG051142.D
Acq On : 20 Nov 2021 4:10
Operator : CG/JU
Sample : M4758-19
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP10

Quant Time: Nov 20 04:51:24 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 19 16:26:22 2021
Response via : Initial Calibration

