

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111925\
 Data File : BG064817.D
 Acq On : 19 Nov 2025 15:57
 Operator : RC/JU
 Sample : Q3530-08
 Misc : LOQ-WATER-2.5PPM
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-02-QT4-2025

Quant Time: Nov 19 16:33:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 12 17:10:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.164	152	44497	20.000	ng	0.00
21) Naphthalene-d8	10.984	136	177607	20.000	ng	0.00
39) Acenaphthene-d10	14.774	164	150709	20.000	ng	0.00
64) Phenanthrene-d10	17.500	188	371324	20.000	ng	0.00
76) Chrysene-d12	21.748	240	443925	20.000	ng	# 0.00
86) Perylene-d12	25.009	264	425566	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.702	112	291003	114.200	ng	0.00
7) Phenol-d6	7.306	99	386865	114.146	ng	0.00
23) Nitrobenzene-d5	9.327	82	258462	71.952	ng	0.00
42) 2,4,6-Tribromophenol	16.248	330	341932	116.331	ng	0.00
45) 2-Fluorobiphenyl	13.405	172	748587	69.308	ng	0.00
79) Terphenyl-d14	20.091	244	1720879	76.254	ng	0.00
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.963	70	5039	2.199	ng	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111925\
Data File : BG064817.D
Acq On : 19 Nov 2025 15:57
Operator : RC/JU
Sample : Q3530-08
Misc : LOQ-WATER-2.5PPM
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-02-QT4-2025

Quant Time: Nov 19 16:33:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 12 17:10:56 2025
Response via : Initial Calibration

