

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG112018\  
 Data File : BG038087.D  
 Acq On : 20 Nov 2018 16:10  
 Operator : JU/SJ  
 Sample : J5993-10  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SPLP-LEACHATE-AMENDED-WCC-

Quant Time: Nov 20 18:15:04 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Nov 13 16:39:28 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)       |
|-----------------------------|-------|------|----------|--------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4   | 8.08  | 152  | 48623    | 20.00  | ng    | 0.00           |
| 21) Naphthalene-d8          | 10.90 | 136  | 209653   | 20.00  | ng    | 0.00           |
| 39) Acenaphthene-d10        | 14.72 | 164  | 134454   | 20.00  | ng    | 0.00           |
| 64) Phenanthrene-d10        | 17.47 | 188  | 291119   | 20.00  | ng    | 0.00           |
| 76) Chrysene-d12            | 21.75 | 240  | 280548   | 20.00  | ng    | 0.00           |
| 87) Perylene-d12            | 25.07 | 264  | 288421   | 20.00  | ng    | 0.00           |
| System Monitoring Compounds |       |      |          |        |       |                |
| 5) 2-Fluorophenol           | 5.63  | 112  | 139911   | 49.68  | ng    | 0.00           |
| 7) Phenol-d6                | 7.23  | 99   | 122420   | 30.45  | ng    | 0.00           |
| 23) Nitrobenzene-d5         | 9.26  | 82   | 317289   | 81.01  | ng    | 0.00           |
| 42) 2,4,6-Tribromophenol    | 16.21 | 330  | 170419   | 111.14 | ng    | 0.00           |
| 45) 2-Fluorobiphenyl        | 13.34 | 172  | 725931   | 78.66  | ng    | 0.00           |
| 79) Terphenyl-d14           | 20.07 | 244  | 1043867  | 87.55  | ng    | 0.00           |
| Target Compounds            |       |      |          |        |       |                |
| 31) Naphthalene             | 10.96 | 128  | 22287    | 2.161  | ng    | Qvalue<br># 96 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG112018\  
Data File : BG038087.D  
Acq On : 20 Nov 2018 16:10  
Operator : JU/SJ  
Sample : J5993-10  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
SPLP-LEACHATE-AMENDED-WCC-

Quant Time: Nov 20 18:15:04 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Nov 13 16:39:28 2018  
Response via : Initial Calibration

