


Quantitation Report (Qedit)




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112025\
 Data File : BG064826.D
 Acq On : 20 Nov 2025 9:18
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 20 09:55:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 12 17:10:56 2025
 Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112025\
Data File : BG064826.D
Acq On : 20 Nov 2025 9:18
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 20 09:55:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 12 17:10:56 2025
Response via : Initial Calibration



TIC: BG064826.D\data.ms

(70) Pentachlorophenol (C)

17.148min (-0.009) 274268.75 ng

response 397178

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	70.10	63.18
264.00	64.00	63.59
0.00	0.00	0.00