

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112123\
 Data File : BG059804.D
 Acq On : 21 Nov 2023 23:07
 Operator : MA/JU
 Sample : 05378-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MH-3

Quant Time: Nov 22 00:34:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 01:51:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.334	152	24057	20.000	ng	0.00
21) Naphthalene-d8	11.184	136	101668	20.000	ng	0.00
39) Acenaphthene-d10	14.962	164	105869	20.000	ng	0.00
64) Phenanthrene-d10	17.717	188	258698	20.000	ng	# 0.00
76) Chrysene-d12	22.042	240	199585	20.000	ng	# 0.00
86) Perylene-d12	25.626	264	222229	20.000	ng	# 0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.831	112	168111	117.618	ng	0.00
7) Phenol-d6	7.459	99	233851	110.407	ng	0.00
23) Nitrobenzene-d5	9.539	82	265830	99.860	ng	0.00
42) 2,4,6-Tribromophenol	16.454	330	231385	192.626	ng	0.00
45) 2-Fluorobiphenyl	13.587	172	786322	99.975	ng	0.00
79) Terphenyl-d14	20.291	244	1707833	126.321	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112123\
Data File : BG059804.D
Acq On : 21 Nov 2023 23:07
Operator : MA/JU
Sample : 05378-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MH-3

Quant Time: Nov 22 00:34:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 01:51:38 2023
Response via : Initial Calibration

